



US006770071B2

(12) **United States Patent**
Woloszko et al.

(10) **Patent No.:** **US 6,770,071 B2**
(45) **Date of Patent:** **Aug. 3, 2004**

(54) **BLADED ELECTROSURGICAL PROBE**

(75) Inventors: **Jean Woloszko**, Mountain View, CA (US); **Craig Tsuji**, San Jose, CA (US); **Hira V. Thapliyal**, Los Altos, CA (US); **Philip E. Eggers**, Dublin, OH (US)

(73) Assignee: **Arthrocare Corporation**, Sunnyvale, CA (US)

(*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 0 days.

(21) Appl. No.: **09/780,745**

(22) Filed: **Feb. 9, 2001**

(65) **Prior Publication Data**

US 2001/0020167 A1 Sep. 6, 2001

Related U.S. Application Data

(63) Continuation-in-part of application No. 09/162,117, filed on Sep. 28, 1998, now Pat. No. 6,117,109, and a continuation-in-part of application No. 09/041,934, filed on Mar. 13, 1998, now Pat. No. 6,391,025, which is a continuation-in-part of application No. 08/977,845, filed on Nov. 25, 1997, now Pat. No. 6,210,402, which is a continuation-in-part of application No. 08/562,332, filed on Nov. 22, 1995, now Pat. No. 6,024,733, which is a continuation-in-part of application No. 08/485,219, filed on Jun. 7, 1995, now Pat. No. 5,697,281.

(60) Provisional application No. 60/182,751, filed on Feb. 16, 2000.

(51) **Int. Cl.**⁷ **A61B 18/18**

(52) **U.S. Cl.** **606/41; 606/45; 606/46; 606/48; 606/49; 606/50**

(58) **Field of Search** **606/41, 45, 46, 606/48-60**

(56) **References Cited**

U.S. PATENT DOCUMENTS

2,050,904 A 8/1936 Trice
2,056,377 A 10/1936 Wappler 125/303

(List continued on next page.)

FOREIGN PATENT DOCUMENTS

DE 3930451 3/1991 A61B/17/39

(List continued on next page.)

OTHER PUBLICATIONS

Pearce, John A. (1986) *Electrosurgery*, pp. 17, 67-75, 87, John Wiley & Sons, New York.

(List continued on next page.)


Primary Examiner—Lee Cohen

(74) *Attorney, Agent, or Firm*—John T. Raffle; Richard R. Batt

(57) **ABSTRACT**

The present invention provides systems, apparatus and methods for selectively applying electrical energy to body tissue in order to, ablate, contract, coagulate, or otherwise modify a tissue or organ of a patients. An electrosurgical apparatus includes an electrode support bearing an active electrode in the form of a plasma blade or hook having an active edge and first and second blade sides. The active edge is adapted for severing a target tissue via localized molecular dissociation of tissue components. The first and second blade sides are adapted for engaging against, and coagulating, the severed tissue. s. A method of the present invention comprises positioning an electrosurgical probe adjacent to the target tissue so that a blade- or hook-like active electrode is brought into at least close proximity to the target tissue in the presence of an electrically conductive fluid. A high frequency voltage is applied between the active electrode and a return electrode to effect cool ablation or other modification of the target tissue. During application of the high frequency voltage, the electrosurgical apparatus may be translated, reciprocated, or otherwise manipulated such that the active edge is moved with respect to the tissue. The present invention volumetrically ablates or otherwise modifies the target tissue with minimal or no damage to surrounding, non-target tissue.

12 Claims, 41 Drawing Sheets



Serendia, LLC
Exhibit 2050
Jeisys v. Serendia
IPR2024-00386
Ex. 2050-001

RA0003031

CX-2648.1

U.S. PATENT DOCUMENTS

3,815,604 A	6/1974	O'Malley et al.	128/305	5,122,138 A	6/1992	Manwaring	606/46
3,828,780 A	8/1974	Morrison, Jr.	128/275	5,125,928 A	6/1992	Parins et al.	606/48
3,901,242 A	8/1975	Storz	128/303	5,167,659 A	12/1992	Ohtomo et al.	606/40
3,920,021 A	11/1975	Hiltebrandt	128/303	5,171,311 A	12/1992	Rydell et al.	606/48
3,939,839 A	2/1976	Curtiss	128/303	5,178,620 A	1/1993	Eggers et al.	606/41
3,970,088 A	7/1976	Morrison	128/303	5,190,517 A	3/1993	Zieve et al.	604/22
4,033,351 A	7/1977	Hetzel	128/303	5,190,541 A	* 3/1993	Abele et al.	606/46
4,040,426 A	8/1977	Morrison, Jr.	128/303	5,192,280 A	3/1993	Parins	606/48
4,043,342 A	8/1977	Morrison, Jr.	128/303	5,195,959 A	3/1993	Smith	604/34
4,074,718 A	2/1978	Morrison, Jr.	128/303	5,197,963 A	3/1993	Parins	606/46
4,092,986 A	6/1978	Schneiderman	128/303	5,207,675 A	5/1993	Canady	606/40
4,116,198 A	9/1978	Roos	128/303	5,217,457 A	6/1993	Delahuerga et al.	606/42
4,181,131 A	1/1980	Ogiu	128/303	5,217,459 A	6/1993	Kamerling	606/48
4,184,492 A	1/1980	Meinke et al.	128/303	5,217,460 A	6/1993	Knocpfier	
4,202,337 A	5/1980	Hren et al.	128/303	5,249,585 A	10/1993	Turner et al.	607/99
4,228,800 A	10/1980	Degler, Jr. et al.	128/303	5,261,905 A	* 11/1993	Doresey, III	606/45
4,232,676 A	11/1980	Herczog	128/303	5,267,994 A	12/1993	Gentelia et al.	606/15
4,240,441 A	12/1980	Khalil	128/692	5,267,997 A	12/1993	Farin et al.	606/38
4,248,231 A	2/1981	Herczog et al.	128/303	5,273,524 A	12/1993	Fox et al.	604/21
4,326,529 A	4/1982	Doss et al.	128/303	5,277,201 A	1/1994	Stern	607/98
4,381,007 A	4/1983	Doss	128/303	5,281,216 A	1/1994	Kliceck	606/42
4,476,862 A	10/1984	Pao	128/303	5,281,218 A	1/1994	Imran	606/41
4,532,924 A	8/1985	Auth et al.	128/303	5,282,799 A	2/1994	Rydell	
4,548,207 A	10/1985	Reimels	128/303	5,290,282 A	3/1994	Casscells	606/29
4,567,890 A	2/1986	Ohta et al.	128/303	5,300,069 A	4/1994	Hunsberger et al.	
4,590,934 A	5/1986	Malis et al.	128/303	5,306,238 A	4/1994	Fleenor	606/42
4,593,691 A	6/1986	Lindstrom et al.	128/303	5,312,400 A	5/1994	Bales et al.	
4,660,571 A	4/1987	Hess et al.	128/784	5,314,406 A	5/1994	Arias et al.	604/21
4,674,499 A	6/1987	Pao	128/303	5,318,563 A	6/1994	Malis et al.	606/38
4,682,596 A	7/1987	Bales et al.	128/303	5,324,254 A	6/1994	Phillips	604/21
4,706,667 A	11/1987	Roos	128/303	5,330,470 A	7/1994	Hagen	606/42
4,709,698 A	12/1987	Johnston et al.	128/303	5,334,140 A	8/1994	Phillips	604/35
4,727,874 A	3/1988	Bowers et al.	128/303	5,334,183 A	8/1994	Wuchinich	606/46
4,765,331 A	8/1988	Petruzzi et al.	128/303	5,336,220 A	8/1994	Ryan et al.	
4,785,807 A	* 11/1988	Blanch	606/45	5,342,357 A	8/1994	Nardella	
4,785,823 A	11/1988	Eggers et al.	128/692	5,366,443 A	11/1994	Eggers et al.	604/114
4,805,616 A	2/1989	Pao	128/303	5,370,675 A	12/1994	Edwards et al.	607/101
4,823,791 A	4/1989	D'Amelio et al.	123/303	5,380,277 A	1/1995	Phillips	604/33
4,832,048 A	5/1989	Cohen	128/786	5,383,876 A	1/1995	Nardella	606/49
4,860,752 A	8/1989	Turner	128/422	5,383,917 A	1/1995	Desai et al.	607/702
4,920,978 A	5/1990	Colvin	128/784	5,395,312 A	3/1995	Desai	604/22
4,931,047 A	6/1990	Broadwin et al.	604/22	5,417,687 A	5/1995	Nardella et al.	606/32
4,936,281 A	6/1990	Stasz	128/660	5,419,767 A	5/1995	Eggers et al.	604/114
4,936,301 A	6/1990	Rexroth et al.	606/45	5,423,810 A	6/1995	Goble et al.	606/40
4,943,290 A	7/1990	Rexroth et al.	606/45	5,423,882 A	6/1995	Jackman et al.	607/122
4,955,377 A	9/1990	Lennox et al.	128/401	5,436,566 A	7/1995	Thompson et al.	624/713
4,966,597 A	10/1990	Cosman	606/50	5,438,302 A	8/1995	Goble	331/167
4,967,765 A	11/1990	Turner et al.	128/785	5,441,499 A	8/1995	Fritzsche	606/45
4,976,711 A	12/1990	Parins et al.	606/48	5,451,224 A	9/1995	Goble et al.	606/48
4,979,948 A	12/1990	Geddes et al.	606/33	5,454,809 A	10/1995	Janssen	606/41
4,998,933 A	3/1991	Eggers et al.	606/41	5,490,850 A	2/1996	Ellman et al.	
5,007,908 A	4/1991	Rydell	606/47	5,496,312 A	3/1996	Kliceck	606/34
5,009,656 A	4/1991	Reimels	606/48	5,514,130 A	5/1996	Baker	606/41
5,035,696 A	7/1991	Rydell	606/47	5,556,397 A	9/1996	Long et al.	606/48
5,047,026 A	9/1991	Rydell	606/48	5,562,703 A	10/1996	Desai	
5,047,027 A	9/1991	Rydell	606/48	5,569,242 A	10/1996	Lax et al.	606/42
5,057,105 A	10/1991	Malone et al.	606/28	5,584,872 A	12/1996	LaFontaine et al.	607/117
5,057,106 A	10/1991	Kasevich et al.	606/33	5,609,151 A	3/1997	Mulier et al.	128/642
5,078,717 A	1/1992	Parins et al.	606/48	5,647,869 A	7/1997	Goble et al.	606/37
5,080,660 A	1/1992	Buelna	606/45	5,662,680 A	9/1997	Desai	606/210
5,083,565 A	1/1992	Parins	128/642	5,676,693 A	10/1997	LaFontaine et al.	607/116
5,084,044 A	1/1992	Quint	606/27	5,681,282 A	10/1997	Eggers et al.	604/114
5,085,659 A	2/1992	Rydell	606/47	5,683,366 A	11/1997	Eggers et al.	604/114
5,088,997 A	2/1992	Delahuerga et al.	606/42	5,697,281 A	12/1997	Eggers et al.	604/114
5,092,339 A	3/1992	Geddes et al.	128/692	5,697,536 A	12/1997	Eggers et al.	604/114
5,098,431 A	3/1992	Rydell	606/48	5,697,882 A	12/1997	Eggers et al.	604/114
5,099,840 A	3/1992	Goble et al.	128/422	5,697,909 A	12/1997	Eggers et al.	604/114
5,100,402 A	* 3/1992	Fan	606/41	5,700,262 A	12/1997	Acosta et al.	606/48
5,108,391 A	4/1992	Flachenecker et al.	606/38	5,702,390 A	12/1997	Austin et al.	606/48
5,112,330 A	5/1992	Nishigaki et al.	606/46	5,725,524 A	3/1998	Mulier et al.	606/41
				5,749,869 A	5/1998	Lindenmeier et al.	606/34

5,766,153 A	6/1998	Eggers et al.	604/114
5,800,431 A	9/1998	Brown	606/42
5,807,392 A *	9/1998	Eggers	606/31
5,807,395 A	9/1998	Mulier et al.	606/41
5,810,764 A	9/1998	Eggers et al.	604/23
5,810,809 A	9/1998	Rydell	606/49
5,843,019 A	12/1998	Eggers et al.	604/22
5,871,469 A	2/1999	Fggers et al.	604/114
5,873,855 A	2/1999	Eggers et al.	604/114
5,885,277 A	3/1999	Korth	606/35
5,888,198 A	3/1999	Eggers et al.	604/114
5,891,095 A	4/1999	Eggers et al.	604/114
5,893,849 A *	4/1999	Weaver	606/41
5,897,553 A	4/1999	Mulier et al.	606/41
5,902,272 A	5/1999	Eggers et al.	604/114
5,944,715 A	8/1999	Goble et al.	606/41
6,004,319 A	12/1999	Goble et al.	606/48
6,013,076 A	1/2000	Goble et al.	606/41
6,015,406 A	1/2000	Goble et al.	606/41
6,024,733 A	2/2000	Eggers et al.	604/500
6,027,501 A	2/2000	Goble et al.	606/41
6,039,734 A	3/2000	Goble	606/41
6,056,746 A *	5/2000	Goble et al.	606/48
6,063,079 A	5/2000	Hovda et al.	606/41
6,063,083 A *	5/2000	Duong-Van	606/45
6,066,134 A	5/2000	Eggers et al.	606/32
6,066,137 A *	5/2000	Greep	606/45
6,068,628 A	5/2000	Fanton et al.	606/41
6,074,386 A	6/2000	Goble et al.	606/34
6,090,106 A	7/2000	Goble et al.	606/41
6,093,186 A	7/2000	Goble et al.	606/34
6,096,037 A	8/2000	Mulier et al.	
6,109,268 A	8/2000	Thapliyal et al.	128/898
6,117,109 A	9/2000	Eggers et al.	604/114
6,126,682 A	10/2000	Sharkey et al.	607/96
6,142,992 A	11/2000	Cheng et al.	606/34
6,149,620 A	11/2000	Baker et al.	604/22
6,159,194 A	12/2000	Eggers et al.	604/500
6,159,208 A	12/2000	Hovda et al.	606/41
6,168,593 B1	1/2001	Sharkey et al.	606/34
6,179,824 B1	1/2001	Eggers et al.	604/500
6,179,836 B1	1/2001	Eggers et al.	606/45
6,183,469 B1	2/2001	Thapliyal et al.	606/41
6,190,381 B1	2/2001	Olsen et al.	606/32
6,203,542 B1	3/2001	Ellsberry et al.	606/41
6,210,402 B1	4/2001	Olsen et al.	606/32
6,217,575 B1	4/2001	DeVore et al.	
6,224,592 B1	5/2001	Eggers et al.	606/32
6,228,081 B1	5/2001	Goble	606/34
6,235,020 B1	5/2001	Cheng et al.	606/34
6,238,391 B1	5/2001	Olsen et al.	606/41
6,254,600 B1	7/2001	Willink et al.	606/41
6,261,286 B1	7/2001	Goble et al.	606/34
6,264,652 B1	7/2001	Eggers et al.	606/41
6,273,887 B1	8/2001	Yamauchi et al.	
6,293,942 B1	9/2001	Goble et al.	606/38
6,296,636 B1	10/2001	Cheng et al.	606/32
6,296,638 B1	10/2001	Davison et al.	606/41
6,306,134 B1	10/2001	Goble et al.	606/42
6,312,408 B1	11/2001	Eggers et al.	604/114
6,322,549 B1	11/2001	Eggers et al.	604/500
6,355,032 B1	3/2002	Hovda et al.	606/32
6,363,937 B1	4/2002	Hovda et al.	128/898
6,379,350 B1	4/2002	Sharkey et al.	
6,379,351 B1	4/2002	Thapliyal et al.	606/41
6,391,025 B1	5/2002	Weinstein et al.	606/41
6,416,507 B1	7/2002	Eggers et al.	606/32
6,416,508 B1	7/2002	Fggers et al.	606/32
6,432,103 B1	8/2002	Ellsberry et al.	606/41
6,468,274 B1	10/2002	Alleyne et al.	
6,632,193 B1	10/2003	Davison et al.	

6,632,220 B1 10/2003 Eggers et al.

FOREIGN PATENT DOCUMENTS

EP	0 703 461	3/1996		G01R/27/02
EP	0 740 926 A2	11/1996		A61B/17/39
EP	0 754 437	1/1997		A61B/17/39
EP	0 623 316	3/1999		
EP	0 717 966	4/2003		
FR	2313949	1/1977		A61N/3/02
GB	2 308 979	7/1997		A61B/17/36
GB	2 308 980	7/1997		A61B/17/36
GB	2 308 981	7/1997		A61B/17/36
GB	2 327 350	1/1999		A61B/17/39
GB	2 327 351	1/1999		A61B/17/39
GB	2 327 352	1/1999		A61B/17/39
GB	2 379 878	3/2003		
JP	57-57802	4/1982		A61B/1/00
JP	57-117843	7/1982		A61B/17/39
WO	90/03152	4/1990		A61B/17/39
WO	WO 90/07303	7/1990		A61B/17/39
WO	92/21278	12/1992		A61B/5/04
WO	WO 93/13816	7/1993		A61B/17/36
WO	93/20747	10/1993		A61B/5/00
WO	WO 94/04220	3/1994		A61N/1/06
WO	94/08654	4/1994		A61M/37/00
WO	WO95/34259	12/1995		A61F/5/48
WO	96/00042	1/1996		A61B/17/39
WO	97/00646	1/1997		A61B/17/39
WO	97/00647	1/1997		A61B/17/39
WO	97/24073	7/1997		A61B/17/39
WO	WO 97/24074	7/1997		A61B/17/39
WO	97/24993	7/1997		A61B/17/39
WO	97/24994	7/1997		A61B/17/39
WO	97/48345	12/1997		A61B/17/39
WO	97/48346	12/1997		A61B/17/39
WO	98/27879	7/1998		A61B/17/36
WO	98/27880	7/1998		A61B/17/39
WO	98/07468	2/1999		A61N/1/40
WO	99/51155	10/1999		A61B/17/36
WO	99/51158	10/1999		A61B/17/39
WO	0 324 339	3/2003		

OTHER PUBLICATIONS

Costello, A. J. et al. "Nd:YAG Laser Ablation of the Prostate as a Treatment of Benign Prostatic Hypertrophy," (1992) *Lasers in Surger and Medicine* 12:121-124.

J.W. Ramsey et al. *Urological Research* vol. 13, pp. 99-102 (1985).

V.E. Elsasser et al. *Acta Medico Technica* vol. 24, No. 4, pp. 129-134 (1976).

P.C. Nardella (1989) *SPIE* 1068:42-49 Radio Frequency Energy and Impedance Feedback.

Rand et al. (1985) *J. Arthro. Surg.* 1:242-246 Effect of Electrocautery on Fresh Human Articular Cartilage.

Buchelt, M. et al. "Excimer Laser Ablation of Fibrocartilage: An In Vitro and In Vivo Study," (1991) *Lasers in Surgery and Medicine* 11:271-279.

R. Tucker et al., Abstract P14-11, p. 248, "A Bipolar Electrosurgical Turp Loop".

R. Tucker et al. *J. of Urology* vol. 141, pp. 662-665, (1989).

R. Tucker et al. *Urological Research* vol. 18, pp. 291-294 (1990).

Kramolowsky et al. *J. of Urology* vol. 143, pp. 275–277 (1990).

Kramolowsky et al. *J. of Urology* vol. 146, pp. 669–674 (1991).

Slager et al. *Z. Kardiol.* 76:Suppl. 6, 67–71 (1987).

Slager et al. *JACC* 5(6):1382–6 (1985).

A.K. Dobbie *Bio-Medical Engineering*. vol. 4, pp. 206–216 (1969).

W. Honig *IEEE* pp. 58–65 (1975).

M.B. Dennis et al. *Digestive Diseases and Sciences* vol. 24(11), pp. 845–848 (1979).


K. Barry et al. *American Heart Journal* vol. 117, pp. 332–341 (1982).

C.P. Swain *Gut* vol. 25, pp. 1424–1431 (1984).

B. Lee et al. *JACC* vol. 13(5), pp. 1167–1175 (1989).

Piercey et al. *Gastroenterology* vol. 74(3), pp. 527–534 (1978).

* cited by examiner






FIG. 2




FIG. 3





FIG. 4






FIG. 6




FIG. 7




FIG. 8




FIG. 9A




FIG. 10A




FIG. 9B




FIG. 10B




FIG. 11A




FIG. 12A




FIG. 13




FIG. 11B




FIG. 12B




FIG. 14




FIG. 15




FIG. 16




FIG. 17A




FIG. 17B




FIG. 18




FIG. 19




FIG. 20A





FIG. 20B

*FIG. 25*




FIG. 21




FIG. 22




FIG. 23





FIG. 24






FIG. 27A




FIG. 27B




FIG. 27C




FIG. 28





FIG. 29






FIG. 31A




FIG. 31B




FIG. 32A




FIG. 32B




FIG. 32C




FIG. 33A




FIG. 33B




FIG. 33C




FIG. 33D




FIG. 33E




FIG. 34A




FIG. 34B




FIG. 34C




FIG. 35A




FIG. 35B




FIG. 35C




FIG. 36A




FIG. 36B




FIG. 37A




FIG. 37B




FIG. 38A




FIG. 38B




FIG. 39A




FIG. 39B




FIG. 39C




FIG. 39D




FIG. 40A




FIG. 40B




FIG. 40C




FIG. 41A




FIG. 41B




FIG. 41C




FIG. 42A







FIG. 42B

*FIG. 43A*

*FIG. 43B*

*FIG. 44*

BLADED ELECTROSURGICAL PROBE**CROSS-REFERENCES TO RELATED APPLICATIONS**

The present invention claims priority from U.S. Provisional Patent Application No. 60/182,751 filed Feb. 16, 2000, and is a continuation in part of U.S. patent application Ser. No. 08/977,845, filed Nov. 25, 1997, now U.S. Pat. No. 6,210,402, which is a continuation-in-part of application Ser. No. 08/562,332, filed Nov. 22, 1995, now U.S. Pat. No. 6,024,733, and U.S. patent application Ser. No. 09/041,934, filed Mar. 13, 1998, now U.S. Pat. No. 6,391,025 which is a continuation-in-part of U.S. patent application Ser. No. 08/485,219 (now U.S. Pat. No. 5,697,281), filed on Jun. 7, 1995.

The present invention is also related to commonly assigned copending U.S. Provisional Patent Application No. 60/062,996, filed Oct. 23, 1997, U.S. patent application Ser. No. 08/990,374, filed Dec. 15, 1997, which is a continuation-in-part of U.S. patent application Ser. No. 08/485,219, filed on Jun. 7, 1995, now U.S. Pat. No. 5,697,281, patent application Ser. Nos. 09/109,219, 09/058,571, 08/874,173 and 09/002,315, filed on Jun. 30, 1998, Apr. 10, 1998, Jun. 13, 1997, and Jan. 2, 1998, respectively, and U.S. patent application Ser. No. 09/054,323, filed on Apr. 2, 1998, U.S. patent application Ser. No. 09/010,382, filed Jan. 21, 1998, and U.S. patent application Ser. No. 09/032,375, filed Feb. 27, 1998, U.S. patent application Ser. Nos. 08/977,845, filed on Nov. 25, 1997, 08/942,580, filed on Oct. 2, 1997, U.S. application Ser. No. 08/753,227, filed on Nov. 22, 1996, U.S. application Ser. No. 08/687,792, filed on Jul. 18, 1996, and PCT International Application, U.S. National Phase Ser. No. PCT/US94/05168, filed on May 10, 1994, now U.S. Pat. No. 5,697,909, which was a continuation-in-part of U.S. patent application Ser. No. 08/059,681, filed on May 10, 1993, which was a continuation-in-part of U.S. patent application Ser. No. 07/958,977, filed on Oct. 9, 1992 which was a continuation-in-part of U.S. patent application Ser. No. 07/817,575, filed on Jan. 7, 1992, the complete disclosures of which are incorporated herein by reference for all purposes. The present invention is also related to commonly assigned U.S. Pat. No. 5,683,366, filed Nov. 22, 1995, the complete disclosure of which is incorporated herein by reference for all purposes.

BACKGROUND OF THE INVENTION

The present invention generally relates to electrosurgical systems and methods for ablating, severing, dissecting, contracting, or otherwise modifying target tissues or organs. The invention relates more particularly to electrosurgical apparatus and methods for modifying a tissue or organ via molecular dissociation of tissue components, wherein the apparatus includes an active electrode in the form of a plasma blade or hook. The present invention further relates to electrosurgical instruments and methods for harvesting blood vessels such as the internal mammary artery, the saphenous vein, or the like, for use in coronary artery bypass graft procedures.

Conventional electrosurgical instruments and techniques are widely used in surgical procedures because they generally reduce patient bleeding and trauma associated with cutting operations, as compared with mechanical cutting and the like. Conventional electrosurgical procedures may be classified as operating in monopolar or bipolar mode. Monopolar techniques rely on external grounding of the patient, where the surgical device defines only a single

electrode pole. Bipolar devices have two electrodes for the application of current between their surfaces. Conventional electrosurgical devices and procedures, however, suffer from a number of disadvantages. For example, conventional electrosurgical cutting devices typically operate by creating a voltage difference between the active electrode and the target tissue, causing an electrical arc to form across the physical gap between the electrode and the tissue. At the point of contact of the electric arcs with the tissue, rapid tissue heating occurs due to high current density between the electrode and the tissue. This high current density causes cellular fluids to rapidly vaporize into steam, thereby producing a "cutting effect" along the pathway of localized tissue heating. Thus, the tissue is parted along the pathway of evaporated cellular fluid, inducing undesirable collateral tissue damage in regions surrounding the target tissue.

Further, monopolar electrosurgical devices generally direct electric current along a defined path from the exposed or active electrode through the patient's body to the return electrode, the latter externally attached to a suitable location on the patient. This creates the potential danger that the electric current will flow through undefined paths in the patient's body, thereby increasing the risk of unwanted electrical stimulation to portions of the patient's body. In addition, since the defined path through the patient's body has a relatively high electrical impedance, large voltage differences must typically be applied between the return and active electrodes in order to generate a current suitable for ablation or cutting of the target tissue. This current, however, may inadvertently flow along body paths having less impedance than the defined electrical path, which will substantially increase the current flowing through these paths, possibly causing damage to or destroying surrounding tissue.

Bipolar electrosurgical devices have an inherent advantage over monopolar devices because the return current path does not flow through the patient. In bipolar electrosurgical devices, both the active and return electrode are typically exposed so that both electrodes may contact tissue, thereby providing a return current path from the active to the return electrode through the tissue. One drawback with this configuration, however, is that the return electrode may cause tissue desiccation or destruction at its contact point with the patient's tissue. In addition, the active and return electrodes are typically positioned close together to ensure that the return current flows directly from the active to the return electrode. The close proximity of these electrodes generates the danger that the current will short across the electrodes, possibly impairing the electrical control system and/or damaging or destroying surrounding tissue.

In addition, conventional electrosurgical methods are generally ineffective for ablating certain types of tissue, and in certain types of environments within the body. For example, loose or elastic connective tissue, such as the synovial tissue in joints, is extremely difficult (if not impossible) to remove with conventional electrosurgical instruments because the flexible tissue tends to move away from the instrument when it is brought against this tissue. Since conventional techniques rely mainly on conducting current through the tissue, they are not effective when the instrument cannot be brought adjacent to, or in contact with, the elastic tissue for a sufficient period of time to energize the electrode and conduct current through the tissue.

A prevalent form of cardiovascular disease is atherosclerosis in which the cardiovascular system leading to the heart is damaged or obstructed as a result of occluding material in the blood stream. Vascular complications produced by

atherosclerosis, such as stenosis, aneurysm, rupture, or occlusion, increase the likelihood of angina, stroke, and heart attacks. In many cases, the obstruction of the blood stream leading to the heart can be treated by a coronary artery bypass graft (CABG) procedure.

In a conventional CABG procedure, the obstruction is bypassed by a vascular conduit established between an arterial blood source and the coronary artery to a location beyond the obstruction. The vascular conduit is typically a non-critical artery or vein harvested from elsewhere in the body. In a procedure known as "free bypass graft", the saphenous vein is harvested from the patient's leg and is used as the vascular conduit. One end of the saphenous vein is anastomosed to the aorta and the other end is anastomosed to the diseased coronary artery at a location past the obstruction. In a procedure known as "in situ bypass graft", an internal mammary artery (IMA) is used as the bypass conduit. In an in situ bypass graft procedure, the surgeon dissects a sufficient length of the artery from its connective tissue, then transects the artery and connects the transected end to the diseased coronary past the obstruction, and leaves the other end of the IMA attached to the arterial supply.

The internal mammary arteries are particularly desirable for use as in situ bypass grafts, as they are conveniently located, have diameters and blood flow volumes that are comparable to those of coronary arteries, and have superior patency rates. Use of the left or right IMA as a bypass graft first involves harvesting the IMA from the inside chest wall.

In conventional CABG procedures, access to the IMA is typically obtained either through a sternotomy or a gross thoracotomy. In the sternotomy or gross thoracotomy, the surgeon typically uses a saw or other cutting instrument to cut the sternum longitudinally to allow two opposing halves of the anterior portion of the rib cage to be spread apart. The opening into the thoracic cavity is created so that the surgeon may directly visualize the heart and thoracic cavity. However, such methods suffer from numerous drawbacks. For example, the longitudinal incision in the sternum often results in bone bleeding, which is difficult to stop. The bone bleeding can produce a high degree of trauma, a larger risk of complications, an extended hospital stay, and a painful recovery period for the patient. Once the surgeon has accessed the thoracic cavity, the conventional method of harvesting the IMA involves the use of scalpels or conventional electrosurgical devices. A number of disadvantages inherent in conventional electrosurgical devices have been set forth hereinabove.

Thus, there is a need for an electrosurgical apparatus which can be used for the precise removal or modification of tissue at a specific location, wherein a target tissue or organ can be dissected, transected, incised, contracted, and/or coagulated, with minimal, or no, collateral tissue damage. The instant invention provides such an apparatus and related methods, wherein tissue may be ablated or otherwise modified by a first region or element of a blade active electrode, and the modified tissue can be further modified by a second region or element of the blade active electrode, and wherein the quantity and quality of the tissue modification can be accurately controlled.

SUMMARY OF THE INVENTION

The present invention generally provides systems, apparatus, and methods for selectively applying electrical energy to cut, incise, ablate, or otherwise modify a tissue or organ of a patient. In one aspect, the electrosurgical systems and methods of the invention are useful for harvesting and

dissecting veins and arteries of a patient, such as the saphenous vein or an IMA for use in a CABG procedure.

In one aspect, the present invention provides a method of creating an incision in a body structure. An electrosurgical probe is positioned adjacent the target tissue so that one or more active electrode(s) are brought into at least partial contact or close proximity with the target tissue. High frequency voltage is then applied between the active electrode(s) and one or more return electrode(s) and the active electrode(s) are moved, translated, reciprocated, or otherwise manipulated to cut through a portion of the tissue. In some embodiments, an electrically conductive fluid, e.g., isotonic saline or conductive gel, is delivered or applied to the target site to substantially surround the active electrode(s) with the fluid. In other embodiments, the active electrode(s) are immersed within the electrically conductive fluid. In both embodiments, the high frequency voltage may be selected to locally ablate or sever a target tissue, and/or to effect a controlled depth of hemostasis of severed blood vessels within the tissue. In another aspect, the electrosurgical systems and methods of the invention are useful for harvesting and dissecting veins and arteries of a patient, such as the saphenous vein or the IMA for use in a CABG procedure.

In one aspect, tissue is cut or otherwise modified by molecular dissociation or disintegration processes. (In contrast, in conventional electrosurgery tissue is cut by rapidly heating the tissue until cellular fluids explode, producing a cutting effect along the pathway of localized heating.) The present invention volumetrically removes the tissue along the cutting pathway in a cool ablation process that minimizes thermal damage to surrounding tissue. In these embodiments, the high frequency voltage applied to the active electrode(s) is sufficient to vaporize the electrically conductive fluid (e.g., gel or saline) between the active electrode(s) and the tissue. Within the vaporized fluid, a plasma is formed and charged particles (e.g., electrons) cause the molecular breakdown or disintegration of the tissue, perhaps to a depth of several cell layers. This molecular dissociation is accompanied by the volumetric removal of the tissue, e.g., along the incision of the tissue. This process can be precisely controlled to effect the volumetric removal of tissue as thin as 10 microns to 150 microns with minimal heating of, or damage to, surrounding or underlying tissue structures. A more complete description of this phenomenon is described in commonly assigned U.S. Pat. No. 5,683,366, the complete disclosure of which is incorporated herein by reference.

In a specific embodiment, the present invention provides a method of accessing a patient's thoracic cavity. The active electrode(s) are positioned in contact or in close proximity to a surface of the sternum. A high frequency voltage is applied between the active electrode(s) and a return electrode. The active electrodes are moved across the sternum to create an incision. In a specific configuration, the sides of the active electrode are slidingly engaged with the sternum as the incision is being made, so as to cause coagulation and hemostasis within the sternum.

In another exemplary embodiment, the present invention provides a method for harvesting the IMA from a patient. The electrosurgical probe is positioned adjacent the IMA and high frequency electrical energy is applied between one or more active electrode(s) and one or more return electrode(s). The probe is then moved so that the active electrode(s) volumetrically removes connective tissue adjacent to the IMA so that the IMA is free from connective tissue along a portion of its length. In an exemplary embodiment, the probe

5

is positioned adjacent to the IMA, and advanced along the length of the IMA while high frequency electrical energy is applied between the active electrode(s) and a return electrode to remove or cut the connective tissue or other structures surrounding the IMA. The residual heat from the electrical energy also provides simultaneous hemostasis of severed blood vessels, which increases visualization and improves recovery time for the patient. In addition, the ability to simultaneously cut through tissue on either side of the IMA decreases the length of the procedure, which further improves patient recovery time. After a suitable length of the IMA has been dissected, it may be transected, and anastomosed to a diseased coronary artery using known methods. In some embodiments, an electrically conductive fluid (liquid, gas, or gel) is placed at the target site adjacent to the IMA so as to provide a current flow path between the return electrode and the active electrode.

Apparatus according to the present invention generally include an electrosurgical instrument, such as a probe or catheter, having a shaft with proximal and distal ends, one or more active electrode(s) at the distal end and one or more connectors coupling the active electrode(s) to a source of high frequency electrical energy. The active electrode(s) are preferably designed for cutting tissue; i.e., they typically have a distal edge or point. In one embodiment, a plurality of active electrodes are aligned with each other to form a linear electrode array for cutting a path through the tissue. In another exemplary embodiment, the active electrode(s) include a sharp distal point to facilitate the cutting of the target tissue. In one specific configuration, the active electrode is a blade having a sharp distal point and sides. As the sharp distal point incises the tissue, the sides of the blade slidably contact the incised tissue. The electrical current flows through that portion of the tissue in the vicinity of the active electrode and/or the conductive fluid to the return electrode, such that the target tissue is first severed, and then the severed tissue is coagulated.

The apparatus can further include a fluid delivery element for delivering electrically conductive fluid to the active electrode(s) and the target site. The fluid delivery element may be located on the probe, e.g., a fluid lumen or tube, or it may be part of a separate instrument. Alternatively, an electrically conductive gel or spray, such as a saline electrolyte or other conductive gel, may be applied to the target site. In this embodiment, the apparatus may not have a fluid delivery element. In both embodiments, the electrically conductive fluid preferably provides a current flow path between the active electrode(s) and one or more return electrode(s). In an exemplary embodiment, the return electrode is located on the probe and spaced a sufficient distance from the active electrode(s) to substantially avoid or minimize current shorting therebetween and to shield the return electrode from tissue at the target site.

In a specific configuration, the electrosurgical probe includes an electrically insulating electrode support member having a tissue treatment surface at the distal end of the probe. One or more active electrode(s) are coupled to, or integral with, the electrode support member such that the active electrode(s) are spaced from the return electrode. In one embodiment, the probe includes a plurality of active electrode(s) having distal edges linearly aligned with each other to form a sharp cutting path for cutting tissue. The active electrodes are preferably electrically isolated from each other, and they extend about 0.2 mm to about 10 mm distally from the tissue treatment surface of the electrode support member. In this embodiment, the probe may further include one or more lumens for delivering electrically

6

conductive fluid to one or more openings around the tissue treatment surface of the electrode support member. In an exemplary embodiment, the lumen extends through a fluid tube exterior to the probe shaft that ends proximal to the return electrode.

In another aspect of the invention, the electrode support member comprises a plurality of wafer layers bonded together, e.g., by a glass adhesive or the like. The wafer layers each have conductive strips plated or printed thereon to form the active electrode(s) and the return electrode(s). In one embodiment, the proximal end of the wafer layers will have a number of holes extending from the conductor strips to an exposed surface of the wafer layers for connection to electrical conductor lead traces in the electrosurgical probe or handpiece. The wafer layers preferably comprise a ceramic material, such as alumina, and the electrode will preferably comprise a metallic material, such as gold, platinum, tungsten, palladium, silver or the like.

In another aspect of the invention, there is provided an electrosurgical probe having a blade-like active electrode affixed to an electrically insulating electrode support on the distal end of a shaft. In a specific configuration, the active electrode is in the form of a plasma blade comprising a substantially flat metal blade having at least one active edge and first and second blade sides. In one embodiment, the active electrode comprises a hook. The hook may include a curved portion. One or more portions of the hook may have a serrated edge. The return electrode is typically located on the shaft distal end proximal to the electrode support. In use, the active electrode and the return electrode are coupled to opposite poles of a high frequency power supply. The active edge may have a variety of shapes, and is adapted for generating high current densities thereon, and for precisely severing or ablating tissue or an organ in a highly controlled manner via molecular dissociation of tissue components. The first and second blade sides are adapted for engaging with tissue, such as tissue severed by the active edge, and for coagulating tissue engaged therewith.

The probe may be provided in various configurations, for example, according to a particular procedure to be performed. Thus, the electrode support may be arranged terminally or laterally on the probe, and the blade active electrode may be arranged terminally or laterally on the electrode support. The active electrode may be provided in various forms, such as a metal blade of unitary construction, e.g., a metal disc or portion thereof, a crosspiece supported by at least one electrode arm, or a hook. The shaft distal end may have a beveled end, a distal curve, and/or a laterally compressed region. Each of these features or elements of the probe may facilitate accessing a tissue or organ targeted for treatment or modification by the probe. In addition, the laterally compressed region may be adapted for accommodating the electrode support.

According to one aspect of the invention, there is provided a method for modifying a tissue using an electrosurgical probe having an active electrode in the form of a single blade which includes at least one active edge and first and second blade sides. The method involves positioning the probe such that the active electrode makes contact with, or is in close proximity to, a target tissue; and applying a high frequency voltage between the active and return electrodes sufficient to precisely sever or remove target tissue via molecular dissociation of tissue components adjacent to the active edge. The probe may be manipulated during the application of the high frequency voltage such that the active electrode is moved with respect to the target tissue. According to one aspect of the invention, the configuration of the

7

active electrode (e.g., a hook shaped electrode) is adapted for severing tissue as the probe distal end is drawn or pulled towards the operator of the probe. In this manner, the extent to which the tissue is severed can be precisely controlled. Thereafter, the severed tissue may be coagulated upon engagement of the tissue against the first and second blade sides of the active electrode.

In another aspect of the invention, there is provided a method of harvesting a tissue or organ using an electrosurgical probe having an active electrode in the form of a single blade, wherein the single blade electrode includes an active edge and first and second blade sides. In situations where the tissue to be harvested is concealed by an overlying tissue, the tissue to be harvested must first be accessed by incising or removing the overlying tissue. Removal of the overlying tissue may be performed in various ways, including: 1) mechanically, e.g. using a scalpel, rongeur, surgical saw or drill, etc. or a combination thereof; 2) via conventional electrosurgery, e.g., a Bovie; or 3) using an electrosurgical probe of the instant invention adapted for severing tissue in a cool ablation process. Once the tissue or organ to be harvested is accessible, the tissue or organ to be harvested may be dissected by juxtaposing the active edge of the active electrode against the surrounding connective tissue, and applying a high frequency voltage between the active and return electrodes sufficient to cause molecular dissociation of connective tissue components. In this way, the connective tissue adjacent to the active electrode is ablated at a temperature in the range of 40° C. to 70° C., with no, or minimal, thermal damage to the tissue to be harvested.

The electrosurgical probe of the invention is also applicable to a broad range of other procedures, including without limitation: cutting, resection, ablation, and/or hemostasis of tissues and organs such as prostate tissue, scar tissue, myocardial tissue, and tissues of the knee, shoulder, hip, and other joints; procedures of the head and neck, such as of the ear, mouth, throat, pharynx, larynx, esophagus, nasal cavity, and sinuses; as well as procedures involving skin tissue removal and/or collagen shrinkage in the epidermis or dermis. A more detailed account of various treatments and procedures which may be carried out according to the invention is set forth in enabling detail hereinbelow.

For a further understanding of the nature and advantages of the invention, reference should be made to the following description taken in conjunction with the accompanying drawings.

BRIEF DESCRIPTION OF THE DRAWINGS

FIG. 1 is a perspective view of an electrosurgical system incorporating a power supply and an electrosurgical probe for tissue ablation, resection, incision, contraction, vessel harvesting, and hemostasis, according to the present invention;

FIG. 2 is a side view of an electrosurgical probe according to the present invention;

FIG. 3 is an end view of the distal portion of the probe of FIG. 2;

FIG. 4 is a cross sectional view of the distal portion of the electrosurgical probe of FIG. 2;

FIG. 5 is an exploded view of a proximal portion of the electrosurgical probe;

FIG. 6 is an end view of an exemplary electrode support comprising a multi-layer wafer with plated conductors for electrodes;

FIGS. 7 and 8 are side views of the electrode support of FIG. 7;

8

FIGS. 9A–13 are side views of the individual wafer layers of the electrode support;

FIGS. 9B–12B are cross-sectional views of the individual wafer layers;

FIGS. 14 and 15 illustrate an alternative multi-layer wafer design according to the present invention;

FIG. 16 is a perspective view of an electrosurgical probe having an elongated, blade-like active electrode;

FIGS. 17A–17C are cross-sectional views of the distal portions of three different embodiments of an electrosurgical probe according to the present invention;

FIG. 18 illustrates an electrosurgical probe with a 90° distal bend and a lateral fluid lumen;

FIG. 19 illustrates an electrosurgical system with a separate fluid delivery instrument according to the present invention;

FIGS. 20A and 20B are cross-sectional and end views, respectively, of yet another electrosurgical probe incorporating flattened active electrodes;

FIG. 21 is a detailed end view of an electrosurgical probe having an elongate, linear array of active electrodes suitable for use in surgical cutting;

FIG. 22 is a detailed view of a single active electrode having a flattened end at its distal tip;

FIG. 23 is a detailed view of a single active electrode having a pointed end at its distal tip;

FIG. 24 is a perspective view of the distal portion of another electrosurgical probe according to the present invention;

FIG. 25 illustrates another embodiment of the probe of the present invention, specifically designed for creating incisions in external skin surfaces;

FIG. 26 is a perspective view of another embodiment of an electrosurgical probe for use in dermatology procedures;

FIGS. 27A–27C are exploded, isometric views of the probe of FIG. 26;

FIG. 28 is a cross-sectional view of another alternative electrosurgical probe;

FIG. 29 illustrates another embodiment of the electrosurgical probe of the present invention, incorporating additional active electrodes;

FIG. 30 is a perspective view of an electrosurgical probe having a blade electrode;

FIG. 31A is a perspective view, and FIG. 31B is a lateral view, of a blade electrode, according to one embodiment of the invention;

FIGS. 32A, 32B, and 32C are a side view, a plan view, and an end view, respectively, of an electrosurgical probe having a blade electrode;

FIGS. 33A and 33B are a side view and a plan view, respectively, of the distal end of an electrosurgical probe having a terminal blade electrode, according to one embodiment of the invention;

FIGS. 33C–33E each show a side view of the distal end of an electrosurgical probe having a terminal blade electrode, according to three different embodiments of the invention;

FIGS. 34A, 34B, and 34C are a side view, a plan view, and an end view, respectively, of an electrosurgical probe having a terminal electrode support and a lateral blade electrode, according to another embodiment of the invention;

FIGS. 35A, 35B, and 35C are a side view, a plan view, and an end view, respectively, of an electrosurgical probe having

a lateral electrode support and a lateral blade electrode, according to another embodiment of the invention;

FIGS. 36A and 36B each show a side view of the distal end of an electrosurgical probe having a blade electrode, according to two different embodiments of the invention;

FIGS. 37A, and 37B are a side view and an end view, respectively, of an electrosurgical probe having a lumen external to the probe shaft, according to one embodiment of the invention;

FIGS. 38A, and 38B are a side view and an end view, respectively, of an electrosurgical probe having an outer sheath surrounding the probe shaft, according to another embodiment of the invention;

FIGS. 39A, 39B, and 39C schematically represent a perspective view, a longitudinal sectional view, and an end view, respectively, of an electrosurgical probe, according to another embodiment of the invention;

FIG. 39D shows detail of the distal portion of the probe of FIGS. 39A-C;

FIGS. 40A and 40B schematically represent a longitudinal sectional view, and an end view, respectively, of an electrosurgical probe, according to another embodiment of the invention;

FIG. 40C shows detail of the distal portion of the probe of FIGS. 40A, 40B;

FIGS. 41A, 41B, and 41C each show detail of the distal portion of an electrosurgical probe, according to three different embodiments of the invention;

FIGS. 42A and 42B schematically represent a procedure for incising and coagulating tissue with an electrosurgical probe having a blade electrode, according to one embodiment of the invention;

FIG. 43A schematically represents a number of steps involved in a method of treating a patient with an electrosurgical probe having a blade electrode, according to one embodiment of the invention;

FIG. 43B schematically represents a number of steps involved in a method of concurrently severing and coagulating tissue, according to one embodiment of the invention; and

FIG. 44 schematically represents a number of steps involved in a method of dissecting a tissue or organ of a patient with an electrosurgical probe, according to another embodiment of the invention.

DESCRIPTION OF THE SPECIFIC EMBODIMENTS

The present invention provides systems and methods for selectively applying electrical energy to a target location within or on a patient's body, particularly for cutting, ablating, and/or coagulating a tissue using a blade-like electrode. The instant invention also provides apparatus and methods for making incisions to access a tissue or organ within a patient's body, to dissect or harvest the tissue or organ from the patient, and to transect or otherwise modify the tissue or organ. In one aspect, the invention provides apparatus and methods for dissecting and harvesting blood vessels from a patient.

The present invention is useful in procedures where the target tissue or organ is, or can be, flooded or submerged with an electrically conductive fluid, such as isotonic saline. In addition, tissues which may be treated by the system and method of the present invention further include, but are not limited to, tissues of the heart, chest, knee, shoulder, ankle,

hip, elbow, hand or foot; as well as prostate tissue, leiomyomas (fibroids) located within the uterus, gingival tissues and mucosal tissues located in the mouth, tumors, scar tissue, myocardial tissue, collagenous tissue within the eye; together with epidermal and dermal tissues on the surface of the skin. The present invention is also useful for resecting tissue within accessible sites of the body that are suitable for electrode loop resection, such as the resection of prostate tissue, leiomyomas (fibroids) located within the uterus, or other tissue to be removed from the body.

The present invention is also useful for procedures in the head and neck, such as the ear, mouth, throat, pharynx, larynx, esophagus, nasal cavity, and sinuses. These procedures may be performed through the mouth or nose using speculae or gags, or using endoscopic techniques, such as functional endoscopic sinus surgery (FESS). These procedures may include the removal of swollen tissue, chronically-diseased inflamed and hypertrophic mucus linings, polyps and/or neoplasms from the various anatomical sinuses of the skull, the turbinates and nasal passages, in the tonsil, adenoid, epi-glottic and supra-glottic regions, and salivary glands, submucosa resection of the nasal septum, excision of diseased tissue and the like. In other procedures, the present invention may be useful for cutting, resection, ablation and/or hemostasis of tissue in procedures for treating snoring and obstructive sleep apnea (e.g., UPPP procedures), for gross tissue removal, such as tonsillectomies, adenoidectomies, tracheal stenosis and vocal cord polyps and lesions, or for the resection or ablation of facial tumors or tumors within the mouth and pharynx, such as glossectomies, laryngectomies, acoustic neuroma procedures and nasal ablation procedures. In addition, the present invention is useful for procedures within the ear, such as stapedotomies, tympanotomies, myringotomies, or the like.

The present invention may also be useful for cosmetic and plastic surgery procedures in the head and neck. For example, the present invention is particularly useful for ablation and sculpting of cartilage tissue, such as the cartilage within the nose that is sculpted during rhinoplasty procedures. The present invention may also be employed for skin tissue removal and/or collagen shrinkage in the epidermis or dermis tissue in the head and neck region, e.g., the removal of pigmentations, vascular lesions, scars, tattoos, etc., and for other surgical procedures on the skin, such as tissue rejuvenation, cosmetic eye procedures (blepharoplasties), wrinkle removal, tightening muscles for facelifts or browlifts, hair removal and/or transplant procedures, etc.

The present invention is also useful for harvesting blood vessels, such as a blood vessel to be used as a graft vessel during the CABG procedure, e.g., the saphenous vein and the internal mammary artery (IMA). One or more embodiments of the invention may be used as follows: i) to access the blood vessel to be harvested, e.g., by opening the leg to access the saphenous vein, or opening the chest (either via a longitudinal incision of the sternum during an open-chest procedure, or during a minimally invasive inter-costal procedure); ii) to dissect the blood vessel to be harvested from the surrounding connective tissue along at least a portion of its length; and iii) to transect the dissected blood vessel at a first position only in the case of a pedicled graft (IMA), or at the first position and at a second position in the case of a free graft (saphenous vein). In each case i) to iii), as well as for other embodiment of the invention, the procedure involves removal of tissue by a cool ablation procedure in which a high frequency voltage is applied to an

active electrode in the vicinity of a target tissue, typically in the presence of an electrically conductive fluid. The cool ablation procedure of the invention is described fully elsewhere herein. The electrically conductive fluid may be a bodily fluid such as blood or synovial fluid, intracellular fluid of the target tissue, or isotonic saline delivered to the target tissue during the procedure. The present invention is also useful for coagulating blood or blood vessels, for example, to minimize bleeding in the sternum during an open-chest procedure.

Although certain parts of this disclosure are directed specifically to creating incisions for accessing a patient's thoracic cavity and the harvesting and dissection of blood vessels within the body during a CABG procedure, systems and methods of the invention are equally applicable to other procedures involving other organs or tissues of the body, including minimally invasive procedures (e.g., minimally invasive CABG procedures), other open procedures, intravascular procedures, urological procedures, laparoscopy, arthroscopy, thoracoscopy or other cardiac procedures, cosmetic surgery, orthopedics, gynecology, otorhinolaryngology, spinal and neurologic procedures, oncology, and the like.

In methods of the present invention, high frequency (RF) electrical energy is usually applied to one or more active electrodes in the presence of an electrically conductive fluid to remove and/or modify target tissue, an organ, or a body structure. Depending on the specific procedure, the present invention may be used to: (1) create incisions in tissue; (2) dissect or harvest tissue; (3) volumetrically remove tissue or cartilage (i.e., ablate or effect molecular dissociation of the tissue); (4) cut, transect, or resect tissue or an organ (e.g., a blood vessel); (5) create perforations or holes within tissue; and/or (6) coagulate blood and severed blood vessels.

In one method of the present invention, the tissue structures are incised by volumetrically removing or ablating tissue along a cutting path. In this procedure, a high frequency voltage difference is applied between one or more active electrode(s) and one or more return electrode(s) to develop high electric field intensities in the vicinity of the target tissue site. The high electric field intensities lead to electric field induced molecular breakdown of target tissue through molecular dissociation (rather than thermal evaporation or carbonization). Applicant believes that the tissue structure is volumetrically removed through molecular disintegration of larger organic molecules into smaller molecules and/or atoms, such as hydrogen, oxides of carbon, hydrocarbons and nitrogen compounds. This molecular disintegration completely removes the tissue structure, as opposed to dehydrating the tissue material by the removal of liquid within the cells of the tissue, as is typically the case with electrosurgical desiccation and vaporization.

The high electric field intensities may be generated by applying a high frequency voltage that is sufficient to vaporize an electrically conductive fluid over at least a portion of the active electrode(s) in the region between the tip of the active electrode(s) and the target tissue. The electrically conductive fluid may be a gas or liquid, such as isotonic saline, delivered to the target site, or a viscous fluid, such as a gel, that is located at the target site. In the latter embodiment, the active electrode(s) are submersed in the electrically conductive gel during the surgical procedure. Since the vapor layer or vaporized region has a relatively high electrical impedance, it minimizes the current flow into the electrically conductive fluid. Within the vaporized fluid a plasma is formed, and charged particles (e.g., electrons) cause the localized molecular dissociation or disintegration

of components of the target tissue, to a depth of perhaps several cell layers. This molecular dissociation results in the volumetric removal of tissue from the target site. This ablation process, which typically subjects the target tissue to a temperature in the range of 40° C. to 70° C., can be precisely controlled to effect the removal of tissue to a depth as little as about 10 microns, with little or no thermal or other damage to surrounding tissue. This cool ablation phenomenon has been termed Coblation®.

While not being bound by theory, applicant believes that the principle mechanism of tissue removal in the Coblation® mechanism of the present invention is energetic electrons or ions that have been energized in a plasma adjacent to the active electrode(s). When a liquid is heated sufficiently that atoms vaporize from the liquid at a greater rate than they recondense, a gas is formed. When the gas is heated sufficiently that the atoms collide with each other and electrons are removed from the atoms in the process, an ionized gas or plasma is formed. (A more complete description of plasmas (the so-called "fourth state of matter") can be found in Plasma Physics, by R. J. Goldston and P. H. Rutherford of the Plasma Physics Laboratory of Princeton University (1995), the complete disclosure of which is incorporated herein by reference.) When the density of the vapor layer (or within a bubble formed in the electrically conductive liquid) becomes sufficiently low (i.e., less than approximately 10^{20} atoms/cm³ for aqueous solutions), the electron mean free path increases to enable subsequently injected electrons to cause impact ionization within these regions of low density (i.e., vapor layers or bubbles). Once the ionic particles in the plasma layer have sufficient energy, they accelerate towards the target tissue. Energy evolved by the energetic electrons (e.g., 3.5 eV to 5 eV) can subsequently bombard a molecule and break its bonds, dissociating a molecule into free radicals, which then combine into final gaseous or liquid species.

Plasmas may be formed by heating and ionizing a gas by driving an electric current through it, or by transmitting radio waves into the gas. Generally, these methods of plasma formation give energy to free electrons in the plasma directly, and then electron-atom collisions liberate more electrons, and the process cascades until the desired degree of ionization is achieved. Often, the electrons carry the electrical current or absorb the radio waves and, therefore, are hotter than the ions. Thus, in applicant's invention, the electrons, which are carried away from the tissue towards the return electrode, carry most of the plasma's heat with them, allowing the ions to break apart the tissue molecules in a substantially non-thermal manner.

The energy evolved by the energetic electrons may be varied by adjusting a variety of factors, such as: the number of active electrodes; electrode size and spacing; electrode surface area; asperities and sharp edges on the electrode surfaces; electrode materials; applied voltage and power; current limiting means, such as inductors; electrical conductivity of the fluid in contact with the electrodes; density of the fluid; electrical insulators over the electrodes; and other factors. Accordingly, these factors can be manipulated to control the energy level of the excited electrons. Since different tissue structures have different molecular bonds, the present invention can be configured to break the molecular bonds of certain tissue, while having too low an energy to break the molecular bonds of other tissue. For example, fatty tissue, (e.g., adipose tissue) contains a large amount of lipid material having double bonds, the breakage of which requires an energy level substantially higher than 4 eV to 5 eV. Accordingly, the present invention can be configured

such that lipid components of adipose tissue are selectively not ablated. Of course, the present invention may be used to effectively ablate cells of adipose tissue such that the inner fat content of the cells is released in a liquid form. Alternatively, the invention can be configured (e.g., by increasing the voltage or changing the electrode configuration to increase the current density at the electrode tips) such that the double bonds of lipid materials are readily broken leading to molecular dissociation of lipids into low molecular weight condensable gases, generally as described hereinabove. A more complete description of the Coblation® phenomenon can be found in commonly assigned U.S. Pat. No. 5,683,366 and co-pending U.S. patent application Ser. No. 09/032,375, filed Feb. 27, 1998, the complete disclosures of which are incorporated herein by reference.

Methods of the present invention typically involve the application of high frequency (RF) electrical energy to one or more active electrodes in the presence of an electrically conductive fluid to remove (i.e., resect, incise, perforate, cut, or ablate) a target tissue, structure, or organ; and/or to seal transected vessels within the region of the target tissue. The present invention is particularly useful for sealing larger arterial vessels, e.g., having a diameter on the order of 1 mm or greater. In some embodiments, a high frequency power supply is provided having an ablation mode, wherein a first voltage is applied to an active electrode sufficient to effect molecular dissociation or disintegration of the tissue; and a coagulation mode, wherein a second, lower voltage is applied to an active electrode (either the same or a different electrode) sufficient to achieve hemostasis of severed vessels within the tissue. In other embodiments, an electrosurgical probe is provided having one or more coagulation electrode(s) configured for sealing a severed vessel, such as an arterial vessel, and one or more active electrodes configured for either contracting the collagen fibers within the tissue or removing (ablating) the tissue, e.g., by applying sufficient energy to the tissue to effect molecular dissociation. In the latter embodiments, the coagulation electrode(s) may be configured such that a single voltage can be applied to both coagulate with the coagulation electrode(s), and to ablate or contract tissue with the active electrode(s). In other embodiments, the power supply is combined with the coagulation probe such that the coagulation electrode is used when the power supply is in the coagulation mode (low voltage), and the active electrode(s) are used when the power supply is in the ablation mode (higher voltage).

In one method of the present invention, one or more active electrodes are brought into close proximity to tissue at a target site, and the power supply is activated in the ablation mode such that sufficient voltage is applied between the active electrodes and the return electrode to volumetrically remove the tissue through molecular dissociation, as described above. During this process, vessels within the tissue are severed. Smaller vessels may be automatically sealed with the system and method of the present invention. Larger vessels and those with a higher flow rate, such as arterial vessels, may not be automatically sealed in the ablation mode. In these cases, the severed vessels may be sealed by actuating a control (e.g., a foot pedal) to reduce the voltage of the power supply into the coagulation mode. In this mode, the active electrodes may be pressed against the severed vessel to provide sealing and/or coagulation of the vessel. Alternatively, a coagulation electrode located on the same or a different probe may be pressed against the severed vessel. Once the vessel is adequately sealed, the surgeon may activate a control (e.g., another foot pedal) to increase the voltage of the power supply back into the ablation mode.

The present invention is also useful for removing or ablating tissue around nerves, such as spinal, or cranial nerves, e.g., the hypoglossal nerve, the optic nerve, facial nerves, vestibulocochlear nerves and the like. This is particularly advantageous when removing tissue that is located close to nerves. One of the significant drawbacks with the conventional RF devices, scalpels, and lasers is that these devices do not differentiate between the target tissue and the surrounding nerves or bone. Therefore, the surgeon must be extremely careful during these procedures to avoid damage to the nerves within and around the target tissue. In the present invention, the Coblation® process for removing tissue results in no, or extremely small amounts, of collateral tissue damage, as described above. This allows the surgeon to remove tissue close to a nerve without causing collateral damage to the nerve fibers and surrounding tissue.

In addition to the generally precise nature of the novel mechanisms of the present invention, applicant has discovered an additional method of ensuring that adjacent nerves are not damaged during tissue removal. According to the present invention, systems and methods are provided for distinguishing between the fatty tissue immediately surrounding nerve fibers and the normal tissue that is to be removed during the procedure. Peripheral nerves usually comprise a connective tissue sheath, or epineurium, enclosing the bundles of nerve fibers, each bundle being surrounded by its own sheath of connective tissue (the perineurium) to protect these nerve fibers. The outer protective tissue sheath or epineurium typically comprises a fatty tissue (e.g., adipose tissue) having substantially different electrical properties than the normal target tissue that is treated. The system of the present invention measures the electrical properties of the tissue at the tip of the probe with one or more active electrode(s). These electrical properties may include electrical conductivity at one, several, or a range of frequencies (e.g., in the range from 1 kHz to 100 MHz), dielectric constant, capacitance or combinations of these. In this embodiment, an audible signal may be produced when the sensing electrode(s) at the tip of the probe detects the fatty tissue surrounding a nerve, or direct feedback control can be provided to only supply power to the active electrode(s) either individually or to the complete array of electrodes, if and when the tissue encountered at the tip or working end of the probe is normal tissue based on the measured electrical properties.

In one embodiment, the current limiting elements are configured such that the active electrodes will shut down or turn off when the electrical impedance reaches a threshold level. When this threshold level is set to the impedance of the fatty tissue surrounding nerves, the active electrodes will shut off whenever they come in contact with, or in close proximity to, nerves. Meanwhile, the other active electrodes, which are in contact with or in close proximity to target tissue, will continue to conduct electric current to the return electrode. This selective ablation or removal of lower impedance tissue in combination with the Coblation® mechanism of the present invention allows the surgeon to precisely remove tissue around nerves or bone. Applicant has found that the present invention is capable of volumetrically removing tissue closely adjacent to nerves without impairing the function of the nerves, and without significantly damaging the tissue of the epineurium.

The present invention can be also be configured to create an incision in a bone of the patient. For example, the systems of the present invention can be used to create an incision in the sternum for access to the thoracic cavity. Applicant has found that the Coblation® mechanism of the present inven-

15

tion allows the surgeon to precisely create an incision in the sternum while minimizing or preventing bone bleeding. The high frequency voltage is applied between the active electrode(s) and the return electrode(s) to volumetrically remove the bone from a specific site targeted for the incision. As the active electrode(s) are passed through the incision in the bone, the sides of the active electrodes (or a third coagulation electrode) slidingly contact the bone surrounding the incision to provide hemostasis in the bone. A more complete description of such coagulation electrodes can be found in U.S. patent application Ser. No. 09/162,117, filed Sep. 28, 1998, the complete disclosure of which is incorporated herein by reference.

The present invention can also be used to dissect and harvest blood vessels from the patient's body during a CABG procedure. The system of the present invention allows a surgeon to dissect and harvest blood vessels, such as the right or left IMA or saphenous vein, while concurrently providing hemostasis at the harvesting site. In some embodiments, a first high frequency voltage, can be delivered in an ablation mode to effect molecular disintegration of connective tissue adjacent to the blood vessel targeted for harvesting; and a second, lower voltage can be delivered to achieve hemostasis of the connective tissue adjacent to the blood vessel. In other embodiments, the targeted blood vessel can be transected at one or more positions along its length, and one or more coagulation electrode(s) can be used to seal the transected blood vessel at the site of transection. The coagulation electrode(s) may be configured such that a single voltage can be applied to the active electrodes to ablate the tissue and to coagulate the blood vessel and target site.

The present invention also provides systems, apparatus, and methods for selectively removing tumors or other undesirable body structures while minimizing the spread of viable cells from the tumor. Conventional techniques for removing such tumors generally result in the production of smoke in the surgical setting, termed an electrosurgical or laser plume, which can spread intact, viable bacterial or viral particles from the tumor or lesion to the surgical team, or viable cancerous cells to other locations within the patient's body. This potential spread of viable cells or particles has resulted in increased concerns over the proliferation of certain debilitating and fatal diseases, such as hepatitis, herpes, HIV and papillomavirus. In the present invention, high frequency voltage is applied between the active electrode(s) and one or more return electrode(s) to volumetrically remove at least a portion of the tissue cells in the tumor or lesion by the molecular dissociation of tissue components into non-condensable gases. The high frequency voltage is preferably selected to effect controlled removal of these tissue cells while minimizing substantial tissue necrosis to surrounding or underlying tissue. A more complete description of this phenomenon can be found in copending U.S. patent application Ser. No. 09/109,219, filed Jun. 30, 1998, the complete disclosure of which is incorporated herein by reference.

A current flow path between the active electrode(s) and the return electrode(s) may be generated by submerging the tissue site in an electrically conductive fluid (e.g., within a viscous fluid, such as an electrically conductive gel) or by directing an electrically conductive fluid along a fluid path to the target site (i.e., a liquid, such as isotonic saline, or a gas, such as argon). This latter method is particularly effective in a dry field procedure (i.e., the tissue is not submersed in fluid). The use of a conductive gel allows a slower, more controlled delivery rate of conductive fluid as compared

16

with a liquid or a gas. In addition, the viscous nature of the gel may allow the surgeon to more easily contain the gel around the target site (e.g., as compared with containment of isotonic saline). A more complete description of an exemplary method of directing electrically conductive fluid between the active and return electrodes is described in U.S. Pat. No. 5,697,281, the full disclosure of which is incorporated herein by reference. Alternatively, the body's natural conductive fluids, such as blood, may be sufficient to establish a conductive path between the return electrode(s) and the active electrode(s), and to provide the conditions for establishing a vapor layer, as described above. However, conductive fluid that is introduced into the patient is generally preferred over blood because blood will tend to coagulate at certain temperatures. Advantageously, a liquid electrically conductive fluid (e.g., isotonic saline) may be used to concurrently "bathe" the target tissue surface to provide an additional means for removing any tissue, and to cool the tissue at or adjacent to the target site.

In some embodiments of the invention, an electrosurgical probe includes an electrode support for electrically isolating the active electrode(s) from the return electrode, and a fluid delivery port or outlet for directing an electrically conductive fluid to the target site or to the distal end of the probe. The electrode support and the fluid outlet may be recessed from an outer surface of the instrument to confine the electrically conductive fluid to the region immediately surrounding the electrode support. In addition, a shaft of the instrument may be shaped so as to form a cavity around the electrode support and the fluid outlet. This helps to assure that the electrically conductive fluid will remain in contact with the active electrode(s) and the return electrode(s) to maintain the conductive path therebetween. In addition, this will help to maintain a vapor layer and subsequent plasma layer between the active electrode(s) and the tissue at the treatment site throughout the procedure, thereby reducing any thermal damage that might otherwise occur if the vapor layer were extinguished due to a lack of conductive fluid. Provision of the electrically conductive fluid around the target site also helps to maintain the tissue temperature at desired levels.

The electrically conductive fluid should have a threshold conductivity to provide a suitable conductive path between the return electrode and the active electrode(s). The electrical conductivity of the fluid (in units of milliSiemens per centimeter or mS/cm) will usually be greater than 0.2 mS/cm, preferably will be greater than 2 mS/cm and more preferably greater than 10 mS/cm. In an exemplary embodiment, the electrically conductive fluid is isotonic saline, which has a conductivity of about 17 mS/cm.

An electrosurgical probe or instrument of the invention typically includes a shaft having a proximal end and a distal end, and one or more active electrode(s) disposed at the shaft distal end. The shaft serves to mechanically support the active electrode(s) and permits the treating physician to manipulate the shaft distal end via a handle attached to the proximal end of the shaft. The shaft may be rigid or flexible, with flexible shafts optionally being combined with a generally rigid external tube for mechanical support. Flexible shafts may be combined with pull wires, shape memory actuators, and other known mechanisms for effecting selective deflection of the distal end of the shaft to facilitate positioning of the electrode array. The shaft will usually have one or more wires, electrode connectors, leads, or other conductive elements running axially therethrough, to permit connection of the electrode(s) to a connection block located at the proximal end of the instrument. The connection block

is adapted for coupling the electrode(s) to the power supply or controller. Typically, the connection block is housed within the handle of the probe.

The shaft of an instrument under the invention may assume various configurations. Generally, the shaft will have a suitable diameter and length to allow the surgeon to access the target site with the distal or working end of the shaft. Thus, the shaft may be provided in a range of sizes according to the particular procedure or tissue targeted for treatment. Typically, the shaft will have a length in the range of from about 5 cm to 30 cm, and have a diameter in the range of from about 0.5 mm to 10 mm. Specific shaft designs will be described in detail in connection with the drawings hereinafter.

The present invention may use a single active electrode or a plurality of electrodes distributed across a contact surface of a probe (e.g., in a linear fashion). In the latter embodiment, the electrode array usually includes a plurality of independently current-limited and/or power-controlled active electrodes to apply electrical energy selectively to the target tissue while limiting the unwanted application of electrical energy to the surrounding tissue and environment resulting from power dissipation into surrounding electrically conductive liquids, such as blood, normal saline, electrically conductive gel and the like. The active electrodes may be independently current-limited by isolating the terminals from each other and connecting each terminal to a separate power source that is isolated from the other active electrodes. Alternatively, the active electrodes may be connected to each other at either the proximal or distal ends of the probe to form a single wire that couples to a power source.

In one configuration, each individual active electrode is electrically insulated from all other active electrodes within the probe and is connected to a power source which is isolated from each of the other active electrodes in the array, or to circuitry which limits or interrupts current flow to the active electrode when low resistivity material causes a low impedance path between the return electrode and the individual active electrode. The isolated power sources for each individual active electrode may be separate power supply circuits having internal impedance characteristics which limit power to the associated active electrode when a low impedance return path is encountered. By way of example, the isolated power source may be a user selectable constant current source. In this embodiment, lower impedance paths will automatically result in lower resistive heating levels since the heating is proportional to the square of the operating current times the impedance. Alternatively, a single power source may be connected to each of the active electrodes through independently actuatable switches, or by independent current limiting elements, such as inductors, capacitors, resistors and/or combinations thereof. The current limiting elements may be provided in the probe, connectors, cable, power supply or along the conductive path from the power supply to the distal tip of the probe. Alternatively, the resistance and/or capacitance may occur on the surface of the active electrode(s) due to oxide layers which form selected active electrodes (e.g., titanium or a resistive coating on the surface of metal, such as platinum).

The distal end of the probe may comprise many independent active electrodes designed to deliver electrical energy in the vicinity of the distal end. The selective application of electrical energy to the conductive fluid is achieved by connecting each individual active electrode and the return electrode to a power source having independently controlled or current limited channels. The return electrode(s) may

comprise a single tubular member of electrically conductive material at the distal end of the probe proximal to the active electrode(s). The same tubular member of electrically conductive material may also serve as a conduit for the supply of the electrically conductive fluid between the active and return electrodes. The application of high frequency voltage between the return electrode(s) and the active electrode(s) results in the generation of high electric field intensities at the distal tip of the active electrode(s), with conduction of high frequency current from each active electrode to the return electrode. The current flow from each active electrode to the return electrode(s) is controlled by either active or passive means, or a combination thereof, to deliver electrical energy to the surrounding conductive fluid while minimizing energy delivery to surrounding (non-target) tissue.

The application of a suitable high frequency voltage between the return electrode(s) and the active electrode(s) for appropriate time intervals effects cutting, removing, ablating, shaping, contracting or otherwise modifying the target tissue. In one embodiment, the tissue volume over which energy is dissipated (i.e., over which a high current density exists) may be precisely controlled, for example, by the use of a multiplicity of small active electrodes whose effective diameters or principal dimensions range from about 5 mm to 0.01 mm, preferably from about 2 mm to 0.05 mm, and more preferably from about 1 mm to 0.1 mm. Electrode areas for both circular and non-circular terminals will have a contact area (per active electrode) below 25 mm², preferably being in the range from 0.0001 mm² to 1 mm², and more preferably from 0.005 mm² to 0.5 mm². The circumscribed area of the electrode array is in the range from 0.25 mm² to 75 mm², preferably from 0.5 mm² to 40 mm². In one embodiment the probe may include a plurality of relatively small active electrodes disposed over the distal contact surfaces on the shaft. The use of small diameter active electrodes increases the electric field intensity and reduces the extent or depth of tissue heating as a consequence of the divergence of current flux lines which emanate from the exposed surface of each active electrode.

The portion of the electrode support on which the active electrode(s) are mounted generally defines a tissue treatment surface of the probe. The area of the tissue treatment surface can vary widely, and the tissue treatment surface can assume a variety of geometries, with particular areas and geometries being selected for specific applications. The area of the tissue treatment surface can range from about 0.25 mm² to 75 mm², usually being from about 0.5 mm² to 40 mm². The geometries of the active electrode(s) can be planar, concave, convex, hemispherical, conical, a linear "in-line" array, or virtually any other regular or irregular shape. Most commonly, the active electrode(s) will be located at the shaft distal end of the electrosurgical probe, frequently having planar, disk-shaped, or hemispherical surfaces for use in reshaping procedures, ablating, cutting, dissecting organs, coagulating, or transecting blood vessels. The active electrode(s) may be arranged terminally or laterally on the electrosurgical probe (e.g., in the manner of a scalpel or a blade). However, it should be clearly understood that the active electrode of the invention does not cut or sever tissue mechanically as for a scalpel blade, but rather by the localized molecular dissociation of tissue components due to application of high frequency electric current to the active electrode. In one embodiment, a distal portion of the shaft may be flattened or compressed laterally (e.g., FIGS. 32A-32C). A probe having a laterally compressed shaft may facilitate access to certain target sites or body structures during various surgical procedures.

In embodiments having a plurality of active electrodes, it should be clearly understood that the invention is not limited to electrically isolated active electrodes. For example, a plurality of active electrodes may be connected to a single lead that extends through the probe shaft and is coupled to a high frequency power supply. Alternatively, the probe may incorporate a single electrode that extends directly through the probe shaft or is connected to a single lead that extends to the power source. The active electrode may have a planar or blade shape, a screwdriver or conical shape, a sharpened point, a ball shape (e.g., for tissue vaporization and desiccation), a twizzle shape (for vaporization and needle-like cutting), a spring shape (for rapid tissue debulking and desiccation), a twisted metal shape, an annular or solid tube shape, or the like. Alternatively, the electrode may comprise a plurality of filaments, a rigid or flexible brush electrode (for debulking a tumor, such as a fibroid, bladder tumor or a prostate adenoma), a side-effect brush electrode on a lateral surface of the shaft, a coiled electrode, or the like.

In one embodiment, the probe comprises a single blade active electrode that extends from an insulating support member, spacer, or electrode support, e.g., a ceramic or silicone rubber spacer located at the distal end of the probe. The insulating support member may be a tubular structure or a laterally compressed structure that separates the blade active electrode from a tubular or annular return electrode positioned proximal to the insulating member and the active electrode. The blade electrode may include a distal cutting edge and sides which are configured to coagulate the tissue as the blade electrode advances through the tissue. In yet another embodiment, the catheter or probe includes a single active electrode that can be rotated relative to the rest of the catheter body, or the entire catheter may be rotated relative to the electrode lead(s). The single active electrode can be positioned adjacent the abnormal tissue and energized and rotated as appropriate to remove or modify the target tissue.

The active electrode(s) are preferably supported within or by an insulating support member positioned near the distal end of the instrument shaft. The return electrode may be located on the instrument shaft, on another instrument, or on the external surface of the patient (i.e., a dispersive pad). For certain procedures, the close proximity of nerves and other sensitive tissue makes a bipolar design more preferable because this minimizes the current flow through non-target tissue and surrounding nerves. Accordingly, the return electrode is preferably either integrated with the instrument body, or located on another instrument. The proximal end of the probe typically includes the appropriate electrical connections for coupling the return electrode(s) and the active electrode(s) to a high frequency power supply, such as an electrosurgical generator.

One exemplary power supply of the present invention delivers a high frequency current selectable to generate average power levels ranging from several milliwatts to tens of watts per electrode, depending on the volume of target tissue being treated, and/or the maximum allowed temperature selected for the instrument tip. The power supply allows the user to select the voltage level according to the specific requirements of a particular otologic procedure, neurosurgery procedure, cardiac surgery, arthroscopic surgery, dermatological procedure, ophthalmic procedures, open surgery or other endoscopic surgery procedure. For cardiac procedures and potentially for neurosurgery, the power source may have an additional filter, for filtering leakage voltages at frequencies below 100 kHz, particularly voltages around 60 kHz. Alternatively, a power supply having a higher operating frequency, e.g., 300 kHz to 500 kHz may

be used in certain procedures in which stray low frequency currents may be problematic. A description of one suitable power supply can be found in co-pending patent applications Ser. Nos. 09/058,571 and 09/058,336, filed Apr. 10, 1998, the complete disclosure of both applications are incorporated herein by reference for all purposes.

The voltage difference applied between the return electrode(s) and the active electrode(s) will be at high or radio frequency, typically between about 5 kHz and 20 MHz, usually being between about 30 kHz and 2.5 MHz, preferably being between about 50 kHz and 500 kHz, often less than 350 kHz, and often between about 100 kHz and 200 kHz. The RMS (root mean square) voltage applied will usually be in the range from about 5 volts to 1000 volts, preferably being in the range from about 10 volts to 500 volts depending on the active electrode size, the operating frequency, and the operation mode of the particular procedure or desired effect on the tissue (e.g., contraction, coagulation, cutting or ablation). Typically, the peak-to-peak voltage for ablation or cutting will be in the range of 10 volts to 2000 volts and preferably in the range of 200 volts to 1800 volts, and more preferably in the range of about 300 volts to 1500 volts, often in the range of about 500 volts to 900 volts peak to peak (again, depending on the electrode size, the operating frequency and the operation mode). Lower peak-to-peak voltages will be used for tissue coagulation or collagen contraction and will typically be in the range from 50 to 1500, preferably 100 to 1000, and more preferably 120 to 600 volts peak-to-peak.

The voltage is usually delivered in a series of voltage pulses or alternating current of time varying voltage amplitude with a sufficiently high frequency (e.g., on the order of 5 kHz to 20 MHz) such that the voltage is effectively applied continuously (as compared with e.g., lasers claiming small depths of necrosis, which are generally pulsed about 10 Hz to 20 Hz). In addition, the duty cycle (i.e., cumulative time in any one-second interval that energy is applied) is on the order of about 50% for the present invention, as compared with pulsed lasers which typically have a duty cycle of about 0.0001%.

The power supply may include a fluid interlock for interrupting power to the active electrode(s) when there is insufficient conductive fluid around the active electrode(s). This ensures that the instrument will not be activated when conductive fluid is not present, minimizing the tissue damage that may otherwise occur. A more complete description of such a fluid interlock can be found in commonly assigned, copending U.S. application Ser. No. 09/058,336, filed Apr. 10, 1998, the complete disclosure of which is incorporated herein by reference.

The power supply may also be current limited or otherwise controlled so that undesired heating of the target tissue or surrounding (non-target) tissue does not occur. In a presently preferred embodiment of the present invention, current limiting inductors are placed in series with each independent active electrode, where the inductance of the inductor is in the range of 10 uH to 50,000 uH, depending on the electrical properties of the target tissue, the desired tissue heating rate and the operating frequency. Alternatively, capacitor-inductor (LC) circuit structures may be employed, as described previously in U.S. Pat. No. 5,697,909, the complete disclosure of which is incorporated herein by reference. Additionally, current limiting resistors may be selected. Preferably, these resistors will have a large positive temperature coefficient of resistance so that, as the current level begins to rise for any individual active electrode in contact with a low resistance medium (e.g., saline

irrigant or blood), the resistance of the current limiting resistor increases significantly, thereby minimizing the power delivery from the active electrode into the low resistance medium (e.g., saline irrigant or blood).

In some procedures, it may also be necessary to retrieve or aspirate the electrically conductive fluid and/or the non-condensable gaseous products of ablation. In addition, it may be desirable to aspirate small pieces of tissue or other body structures that are not completely disintegrated by the high frequency energy, or other fluids at the target site, such as blood, mucus, purulent fluid, the gaseous products of ablation, or the like. Accordingly, the system of the present invention may include one or more suction lumen(s) in the instrument, or on another instrument, coupled to a suitable vacuum source for aspirating fluids from the target site. In addition, the invention may include one or more aspiration electrode(s) coupled to the distal end of the suction lumen for ablating, or at least reducing the volume of, non-ablated tissue fragments that are aspirated into the lumen. The aspiration electrode(s) function mainly to inhibit clogging of the lumen that may otherwise occur as larger tissue fragments are drawn therein. The aspiration electrode(s) may be different from the ablation active electrode(s), or the same electrode(s) may serve both functions. A more complete description of instruments incorporating aspiration electrode(s) can be found in commonly assigned, co-pending patent application Ser. No. 09/010,382, filed Jan. 21, 1998, the complete disclosure of which is incorporated herein by reference.

During a surgical procedure, the distal end of the instrument and the active electrode(s) may be maintained at a small distance away from the target tissue surface. This small spacing allows for the continuous flow of electrically conductive fluid into the interface between the active electrode(s) and the target tissue surface. The continuous flow of the electrically conductive fluid helps to ensure that the thin vapor layer will remain between the active electrode(s) and the tissue surface. In addition, dynamic movement of the active electrode(s) over the tissue site allows the electrically conductive fluid to cool the tissue underlying and surrounding the target tissue to minimize thermal damage to this surrounding and underlying tissue. Accordingly, the electrically conductive fluid may be cooled to facilitate the cooling of the tissue. Typically, the active electrode(s) will be about 0.02 mm to 2 mm from the target tissue and preferably about 0.05 mm to 0.5 mm during the ablation process. One method of maintaining this space is to move, translate and/or rotate the probe transversely relative to the tissue, i.e., for the operator to use a light brushing motion, to maintain a thin vaporized layer or region between the active electrode and the tissue. Of course, if coagulation or collagen shrinkage of a deeper region of tissue is necessary (e.g., for sealing a bleeding vessel embedded within the tissue), it may be desirable to press the active electrode(s) against the tissue to effect joulean heating therein.

Referring to FIG. 1, an exemplary electrosurgical system 11 for cutting, ablating, resecting, or otherwise modifying tissue will now be described in detail. Electrosurgical system 11 generally comprises an electrosurgical handpiece or probe 10 connected to a power supply 28 for providing high frequency voltage to a target site, and a fluid source 21 for supplying electrically conductive fluid 50 to probe 10. In addition, electrosurgical system 11 may include an endoscope (not shown) with a fiber optic head light for viewing the surgical site. The endoscope may be integral with probe 10, or it may be part of a separate instrument. The system 11 may also include a vacuum source (not shown) for coupling

to a suction lumen or tube 211 (see FIG. 2) in the probe 10 for aspirating the target site.

As shown, probe 10 generally includes a proximal handle 19 and an elongate shaft 18 having one or more active electrodes 58 at its distal end. A connecting cable 34 has a connector 26 for electrically coupling the active electrodes 58 to power supply 28. In embodiments having a plurality of active electrodes, active electrodes 58 are electrically isolated from each other and the terminal of each active electrode 58 is connected to an active or passive control network within power supply 28 by means of a plurality of individually insulated conductors (not shown). A fluid supply tube 15 is connected to a fluid tube 14 of probe 10 for supplying electrically conductive fluid 50 to the target site.

Power supply 28 has an operator controllable voltage level adjustment 30 to change the applied voltage level, which is observable at a voltage level display 32. Power supply 28 also includes first, second, and third foot pedals 37, 38, 39 and a cable 36 which is removably coupled to power supply 28. The foot pedals 37, 38, 39 allow the surgeon to remotely adjust the energy level applied to active electrode(s) 58. In an exemplary embodiment, first foot pedal 37 is used to place the power supply into the "ablation" mode and second foot pedal 38 places power supply 28 into the "coagulation" mode. The third foot pedal 39 allows the user to adjust the voltage level within the ablation mode. In the ablation mode, a sufficient voltage is applied to the active electrodes to establish the requisite conditions for molecular dissociation of the tissue (i.e., vaporizing a portion of the electrically conductive fluid, ionizing the vapor layer and accelerating charged particles against the tissue). As discussed above, the requisite voltage level for ablation will vary depending on the number, size, shape and spacing of the electrodes, the distance in which the electrodes extend from the support member, etc. When the surgeon is using the power supply in the ablation mode, voltage level adjustment 30 or third foot pedal 39 may be used to adjust the voltage level to adjust the degree or aggressiveness of the ablation.

Of course, it will be recognized that the voltage and modality of the power supply may be controlled by other input devices. However, applicant has found that foot pedals are convenient means for controlling the power supply while manipulating the probe during a surgical procedure.

In the coagulation mode, the power supply 28 applies a low enough voltage to the active electrode(s) (or the coagulation electrode) to avoid vaporization of the electrically conductive fluid and subsequent molecular dissociation of the tissue. The surgeon may automatically toggle the power supply between the ablation and coagulation modes by alternately stepping on foot pedals 37, 38, respectively. This allows the surgeon to quickly move between coagulation and ablation in situ, without having to remove his/her concentration from the surgical field or without having to request an assistant to switch the power supply. By way of example, as the surgeon is sculpting soft tissue in the ablation mode, the probe typically will simultaneously seal and/or coagulation small severed vessels within the tissue. However, larger vessels, or vessels with high fluid pressures (e.g., arterial vessels) may not be sealed in the ablation mode. Accordingly, the surgeon can simply step on foot pedal 38, automatically lowering the voltage level below the threshold level for ablation, and apply sufficient pressure onto the severed vessel for a sufficient period of time to seal and/or coagulate the vessel. After this is completed, the surgeon may quickly move back into the ablation mode by stepping on foot pedal 37. A specific design of a suitable power supply for use with the present invention can be found

in Provisional Patent Application No. 60/062,997, filed Oct. 23, 1997, previously incorporated herein by reference.

FIG. 2 shows an electrosurgical probe 20 according to one embodiment of the invention. Probe 20 may be used in conjunction with a system similar or analogous to system 11 (FIG. 1). As shown in FIG. 2, probe 20 generally includes an elongated shaft 100 which may be flexible or rigid, a handle 204 coupled to the proximal end of shaft 100 and an electrode support member 102 coupled to the distal end of shaft 100. Shaft 100 may comprise a plastic material that is easily molded into the shape shown in FIG. 3, or shaft 100 may comprise an electrically conductive material, usually a metal, such as tungsten, stainless steel alloys, platinum or its alloys, titanium or its alloys, molybdenum or its alloys, and nickel or its alloys. In the latter case (i.e., shaft 100 is electrically conductive), probe 20 includes an electrically insulating jacket 108, which is typically formed as one or more electrically insulating sheaths or coatings, such as polytetrafluoroethylene, polyimide, and the like. The provision of electrically insulating jacket 108 over shaft 100 prevents direct electrical contact between the metal shaft and any adjacent body structure or the surgeon. Such direct electrical contact between a body structure (e.g., heart, bone, nerves, skin, or other blood vessels) and an exposed electrode could result in unwanted heating and necrosis of the structure at the point of contact.

Handle 204 typically comprises a plastic material that is easily molded into a suitable shape for handling by the surgeon. Handle 204 defines an inner cavity (not shown) that houses an electrical connections unit 250 (FIG. 5), and provides a suitable interface for coupling probe 20 to power supply 28 via an electrical connecting cable. Electrode support member 102 extends from the distal end of shaft 100 (usually about 1 mm to 20 mm), and provides support for an active electrode or a plurality of electrically isolated active electrodes 104. In the specific configuration shown in FIG. 2, probe 20 includes a plurality of active electrodes. As shown in FIG. 2, a fluid tube 233 extends through an opening in handle 204, and includes a connector 235 for connection to a fluid supply source for supplying electrically conductive fluid to the target site. Fluid tube 233 is coupled to a distal fluid tube 239 that extends along the outer surface of shaft 100 to an opening 237 at the distal end of the probe 20, as will be discussed in detail below. Of course, the invention is not limited to this configuration. For example, fluid tube 233 may extend through a single lumen (not shown) in shaft 100, it may be coupled to a plurality of lumens (also not shown) that extend through shaft 100 to a plurality of openings at its distal end, or the fluid tube may be completely independent of shaft 100. Probe 20 may also include a valve or equivalent structure for controlling the flow rate of the electrically conductive fluid to the target site.

As shown in FIGS. 3 and 4, electrode support member 102 has a substantially planar tissue treatment surface 212 and comprises a suitable insulating material (e.g., a ceramic or glass material, such as alumina, zirconia and the like) which could be formed at the time of manufacture in a flat, hemispherical or other shape according to the requirements of a particular procedure. The preferred support member material is alumina (Kyocera Industrial Ceramics Corporation, Elk Grove, Ill.), because of its high thermal conductivity, good electrically insulative properties, high flexural modulus, resistance to carbon tracking, biocompatibility, and high melting point. Electrode support member 102 is adhesively joined to a tubular support member (not shown) that extends most or all of the distance between support member 102 and the proximal end of probe

20. The tubular member preferably comprises an electrically insulating material, such as an epoxy or silicone-based material.

In a preferred construction technique, active electrodes 104 extend through pre-formed openings in the support member 102 so that they protrude above tissue treatment surface 212 by the desired distance. The electrodes are then bonded to the tissue treatment surface 212 of support member 102, typically by an inorganic sealing material. The sealing material is selected to provide effective electrical insulation, and good adhesion to both support member 102 and active electrodes 104. In one embodiment, active electrodes 104 comprise an electrically conducting, corrosion resistant metal, such as platinum or titanium. The sealing material additionally should have a compatible thermal expansion coefficient and a melting point well below that of platinum or titanium and alumina or zirconia, typically being a glass or glass ceramic.

In the embodiment shown in FIGS. 2-5, probe 20 includes a return electrode 112 for completing the current path between active electrodes 104 and a high frequency power supply 28 (see FIG. 1). As shown, return electrode 112 preferably comprises an annular conductive band coupled to the distal end of shaft 100 at a location proximal to tissue treatment surface 212 of electrode support member 102, typically about 0.5 mm to 10 mm proximal to surface 212, and more preferably about 1 mm to 10 mm proximal to surface 212. Return electrode 112 is coupled to a connector 258 that extends to the proximal end of probe 20, where it is suitably connected to power supply 28 (FIGS. 1 and 2).

As shown in FIG. 2, return electrode 112 is not directly connected to active electrodes 104. To complete this current path so that active electrodes 104 are electrically connected to return electrode 112, electrically conductive fluid (e.g., isotonic saline) is caused to flow therebetween. In the representative embodiment, the electrically conductive fluid is delivered through an external fluid tube 239 to opening 237, as described above (FIGS. 2 and 4). Alternatively, the fluid may be continuously delivered by a fluid delivery element (not shown) that is separate from probe 20.

In alternative embodiments, the fluid path may be formed in probe 20 by, for example, an inner lumen or an annular gap between the return electrode and a tubular support member within shaft 100 (not shown). This annular gap may be formed near the perimeter of the shaft 100 such that the electrically conductive fluid tends to flow radially inward towards the target site, or it may be formed towards the center of shaft 100 so that the fluid flows radially outward. In both of these embodiments, a fluid source (e.g., a bag of fluid elevated above the surgical site or having a pumping device), is coupled to probe 20 via a fluid supply tube (not shown) that may or may not have a controllable valve. A more complete description of an electrosurgical probe incorporating one or more fluid lumen(s) can be found in U.S. Pat. No. 5,697,281, filed on Jun. 7, 1995, the complete disclosure of which is incorporated herein by reference.

Referring to FIGS. 3 and 4, the electrically isolated active electrodes 104 are preferably spaced from each other and aligned to form a linear array 105 of electrodes for cutting a substantially linear incision in the tissue. The tissue treatment surface and individual active electrodes 104 will usually have dimensions within the ranges set forth above. Active electrodes 104 preferably have a distal edge 107 to increase the electric field intensities around terminals 104, and to facilitate cutting of tissue. Thus, active electrodes 104 have a screwdriver shape in the representative embodiment

of FIGS. 2-4. In one representative embodiment, the tissue treatment surface **212** has a circular cross-sectional shape with a diameter in the range of about 1 mm to 30 mm, usually about 2 mm to 20 mm. The individual active electrodes **104** preferably extend outward from tissue treatment surface **212** by a distance of about 0.1 mm to 8 mm, usually about 1 mm to 4 mm. Applicant has found that this configuration increases the high electric field intensities and associated current densities around active electrodes **104** to facilitate the ablation of tissue as described in detail above.

Probe **20** may include a suction or aspiration lumen **213** (see FIG. 2) within shaft **100** and a suction tube **211** (FIG. 2) for aspirating tissue, fluids and/or gases from the target site. In this embodiment, the electrically conductive fluid generally flows from opening **237** of fluid tube **239** radially inward and then back through one or more openings (not shown) in support member **102**. Aspirating the electrically conductive fluid during surgery allows the surgeon to see the target site, and it prevents the fluid from flowing into the patient's body (e.g., the thoracic cavity). This aspiration should be controlled, however, so that the conductive fluid maintains a conductive path between the active electrode(s) and the return electrode. In some embodiments, the probe **20** will also include one or more aspiration electrode(s) (not shown) coupled to the aspiration lumen for inhibiting clogging during aspiration of tissue fragments from the surgical site. A more complete description of these embodiments can be found in commonly assigned co-pending U.S. patent application Ser. No. 09/010,382, filed Jan. 21, 1998, the complete disclosure of which is incorporated herein by reference for all purposes.

FIG. 5 illustrates the electrical connections **250** within handle **204** for coupling active electrodes **104** and return electrode **112** to the power supply **28**. As shown, a plurality of wires **252** extend through shaft **100** to couple electrodes **104** to a plurality of pins **254**, which are plugged into a connector block **256** for coupling to a connecting cable **22** (FIG. 1). Similarly, return electrode **112** is coupled to connector block **256** via a wire **258** and a plug **260**.

According to the present invention, probe **20** further includes an identification element that is characteristic of the particular electrode assembly so that the same power supply **28** can be used for different electrosurgical operations. In one embodiment, for example, probe **20** includes a voltage reduction element or a voltage reduction circuit for reducing the voltage applied between the active electrodes **104** and the return electrode **112**. The voltage reduction element serves to reduce the voltage applied by the power supply so that the voltage between the active electrodes and the return electrode is low enough to avoid excessive power dissipation into the electrically conductive medium and/or the tissue at the target site. The voltage reduction element primarily allows the electrosurgical probe **10/20** to be compatible with a range of different power supplies that are adapted to apply higher voltages for ablation or vaporization of tissue (e.g., various power supplies or generators manufactured by ArthroCare Corporation, Sunnyvale, Calif.). For contraction of tissue, for example, the voltage reduction element will serve to reduce a voltage of about 100 to 135 volts RMS (which corresponds to a setting of 1 on the ArthroCare Model 970 and 980 (i.e., 2000) Generators) to about 45 to 60 volts RMS, which is a suitable voltage for contraction of tissue without ablation (e.g., molecular dissociation) of the tissue.

Again with reference to FIG. 5, in the representative embodiment the voltage reduction element is a dropping capacitor **262** which has a first leg **264** coupled to the return

electrode wire **258** and a second leg **266** coupled to connector block **256**. Of course, the capacitor may be located in other places within the system, such as in, or distributed along the length of, the cable, the power supply, the connector, etc. In addition, it will be recognized that other voltage reduction elements, such as diodes, transistors, inductors, resistors, capacitors or combinations thereof, may be used in conjunction with the present invention. For example, probe **20** may include a coded resistor (not shown) that is constructed to lower the voltage applied between return electrode **112** and active electrodes **104** to a suitable level for contraction of tissue. In addition, electrical circuits may be employed for this purpose.

Alternatively or additionally, the cable **22** that couples the power supply **28** to probe **10/20** may be used as a voltage reduction element. The cable has an inherent capacitance that can be used to reduce the power supply voltage if the cable is placed into the electrical circuit between the power supply, the active electrodes and the return electrode. In this embodiment, the cable **22** may be used alone, or in combination with one of the voltage reduction elements discussed above, e.g., a capacitor.

Further, it should be noted that various electrosurgical probes of the present invention can be used with a particular power supply that is adapted to apply a voltage within a selected range for a certain procedure or treatment. In which case, a voltage reduction element or circuitry may not be necessary nor desired.

With reference to FIGS. 6-8, electrode support member **70** according to one embodiment includes a multi-layer substrate comprising a suitable high temperature, electrically insulating material, such as ceramic. The multi-layer substrate is a thin or thick-film hybrid having conductive strips that are adhered to the ceramic wafer layers (e.g., thick-film printed and fired onto or plated onto the ceramic wafers). The conductive strips typically comprise tungsten, gold, nickel, silver, platinum or equivalent materials. In the exemplary embodiment, the conductive strips comprise tungsten, and they are co-fired together with the wafer layers to form an integral package. The conductive strips are coupled to external wire connectors by holes or vias that are drilled through the ceramic layers, and plated or otherwise covered with conductive material. A more complete description of such support members **370** can be found in U.S. patent application Ser. No. 08/977,845, filed Nov. 25, 1997, the entire disclosure of which is incorporated herein by reference.

In the representative embodiment, support member **70** comprises five ceramic layers **200**, **202**, **204**, **206**, **208** (see FIGS. 6-10), three gold plated active electrodes **210a**, **210b**, **210c** and first and second gold plated return electrodes **216**, **218**. As shown in FIGS. 9A and 9B, a first ceramic layer **200**, which is one of the outer layers of support **70**, includes first gold plated return electrode **216** on a lateral surface **220** of layer **200**. First ceramic layer **200** further includes a gold conductive strip **222** extending from return electrode **216** to the proximal end of layer **200** for coupling to a lead wire (not shown), and three gold conductive lines **224**, **226**, **228** extending from a mid-portion of layer **200** to its proximal end. Conductive strips **224**, **226**, **228** are each coupled to one of the active electrodes **210a**, **210b**, **210c** by conductive holes or vias **230**, **232**, **234**, respectively. As shown, all three vias **230**, **232**, **234** extend through wafer layer **200**.

Referring to FIGS. 10A and 10B, a second wafer layer **202** is bonded between first outer wafer layer **200** and a middle wafer layer **204** (See FIGS. 11A and 11B). As shown,

first active electrode **210a** is attached to the distal surface of second wafer layer **202**, and a conductive strip **240** extends to via **230** to couple active electrode **210a** to a lead wire. Similarly, wafer layers **204** and **206** (FIGS. 11A, 11B, 12A, and 12B) each have an active electrode **210b**, **210c** plated to their distal surfaces, and a conductive strip **242**, **244**, respectively, extending to one of the vias **232**, **234**, respectively. Note that the vias only extend as far as necessary through the ceramic layers. As shown in FIG. 13, a second outer wafer layer **208** has a second return electrode **218** plated to the lateral surface **250** of layer **208**. The second return electrode **218** is coupled directly to the first return electrode **216** through a via **252** extending through the entire ceramic substrate.

Of course, it will be recognized that a variety of different types of single layer and multi-layer wafers may be constructed according to the present invention. For example, FIGS. 14 and 15 illustrate an alternative embodiment of the multi-layer ceramic wafer, wherein the active electrodes comprise planar strips **280** that are plated or otherwise bonded between the ceramic wafer layers **282**. Each of the planar strips **280** has a different length, as shown in FIG. 15, so that the active electrodes can be electrically isolated from each other, and coupled to lead wires by vias (not shown).

FIG. 16 illustrates an electrosurgical probe **20'** according to another embodiment of the present invention. Probe **20'** generally includes handle **104** attached to shaft **100**, and has a single, thin, elongated active blade electrode **58**. Active electrode **58** is mechanically and electrically separated from return electrode **112** by a support structure **102**. The active blade electrode **58** has a sharp distal edge **59** which helps facilitate the cutting process, and sides **62** which contact the tissue (e.g., bone) as the blade electrode **58** passes through the tissue or body structure. By contacting the sides of the blade electrode **58** directly with the tissue or body structure, the electrical power supplied to electrode **58** by power supply **28** can provide hemostasis to the body structure during the cutting process. Optionally, probe **20'** can further include one or more coagulation electrode(s) (not shown) configured to seal a severed vessel, bone, or other tissue that is being incised. Such coagulation electrode(s) may be configured such that a single voltage can be applied to coagulate with the coagulation electrode(s) while ablating tissue with the active electrode(s). According to one aspect of the invention, probe **20'** is particularly useful for creating an incision in a patient's chest. For example, in an open-chest CABG procedure a median sternotomy is first performed in which the sternum is sectioned longitudinally so as to allow the chest to be opened for access to the thoracic cavity. Active electrodes **58** include distal edge **59** suitable for sectioning the sternum, and sides **62** suitable for arresting bone bleeding within the incised sternum. Sides **62** are configured to slidably engage the sternum as active electrode **58** is moved with respect to the sternum. Return electrode **112** is spaced proximally from active electrode **58** such that the electrical current is drawn away from the surrounding tissue. Alternatively, the return electrode **112** may be a dispersive pad located on the external surface of the patient's body. By minimizing bleeding of the sternum during an open-chest procedure, the patient's recovery time can be substantially shortened and patient suffering is alleviated.

FIGS. 17A–17C schematically illustrate the distal portion of three different embodiments of a probe **90** according to the present invention. As shown in FIG. 17A, active electrodes **104** are anchored in a support **102** of suitable insulating material (e.g., ceramic or glass material, such as

alumina, zirconia and the like) which could be formed at the time of manufacture in a flat, hemispherical or other shape according to the requirements of a particular procedure. In one embodiment, the support material is alumina, available from Kyocera Industrial Ceramics Corporation, Elk Grove, Ill., because of its high thermal conductivity, good electrically insulative properties, high flexural modulus, resistance to carbon tracking, biocompatibility, and high melting point. The support **102** is adhesively joined to a tubular support member **78** that extends most or all of the distance between matrix **102** and the proximal end of probe **90**. Tubular member **78** preferably comprises an electrically insulating material, such as an epoxy or silicone-based material.

According to one construction technique, active electrodes **104** extend through pre-formed openings in the support **102** so that they protrude above tissue treatment surface **212** by the desired distance. The electrodes are then bonded to the tissue treatment surface **212** of support **102**, typically by an inorganic sealing material **80**. Sealing material **80** is selected to provide effective electrical insulation, and good adhesion to both the support **102** and the platinum or titanium active electrodes. Sealing material **80** additionally should have a compatible thermal expansion coefficient, and a melting point well below that of platinum or titanium and alumina or zirconia, typically being a glass or glass ceramic.

In the embodiment shown in FIG. 17A, return electrode **112** comprises an annular member positioned around the exterior of shaft **100** of probe **90**. Return electrode **112** may fully or partially circumscribe tubular member **78** to form an annular gap **54** therebetween for flow of electrically conductive liquid **50** therethrough, as discussed below. Gap **54** preferably has a width in the range of 0.25 mm to 4 mm. Alternatively, probe **90** may include a plurality of longitudinal ribs between tubular member **78** and return electrode **112** to form a plurality of fluid lumens extending along the perimeter of shaft **100**. In this embodiment, the plurality of lumens will extend to a plurality of openings.

Return electrode **112** is disposed within an electrically insulative jacket **17**, which is typically formed as one or more electrically insulative sheaths or coatings, such as polytetrafluoroethylene, polyimide, and the like. The provision of the electrically insulative jacket **17** over return electrode **112** prevents direct electrical contact between return electrode **112** and any adjacent body structure. Such direct electrical contact between a body structure (e.g., the heart) and an exposed electrode member **112** could result in unwanted heating and necrosis of the structure at the point of contact.

As shown in FIG. 17A, return electrode **112** is not directly connected to active electrodes **104**. To complete a current path so that active electrodes **104** are electrically connected to return electrode **112**, electrically conductive liquid **50** (e.g., isotonic saline) is caused to flow along fluid path(s) **83**. Fluid path **83** is formed by annular gap **54** between outer return electrode **112** and tubular support member **78**. The electrically conductive liquid **50** flowing through fluid path **83** provides a pathway for electrical current flow between active electrodes **104** and return electrode **112**, as illustrated by the current flux lines **60** in FIG. 17A. When a voltage difference is applied between active electrodes **104** and return electrode **112**, high electric field intensities will be generated at the distal tips of active electrodes **104** with current flow from electrodes **104** through the target tissue to the return electrode, the high electric field intensities causing ablation of tissue **52** in zone **88**.

FIG. 17B illustrates another alternative embodiment of electrosurgical probe **90** which has a return electrode **112**

positioned within tubular member 78. Return electrode 112 may comprise a tubular member defining an inner lumen 57 for allowing electrically conductive liquid 50 (e.g., isotonic saline) to flow therethrough in electrical contact with return electrode 112. In this embodiment, a voltage difference is applied between active electrodes 104 and return electrode 112 resulting in electrical current flow through the electrically conductive liquid 50 as shown by current flux lines 60. As a result of the applied voltage difference and concomitant high electric field intensities at the tips of active electrodes 104, tissue 52 becomes ablated or transected in zone 88.

FIG. 17C illustrates another embodiment of probe 90 that is a combination of the embodiments in FIGS. 17A and 17B. As shown, this probe includes both an inner lumen 57 and an outer gap or plurality of outer lumens 54 for flow of electrically conductive fluid. In this embodiment, the return electrode 112 may be positioned within tubular member 78 as in FIG. 17B, outside of tubular member 78 as in FIG. 17A, or in both locations.

FIG. 18 illustrates another embodiment of probe 90 where the distal portion of shaft 100 is bent so that active electrodes extend transversely to the shaft. Preferably, the distal portion of shaft 100 is perpendicular to the rest of the shaft so that tissue treatment surface 212 is generally parallel to the shaft axis. In this embodiment, return electrode 112 is mounted to the outer surface of shaft 100 and is covered with an electrically insulating jacket 17. The electrically conductive fluid 50 flows along flow path 83 through return electrode 112 and exits the distal end of electrode 112 at a point proximal of tissue treatment surface 212. The fluid is directed exterior of shaft to surface 212 to create a return current path from active electrodes 104, through the fluid 50, to return electrode 112, as shown by current flux lines 60.

FIG. 19 illustrates another embodiment of the invention where electrosurgical system 11 further includes a liquid supply instrument 64 for supplying electrically conductive fluid 50 between active electrodes 104 and a return electrode 112'. Liquid supply instrument 64 comprises an inner tubular member or return electrode 112' surrounded by an electrically insulating jacket 17. Return electrode 112' defines an inner passage 83 for flow of fluid 50. As shown in FIG. 19, the distal portion of instrument 64 is preferably bent so that liquid 50 is discharged at an angle with respect to instrument 64. This allows the surgical team to position liquid supply instrument 64 adjacent tissue treatment surface 212 with the proximal portion of supply instrument 64 oriented at a similar angle to probe 90.

The present invention is not limited to an electrode array disposed on a relatively planar surface at the distal tip of probe 90, as described above. Referring to FIGS. 20A and 20B, an alternative probe 90 includes a pair of electrodes 105a, 105b mounted to the distal end of shaft 100. Electrodes 105a, 105b are electrically connected to a power supply, as described above, and preferably have tips 107a, 107b having a screwdriver shape. The screwdriver shape provides a greater amount of "edges" to electrodes 105a, 105b, to increase the electric field intensity and current density at tips 107a, 107b, thereby improving the cutting ability as well as the ability to provide hemostasis of the incised tissue.

FIG. 21 illustrates yet another embodiment designed for cutting of body tissue, organs, or structures. In this embodiment, the active electrodes 104 are arranged in a linear or columnar array of one or more closely spaced columns so that as the electrodes 104 are moved along the longer axis (denoted by arrow 160 in FIG. 21), the current

flux lines are narrowly confined at the tip of the active electrodes 104 and result in a cutting effect in the body structure being treated. As before, the current flux lines 60 emanating from the active electrodes 104 pass through the electrically conductive liquid to the return electrode structure 112 located proximal to the probe tip.

Referring now to FIGS. 22 and 23, alternative geometries are shown for the active electrodes 104. These alternative electrode geometries allow the electrical current densities emanating from the active electrodes 104 to be concentrated to achieve an increased ablation rate and/or a more concentrated ablation effect due to the fact that sharper edges (i.e., regions of smaller radii of curvature) result in higher current densities. FIG. 22 illustrates a flattened extension of a round wire active electrode 104 which results in higher current densities at the edges 180. Another example is shown in FIG. 23 in which the active electrode 104 is formed into a cone shaped point 182 resulting in higher current densities at the tip of the cone.

Another embodiment of the electrosurgical probe is illustrated in FIG. 24. The electrosurgical probe 90 comprises a shaft 100 and at least two active electrodes 104 extending from a support 102 at the distal end of the shaft. The active electrodes 104 preferably define a distal edge 600 for making an incision in tissue. The edges 600 of the active electrodes 104 are substantially parallel with each other and usually spaced a distance of about 4 mm to 15 mm apart, preferably about 8 mm to 10 mm apart. The edges 600 extend from the distal end of support 102 by a distance of about 0.5 mm to 10 mm, preferably about 2 mm to 5 mm. In the exemplary embodiment, probe 90 will include a return electrode 112 spaced proximally from the active electrodes 104. In an alternative embodiment (not shown), one of the active electrodes 104 may function as a return electrode, or the return electrode may be a dispersive pad located on an external surface of the patient's body.

FIG. 25 illustrates a distal portion of an electrosurgical probe 500 according to another embodiment of the present invention. The embodiment of FIG. 25 is particularly useful for cutting or creating incisions in tissue structures. Probe 500 comprises a support member 502 coupled to a shaft or disposable tip (not shown) as described in previous embodiments. Support member 502 preferably comprises an inorganic electrically insulating material, such as ceramic, glass or glass-ceramic. In this embodiment, however, support member 502 may comprise an organic material, such as plastic, because the active electrode 506 and return electrode 508 are both spaced away from support member 502. Thus, the high intensity electric fields may be far enough away from support member 502 so as to allow an organic material.

An electrode assembly 504 extends from a distal end of support member 502, preferably by a distance of about 2 mm to 20 mm. Electrode assembly 504 comprises a single, active electrode 506 and a return electrode sleeve 508 spaced proximally from active electrode 506 by an insulation member 510, which preferably comprises an inorganic material, such as ceramic, glass or glass-ceramic. As shown, active electrode 506 preferably tapers to a sharp distal end 512 to facilitate the cutting or incising of tissue. In the exemplary embodiment, active electrode 506 has a proximal diameter of about 0.2 to 20 mm and a distal diameter of less than about 0.2 mm. Return electrode 508 is spaced from active electrode 506 a sufficient distance to prevent shorting or arcing therebetween at sufficient voltages to allow the volumetric removal of tissue. In the representative embodiment, the distal exposed portion of return electrode 508 is spaced about 0.5 to about 5 mm from the proximal exposed portion

31

of active electrode 506. Of course, it will be recognized that the present invention is not limited to the particular dimensions and configuration of the electrode assembly 504 described herein, and a variety of different configurations may be envisioned depending on the surgical application.

As shown, probe 500 includes a fluid lumen 520 passing through support member 502 to a distal opening (not shown) at the distal end of support member 502. Fluid lumen 520 is coupled to a supply of electrically conductive fluid, such as isotonic saline, or other suitable conductive fluid for delivery of such fluid to the target site. In the exemplary embodiment, probe 500 is designed such that lumen 520 will be positioned above electrode assembly 504 during use such that the conductive fluid exiting the distal opening of lumen 520 will naturally pass over return electrode 508 and active electrode 506 thereby creating a current path therebetween. In addition, the conductive fluid will be sufficient to cover the active electrode 506 such that the conditions for plasma formation can be met, as described in detail above.

FIGS. 26, and 27A–C illustrate another exemplary electrosurgical probe 310 for cutting, incising, or removing tissue structures. Probe 310 comprises a shaft or disposable tip 313 removably coupled to a proximal handle 312, and an electrically insulating electrode support member 370 extending from tip 313 for supporting a plurality of active electrodes 358. Tip 313 and handle 312 typically comprise a plastic material that is easily molded into a suitable shape for handling by the surgeon. As shown in FIGS. 27A and 27B, handle 312 defines an inner cavity 372 that houses the electrical connections 374, and provides a suitable interface for connection to electrical connecting cable 34 (see FIG. 1). In the exemplary embodiment, handle 312 is constructed of a steam autoclavable plastic or metal (e.g., polyethylether ketone), or a stable metal alloy containing aluminum and/or zinc) so that it can be re-used by sterilizing handle 312 between surgical procedures. High service temperature materials are preferred, such as a silicone cable jacket and a poly-ether-imide handpiece or ULTEM® that can withstand repeated exposure to high temperatures.

Referring to FIGS. 27A–27C, tip 313 preferably comprises first and second housing halves 500, 502 that snap fit together, and form a recess 404 therebetween for holding electrode support member 370 within the tip 313. Electrode support member 370 extends from the distal end of tip 313, usually by about 0.5 mm to 20 mm, and provides support for a plurality of electrically isolated active electrodes 358 and one or more return electrodes 400. Alternatively, electrode support member 370 may be recessed from the distal end of tip 313 to help confine the electrically conductive fluid around the active electrodes 358 during the surgical procedure, as discussed above. Electrode support member 370 has a substantially planar tissue treatment surface 380 that is usually disposed at an angle of about 10 to 90 degrees relative to the longitudinal axis of handle 312 to facilitate handling by the surgeon. In the exemplary embodiment, this function is accomplished by orienting tip 313 at an acute angle relative to the longitudinal axis of handle 312.

In the embodiment shown in FIGS. 26–27C, probe 310 includes a single annular return electrode 400 for completing the current path between active electrodes 358 and power supply 28 (see FIG. 1). As shown, return electrode 400 preferably has a fluid contact surface slightly proximal to tissue treatment surface 380, typically by about 0.1 mm to 2 mm, and preferably by about 0.2 mm to 1 mm. Return electrode 400 is coupled to a connector 404 that extends to the proximal end of handle 313, where it is suitably connected to power supply 28 (FIG. 1).

32

Referring again to FIGS. 27A–27C, tip 313 further includes a proximal hub 506 for supporting a male electrical connector 508 that holds a plurality of wires 510 each coupled to one of the active electrodes 358 or to return electrode 400 on support member 370. A female connector 520 housed within handle 312 is removably coupled to male connector 508, and a plurality of wires 522 extend from female connector 520 through a strain relief 524 to cable 334. Both sets of wires 510, 522 are insulated to prevent shorting in the event of fluid ingress into the probe 310. This design allows for removable connection of the electrodes in tip 313 with the connector 520 within handle 312 so that the handle can be re-used with different tips 313. Probe 310 will preferably also include an identification element, such as a coded resistor (not shown), for programming a particular voltage output range and mode of operation for the power supply. This allows the power supply to be employed with a variety of different probes for a variety of different applications.

In the representative embodiment, probe 310 includes a fluid tube 410 (FIG. 26) for delivering electrically conductive fluid to the target site. Fluid tube 410 is sized to extend through a groove 414 in handle 313 and through an inner cavity 412 in tip 312 to a distal opening 414 (FIG. 26) located adjacent electrode support member 370. Tube 410 extends all the way through inner cavity 412 to opening 414 to eliminate any possible fluid ingress into cavity 412. Fluid tube 410 includes a proximal connector for coupling to an electrically conductive fluid source 321.

Probe 310 will also include a valve or equivalent structure for controlling the flow rate of the electrically conductive fluid to the target site. In the representative embodiment shown in FIGS. 27A–27C, handle 312 comprises a main body 422 coupled between distal hub 418 and strain relief 420, and a rotatable sleeve 416 around main body 422. Distal hub 418 has an opening 419 for receiving proximal hub 506 of tip 313 for removably coupling the tip 313 to the handle 312. Sleeve 416 is rotatably coupled to strain relief 420 and distal hub 418 to provide a valve structure for fluid tube 410. As shown in FIG. 27A, fluid tube 410 extends through groove 414 from strain relief 420, through main body 422 and distal hub 420 to tip 313. Rotation of sleeve 416 will impede, and eventually obstruct, the flow of fluid through tube 410. Of course, this fluid control may be provided by a variety of other input and valve devices, such as switches, buttons, etc.

In alternative embodiments, the fluid path may be directly formed in probe 310 by, for example, a central inner lumen or an annular gap (not shown) within the handle and the tip. This inner lumen may be formed near the perimeter of the probe 310 such that the electrically conductive fluid tends to flow radially inward towards the target site, or it may be formed towards the center of probe 310 so that the fluid flows radially outward. In addition, the electrically conductive fluid may be delivered from a fluid delivery element (not shown) that is separate from probe 310. In arthroscopic surgery, for example, the body cavity will be flooded with isotonic saline and the probe 310 will be introduced into this flooded cavity. Electrically conductive fluid will be continually resupplied to maintain the conduction path between return electrode 400 and active electrodes 358. A more complete description of alternative electrosurgical probes incorporating one or more fluid lumen(s) can be found in commonly assigned, co-pending application Ser. No. 08/485,219, filed on Jun. 7, 1995, the complete disclosure of which is incorporated herein by reference.

Referring now to FIG. 26, electrically isolated active electrodes 358 are spaced apart over tissue treatment surface

380 of electrode support member 370, preferably in a linear array. In the representative embodiment, three active electrodes 358, each having a substantially conical shape, are arranged in a linear array extending distally from surface 380. Active electrodes 358 will usually extend a distance of about 0.5 mm to 20 mm from tissue treatment surface 380, preferably about 1 mm to 5 mm. Applicant has found that this configuration increases the electric field intensities and associated current densities at the distal edges of active electrodes 358, which increases the rate of tissue cutting. In the representative embodiment, the tissue treatment surface 380 has a circular cross-sectional shape with a diameter in the range of about 0.5 mm to 20 mm (preferably about 2 mm to 10 mm). The individual active electrodes 358 preferably taper outward as shown, or they may form a distal edge, such as the electrodes shown in FIGS. 3 and 24.

Probe 430 of FIG. 28 includes a shaft 432 coupled to a proximal handle 434 for holding and controlling shaft 432. Probe 430 includes an active electrode array 436 at the distal tip of shaft 432, an annular return electrode 438 extending through shaft 432 and proximally recessed from the active electrode array 436, and an annular lumen 442 between return electrode 438 and an outer insulating sheath 446. Probe 430 further includes a liquid supply conduit 444 attached to handle 434 and in fluid communication with lumen 442, and a source of electrically conductive fluid (not shown) for delivering the fluid past return electrode 438 to the target site on the tissue 440. Electrode array 436 is preferably flush with the distal end of shaft 432 or distally extended from the distal end by a small distance (on the order of 0.005 inches) so as to minimize the depth of ablation. Preferably, the distal end of shaft 432 is beveled to improve access and control of probe 430 while treating the target tissue.

Yet another embodiment of the present invention is shown in FIG. 29. Auxiliary active electrodes 458, 459 are positioned at the distal tip 70 of the probe. Auxiliary active electrodes 458, 459 may be the same size as ablation active electrodes 58, or larger as shown in FIG. 29. One operating arrangement is to connect auxiliary active electrodes 458, 459 to two poles of a high frequency power supply to form a bipolar circuit allowing current to flow between the terminals of auxiliary active electrodes 458, 459 as shown by current flux lines 460. Auxiliary active electrodes 458, 459 are electrically isolated from ablation electrodes 58. By proper selection of the inter-electrode spacing, W_2 , and electrode width, W_3 , and the frequency of the applied voltage, the current flux lines 460 can be caused to flow below the target layer as described above.

The voltage will preferably be sufficient to establish high electric field intensities between the active electrode array 436 and the target tissue 440 to thereby induce molecular breakdown or disintegration of several cell layers of the target tissue. As described above, a sufficient voltage will be applied to develop a thin layer of vapor within the electrically conductive fluid and to ionize the vaporized layer or region between the active electrode(s) and the target tissue. Energy in the form of charged particles are discharged from the vapor layer to ablate the target tissue, thereby minimizing necrosis of surrounding tissue and underlying cell layers.

With reference to FIG. 30, there is shown in perspective view an electrosurgical probe 700, according to another embodiment of the invention. Probe 700 includes a shaft 702 having a shaft distal end portion 702a and a shaft proximal end portion 702b. Shaft 702 is affixed at its proximal end 702b to a handle 704. Shaft 702 typically comprises an electrically conductive material, usually a metal, such as

tungsten, stainless steel, platinum or its alloys, titanium or its alloys, molybdenum or its alloys, nickel or its alloys. An electrically insulating electrode support 710 is disposed at shaft distal end 702a. An active electrode 712 is disposed on electrode support 710. Active electrode 712 comprises a blade electrode (e.g., FIGS. 31A, 31B). An electrically insulating sleeve 716 covers a portion of shaft 702, and terminates at sleeve distal end 716a to define an exposed portion of shaft 702 extending between electrode support proximal end 710b and sleeve distal end 716a. This exposed portion of shaft 702 defines a return electrode 718 on shaft distal end portion 702a. (In an alternative embodiment, the return electrode may take the form of an annular band of an electrically conductive material, e.g., a platinum alloy, disposed on the exterior of the shaft distal end.) A cavity within handle 704 accommodates a connection block 706 which is connected to active electrode 712 and return electrode 718 via electrode leads (not shown). Connection block 706 provides a convenient mechanism for coupling active electrode 712 and return electrode 718 to opposite poles of a power supply (e.g., power supply 28, FIG. 1).

FIG. 31A is a perspective view of an active electrode 712 of probe 700, according to one embodiment of the invention. Active electrode 712 is in the form of a single blade electrode which extends from electrode support 710 to a distance, H_b . The distance H_b may vary, for example, according to the intended applications of probe 700, and the value of H_b is at least to some extent a matter of design choice. Typically, for a broad array of electrosurgical procedures, the distance H_b is in the range of from about 0.02 mm to about 5 mm. Active electrode 712 includes an active edge 713 which is adapted for generating high current densities thereat upon application of a high frequency voltage from the power supply between active electrode 712 and return electrode 718. In this way, active edge 713 can efficiently effect localized ablation of tissues via molecular dissociation of tissue components which contact, or are in close proximity to, active edge 713. A process for ablation of tissues via molecular dissociation of tissue components has been described hereinabove.

As best seen in FIG. 31B, the blade-like active electrode 712 further includes first and second blade sides, 714a, 714b, respectively. First and second blade sides 714a, 714b are separated by a maximum distance, W_b . The distance W_b is typically in the range of from about 0.1 mm to about 2.5 mm. In the embodiment of FIG. 31B, first and second blade sides 714a, 714b are substantially parallel to each other. Each of first and second blade sides 714a, 714b are adapted for engaging tissue severed, ablated, or otherwise modified by active edge 713, and for coagulating tissue engaged by first blade side 714a and/or second blade side 714b. In this way, active electrode 712 can precisely and effectively sever, ablate, or otherwise modify a target tissue with active edge 713 to form a first-modified tissue, and at the same time, or shortly thereafter, further modify the first-modified tissue by means of first and second blade sides 714a, 714b. For example, active edge 713 can make an incision in a target tissue via localized molecular dissociation of target tissue components, while first and second blade sides 714a, 714b can effect hemostasis in the severed tissue.

FIGS. 32A, 32B, and 32C are a side view, a plan view, and an end view, respectively, of electrosurgical probe 700 having a blade-like active electrode 712, according to one embodiment of the invention. In the embodiment of FIGS. 32A-C, electrode support 710 is disposed at the terminus of shaft 702, and active electrode 712 is affixed to support distal end 710a (e.g., FIG. 33A). However, other arrangements for

electrode support **710** and active electrode **712** are within the scope of the invention (e.g., FIGS. **34A–C**, **35A–C**). Active electrode **712** is in the form of a substantially flat metal blade. Active electrode **712** is shown as being substantially rectangular as seen from the side (FIG. **32A**). However, various other shapes for active electrode **712** are within the scope of the invention (e.g., FIGS. **33C–E**). FIG. **32C** is an end view of probe **700** as seen along the lines **32C–32C** of FIG. **32B**, showing a laterally compressed region **703** of shaft **702**. Laterally compressed region **703** may be adapted for housing electrode support **710**. Laterally compressed region **703** may also facilitate manipulation of shaft distal end portion **702a** of probe **700** during various surgical procedures, particularly in situations where accessibility of a target tissue is restricted.

FIGS. **33A** and **33B** are a side view and a plan view, respectively, of the distal end of probe **700**, showing details of shaft distal end portion **702a** and terminally disposed blade active electrode **712**, according to one embodiment of the invention. Blade electrode **712** is substantially rectangular in shape as seen from the side (FIG. **33A**). The distal end of shaft **702** includes laterally compressed region **703**. As seen from the side (FIG. **33A**), laterally compressed region **703** appears wider than more proximal portions of shaft **702**. FIG. **33B** is a plan view of probe **700** as seen along the lines **33B–33B** of FIG. **33A**, in which laterally compressed region **703** appears narrower than more proximal portions of shaft **702**. Electrode support **710** is mounted to the distal end of laterally compressed region **703**. Typically, electrode support **710** comprises a durable, electrically insulating, refractory material having a certain amount of flexibility. For example, electrode support **710** may comprise a material such as a silicone rubber, a polyimide, a fluoropolymer, a ceramic, or a glass.

FIGS. **33C–33E** each show a side view of the distal end of probe **700** having a terminal blade active electrode **712**, according to three different embodiments of the invention. Electrode support **710** is mounted terminally on shaft **702**, and includes a support distal end **710a** and a support proximal end **710b**. In the embodiment of FIG. **33C**, active edge **713** of active electrode **712** is arcuate, convex, or substantially semi-circular in shape. In the embodiment of FIG. **33D**, active electrode **712** has a pointed active edge **713**, while in the embodiment of FIG. **33E**, the active edge **713** of active electrode **712** is serrated.

FIG. **34A** shows in side view an electrosurgical probe **700** having electrode support **710** mounted terminally on shaft **702** and blade active electrode **712** disposed laterally on electrode support **710**, according to another embodiment of the invention. FIG. **34B** is a plan view of probe **700** taken along the lines **34B–34B** of FIG. **34A**. FIG. **34C** is an end view taken along the lines **34C–34C** of FIG. **34A**. In the embodiments of FIGS. **34A–C**, electrode **712** is in the form of a substantially flat, metal blade having first and second blade sides **714a**, **714b**, substantially parallel to each other. First and second blade sides **714a**, **714b** are adapted for engaging and coagulating severed or modified tissue, as described hereinabove.

FIG. **35A** shows in side view an electrosurgical probe **700** having electrode support **710** mounted laterally on the distal end of shaft **702**, according to another embodiment of the invention. Blade active electrode **712** is mounted laterally on electrode support **710**. FIG. **35B** is a plan view of probe **700** taken along the lines **35B–35B** of FIG. **35A**. FIG. **35C** is an end view taken along the lines **35C–35C** of FIG. **35A**. Active electrode **712** is in the form of a substantially flat, metal blade having first and second blade sides **714a**, **714b**,

substantially parallel to each other. Electrode support **710** is mounted laterally on laterally compressed region **703** of shaft **702**.

FIG. **36A** shows a side view of the distal end of an electrosurgical probe **700**, wherein shaft **702** includes a beveled end **728** to which electrode support **710** is mounted. Blade active electrode **712** is disposed on electrode support **710**. The arrangement of electrode support **710** and electrode **712** on beveled end **728** may facilitate access of shaft distal end portion **702a** in general, and of electrode **712** in particular, to a target tissue during various surgical procedures, particularly in situations where accessibility is restricted. FIG. **36B** shows a side view of the distal end of an electrosurgical probe **700**, according to another embodiment of the invention. Shaft **702** includes a curved distal end **702a'**. Electrode support **710** is mounted on distal end **702a'**, and blade active electrode **712** is affixed to electrode support **710**. Curved distal end **702a'** facilitates access of electrode **712** to a target tissue during various surgical procedures.

Although in the embodiments of FIGS. **34A–C**, **35A–C**, and **36A–B** active electrode **712** is shown as being substantially rectangular, this representation should not be construed as limiting these embodiments to a rectangular active electrode **712**. Indeed, each of the embodiments of FIGS. **34A–C**, **35A–C**, and **36A–B** may have an active electrode **712** in a broad range of shapes, including those represented in FIGS. **33C–E**.

FIG. **37A** shows in side view an electrosurgical probe **700** having an exterior tube **724** arranged on shaft **702** and coupled at its proximal end to a connection tube **720** at handle **704**. Exterior tube **724** may comprise a plastic tube of suitable length commensurate with the size of probe **700**. Exterior tube **724** defines a lumen **726**, and typically terminates at shaft distal end **702a** at a location somewhat proximal to electrode support **710**. In some embodiments, probe **700** may include two or more exterior tubes **724**, each exterior tube **724** having lumen **726**. Each lumen **726** may serve as a conduit for an aspiration stream, or as a conduit for delivery of an electrically conductive fluid to the shaft distal end, generally as described hereinabove. FIG. **37B** is an end view of probe **700** taken along the lines **37B–37B** of FIG. **37A**, showing exterior tube **724** and lumen **726** in relation to shaft **702**. The diameter of exterior tube **724** is, at least to some extent, a matter of design choice. Exterior tube **724** may comprise a substantially rigid or somewhat flexible plastic tube comprising polyethylene, a polyimide, a fluoropolymer, and the like.

FIG. **38A** shows, in side view, an electrosurgical probe **700** having an outer sheath **722** surrounding the exterior of a portion of shaft **702**, according to another embodiment of the invention. Outer sheath **722** is coupled at its proximal end to a connection tube **720** at handle **704**. Outer sheath **722** may comprise a plastic tube of suitable length and having a diameter larger than that of shaft **702**. Together with the exterior of shaft **702**, outer sheath **722** defines a lumen **726'** in the form of an annular void. Typically, outer sheath **722** terminates at shaft distal end **702a** at a location proximal to electrode support **710**. Lumen **726'** typically serves as a conduit for delivery of an electrically conductive fluid to the shaft distal end. FIG. **38B** is an end view of probe **700** taken along the lines **38B–38B** of FIG. **38A**, showing outer sheath **722** and lumen **726'** in relation to shaft **702**. The diameter of outer sheath **722** is, at least to some extent, a matter of design choice. Outer sheath **722** may comprise a substantially rigid or somewhat flexible plastic tube comprising polyethylene, a polyimide, and the like.

FIG. **39A** schematically represents an electrosurgical probe **700**, according to another embodiment of the inven-

tion. Probe **700** includes shaft **702** and handle **704** affixed at shaft proximal end **702b**. A first electrode support **711a** and a second electrode support **711b** are disposed at shaft proximal end **702a**. A blade active electrode **712** is arranged on first and second electrode supports, **711a**, **711b**. Each of first and second electrode supports **711a**, **711b** may comprise a refractory and electrically insulating material, such as a silicone rubber or the like, as described hereinabove. A return electrode **718** is located at shaft distal end **702** proximal to first and second electrode supports **711a**, **711b**. Return electrode **718** may comprise an exposed portion of shaft distal end **702a** (e.g., FIGS. 32A–C). Blade active electrode **712** typically extends distally from electrode support **710** by a distance in the range of from about 0.1 mm to about 10 mm, an more typically from about 2 mm to 10 mm.

Blade active electrode **712** and return electrode **718** may be independently coupled to opposite poles of a high frequency power supply via electrode leads (not shown) and a connection block (e.g., FIG. 30). In one embodiment, an active electrode lead is coupled to one of first and second electrode arms **715a**, **715b**, and the other arm terminates in a free, electrically isolated end, for example, within first electrode support **711a** or second electrode support **711b**. Blade active electrode **712** includes a crosspiece **715c** (FIGS. 39B–D) located distal to aspiration port **734**. A fluid delivery element or unit including an outer sheath **722'** (e.g., FIG. 39B) is omitted from FIG. 39A for the sake of clarity.

FIG. 39B is a partial sectional view of probe **700** of FIG. 39A as seen from the side. Outer sheath **722'** defines an annular fluid delivery lumen **726'** between sheath **722'** and shaft **702**. Lumen **726'** terminates in an annular fluid delivery port **725** at shaft distal end **702a**. Fluid delivery lumen **726'** is in communication proximally with a fluid delivery tube **721**. Solid arrows indicate the direction of flow of an electrically conductive fluid (e.g., isotonic saline) within fluid delivery lumen **726'**. Aspiration port **734** is in communication proximally with an aspiration lumen **732** and an aspiration tube **730**. Solid arrows within aspiration lumen **732** indicate the direction of flow of an aspiration stream, which flows from aspiration port **734** towards a source of vacuum (not shown), the latter coupled to aspiration tube **730**. FIG. 39C is an end view of probe **700** taken along the lines 39C–39C of FIG. 39B. Active electrode **712** includes crosspiece **715c** extending between first and second electrode arms **715a**, **715b**, respectively (FIG. 39D). Active electrode **712** further includes first and second blade sides **714a**, **714b**. In some embodiments, first and second blade sides **714a**, **714b** are adapted for engaging tissue that has been severed, and for coagulating the severed tissue. Crosspiece **715c** at least partially spans aspiration port **734**. Typically, active electrode **712** comprises a single metal blade, comprising a material such as platinum, tungsten, palladium, iridium, or titanium, or their alloys.

FIG. 39D shows detail of the distal portion of probe **700** of FIGS. 39A–C including blade active electrode **712**. As shown, first and second electrode arms **715a**, **715b** are disposed on first and second electrode supports **711a**, **711b**, respectively. In an alternative embodiment, first and second electrode arms **715a**, **715b** may be disposed on a single annular electrode support having a substantially central void defining aspiration port **734**. In one embodiment, active electrode **712** includes both a distal active edge **713a**, and a proximal active edge **713b**. Distal active edge **713a**, in particular, is adapted for aggressively ablating tissue via molecular dissociation of tissue components and for severing tissue targeted for resection, transection, dissection, or other treatment.

FIG. 40A is a partial sectional view of an electrosurgical probe **700** according to another embodiment of the invention. Probe **700** of FIG. 40A generally includes shaft **702** and handle **704**, together with a fluid delivery element, and an aspiration unit, essentially as for the embodiment described with reference to FIGS. 39A–D. In the interests of brevity these elements and features will not be described in detail with reference to FIGS. 40A–C. The embodiment of FIG. 40A differs from other embodiments described herein in having an active electrode in the form of a plasma hook **712'**. Hook **712'** is in some respects analogous to plasma blade electrodes described hereinabove. For example, in one respect hook **712'** is analogous to a truncated version of electrode **712** of the embodiment of FIGS. 39A–D in which one of arms **715a** or **715b** is omitted leaving one electrode arm affixed to crosspiece **715c**. From a functional standpoint, hook **712'** allows the operator (surgeon) to ablate tissue by drawing the instrument towards himself/herself. In this manner, greater control is exerted over the amount or extent of tissue removed or severed by probe **700**. Hook **712'** includes a first axial portion **712'a** (FIG. 40C) in contact at its proximal end with electrode support **710**. Hook **712'** may further include a second portion **712'b** extending from the distal portion of first axial portion **712'a**. In some embodiments, second portion **712'b** is arranged substantially orthogonal to first axial portion **712'a**. In one embodiment, second portion **712'b** may be structurally similar or analogous to crosspiece **715c** of the embodiment of FIGS. 39A–D. Second portion **712'b** at least partially spans aspiration port **734** (FIG. 40B). Electrode support **710** may comprise a refractory and electrically insulating material, such as a silicone rubber or the like, as described hereinabove.

FIG. 40B shows an end view of probe **700** taken along the lines 40B–40B of FIG. 40A. Hook **712'** includes first and second blade sides **714a**, **714b**. Second portion **712'b** extends at least partially across aspiration port **734**. FIG. 40C shows detail of the distal end portion of probe **700** of FIGS. 40A, 40B, including hook **712'**. Hook **712'** includes a distal active edge **713a**, a proximal active edge **713b**, and an active tip **713c**. Return electrode **718** is located proximal to electrode support **710**. Upon application of a high frequency voltage between hook **712'** and return electrode **718**, a high current density may be generated at each of distal active edge **713a**, proximal active edge **713b**, and active tip **713c**. Each of distal active edge **713a**, proximal active edge **713b**, and active tip **713c** may be adapted for severing tissue via electrosurgical molecular dissociation of tissue components.

FIGS. 41A, 41B, and 41C each show detail of the distal end portion of an electrosurgical probe including a hook electrode **712'**, according to three different embodiments of the invention. In the embodiment of FIG. 41A, hook **712'** is curved, having a convex distal edge **713a**, and a concave proximal edge **713b**. In the embodiment of FIG. 41B, proximal edge **713b** includes serrations thereon. In an alternative embodiment (not shown), distal edge **713a**, and/or active tip **713c** may be similarly serrated. In the embodiment of FIG. 41C, hook **712'** is curved, having a concave distal edge **713a**, and a convex proximal edge **713b**. According to various embodiments of probe **700**, second portion **712'b** may have a length which is less than, equal to, or greater than the diameter of shaft **702**. In the latter case, second portion **712'b** extends laterally beyond the exterior surface of shaft **702** (e.g., FIG. 41C). In each of the embodiments of FIGS. 41A–C, hook **712'** typically comprises a single blade having first and second blade sides **714a**, **714b** (e.g., FIG. 40B). Hook **712'** typically comprises a metal such as platinum, tungsten, palladium, iridium, or titanium, or their alloys.

FIGS. 42A–B schematically represent a process during treatment of a patient with electrosurgical probe **700**. Blade active electrode **712** is affixed to support **710** on shaft **702**. Blade active electrode **712** includes active edge **713** and first and second blade sides, **714a**, **714b** (e.g., FIGS. 31A–B). Referring to FIG. 42A, active edge **713** forms an incision, **I**, in a target tissue, **T**, via localized molecular dissociation of tissue components upon application of a high frequency voltage between active electrode **712** and return electrode **718**. (The localized molecular dissociation may be facilitated by the delivery of a suitable quantity of an electrically conductive fluid (e.g. isotonic saline) to form a current flow path between active electrode **712** and return electrode **718**.) With reference to FIG. 42B, as the incision **I** is deepened within tissue **T**, first and second blade sides, **714a**, **714b** engage severed tissue in regions indicated by the arrows labeled **E**. In this way, the severed tissue is coagulated by first and second blade sides, **714a**, **714b**, thereby effecting hemostasis at the point of incision of the tissue.

FIG. 43A schematically represents a number of steps involved in a method of treating a patient with an electrosurgical probe, wherein step **1000** involves positioning the distal end of the probe adjacent to target tissue such that an active electrode of the probe is in contact with or in close proximity to the target tissue. In one embodiment, the active electrode is spaced a short distance from the target tissue, as described hereinabove. Typically, step **1000** involves positioning the probe such that an active edge of the active electrode makes contact with, or is in close proximity to, the target tissue. Step **1002** involves delivering an electrically conductive fluid to the distal end of the probe in the vicinity of the active electrode and the return electrode, such that the electrically conductive fluid forms a current flow path between the active electrode and the return electrode. The electrically conductive fluid may be delivered via an exterior tube disposed on the outside of the shaft (e.g., FIGS. 37A, 37B), or an outer sheath external to the shaft and forming an annular fluid delivery lumen (e.g., FIGS. 38A, 38B). The electrically conductive fluid may be a liquid, a gel, or a gas. Apart from providing an efficient current flow path between the active and return electrodes, a clear, colorless electrically conductive liquid, such as isotonic saline, exhibits the added advantage of increasing the visibility of the surgeon at the target site. However, in situations where there is an abundance of electrically conductive body fluids (e.g., blood, synovial fluid) already present at the target site, step **1002** may optionally be omitted.

Step **1004** involves applying a high frequency voltage between the active electrode and the return electrode sufficient to ablate or otherwise modify the target tissue via localized molecular dissociation of target tissue components. By delivering an appropriate high frequency voltage to a suitably configured probe, the target tissue can be incised, dissected, transected, contracted, or otherwise modified. In addition, the modified tissue can also be coagulated (e.g., FIG. 42B). The frequency of the applied voltage will generally be within the ranges cited hereinabove. For example, the frequency will typically range from about 5 kHz to 20 MHz, usually from about 30 kHz to 2.5 MHz, and often between about 100 kHz and 200 kHz. The root mean square (RMS) voltage that is applied in step **1004** is generally in the range of from about 5 volts to 1000 volts RMS, more typically being in the range of from about 10 volts to 500 volts RMS. The actual voltage applied may depend on a number of factors, including the size of the active electrode, the operating frequency, and the particular procedure or desired type of modification of the tissue (incision, contraction, etc.), as described hereinabove.

Step **1006** involves manipulating the probe with respect to the tissue at the target site. For example, the probe may be manipulated such that an active edge of a blade or hook electrode reciprocates with respect to the target tissue, such that the target tissue is severed, incised, or transected at the point of movement of the active edge by a process involving molecular dissociation of tissue components. In embodiments where the active electrode is in the form of a hook, step **1006** may involve engaging the hook against the target tissue and drawing the hook towards the operator in order to cut or sever the tissue. In this manner, the extent of cutting or severing can be precisely controlled. In one embodiment, step **1006** involves reciprocating an active edge in a direction parallel to a surface of the target tissue. Typically, step **1006** is performed concurrently with step **1004**. Step **1002** may be performed at any stage during the procedure, and the rate of delivery of the electrically conductive fluid may be regulated by a suitable mechanism, such as a valve.

Step **1008** involves modifying the target tissue as a result of the high frequency voltage applied in step **1004**. The target tissue may be modified in a variety of different ways, as referred to hereinabove. The type of tissue modification achieved depends, inter alia, on the voltage parameters of step **1004**; the shape, size, and composition of the active electrode; and the manner in which the probe is manipulated by the surgeon in step **1006**. At relatively high voltage levels, tissue components typically undergo localized molecular dissociation, whereby the target tissue can be dissected, incised, transected, etc. At a lower voltage, or at a lower current density on the active electrode surface, the target tissue can be contracted (e.g., by shrinkage of collagen fibers in the tissue), or a blood vessel can be coagulated. For example, in step **1010** the first and second blade sides of the active electrode may be engaged against a region of the target tissue which has been modified as a result of localized molecular dissociation of tissue components in step **1008**. The first and second blade sides are substantially flat metal plates having lower current densities than the active edge. In this manner, the lower current densities of the first and second blade sides cause further modification (e.g., coagulation) of the previously modified (e.g., severed) target tissue (step **1012**).

FIG. 43B schematically represents a number of steps involved in a method of severing tissue with an electrosurgical probe via a process involving molecular dissociation of tissue components, and of coagulating the severed tissue with the same electrosurgical probe during a single procedure, according to one embodiment of the invention. The electrosurgical probe typically comprises an active electrode in the form of a single, substantially flat metal hook or blade having at least one active edge adapted for electrosurgically severing the tissue, and first and second blade sides adapted for effecting hemostasis of the severed tissue. Steps **1000'** through **1006'** are substantially the same or analogous to steps **1000** through **1006**, as described hereinabove with reference to FIG. 43A. Step **1008'** involves severing the target tissue via localized molecular dissociation of tissue components due to high current densities generated at the position of an active edge upon execution of step **1004'**. Step **1010'** involves engaging the first and second blade sides against the tissue severed in step **1008'**, whereby blood/blood vessels in the severed tissue are coagulated as a result of the relatively low current densities on the first and second blade sides (step **1012'**).

FIG. 44 schematically represents a number of steps involved in a method of dissecting a tissue or organ of a patient with an electrosurgical probe having a hook or blade

active electrode, according to one embodiment of the invention, wherein step **1100** involves accessing an organ or tissue. Typically, accessing an organ or tissue in step **1100** involves incising an overlying tissue which conceals the organ or tissue to be dissected. As an example, in an open chest procedure involving a median sternotomy, the thoracic cavity is accessed by making a longitudinal incision through the sternum. Incising an overlying tissue in step **1100** may be performed generally according to the methods described with reference to FIG. **43A** or **43B**. Step **1102** involves positioning the distal end of the electrosurgical probe, and in particular an active edge of the hook or blade active electrode, in at least close proximity to connective tissue adjacent to the tissue or organ to be dissected. As an example, the connective tissue may be soft tissue, such as adipose tissue, or relatively hard tissue such as cartilage or bone. Optional step **1104** involves delivering an electrically conductive fluid to the distal end of the probe such that the electrically conductive fluid forms a current flow path between the active electrode and the return electrode, generally as described for step **1002**, supra. Step **1106** involves applying a high frequency voltage between the active electrode and the return electrode, generally as described for step **1004**, supra.

Depending on the type of procedure, e.g., the nature of the tissue or organ to be dissected, optional step **1108** may be performed, in which the probe is manipulated such that an active edge of the active electrode is moved with respect to the connective tissue adjacent to the tissue or organ to be dissected. Where the active electrode comprises a hook, the hook may be engaged against the connective tissue and drawn towards the operator of the probe to precisely control the degree of cutting or tissue removal. Step **1110** involves electrosurgically ablating, via molecular dissociation of connective tissue components, at least a portion of the connective tissue adjacent to the tissue or organ to be dissected. As an example, connective tissue adjacent to the internal mammary artery may be dissected by a process involving molecular dissociation of connective tissue components, in either an open-chest or a minimally invasive procedure, such that the IMA is substantially free from connective tissue over a portion of its length.

It is to be understood that the electrosurgical apparatus of the invention, e.g., probe **700**, is by no means limited to those methods described with reference to FIGS. **42A-44**. Thus, as stated hereinabove, embodiments of an electrosurgical probe having an active electrode in the form of a blade or hook are applicable to a broad range of surgical procedures, such as ablation, incision, contraction, coagulation, or other modification of: connective tissue, including adipose tissue, cartilage, and bone; dermal tissue; vascular tissues and organs, including arteries and veins; and tissues of the shoulder, knee, and other joints. Thus, while the exemplary embodiments of the present invention have been described in detail, by way of example and for clarity of understanding, a variety of changes, adaptations, and

modifications will be apparent to those of skill in the art. Therefore, the scope of the present invention is limited solely by the appended claims.

What is claimed is:

1. An electrosurgical probe, comprising:
 - a shaft having a shaft distal end portion and a shaft proximal end portion;
 - a fluid delivery unit including a fluid delivery lumen and a fluid delivery port;
 - an aspiration lumen and an aspiration port;
 - an electrically nonconductive electrode support affixed to the shaft distal end portion;
 - a blade electrode affixed to the electrode support, wherein the blade electrode comprises a hook having at least one active edge, a first blade side and a second blade side; and
 - a return electrode proximal to said blade electrode.
2. The electrosurgical probe of claim **1**, wherein the hook comprises a substantially flat metal blade having a first axial portion and a second portion extending in a direction substantially orthogonal to the first axial portion.
3. The electrosurgical probe of claim **1**, wherein the hook is curved.
4. The electrosurgical probe of claim **1**, wherein the hook has at least one serrated edge.
5. The electrosurgical probe of claim **1**, wherein the hook includes a distal active edge and a proximal active edge.
6. The electrosurgical probe of claim **5**, wherein the hook is curved and the distal active edge is concave.
7. The electrosurgical probe of claim **1**, wherein the hook is adapted for severing tissue by drawing the probe towards an operator of the probe.
8. The electrosurgical probe of claim **1**, wherein each of the first blade side and the second blade side are adapted for engaging a severed tissue and for coagulating the severed tissue.
9. The probe of claim **1**, further comprising an insulating sleeve covering a portion of the shaft, the insulating sleeve terminating distally in a sleeve distal end at a location proximal to the electrode support to define an exposed portion of the shaft distal end portion, and the exposed portion defining said a return electrode.
10. The probe of claim **1**, wherein the fluid delivery unit includes an outer sheath situated external to the shaft and forming an annular fluid delivery lumen between the outer sheath and the shaft, the annular fluid delivery lumen terminating in an annular fluid delivery port.
11. The probe of claim **10**, wherein the fluid delivery port is located at the shaft distal end portion proximal to said return electrode, said return electrode proximally spaced from said blade electrode.
12. The probe of claim **1**, wherein the blade electrode comprises a material selected from the group consisting of platinum, tungsten, palladium, iridium, and titanium.

* * * * *