cologuard™

PATIENT GUIDE

Indication for Use

Cologuard is intended for the qualitative detection of colorectal neoplasia associated DNA markers and for the presence of occult hemoglobin in human stool. A positive result may indicate the presence of colorectal cancer (CRC) or advanced adenoma (AA) and should be followed by diagnostic colonoscopy. Cologuard is indicated to screen adults of either sex, 50 years or older, who are at typical average-risk for CRC. Cologuard is not a replacement for diagnostic colonoscopy or surveillance colonoscopy in high risk individuals.

exact sciences

IVD

For questions or help call **Exact Sciences at 1-844-870-8878**
www.cologuardtest.com
Rx Only

Exact Sciences
Madison WI USA

LBL-0207rev3

Table of Contents

Contra-Indications 3

Cologuard Warnings and Precautions 4-5

How the Cologuard Collection Kit Works 6

What Does the Cologuard Test Result Mean 7

What Is Inside the Cologuard Collection Kit 8

Who Should Use Cologuard? 9

Caution 10

Risks 10

What Should I Know Before Using the Cologuard Collection Kit?..... 11

When Should I Not Use the Cologuard Collection Kit? 12

How to Use the Cologuard Collection Kit..... 13 -31

Glossary 32

Symbols Used on the Cologuard Collection Kit 33

Clinical Study Results..... 33

Contra-Indications

Cologuard is intended for use with patients, age 50 years and older, at average risk who are typical candidates for CRC screening. Cologuard was not clinically evaluated for the following types of patients:

- Patients with a history of colorectal cancer, adenomas, or other related cancers.
- Patients who have had a positive result from another colorectal cancer screening method within the last 6 months.
- Patients who have been diagnosed with a condition that is associated with high risk for colorectal cancer. These include but are not limited to:
    - Inflammatory Bowel Disease (IBD)
    - Chronic ulcerative colitis (CUC)
    - Crohn's disease
    - Familial adenomatous polyposis (FAP)
    - Family history of colorectal cancer
- Patients who have been diagnosed with a relevant familial (hereditary) cancer syndrome, such as Hereditary non-polyposis colorectal cancer syndrome (HNPCCC or Lynch Syndrome), Peutz-Jeghers Syndrome, MYH-Associated Polyposis (MAP), Gardner's syndrome, Turcot's (or Crail's) syndrome, Cowden's syndrome, Juvenile Polyposis, Cronkhite-Canada syndrome, Neurofibromatosis, or Familial Hyperplastic Polyposis.

Cologuard Warnings and Precautions

- The performance of Cologuard has been established in a cross sectional study (i.e., single point in time). Programmatic performance of Cologuard (i.e., benefits and risks with repeated testing over an established period of time) has not been studied. Performance has not been evaluated in adults who have been previously tested with Cologuard. Non-inferiority or superiority of Cologuard programmatic sensitivity as compared to other recommended screening methods for CRC and AA has not been established.
- CRC screening guideline recommendations vary for persons over the age of 75. The decision to screen persons over the age of 75 should be made on an individualized basis in consultation with a healthcare provider. Cologuard test results should be interpreted with caution in older patients as the rate of false positive results increases with age.
- A negative Cologuard test result does not guarantee absence of cancer or advanced adenoma. Patients with a negative Cologuard test result should be advised to continue participating in a colorectal cancer screening program with another recommended screening method. The screening interval for this follow-up has not been established.
- Cologuard may produce false negative or false positive results. A false positive result occurs when Cologuard produces a positive result, even though a colonoscopy will not find cancer or precancerous polyps. A false negative result occurs when Cologuard does not detect a precancerous polyp or colorectal cancer even when a colonoscopy identifies the positive result.

Cologuard Warnings and Precautions

- Patients should not provide a sample for Cologuard if they have diarrhea or if they have blood in their urine or stool (e.g., from bleeding hemorrhoids, bleeding cuts or wounds on their hands, rectal bleeding, or menstruation).
- To ensure the integrity of the sample, the laboratory must receive the patient specimens within 72 hours of collection. Patients should send stool samples to the laboratory according to the instructions stated in the Cologuard Patient Guide.
- Patients should be advised of the caution listed in the Cologuard Patient Guide. Patients should NOT drink the preservative liquid.
- The risks related to using the Cologuard Collection Kit are low, with no serious adverse events reported among people in a clinical trial. Patients should be careful when opening and closing the lids to avoid the risk of hand strain.

How the Cologuard Collection Kit Works

Your doctor has ordered Cologuard to screen for colorectal cancer and precancer.

Cologuard is a screening test that uses a stool sample (your bowel movement) to detect colorectal cancer and precancer. Every day, your colon sheds cells that line the colon. As part of this process, normal cells along with abnormal cells from precancer or cancers are shed into the colon. Your stool picks up those cells as it passes through your colon. Cologuard is designed to detect the DNA and hemoglobin released from these abnormal cells in your stool.

The Cologuard collection kit is used to collect your stool sample. After you collect your stool sample following the instructions in this Patient Guide, the collection kit will be delivered to a lab. The lab will test your stool sample and send the results to your doctor. Your doctor will contact you with the test results.

Note: You are not required to change your diet or medications to use this screening test.

How Do I Store the *Cologuard* Collection Kit When It Arrives?

You can store your kit until you are ready to use it.

- Store at room temperature.
- Keep away from heat and direct sunlight.
- Keep out of the reach of children.

What Does the Cologuard Test Result Mean?

Your doctor will talk with you about your results. The test result can be **Positive**, **Negative** or **No result obtained**.

What Does a Positive Result Mean?

- A Positive result means the test detected abnormal DNA and/or blood that could be caused by precancer or cancer in the colon.
- The test can also have a Positive result that is incorrect (false positive). This means the test result is Positive, but no cancer or precancer is actually present.
- Any Positive result should be followed by a colonoscopy.
- Talk about your test result with your doctor to find out if additional testing is needed.

What Does a Negative Result Mean?

- A Negative test result means the test did not detect abnormal DNA and/or blood in the sample.
- The test can also have a Negative result that is incorrect (false negative). This means the test result missed a potential cancer or precancer. For that reason, it is recommended that you schedule regular screenings.
- Talk to your doctor about your test result. Your doctor will recommend a screening schedule that is best for you.

What Does No Result Obtained Mean?

- A "No Result Obtained" means the test was not able to provide a result.
- If this happens, your doctor will talk with you about the next steps. For example, you may be asked to provide another stool sample to test.

What Is Inside the Cologuard Collection Kit



- The bottle of liquid contains a preservative (less than 10% EDTA in Tris buffered solution).
- The tube contains a 10% albumin in Tris buffered detergent solution with an antimicrobial agent.

Who Should Use Cologuard?

Cologuard is intended for people who are typical candidates for colorectal cancer screening, which include:

- Men and women 50 years or older
- At average risk for colorectal cancer

Is Cologuard Right for You?

Talk with your doctor about using Cologuard if any of the following apply to you:

- A history of colorectal cancer, adenomas, or other related cancers.
- Had a positive result from another colorectal cancer screening method within the last 6 months.
- Been diagnosed with a condition that is associated with high risk for colorectal cancer. These include but are not limited to:
  - Inflammatory Bowel Disease (IBD)
  - Chronic ulcerative colitis (CUC)
  - Crohn's disease
  - Familial adenomatous polyposis (FAP)
  - Family history of colorectal cancer
- Been diagnosed with a relevant familial (hereditary) cancer syndrome, such as Hereditary non-polyposis colorectal cancer syndrome (HNPCCC or Lynch Syndrome), Peutz-Jeghers Syndrome, MYH-Associated Polyposis (MAP), Gardner's syndrome, Turcot's (or Crail's) syndrome, Cowden's syndrome, Juvenile Polyposis, Cronkhite-Canada syndrome, Neurofibromatosis, Familial Hyperplastic Polyposis.

Caution

Caution

- **Do not drink the bottle of preservative liquid.**
- Do **not** let the liquid touch your skin or eyes.
- If the liquid touches your skin or eyes, wash the area with water.

Risks

The risks related to using the Cologuard collection kit are low. No serious adverse events were reported among 10,023 people in a clinical trial.

- Opening or closing the lids of items in the kit may be difficult for some people.
  - Be careful when opening and closing the lids to avoid the risk of hand strain.
  - Close all sample containers tightly.
- Using this kit requires sitting down on the toilet and standing up from the toilet.
  - Have someone who can help you if needed.
- There is a chance that a stool sample sent to the lab may have no result.
  - If this happens, you will be contacted. You may be asked to provide another sample.

What Should I Know Before Using the Cologuard Collection Kit?

Only remove the items you need to collect your sample following the steps in this guide.

- **Leave the plastic bag inside the box.** The box, zippered bag and cardboard tray inside the bag will be used to send your sample to the lab.
- Use the kit before the expiration date printed on the side of the box.
- Avoid getting urine in the container when collecting your stool sample.
- Avoid getting toilet paper or other materials in the container when collecting your stool sample.
- **The lab must receive samples within 3 days.**
  - Collect a sample when you can ship your sample within a day of collection.
  - Make sure a holiday will not delay your shipment.
  - Review **Step 8: Ship Your Sample Using UPS** in this guide for detailed shipping information.
- You are not required to change your diet or medications to use this screening test.

When Should I Not Use the Cologuard Collection Kit?

Certain conditions may cause an incorrect test result or no result. Do **not** use this kit to collect a stool sample if you have:

- Bleeding hemorrhoids
- Bleeding cuts or wounds on your hands
- Rectal bleeding
- Menstrual period
- Diarrhea

Step 1: Check the Expiration Date and the Kit

Check your kit to make sure the kit has not expired and you have all the parts of the kit.



a. Check the expiration date on the outside of the box.

b. Use the kit before the expiration date printed on the side of the box.

- If the date has passed, do **not** use the kit. Close the box and contact Exact Sciences at 1-844-870-8878 for a new kit.



Leave the plastic bag and gray cardboard insert inside the box.

c. Check the items in your box and make sure you have the following:

- Bracket
- Sample container
- Tube
- Bottle of liquid (preservative)
- Shipping labels (attached to top of box)
- Sample labels

If any items are missing from your kit, do **not** use the kit. Contact Exact Sciences at 1-844-870-8878 for a new kit.

d. If all items are included in your kit, your kit is ready to use.

Follow the next steps when you feel ready to have a bowel movement.

Step 2: Prepare to Collect Stool Sample

Follow these steps in your bathroom to set up the bracket and stool sample container.



Timing Your Collection

- ✓ **Ready to have a bowel movement**
- ✓ **Can ship sample <u>within a day</u>**
- ✓ **Use shipping schedule in Step 8**

a. Decide the best time to collect a stool sample.

- Follow these steps when you feel ready to have a bowel movement.
- Collect your sample on a day when you can ship your sample to arrive at the lab within 3 days.
  - o **Ship your sample <u>within a day</u> of collection.**
  - o See **Step 8: Ship Your Sample Using UPS** for a shipping schedule.
- Do **<u>not</u>** collect a stool sample if you have:
  - o Bleeding hemorrhoids
  - o Bleeding cuts or wounds on your hands
  - o Rectal bleeding
  - o Menstrual period
  - o Diarrhea



b.

b. Remove the bracket from the box and unfold the sides of the bracket

c.

c. Raise the toilet lid **and** seat.



d.

d. Place the bracket on the toilet as shown.

- Place the bracket toward the back of the toilet.



e.

e. Lower the toilet seat onto the bracket.

- The entire opening of the bracket should be visible.

f. Lift the stool sample container out of the box and place on a hard, flat surface.



g.

g. Turn the container lid and unscrew it.

h. Set the container lid down.

i.

i. Place the container into the top of the bracket.

Step 3: Collect the Stool Sample

- Make sure you have help if you have trouble sitting and standing when using the toilet.



a.

c.

a. Sit on the toilet and have a bowel movement in the container.
   - **Try to keep urine from going into the container.**
   - **Do not put toilet paper or other items into the container.**

b. When your bowel movement is complete, stand up.

c. Lift the stool sample container from the bracket and set the container on a flat, stable surface.
   - **Leave the container open.**

d.

d. Remove the bracket from the toilet.

- The bracket can be recycled or thrown in the trash.

e.

e. Finish using the bathroom if needed.

- Follow the rest of the steps in this guide **immediately** after collecting your stool sample.

IMPORTANT: Complete the next step **before** you close the container.

Next Page

Step 4: Scrape the Stool Sample

You must scrape the stool sample with the probe to get another small sample for the Cologuard test.



a. Lift the tube out of the box.

b. Turn the white tube cap and unscrew it.



c. Pull the probe from the tube.
   - You may have to pull hard.



d. Scrape the surface of your stool sample until the end of the probe has stool on it.
   - Your stool sample may look different from the stool sample pictured.



e. Place the probe back into the open end of the tube.



f. Turn the cap to close.

g. Set the tube down.

Step 5: Prepare Stool Sample Container for Shipping

The stool sample must have a preservative poured onto it to make sure the lab can test it. Then, the container must be closed tightly for shipping.



a. Lift the bottle of liquid preservative out of the box.

b. Hold the bottle and turn the cap to unscrew it.

Caution

- Do **not** drink the liquid.
- If the liquid from the bottle touches your skin or eyes, wash with water.



c. Pour **all** the liquid in the bottle into the container with the stool.

- The empty bottle and cap can be recycled or thrown in the trash.



e.



f.

d. Hold the stool sample container on a hard, flat surface.

e. Place the lid on top of the container.
   - Make sure the lid is straight so it will close tightly.

f. Turn the lid to tighten until it does not tighten anymore.

Step 6: Label Your Samples

You **must** label your samples to identify them. For best results, use a ballpoint pen and write the labels on a hard, flat surface.


Step 6: Label Your Samples

You **must** label your samples with the information listed below. Use a ballpoint pen and write on a hard, flat surface.

EXAMPLE

John
First Name
DOE
Last Name
05/16/58
Date of Birth (mm/dd/yy)
03/12/13
Date of Collection (mm/dd/yy)
08:15 AM
Time of Collection (00:00) AM/PM

a. Fill out both labels. **Print** your information as shown in the example.

First Name
Last Name
Date of Birth (mm/dd/yy)
Date of Collection (mm/dd/yy)
Time of Collection (00:00) AM/PM

First Name
Last Name
Date of Birth (mm/dd/yy)
Date of Collection (mm/dd/yy)
Time of Collection (00:00) AM/PM

b. Peel the label away from the paper.

c. Wrap one label around the probe tube as shown.

d. Place one label on the lid of the stool sample container as shown.

Exact Sciences
Madison, WI USA
REF 300846

For questions or help call Exact Sciences at 1-8XX-XXX-XXXX
www.cologuardtest.com

Page 1 of 1
IVD

a. a. Find the page **Label Your Samples** that is included with this guide.


EXAMPLE

John
First Name
DOE
Last Name
05/16/58
Date of Birth (mm/dd/yy)
03/12/13
Date of Collection (mm/dd/yy)
08:15 AM
Time of Collection (00:00) AM/PM

b. b. Fill out **both** labels. **Print** the information in this order:

- Your first name
- Your last name
- Your birthdate (MM/DD/YY)
- The date you collected your stool sample (MM/DD/YY)
- The time you collected your stool sample including AM/PM

c.

c. Peel the label away from the paper.



d.

d. Wrap one label around the tube.

e.

e. Place one label on the lid of the stool sample container.

Step 7: Prepare Your Samples for Shipment

- For best results, prepare your box on a hard, flat surface.
- When you start, you should have the zippered bag inside the box. The tray for holding the container and tube should be inside the bag.



a. Place the tube and container into the box.

b. Fold the bag and push extra air out of the bag.
   - Too much air in the bag will make it hard to close the box.

c. Slide the zipper across the top of the bag to close and seal the bag.



d.

d. Fold the bag until it fits inside the box.

e.

e. Fold the smaller flaps over the top of the box.



f.

f. Fold the long flap over the two smaller flaps.

- This is the flap with the STOP sign printed on it.



g.

g. Peel the paper backing off the tape on the box lid.

- Discard paper backing in garbage.

h.

h. Fold the box top over the other flaps and press down firmly to seal the box.

i.

i. Locate the shipping label on top of your box.
   - Shipping labels on the box may differ.

j.



j. If the shipping label has a green tear strip, it must be removed.
   - Lift the left corner of the tear strip and pull it off.
   - Lift the bottom edge of the label and pull up.
   - Tear the top label off the box.
   - A new label now shows.
   - You are ready to ship your box.
   - Continue with the next step.

k. If the shipping label has **no** green tear strip, you are ready to ship your box.

This page intentionally left blank.
Turn the page to continue.



Step 8: Ship Your Samples Using UPS

Samples must be shipped to arrive at the lab within 3 days.

- The lab **must** receive samples within 3 days for samples to be tested. The lab will not test samples not received in time.
- UPS will ship your samples to arrive at the lab the next day from the day it shipped.
- Ship your samples **within a day** of collection for quickest delivery to the lab.

Ship your samples within a day of collection and no later than the day listed on the schedule below.

If you collected your samples on:	You must ship your samples by:
Saturday	Monday
Sunday	Tuesday
Monday	Wednesday
Tuesday	Thursday
Wednesday	Friday
Thursday	Friday
Friday	Saturday

Make sure a holiday will not delay your shipment.

If you have questions about when to ship your samples, call Exact Sciences at 1-844-870-8878.

Storing Your Samples Before Shipment

To store the samples while you plan for shipping the box:

- Store your boxed and sealed samples at room temperature.
- Keep the samples away from heat and direct sunlight.
- Keep samples out of the reach of children.
- Keep the box upright.

Ways to Ship the Samples

You have the following options for shipping your samples:

- Drop off at a UPS location. Check **www.ups.com** for UPS drop-off locations.
  - UPS Stores
  - UPS Customer Center
  - UPS Alliance Location (For Example: Staples, Office Depot)
  - UPS Authorized Shipping Outlet
  - Any UPS Driver
- Call UPS at 1-800-742-5877 for pick up or to find a drop off location near you (M-F 7:30am-12:00am, Sa 7:30am-9:00pm Eastern Standard Time).
  - You can also schedule a pick up online at www.ups.com any time.
  - Be sure to provide the tracking number shown on your prepaid return shipping label, which starts with 1Z.

Once you have shipped the samples to the lab, you are finished! Your doctor will contact you with your result.

Glossary

Adjunctive: A test used together with another test.

Colon: A part of the human digestive system sometimes known as the large intestine.

Colorectal Cancer: A disease of abnormal growths in the colon or rectum that, if left untreated, may spread throughout the body. Colon cancer generally develops over a number of years.

Colonoscopy: A medical test where a flexible tube is placed into the colon so the surface of the colon can be seen by a camera. Instruments can be placed through the tube to remove growths found on the colon wall.

DNA: Deoxyribonucleic acid. Changes in the stool DNA (sDNA) can show your doctor if there is cancer or a possibility of cancer present.

Hemoglobin: A compound from blood. Blood or hemoglobin in the stool that cannot be seen by the naked eye is referred to as occult blood or occult hemoglobin.

Neoplasia: A growth that is not normal.

Precancerous Polyp (Adenoma): A growth on the wall of the colon that may become cancer.

Symbols Used on the Cologuard Collection Kit

Collection of harmonized symbols used on labeling for the kit.



Clinical Study Results

Cologuard was studied in a large clinical trial to determine the effectiveness of the test. The trial included more than 10,000 patients at 90 sites in the US and Canada. Each person in the study completed Cologuard and a fecal occult blood test before having a standard colonoscopy. The main purpose of the study was to find out the performance of Cologuard for finding cancer and precancer compared to a colonoscopy.

In the clinical study, Cologuard correctly detected 92% of colorectal cancers and 42% of advanced adenomas in the study population that had disease. The Cologuard test correctly gave a negative screening result for 87% of the study subjects that did not have colorectal cancer or advanced adenomas. In other words, 13% of people without cancer or precancer tested positive.

Cologuard is a trademark of Exact Sciences Corporation. Refer to www.exactsciences.com/intellectual-property/ for additional information regarding Exact Sciences intellectual property holdings and policies.

©2014 Exact Sciences Corporation. All rights reserved.

EXACT SCIENCES CORPORATION
441 Charmany Drive | Madison, WI 53719 | www.exactsciences.com

For questions or help call **Exact Sciences at 1-844-870-8878**
www.cologuardtest.com
For Rx Only

PATIENT BROCHURE

LBL-0259rev1

441 Charmany Dr
Madison WI, 53719
844-870-8870
RX Only

Cologuard® colorectal cancer screening test is a registered trademark of Exact Sciences Corporation.

Indications for Use

Cologuard® is intended for the qualitative detection of colorectal neoplasia associated DNA markers and for the presence of occult hemoglobin in human stool. A positive result may indicate the presence of colorectal cancer (CRC) or advanced adenoma (AA) and should be followed by diagnostic colonoscopy. Cologuard is indicated to screen adults of either sex, 50 years or older, who are at typical average risk for CRC. Cologuard is not a replacement for diagnostic colonoscopy or surveillance colonoscopy in high risk individuals.

Contraindications

Cologuard is intended for use with patients, age 50 years and older, at average risk who are typical candidates for CRC screening. Cologuard was not clinically evaluated for the following types of patients:

- Patients with a history of colorectal cancer, adenomas, or other related cancers.
- Patients who have had a positive result from another colorectal cancer screening method within the last 6 months.
- Patients who have been diagnosed with a condition that is associated with high risk for colorectal cancer. These include but are not limited to:
    - Inflammatory Bowel Disease (IBD)
    - Chronic ulcerative colitis (CUC)
    - Crohn's disease
    - Familial adenomatous polyposis (FAP)
    - Family history of colorectal cancer
- Patients who have been diagnosed with a relevant familial (hereditary) cancer syndrome, such as Hereditary non-polyposis colorectal cancer syndrome (HNPCCC or Lynch Syndrome), Peutz-Jeghers Syndrome, MYH-Associated Polyposis (MAP), Gardner's syndrome, Turcot's (or Crail's) syndrome, Cowden's syndrome, Juvenile Polyposis, Cronkhite-Canada syndrome, Neurofibromatosis, or Familial Hyperplastic Polyposis.

Warnings and Precautions

- The performance of Cologuard has been established in a cross sectional study (i.e., single point in time). Programmatic performance of Cologuard (i.e., benefits and risks with repeated testing over an established period of time) has not been studied. Performance has not been evaluated in adults who have been previously tested with Cologuard. Non-inferiority or superiority of Cologuard programmatic sensitivity as compared to other recommended screening methods for CRC and AA has not been established.
- CRC screening guideline recommendations vary for persons over the age of 75. The decision to screen persons over the age of 75 should be made on an individualized basis in consultation with a healthcare provider. Cologuard test results should be interpreted with caution in older patients as the rate of false positive results increases with age.
- A negative Cologuard test result does not guarantee absence of cancer or advanced adenoma. Patients with a negative Cologuard test result should be advised to continue participating in a colorectal cancer screening program with another recommended screening method. The screening interval for this follow-up has not been established.
- Cologuard may produce false negative or false positive results. A false positive result occurs when Cologuard produces a positive result, even though a colonoscopy will not find cancer or precancerous polyps. A false negative result occurs when Cologuard does not detect a precancerous polyp or colorectal cancer even when a colonoscopy identifies the positive result.
- Patients should not provide a sample for Cologuard if they have diarrhea or if they have blood in their urine or stool (e.g., from bleeding hemorrhoids, bleeding cuts or wounds on their hands, rectal bleeding, or menstruation).
- To ensure the integrity of the sample, the laboratory must receive the patient specimens within 72 hours of collection. Patients should send stool samples to the laboratory according to the instructions stated in the Cologuard Patient Guide.
- Patients should be advised of the caution listed in the Cologuard Patient Guide. Patients should NOT drink the preservative liquid.
- The risks related to using the Cologuard Collection Kit are low, with no serious adverse events reported among people in a clinical trial. Patients should be careful when opening and closing the lids to avoid the risk of hand strain..

What is cancer screening?

Some types of cancer can be found before symptoms are present or when the disease is in an early stage of development. Checking for cancer (or for conditions that may lead to cancer) in people who have no symptoms is called **screening**. Screening can help doctors find and treat some types of cancer early. Generally, the earlier colorectal cancer is detected, the easier it is to treat.

Being screened for colorectal cancer is the first and most important step in finding and preventing colorectal cancer for all adults 50 years of age and older.

Important facts about colorectal cancer

Colorectal cancer is one of the most preventable, yet least prevented, cancers in the US today.[1] It is the third most diagnosed cancer, and the second leading cause of cancer deaths in both men and women 50 years of age and older. [2] Despite these facts, colorectal cancer is one of the most treatable cancers if it is found early through screening. [1] Yet, one in 3 adults 50 years of age or older are still not getting screened as recommended. [3]

Colorectal cancer can be detected early, if you are looking for it. [1] Colorectal cancer grows slowly, generally starting from small, noncancerous polyps in the colon or rectum. [1] A polyp is simply an abnormal growth in the inner wall of your colon or rectum, and is relatively common in people over 50 years of age. [1] While polyps are common and typically don't cause symptoms, some are dangerous and can turn into cancer over time. [4] If polyps are found and removed early, the chance of developing colorectal cancer can be reduced dramatically. [5] If colorectal cancer is detected in its early stages through screening, treatment is most likely to be successful. [5]

Why should I get screened for colorectal cancer?

The earlier colorectal cancer is detected, the easier it is to treat. Regular colorectal cancer screening for all adults 50 years of age and older is worth doing because it has the potential to save lives. Choosing among colorectal cancer screening tests isn't always an easy decision. That is why it is important to talk to your health care provider about when to begin screening for colorectal cancer and how to choose among the different colorectal cancer screening methods and procedures available today. There are many colorectal cancer screening methods, both invasive and noninvasive, with newer noninvasive methods that are available to everyone considered at "normal risk" for colorectal cancer.

Choosing the best colorectal cancer screening test for you

Be certain to consult your healthcare provider about your colorectal cancer screening options when choosing a colorectal cancer screening test that's right for you– you have multiple choices. Factors to discuss include:

- Your age, medical history, family history, general health
- The ability of the test to find both precancer and cancer
- How the test is performed
- If sedation is necessary
- The preparation and amount of time required to take the test
- The convenience of the test
- The potential harms of the test
- Follow-up care after the test

What is Cologuard? Why is it different?

Cologuard is an accurate noninvasive colorectal cancer screening test for men and women, 50 years of age and older, who are at average risk for colorectal cancer. Cologuard is the only colorectal screening test that uses advanced stool DNA technology and is effective in finding both precancer and cancer.

What is stool DNA technology?

Every day your colon sheds cells that line the inside of the colon. As part of this process, if precancer or cancer is present, abnormal cells will shed into the colon, along with normal cells. When you have a bowel movement, your stool picks up the shedding cells as it passes through your colon. Cologuard utilizes advanced stool DNA technology to detect the DNA and hemoglobin (red blood cells) released from abnormal cells, if present. Stool DNA technology looks for specific markers from abnormal cells and does not require analysis of your personal genetic information. Unlike other noninvasive colorectal cancer screening tests, Cologuard can detect both precancer and cancer.

Cologuard is easy to use:

- Cologuard allows you to easily collect a stool sample for testing in the privacy of your own bathroom.
- The test does not require you to follow a special diet or change your medications

Is Cologuard accurate and effective in finding precancer and cancer?

Cologuard is an effective noninvasive colorectal cancer screening test.

- Cologuard finds 92% of colon cancers.
- Cologuard is effective because it finds both advanced adenomas (precancer) and cancer
- In a large clinical study, Cologuard found more cancers and precancers than an ordinary fecal blood test (e.g. FIT).

Cologuard does produce some false positive results, so any positive should be discussed with your doctor and followed by a diagnostic colonoscopy. In the clinical study of Cologuard, Cologuard detected 92% of colorectal cancers and 42% of precancers while an ordinary fecal blood test (e.g., FIT) detected 74% of cancers and 24% of precancers. The Cologuard test correctly gave a negative screening result for 87% of the study subjects without disease, while the FIT provided accurate negative screening results for 95% of the study population without disease.

How was the effectiveness of Cologuard determined?

Cologuard was studied in a large clinical trial to determine the effectiveness of the test. The trial included more than 10,000 patients at 90 sites in the US and Canada. Individuals in the study completed Cologuard and a fecal immunochemical test before having a standard colonoscopy. The main purpose of the study was to find out how well Cologuard detects cancer and precancer compared to a colonoscopy.

I want to be screened for colorectal cancer. How do I get Cologuard?

Cologuard is prescribed through your health care provider and cannot be purchased over the counter. Once a health care provider prescribes Cologuard, Exact Sciences Laboratories makes it easy to complete the Cologuard sample collection process:

- A collection kit is sent directly to your preferred mailing address. You can store your kit until you are ready to use it - just store at room temperature in a cool, dry place. Keep away from heat and direct sunlight and out of reach of children.
- When you are ready, a sample collection can be done at home. Caution, do not collect your sample if you are experiencing any of the following:
    - Bleeding hemorrhoids
    - Bleeding cuts or wounds on your hands
    - Rectal bleeding
    - Menstrual period
    - Diarrhea

- After you have completed your stool sample collection, you drop off the collection kit at any UPS store or call to schedule a pick-up where it will be sent directly back to Exact Sciences Laboratories within 3 days of your sample collection. All postage is pre-paid.
- Exact Sciences Laboratories will test your stool sample and send the results directly back to your doctor. **Your doctor will contact you with your test results**.



Should you have any questions or concerns, Exact Sciences Laboratories offers a dedicated Customer Support Center that will be happy to assist you. You can call, toll free, and speak with a representative at 844-870-8870. More resources are also available online at www.CologuardTest.com or at your healthcare provider's office.

Understanding your results

Your doctor will talk with you about your results. The test result can be **POSITIVE or NEGATIVE.**

What does a **POSITIVE** result mean?

- A Positive result means the test detected abnormal DNA and/or blood that could be caused by precancer or cancer in the colon.
- Any Positive result should be followed by a colonoscopy.

- The test can also have a Positive result that is incorrect (false positive). This means the test result is Positive, but no cancer or precancer is actually present. Talk about your test result with your doctor to find out if additional testing is needed.

What does a **NEGATIVE** result mean?

- A Negative result means the test did not detect abnormal DNA and/or blood that could be caused by precancer or cancer in the colon.
- The test can also have a Negative result that is incorrect (false negative). This means the test result missed a potential cancer or precancer. For that reason, it is recommended that you schedule regular screenings. Your doctor may recommend an alternative screening method.
- Talk to your doctor about your test result. Your doctor will recommend a screening schedule that is best for you.

Please note that in some cases Cologuard may not generate a result. If this happens, you will be contacted and may be asked to provide another stool sample.

Cologuard may not be for everyone

Talk with your doctor if any of the following apply to you:

- A history of colorectal cancer, adenomas, or other related cancers.
- If you had a positive result from another colorectal cancer screening method within the last 6 months.
- If you have been diagnosed with a condition that places you at high risk for colorectal cancer. These include but are not limited to: Inflammatory Bowel Disease (IBD), Chronic ulcerative colitis (CUC), Crohn's disease, Familial adenomatous polyposis (FAP), Family history of colorectal cancer.
- Been diagnosed with a relevant cancer syndrome passed on from your family (hereditary), such as Hereditary non-polyposis colorectal cancer syndrome (HNPCCC or Lynch Syndrome), Peutz-Jeghers Syndrome, MYH-Associated Polyposis (MAP),Gardner's syndrome, Turcot's (or Crail's) syndrome, Cowden's syndrome, Juvenile Polyposis, Cronkhite-Canada syndrome, Neurofibromatosis, Familial Hyperplastic Polyposis.

What are the risks associated with using Cologuard?

The risks related to using the Cologuard collection kit are low. No serious adverse events were reported in the clinical trial.

- Opening or closing the lids of items in the kit may be difficult for some people. Be careful when opening and closing the lids to avoid the risk of hand strain. Close all sample containers tightly.

- Using this kit requires sitting down on the toilet and standing up from the toilet. Have someone who can help you sit down or stand up if needed.
- There is a chance that a stool sample sent to the lab may have no result. If this happens, you will be contacted. You may be asked to provide another sample.

Cologuard Precautions

- Do not drink the bottle of preservative liquid & keep away from children. Do not let the liquid touch your skin or eyes. If the liquid touches your skin or eyes, wash the area with water.

Who can I call if I have questions?

If you have questions, please call Exact Sciences Laboratories Customer Support Center. You may also find helpful resources online.

Contact Center: 1-844-870-8870
Hours of Operations: 24 hours a day, 7 days a week
www.cologuardtest.com

References

1. American Cancer Society. Can colon cancer be prevented? Available at http://www.cancer.org/cancer/colonandrectumcancer/detailedguide (accessed February 19, 2014).
2. American Cancer Society. Cancer Facts & Figures 2014. Available at http://www.cancer.org/research/cancerfactsstatistics/cancerfactsfigures2014/ (accessed February 19, 2014).
3. CDC Vital Signs. Colorectal Cancer Tests Save Lives. Available at http://www.cdc.gov/vitalsigns/colorectalcancerscreening/index.html (which references http://seer.cancer.gov) (accessed February 19, 2014).
4. Pickhart PJ et al. Clin Gastroenterol Hepatol. 2010;8(7):610-615; Amersi F et al. Clin Colon Rectal Surg. 2005;18(3):133-140; Leggett B, Whitehall V. Gastroenterology. 2010;138(6):2088-2100.
5. Zauber AG et al. N Engl J Med. 2012;366(8):687-696;

©2014 Exact Sciences Corporation

Cologuard® colorectal cancer screening test is a registered trademark of Exact Sciences Corporation.

Cologuard® Physician Brochure

Indications for Use

Cologuard is intended for the qualitative detection of colorectal neoplasia associated DNA markers and for the presence of occult hemoglobin in human stool. A positive result may indicate the presence of colorectal cancer (CRC) or advanced adenoma (AA) and should be followed by diagnostic colonoscopy. Cologuard is indicated to screen adults of either sex, 50 years or older, who are at typical average-risk for CRC. Cologuard is not a replacement for diagnostic colonoscopy or surveillance colonoscopy in high risk individuals.

Contraindications

Cologuard is intended for use with patients, age 50 years and older, at average risk who are typical candidates for CRC screening. Cologuard was not clinically evaluated for the following types of patients:

- Patients with a history of colorectal cancer, adenomas, or other related cancers.
- Patients who have had a positive result from another colorectal cancer screening method within the last 6 months.
- Patients who have been diagnosed with a condition that is associated with high risk for colorectal cancer. These include but are not limited to:
    - Inflammatory Bowel Disease (IBD)
    - Chronic ulcerative colitis (CUC)
    - Crohn's disease
    - Familial adenomatous polyposis (FAP)
    - Family history of colorectal cancer
- Patients who have been diagnosed with a relevant familial (hereditary) cancer syndrome, such as Hereditary non-polyposis colorectal cancer syndrome (HNPCCC or Lynch Syndrome), Peutz-Jeghers Syndrome, MYH-Associated Polyposis (MAP), Gardner's syndrome, Turcot's (or Crail's) syndrome, Cowden's syndrome, Juvenile Polyposis, Cronkhite-Canada syndrome, Neurofibromatosis, or Familial Hyperplastic Polyposis.

Warnings and Precautions

- The performance of Cologuard has been established in a cross sectional study (i.e., single point in time). Programmatic performance of Cologuard (i.e., benefits and risks with repeated testing over an established period of time) has not been studied. Performance has not been evaluated in adults who have been previously tested with Cologuard. Non-inferiority or superiority of Cologuard programmatic sensitivity as compared to other recommended screening methods for CRC and AA has not been established.
- CRC screening guideline recommendations vary for persons over the age of 75. The decision to screen persons over the age of 75 should be made on an individualized basis in consultation with a healthcare provider. Cologuard test results should be interpreted with caution in older patients as the rate of false positive results increases with age.
- A negative Cologuard test result does not guarantee absence of cancer or advanced adenoma. Patients with a negative Cologuard test result should be advised to continue participating in a colorectal cancer screening program with another recommended screening method. The screening interval for this follow-up has not been established.
- Cologuard may produce false negative or false positive results. A false positive result occurs when Cologuard produces a positive result, even though a colonoscopy will not find cancer or precancerous

polyps. A false negative result occurs when Cologuard does not detect a precancerous polyp or colorectal cancer even when a colonoscopy identifies the positive result.

- Patients should not provide a sample for Cologuard if they have diarrhea or if they have blood in their urine or stool (e.g., from bleeding hemorrhoids, bleeding cuts or wounds on their hands, rectal bleeding, or menstruation).
- To ensure the integrity of the sample, the laboratory must receive the patient specimens within 72 hours of collection. Patients should send stool samples to the laboratory according to the instructions stated in the Cologuard Patient Guide.
- Patients should be advised of the caution listed in the Cologuard Patient Guide. Patients should NOT drink the preservative liquid.
- The risks related to using the Cologuard Collection Kit are low, with no serious adverse events reported among people in a clinical trial. Patients should be careful when opening and closing the lids to avoid the risk of hand strain.

RX Only

Table of Contents

Indications for Use 1
Contraindications 1
Warnings and Precautions 1
Table of Contents 2
Cologuard Overview 3
Patient Samples for Cologuard 3
Cologuard Compliance Program 3
Colorectal Cancer Overview 3
Device Description 4
Assay Technology 4
Clinical Study: Multi-Target Colorectal Cancer Screening Test for the Detection of Colorectal Advanced Adenomatous Polyps and Cancer (DeeP-C) 4
Overview 4
Study Design 5
Study Population and Baseline Demographics 5
Clinical Performance Measures 6
Summary of Clinical Study Results 6
Cologuard and FIT Performance Comparison 7
Cologuard Subgroup Analysis: please note that the clinical study was not designed to evaluate subgroups and subgroup analysis should be interpreted with that in mind. 8
Ordering Cologuard 10
Sample Collection 10
Interfering Substances 10
Instructions for Sample Collection 10
Interpretation of Cologuard Assay Results 11
References 12

Cologuard Overview

Cologuard uses advanced multiple-marker, stool DNA technology to detect colorectal cancer (CRC) and advanced adenomas (AA). Cologuard is 92% sensitive for detection of CRC. Cologuard is a statistically superior noninvasive stool test for detecting CRC and AA, as shown in a head-to-head, cross-sectional clinical study of Cologuard and a commercially available fecal immunochemical test (OC FIT-CHEK, Polymedco, Inc.) ("FIT"). In the study, Cologuard specificity was 87% (the specificity calculation excluded both CRC and AA), which is lower than that of FIT.

Cologuard is designed to analyze patients' stool for the presence of 11 molecular markers, including hemoglobin and DNA markers, which may indicate the presence of colorectal cancer or advanced adenomas. Because cellular exfoliation of DNA into stool occurs continuously, Cologuard can detect pre-malignant neoplasia at early onset of abnormality.

Based on combined results of all of the DNA markers and hemoglobin, a single Cologuard result is determined. Cologuard results are qualitative, positive or negative. A patient with a positive result should be referred to a diagnostic colonoscopy. A patient with a negative result should continue with a regular screening schedule. If no result is obtained, a second stool collection may be requested.

Patient Samples for Cologuard

Patients are not required to undergo bowel preparation or follow dietary or medication restrictions in order to complete the test. Patients follow the detailed instructions in the Cologuard Patient Guide received with the collection kit, consisting of a container for collection of stool for DNA testing and a separate sampler for collection of stool for hemoglobin testing. Both of these stool samples are required to obtain a Cologuard result. Samples are sent to a qualified laboratory for processing and testing.

Cologuard Compliance Program

Cologuard includes a compliance program to handle collection kit shipment to the patient's home in addition to live representatives for patient support, patient reminders, and billing and reimbursement questions. The compliance program also provides compliance tracking for physicians to measure and improve patient compliance.

Colorectal Cancer Overview

Colorectal cancer (CRC) is the second leading cause of death from cancers affecting both men and women in the United States. One in 17 Americans will suffer from CRC during their lifetime; the lifetime risk is 35% higher for men than for women.[1] Early detection by screening has been shown to reduce CRC mortality.[2,3,4] Current guidelines for CRC screening in the average-risk population recommend initiation of screening at age 50 (age 45 for African Americans), as the incidence of both CRC and premalignant lesions increases sharply after this age.[5]

Detection of potentially pre-malignant lesions, also known as advanced adenomas (AA), is essential for CRC prevention. Advanced adenomas include any size adenomas with carcinomas in situ or high grade dysplasia (HGD), adenomas with villous growth patterns (>25%), or adenoma ≥1.0 cm in size.[6,7,8,9] Serrated lesions (polyps and sessile serrated adenoma) are typically found in the proximal colon, occur more frequently in the elderly, are often flat and inconspicuous endoscopically, and may have a more aggressive natural history than classic colorectal adenomas.[9]

Device Description

Cologuard utilizes a multi-target approach to detect DNA and hemoglobin markers associated with CRC, as well as pre-malignant colorectal neoplasia (i.e., AA). Three independent categories of biomarkers are targeted and provide an additive association with CRC and pre-malignant colorectal neoplasia

The first category of biomarkers involves epigenetic DNA changes characterized by aberrant gene promoter region methylation. The specific methylated gene targets include N-Myc Downstream-Regulated Gene 4 (NDRG4) and the Bone Morphogenetic Protein 3 (BMP3).[10,11] NDRG4 and BMP3 have been shown to be hypermethylated in colorectal cancer.[1,10] The Cologuard procedure incorporates bisulfite conversion of non-methylated cytosine residues to uracil in the DNA sequence to enable sensitive detection of hypermethylated NDRG4 and BMP3.

The second category targets specific DNA point mutations in the v-Ki-ras2 Kirsten rat sarcoma viral oncogene homolog (KRAS) gene, which encodes a small GTPase that is activated transiently as a response to extracellular stimuli or signals.[12,13,14] KRAS mutations have been detected in up to 35% of colorectal cancers and the 7 mutations in Exon 2 detected by Cologuard account for 98% of KRAS mutations.[16] KRAS mutations, along with NDRG4 and BMP3 methylation markers, have been shown to be detected in the stool of subjects with colorectal neoplasia, including subjects with colorectal cancer and pre-malignant lesions. [15,16]

The third category of biomarker is non-DNA based and detects hemoglobin, which can be associated with colonic bleeding. Results from the methylation, mutation, and hemoglobin assays are combined in the laboratory analysis to determine a positive or negative reportable result or no result.

Assay Technology

The patient stool samples are processed at the laboratory to isolate the DNA for testing. Amplification and detection of methylated target DNA (NDRG4, BMP3), KRAS point mutations, and ACTB (a reference gene for quantitative estimation of the total amount of human DNA in each sample) is performed using the Quantitative Allele-specific Real-time Target and Signal Amplification (QuARTS™) technology. Multi-plexed QuARTS reactions are processed using a real-time cycler with each marker (NDRG4, BMP3, KRAS, and ACTB) monitored separately through independent fluorescent detection channels. The hemoglobin stool sample is prepared and analyzed in a quantitative Enzyme-Linked Immunosorbent Assay (ELISA) that determines the concentration of hemoglobin in the sample.

Run control samples for both the QuARTS assays and hemoglobin assay are tested along with patient samples to show that the process has been performed appropriately. Results from the methylation, mutation, and hemoglobin assays are combined during analysis to determine a positive result, negative result, or no result.

Clinical Study: Multi-Target Colorectal Cancer Screening Test for the Detection of Colorectal Advanced Adenomatous Polyps and Cancer (DeeP-C)

Overview

Cologuard was the subject of a prospective, multi-centered, pivotal trial ("Multi-Target Colorectal Cancer Screening Test for the Detection of Colorectal Advanced Adenomatous Polyps and Cancer: DeeP-C Study"). A total of 12,776 patients were enrolled from 90 sites, including both colonoscopy centers and primary care sites. The results of the study demonstrated the safety and effectiveness of Cologuard as a screening test for the detection of markers associated with the presence of CRC and colorectal neoplasia. Cologuard demonstrated 92.3% CRC sensitivity and 86.6% specificity (specificity in this study excludes CRC and AA), using

colonoscopy with histopathological confirmation as the reference method. These results met the protocol-specified criteria for primary performance measures and study success. The study results exceeded the prospectively specified sensitivity threshold by nearly 20%. The study further compared CRC and AA detection by Cologuard to a commercially available fecal immunochemical test (OC FIT-CHEK, Polymedco, Inc.) ("FIT"), successfully demonstrating superiority for CRC (p=0.0018) and AA (p<0.0001) sensitivity.

Study Design

The study was designed to enroll subjects of either sex between the ages of 50 and 84 years (inclusive), who were at average risk for development of colorectal cancer and asymptomatic for gastrointestinal symptoms warranting diagnostic colonoscopy. In addition, subject enrollment was age-weighted toward a slightly older population to increase the point prevalence of colorectal cancer in this study. 64% of subjects in the actual study population were of age 65-84.

Subjects participating in the pivotal trial provided a stool sample and subsequently underwent colonoscopy within 90 days of study enrollment. Subjects collected stool samples for Cologuard and FIT testing at home. Subjects then underwent colonoscopy per standard of care. Subjects and physicians remained blinded to the results of Cologuard and the FIT. Results from Cologuard and the FIT test were compared to the results of the colonoscopy examination and histopathologic diagnosis of all significant lesions either biopsied or removed.

Negative colonoscopy findings were categorized as negative (Table 1, category 6.2). Histopathological results from biopsied tissue or excised lesions were categorized based on the most clinically significant lesion present (i.e. the index lesion) by a central pathologist according to the pre-specified standards outlined in Table 1. Sensitivity analysis was performed using positive findings in categories 1 and 2 while specificity was calculated using categories 3 through 6 (all findings excluding CRC and AA).

Table 1: Category definitions

Category	Findings
1	CRC, all stages (I-IV)
2	Advance adenoma, including the following subcategories: 2.1 – Adenoma with carcinoma in situ/high grade dyplasia, any size 2.2 – Adenoma, villous growth pattern (>25%), any size 2.3 – Adenoma > 1.0 cm in size, or 2.4 – Serrated lesion, > 1.0 cm in size
3	1 or 2 adenoma (s), >5 mm in size, or < 10 mm size, non-advanced
4	> 3 adenomas, <10mm, non-advanced
5	1 or 2 adenoma(s), ≤5 mm in size, non-advanced
6	Negative – No neoplastic findings 6.1 – negative upon histopathological review 6.2 – no findings on colonoscopy, no histopathological review

Study Population and Baseline Demographics

Study enrollment and population demographics are summarized in Figure 1. A total of 10,023 subjects with colonoscopy and Cologuard data were included in the primary analysis population. This population included 65 subjects with CRC. Analysis was conducted to rule out bias associated with the subjects excluded from the analysis population.

The average age of subjects included in the primary analysis was 64.2 years, and there were a slightly higher percentage of female subjects (5,378/10,023, 53.7%) as compared with male subjects (4,645 /10,023, 46.3%). Two 49-year-old subjects and one 44-year old subject were included in the study, which is inconsistent with the intended user population. Each of these subjects was a true negative on Cologuard and their inclusion did not

notably impact data analyses. The majority of subjects were White (8,422/10,023, 84.1%), although 10.7% of the population were Black or African American subjects (1,071/10,023). Nearly 10% of subjects were Hispanic or Latino (991/10,023, 9.9%). Average BMI was 28.8 and the majority of subjects never smoked (5.531 /10,023, 55.2%).

Figure 1: Clinical Study Demographics

Total Enrollment	**12,776**
Primary Endpoint Valid Cologuard + Colonoscopy	**10,023**
Secondary Endpoint Valid Cologuard + FIT + Colonoscopy	**9,989**

Demographics		
Age (Years)	Average	64.2
	Range	44-84
Gender	Male	46.3%
	Female	53.7%
BMI	Average	28.8
	Range	13.3-68.2
Ethnicity	Hispanic/Latino	9.9%
	Non Hispanic/Latino	90.1%
Race	White	84.1%
	Black	10.7%
	Asian	2.6%
	Amer. Ind/Alaskan Native	0.4%
	Native Hawaiian/Pacific Islands	0.2%
	Other	2.1%
Smoking History	Never Smoked	55.2%
	Former Smoker	35.8%
	Smoker	9.0%

Clinical Performance Measures

The primary and secondary performance measures for the clinical study are summarized in Table 2 below. The primary performance measures were the sensitivity and specificity of Cologuard for CRC, using colonoscopy with histopathology as the reference method. The primary analysis required that the lower bound of the 95% one-sided confidence interval for the sensitivity of Cologuard for CRC exceed 65%. The specificity analysis for CRC required that the lower bound of the one-sided 95% confidence interval exceed 85%.

With respect to the secondary performance measure, Cologuard was compared to FIT using a non-inferiority test for CRC sensitivity and using a superiority test for advanced adenoma (AA) sensitivity. In order for Cologuard to be deemed non-inferior to FIT, the one-sided 95% confidence interval lower bound for the Cologuard – FIT difference in percentages with a positive test among subjects with CRC was required to exceed -5%. Establishing superiority required a one-sided p-value <0.025 (exact McNemar's comparison test).

Table 2: Clinical Study Primary and Secondary Performance Measures

Primary Performance measures	• Determine the CRC sensitivity and specificity of Cologuard.
Secondary Performance measures	• Compare Cologuard to FIT for CRC and AA sensitivity.

Summary of Clinical Study Results

Results from the clinical study demonstrated that Cologuard successfully met the primary performance measure of the study, establishing a clinically meaningful sensitivity and specificity for CRC. Sensitivity of Cologuard for CRC was 92.3% (60/65) with a one-sided 95% confidence interval lower bound of 84.5,

substantially exceeding the protocol-specified threshold of 65%. In addition, Cologuard successfully demonstrated a clinically meaningful specificity according to the protocol-specified criteria. The specificity of Cologuard was 86.6%, with a one-sided 95% confidence interval lower bound of >86.0%.

Clinical study results demonstrated superiority of Cologuard to FIT for sensitivity in detecting CRC. Secondary performance measures included an analysis of performance Cologuard and FIT using colonoscopy as a reference. Cologuard correctly detected 60 of the 65 total CRC cases identified by colonoscopy (92.3%). FIT captured only 48 of the 65 CRC cases identified by colonoscopy (73.8%). FIT identified only a single cancer that was not identified by Cologuard. Cologuard, meanwhile, identified 13 cancers that were missed by FIT. Cologuard was compared to FIT using a non-inferiority test for CRC sensitivity. In addition, Cologuard demonstrated superiority over FIT with respect to sensitivity for CRC using an exact McNemar's comparison test as the one-sided p-value (p=0.0018) was well below the p <0.025 threshold for superiority. The lower bound of the one-sided confidence interval for the Cologuard – FIT difference was 0.080, substantially exceeding the protocol-specified non-inferiority threshold of -0.05.

Establishing superiority for AA sensitivity required a one-sided p-value <0.025 (exact McNemar's comparison test). Cologuard demonstrated superiority for AA sensitivity, with a p-value of <0.0001, substantially below the threshold for superiority of p<0.025. FIT identified only 29 AA cases that were not captured by Cologuard, while Cologuard identified 170 AA cases that were not positive on the FIT test.

Analysis was also performed to calculate the Cologuard negative predictive value (NPV) for Category 1 (CRC) and Category 2 (AA). Clinical results show that a negative patient result for Cologuard gives 99.94% assurance that the patient does not have cancer and a 94.79% chance that the patient does not have an advanced adenoma.

Cologuard and FIT Performance Comparison

Cologuard was superior to FIT for both CRC and AA detection. Cologuard also demonstrated high sensitivity for detection of lesions and polyps which historically have been difficult to capture with FIT, including early stage CRC, proximal lesions, and higher risk precancerous lesions. Cologuard demonstrated a numerically greater sensitivity than FIT for detection of CRC and AA across lesion subgroups. Sensitivity results are summarized in Table 3 and Table 4 below. As noted above, Cologuard specificity was 86.6% and FIT specificity was 95%. These specificity measures excluded CRC and AA for both tests.

Cologuard sensitivity for stage I cancer was 89.7% compared to 65.5% for FIT (p=0.039). Sensitivity for stage II cancer was 100.0% for Cologuard compared to 76.2% for FIT (p=0.062). CRC sensitivity was also compared to FIT by size of the lesion, with higher detection at each lesion size than FIT. When analyzed by lesion location, Cologuard showed 90.0% sensitivity for proximal cancer compared to 66.7% for FIT (p=0.039). Cologuard also detected higher risk precancerous lesions, including high grade dysplasia (69.2% Cologuard, 46.2% FIT, p=0.004) and sessile serrated polyps (43.0% Cologuard, 5.1% FIT, p<0.001). Cologuard and FIT were both better at detecting precancerous lesions as lesion size increased from 0.5 cm to ≥3 cm (value for trend for both was p<0.001).

Table 3: Cologuard and FIT Cancer Sensitivity

Subgroup	n=	Cologuard Sensitivity	FIT Sensitivity
Cancer Stage			
CRC, all stages (p=0.018)	65	92.3%	73.8%
Stage I (p=0.039)	29	89.7%	65.5%

Stage II (p=0.062)	21	100.0%	76.2%
Stage III	10	90.0%	90.0%
Stage IV	4	75.0%	75.0%
Stage I-III (p=0.002)	60	93.3%	73.3%
Cancer Size			
< 5 mm	0	0	0
5-9 mm	5	80.0%	60.0%
10-19 mm	14	92.9%	85.7%
20-29 mm	12	91.7%	66.7%
≥30 mm	34	94.1%	73.5%
Cancer location			
Proximal (p=0.039)	30	90.0%	66.7%
Distal (p=0.062)	35	94.3%	80.0%

*Cologuard specificity was 86.6% and FIT specificity was 95%. These specificity measures excluded CRC and AA for both tests.

Table 4: Cologuard and FIT Advanced Adenoma Sensitivity

Subgroup	Cologuard n=	Cologuard Sensitivity	FIT n=	FIT Sensitivity
Pre-malignant Neoplasia				
AA, all subcategories (p<0.001)	760	42.4%	757	23.8%
High grade dysplasia (p-0.004)	39	69.2%	39	46.2%
Sessile serrated ≥10 mm (p<0.001)	100	43.0%	99	5.1%
AA location				
Proximal (p<0.001)	433	33.0%	431	15.5%
Distal (p<0.001)	326	54.6%	325	34.8%
Lesion Size	p value for trend<0.001		p value for trend<0.001	
< 5 mm	10	20.0%	10	20.0%
5-9 mm	56	32.1%	56	14.3%
10-19 mm	577	39.0%	574	20.9%
20-29 mm	79	64.6%	79	43.0%
≥30 mm	38	68.4%	38	42.1%

*Cologuard specificity was 86.6% and FIT specificity was 95%. These specificity measures excluded CRC and AA for both tests.

Cologuard Subgroup Analysis: please note that the clinical study was not designed to evaluate subgroups and subgroup analysis should be interpreted with that in mind.

The clinical study results were analyzed according to various demographic characteristics, including gender, age, and race/ethnicity as summarized in Table 5 below. Although CRC sensitivity was higher for males versus females and higher in Whites and Asians compared to Black/African Americans, AA sensitivity and specificity remained consistent across subgroups, with only a few differences likely attributed to a lower number of subjects from all subpopulations in the study.

Cologuard CRC sensitivity was higher for males versus females. Meanwhile, specificity of Cologuard was similar for females as compared with males. Specificity was 87.3% (4,398/5,037) for females, compared with

85.8% (3,569/4,161) for male subjects. Advanced adenoma detection showed similar results between males and females.

For age, Cologuard sensitivity for CRC was consistently high across all age groups. Sensitivity for patients 65 years of age and older ranged from 88.9% to 100.0%. Although sensitivity was 75% for subjects age 60-64, the number of CRC cases was particularly small in this age group (n = 4); only one CRC case was not detected by Cologuard. With respect to AA, sensitivity was similar across all age groups, with sensitivity as high as 46.8% for subjects between the ages of 70 and 79. Cologuard specificity for CRC was also high across all age groups. Specificity was in the 80% range or above for most age groups, aside from subjects > 75 years old. Specificity for AA was also similar across age groups.

Cologuard CRC sensitivity was very high among White subjects, but lower among Black or African-American subjects) and high among the small number of Asian CRC cases. However, the results observed in Black/African-American subjects may have been affected by the low overall number of cancer cases in that subpopulation. Sensitivity among Hispanic or Latino subjects was high, although the sample size was small.

Cologuard sensitivity for AA was similar for White and Black/African-American subjects. Sensitivity was also similar among Hispanic/Latino subjects. AA sensitivity was lower among Asian subjects and very high for American Indian or Alaskan Natives, compared with other groups. Only the American Indian and Alaskan Native subpopulations showed higher sensitivity in AA detection. Differences between racial and ethnic subpopulation results may be affected by the small number of subjects in the African American and American Indian or Alaska Native subpopulations. Cologuard specificity was high across all racial and ethnic groups, with rates > 85% for most groups.

Table 5: Cologuard Performance by Subgroup

Subgroup	CRC Sensitivity	AA sensitivity	Specificity
Gender			
Male	34/34 (100%)	201/450 (44.7%)	3569/4161 (85.8%)
Female	26/31 (83.9%)	121/310 (39%)	4398/5037 (87.3%)
Age			
<60 yrs	7/7 (100.0%)	65/171 (38.0%)	2491/2703 (92.2%)
60-64 yrs	3/4 (75.0%)	24/57 (42.1%)	681/765 (89.0%)
65-69 yrs	19/20 (95.0%)	125/301 (41.5%)	2871/3352 (85.7%)
70-74 yrs	16/18 (88.9%)	72/154 (46.8%)	1292/1566 (82.5%)
75-79 yrs	6/6 (100.0%)	29/62 (46.8%)	480/617 (77.8%)
>79 yrs	9/10 (90.0%)	7/15 (46.7%)	152/195 (77.9%)
Race			
White	53/55 (96.4%)	271/641 (42.3%)	6639/7726 (85.9%)
Black or African American	5/8 (62.5%)	36/85 (42.4%)	879/978 (89.9%)
Asian	1/1 (100.0%)	4/13 (30.8%)	229/245 (93.5%)
American Indian or Alaska Native	0/0	3/4 (75.0%)	24/32 (75.0%)
Native Hawaiian or Other Pacific Islander	0/0	0/0	21/23 (91.3%)
Other	1/1 (100.0%)	7/16 (43.8%)	171/189 (90.5%)
Ethnicity			
Hispanic or Latino	8/9 (88.9%)	23/59 (39.0%)	837/923 (90.7%)
Not Hispanic or Latino	52/56 (92.9%)	298/700 (42.6%)	7127/8272 (86.2%)

Ordering Cologuard

Cologuard is available for physicians to order through the Exact Sciences Laboratories online portal at www.CologuardTest.com or through paper requisition. Cologuard includes a compliance program and provides attentive service to physicians and patients with live operators. For any questions about Cologuard or specific questions on how to order the test, please contact Exact Sciences Laboratories.

Exact Sciences Laboratories
145 E. Badger Rd, Suite 100
Madison, WI 53713
844-870-8870

Sample Collection

- Samples for use with Cologuard must be collected with the Cologuard Collection Kit (Exact Sciences, 100026), including a stool sample for DNA testing (Container) and a stool sample for Hemoglobin testing (Tube).
- Patients should not provide a sample if they have diarrhea or blood in their urine or stool from bleeding hemorrhoids, bleeding cuts or wounds on their hands, rectal bleeding, or menstruation.
- Patients should familiarize themselves with detailed information contained in Cologuard Patient Guide and collection instructions before completing sample collection.
- The use of this kit requires sitting down on the toilet and standing up from the toilet. Patients should have someone available to help them sit down or stand up if needed.
- To ensure the integrity of the sample, the laboratory must receive patient specimens within 72 hours of collection. Patients should send stool samples to the laboratory according to the instructions stated in the Cologuard Patient Guide.

Interfering Substances

There are no known interfering substances with Cologuard. The molecular and hemoglobin assays of the test were challenged independently with the substances that could potentially be found in patient samples, including common lotions and creams, feminine over the counter products, stool softeners, anti-diarrhea products, laxatives, anti-acids, upset stomach relief products, urine, alcohol, common vegetables and fruits, fats, and lipids. There was no observed interference with any substance in either assay. The hemoglobin assay was also tested with antibiotics, anti-inflammatories, anti-fungal drugs, pain relievers, and decongestants with no observed interference. The molecular assay was additionally tested with animal genomic DNA of commonly edible animals (both high and low levels) with no observed interference.

Instructions for Sample Collection

Once the Cologuard test has been ordered, the collection kit will be sent to the patient at their home. Detailed instructions for patient specimen collection are provided in the Cologuard Patient Guide as part of the collection kit. Full closure of the stool collection container should be emphasized to patients to ensure receipt of a usable sample for testing. A toll free number is also provided with the patient guide to ensure that any patient questions are addressed. An overview of the collection process is provided in the figure below.

Figure 5: Patient sample collection process



Interpretation of Cologuard Results

A negative test result means that the test did not detect abnormal DNA and/or blood in the sample. A test can also have a negative result that is incorrect (false negative). For that reason, it is important to continue a regular screening schedule with your patients. A positive Cologuard test means that the test detected abnormal DNA and/or blood that could be caused by precancer or cancer in the colon or rectum. A test can also have a positive test that is incorrect (false positive). Any positive result should be followed by a diagnostic colonoscopy. In some cases Cologuard may not generate a result. If this occurs a new patient sample may be requested.

References

1. Jemal A, R Siegel, E Ward, et al. 2010. Cancer statistics, 2010. CA Cancer J Clin, 60: 277-300.
2. Mandel JS, JH Bond, TR Church, et al. 1993. Reducing mortality from colorectal cancer by screening for fecal occult blood. Minnesota colon cancer control study. New Engl J Med, 328: 1365-1371.
3. Hardcastle JD, JO Chamberlain, MH Robinson, et al. 1996. Randomized controlled trial of fecal-occult-blood screening for colorectal cancer. Lancet, 348: 1467-1471.
4. Olson J, DH Whitney, K Durkee, and AP Shuber. 2005. DNA stabilization is critical for maximizing performance of fecal DNA-based colorectal cancer tests. DiagnMolPathol, 14: 183-191.
5. Rex D, Johnson D, Anderson J, et al. 2009. American College of Gastroenterology Guidelines for Colorectal Cancer Screening 2008. Am J Gastroenterol 2009;104:739–750.
6. Hussain FT, TC Yab, JJ Harrington, et al. 2010. Noninvasive detection of serrated colorectal polyps by stool assay of methylated vimentin and mutant BRAF genes. In DDW 2010, New Orleans, 2010.
7. Legett, B and Whitehall A. 2010. Role of the serrated pathway in colorectal cancer pathogenesis, Gastroenterology,138:2088-2100.
8. Kim DH, Pickhardt PJ, Taylor AJ. 2007. Characteristics of Advanced Adenomas Detected at CT Colonographic Screening: Implications for Appropriate Polyp Size Thresholds for Polypectomy Versus Surveillance *American J Radiology*, 188:940-944.
9. Noffsinger AE. 2009. Serrated polyps and colorectal cancer: new pathway to malignancy. Annual Rev Pathol, 4: 343-364.
10. Melotte, V, Lentjes, MHFM, van den Bosch SM , et al. 2009. N-Myc Downstream-Regulated Gene 4 (*NDRG4*): A Candidate Tumor Suppressor Gene and Potential Biomarker for Colorectal Cancer V, JNCI J Natl Cancer Inst, Volume101, Issue13.916-927.
11. Zou H, Harrington JH, Abdirashid M, et al. 2007. Highly Methylated Genes in Colorectal Neoplasia: Implications for Screening, Cancer Epidemiology, Biomarkers and Prevention;16:2686.
12. Bos J, Fearon E, Hamilton S, et al. 1987. Prevalence of ras gene mutations in human colorectal cancers. Nature 327, 293-297.
13. Malumbres M, Barbacid M. 2003. RAS oncogenes: the first 30 years. Nat Rev Cancer, 3:459
14. Downward J. 2033. Targeting RAS signaling pathways in cancer therapy. Nat Rev Cancer, 3:11-22
15. Ahlquist DA, Zou H, Domanico M, et al. 2010. Next generation stool DNA testing for detection of colorectal neoplasia: early marker evaluation. Presented at American Association of Cancer Research, Oct. 29, 2010, Philadelphia, PA.
16. miRBase Sequence database, release 13.0. ftp://ftp.sanger.ac.uk/pub/mirbase/sequences/.

Cologuard™

sDNA-based Colorectal Cancer Screening Test

Instructions for Use

IVD

For *in vitro* diagnostic use

Exact Sciences Corporation

441 Charmany Drive

Madison, WI 53719

www.exactsciences.com

(800) XXX-XXXX

Table of Contents

Intended Use and Indications for Use 6
Contraindications 6
Warnings and Precautions 6
Summary and Explanation of the Test 7
Principles of the Procedure 8
Reagents 9
DNA Capture Reagents (Exact Sciences, 100028) 9
DNA Preparation Reagents (Exact Sciences, 100029) 10
QuARTS Assay Reagents (Exact Sciences, 100030) 10
Cologuard DNA Control Kit (Exact Sciences, 100074) 10
Hemoglobin Assay Reagents (Exact Sciences, 100031) 11
Cologuard Hemoglobin Control Kit (Exact Sciences, 100073) 11
Ancillary Materials and Bulk Assay Reagents 11
Procedural Warnings and Precautions 12
Reagent Warnings and Precautions 12
Instrument Warnings and Precautions 13
Reagent Storage and Handling Requirements 13
Instrumentation 14
Specimen Collection and Preparation for Analysis 15
Cologuard Laboratory Procedure 15
Receipt of *Cologuard* Collection Kit 15
Preparation of Stool Homogenate for DNA Testing 16
Assay Overview 17
Example *Cologuard* Setup 18
DNA Capture 19
Prepare Capture Beads 19
Prepare Samples and Perform DNA Capture 19
Prepare and Label Sample Tubes 19
Prepare Supernatant 20
Capture Incubation 21

DNA Preparation and *QuARTS* Assay 22
Automated DNA Preparation and *QuARTS* Assay Setup 23
Reagent Preparation 23
STARlet Setup 23
DNA Preparation 28
QuARTS Plate Setup 28
Run the *QuARTS* Plate 29
Hemoglobin Assay 33
Preparation of Hemoglobin Samples and Reagents 33
Prepare Samples 33
Prepare Hemoglobin Assay Reagents and STARlet 34
STARlet Setup for Hemoglobin Assay 35
Wash Hemoglobin Assay Plate 35
Automated Hemoglobin Plate Setup 36
Hemoglobin Assay Procedure 36
Read Hemoglobin Plate 37
Hemoglobin Sample Storage 38
Data Handling and Analysis 39
Review and Release Methylation, Mutation, or Hemoglobin Assay Results 39
Review and Release Overall *Cologuard* Results 42
Interpretation of Results 46
Procedural Notes and Precautions 46
Additional Stool Homogenate Aliquots 46
Insufficient Supernatant 46
Enter Supplementary Lot Information 47
Prepare Capture Beads (for >6 tubes) 47
Procedure for Performing Partial Runs 47
Quality Control 48
Process Controls 48

Lot Matching and Sample Tracking during Processing 48
Review and Release *Cologuard* Results.....48
Troubleshooting Guide.....49
DNA Capture 49
Centrifuged Stool Sample Abnormalities 49
Incomplete Dispersion of Inhibitor Removal Tablet.....49
Broken Spin Filter.....49
Capture Incubator Produces Error Message.....49
Capture Aspirator Produces Error Message 50
Incomplete Aspiration of Supernatant.....50
Sample Appears Gelatinous After Aspiration.....50
Beads Not Fully Suspended After Shaking 50
DNA Preparation and *QuARTS* Assay 50
Methylation/Mutation Setup Run Aborts.....50
Hemoglobin Assay 51
Hemoglobin Setup Run Aborts 51
Retest Hemoglobin Sample Tubes 51
Procedural Limitations.....51
Performance Characteristics.....52
Clinical Cutoff 52
Analytical Sensitivity.....52
Specificity and Cross-Reactivity.....53
Precision and Reproducibility 55
Robustness 59
Interfering Substances.....60
Carry-over and Cross-contamination Cologuard Testing 61
Stability Studies.....62
Clinical Sensitivity and Specificity.....64
Abbreviations Used.....83

Key Symbols Used....83
References....84
Contact Information....84
Trademarks....84
Patents....84
Limited License....85
Warranty....85
Copyright....85

Intended Use and Indications for Use

Intended Use

Cologuard is intended for the qualitative detection of colorectal neoplasia associated DNA markers and for the presence of occult hemoglobin in human stool. Cologuard is for use with the Cologuard collection kit and the following instruments: BioTek ELx808 Absorbance Microplate Reader; Applied Biosystems® 7500 Fast Dx Real-Time PCR; Hamilton Microlab® STARlet; and the Exact Sciences System Software with Cologuard Test Definition.

Indications for Use

Cologuard is intended for the qualitative detection of colorectal neoplasia associated DNA markers and for the presence of occult hemoglobin in human stool. A positive result may indicate the presence of colorectal cancer (CRC) or advanced adenoma (AA) and should be followed by diagnostic colonoscopy. Cologuard is indicated to screen adults of either sex, 50 years or older, who are at typical average-risk for CRC. Cologuard is not a replacement for diagnostic colonoscopy or surveillance colonoscopy in high risk individuals.

Contraindications

Cologuard is intended for use with patients, age 50 years and older, at average risk who are typical candidates for CRC screening. Cologuard was not clinically evaluated for the following types of patients:

- Patients with a history of colorectal cancer, adenomas, or other related cancers.
- Patients who have had a positive result from another colorectal cancer screening method within the last 6 months.
- Patients who have been diagnosed with a condition that is associated with high risk for colorectal cancer. These include but are not limited to:
    - Inflammatory Bowel Disease (IBD)
    - Chronic ulcerative colitis (CUC)
    - Crohn's disease
    - Familial adenomatous polyposis (FAP)
    - Family history of colorectal cancer
- Patients who have been diagnosed with a relevant familial (hereditary) cancer syndrome, such as Hereditary non-polyposis colorectal cancer syndrome (HNPCCC or Lynch Syndrome), Peutz-Jeghers Syndrome, MYH-Associated Polyposis (MAP), Gardner's syndrome, Turcot's (or Crail's) syndrome, Cowden's syndrome, Juvenile Polyposis, Cronkhite-Canada syndrome, Neurofibromatosis, or Familial Hyperplastic Polyposis.

Warnings and Precautions

- The performance of Cologuard has been established in a cross sectional study (i.e., single point in time). Programmatic performance of Cologuard (i.e., benefits and risks with repeated testing over an established period of time) has not been studied.

Performance has not been evaluated in adults who have been previously tested with Cologuard. Non-inferiority or superiority of Cologuard programmatic sensitivity as compared to other recommended screening methods for CRC and AA has not been established.

- CRC screening guideline recommendations vary for persons over the age of 75. The decision to screen persons over the age of 75 should be made on an individualized basis in consultation with a healthcare provider. Cologuard test results should be interpreted with caution in older patients as the rate of false positive results increases with age.
- A negative Cologuard test result does not guarantee absence of cancer or advanced adenoma. Patients with a negative Cologuard test result should be advised to continue participating in a colorectal cancer screening program with another recommended screening method. The screening interval for this follow-up has not been established.
- Cologuard may produce false negative or false positive results. A false positive result occurs when Cologuard produces a positive result, even though a colonoscopy will not find cancer or precancerous polyps. A false negative result occurs when Cologuard does not detect a precancerous polyp or colorectal cancer even when a colonoscopy identifies the positive result.
- Patients should not provide a sample for Cologuard if they have diarrhea or if they have blood in their urine or stool (e.g., from bleeding hemorrhoids, bleeding cuts or wounds on their hands, rectal bleeding, or menstruation).
- To ensure the integrity of the sample, the laboratory must receive the patient specimens within 72 hours of collection. Patients should send stool samples to the laboratory according to the instructions stated in the Cologuard Patient Guide.
- Patients should be advised of the caution listed in the Cologuard Patient Guide. Patients should NOT drink the preservative liquid.
- The risks related to using the Cologuard Collection Kit are low, with no serious adverse events reported among people in a clinical trial. Patients should be careful when opening and closing the lids to avoid the risk of hand strain.

Summary and Explanation of the Test

Cologuard utilizes a multi-target approach to detect DNA and hemoglobin markers associated with colorectal cancer (CRC), as well as pre-malignant colorectal neoplasia. Three independent categories of biomarkers are targeted and provide an additive association with CRC and pre-malignant neoplasms.

The first category of biomarkers involves epigenetic DNA changes characterized by aberrant gene promoter region methylation. The specific methylated gene targets include N-Myc Downstream-Regulated Gene 4 (*NDRG4*) and Bone Morphogenetic Protein 3 (*BMP3*).[6,7] *NDRG4* and *BMP3* have been shown to be hypermethylated in colorectal cancer.[5,6] The *Cologuard* procedure incorporates bisulfite conversion of non-methylated cytosine residues to uracil in the DNA sequence to enable sensitive detection of hypermethylated *NDRG4* and *BMP3*.

The second category targets specific DNA point mutations in the v-Ki-ras2 Kirsten rat sarcoma viral oncogene homolog (*KRAS*) gene, which encodes a small GTPase that is activated transiently as a response to extracellular stimuli or signals.[8,9,10] *KRAS* mutations have been detected in up to 35% of colorectal cancers and the 7 mutations in Exon 2 detected by *Cologuard* account for 98% of these *KRAS* mutations.[12] *KRAS* mutations, along with *NDRG4* and *BMP3* methylation markers, have been shown to be detected in the stool of subjects with colorectal neoplasia, including subjects with colorectal cancer and pre-malignant lesions.[4,11]

The third category of biomarker is non-DNA based and detects hemoglobin, which can be associated with colonic bleeding. Results from the methylation, mutation, and hemoglobin assays are integrated by the Exact Sciences Analysis Software to determine a Positive or Negative reportable result or invalid result.

Principles of the Procedure

Cologuard is designed to analyze patients' stool for the presence of DNA and hemoglobin markers, which may indicate the presence of colorectal cancer or pre-cancerous lesions. Patients use the *Cologuard* Collection Kit, consisting of a Container for collection of stool for DNA testing and a separate sampler (Tube) for collection of stool for hemoglobin testing. Both of these stool samples are required to obtain a *Cologuard* result.

In the processing procedure for DNA testing, the stool sample is mixed with buffer in the Container using the Sample Mixer. An aliquot of the buffered stool sample is centrifuged to pellet solids and generate supernatant. The assay procedure begins with treatment of the supernatant with an Inhibitor Removal Tablet to remove inhibitors that may affect the detection of the DNA biomarkers. Treated supernatant is then combined with denaturing reagents and incubated with target-specific magnetic particles using the Capture Incubator instrument to capture sequences for *NDRG4*, *BMP3*, *KRAS,* and *ACTB* (reference gene).

Using automated processes in the Capture Aspirator and Hamilton Microlab® STARlet (STARlet) instruments, targeted sequences are separated from the solution, washed, and eluted from the particles. Eluted DNA is split to provide two separate DNA aliquots for performing the methylation and mutation assays. The aliquot for the methylation assay is treated with bisulfite conversion reagents. Both aliquots are further purified with silica-coated magnetic beads from which DNA is eluted.

The Quantitative Allele-specific Real-time Target and Signal Amplification (*QuARTS*™) technology combines real-time PCR and invasive cleavage to perform allele-specific amplification and detection of methylated target DNA (*NDRG4*, *BMP3*) and specific DNA point mutations (*KRAS*) in the molecular assays. Each purified DNA aliquot is mixed with the appropriate *QuARTS* reaction master mix. The bisulfite-converted DNA is mixed with a master mix for the *NDRG4*, *BMP3*, and A*CTB QuARTS* reaction. The unconverted, purified DNA for *KRAS* detection is mixed with a master mix for the 7 *KRAS* mutations and *ACTB*. Both *QuARTS* reactions are processed using a real-time cycler in the same assay plate with the same cycling and detection program. Each assay for the *NDRG4*, *BMP3*, *ACTB*, and *KRAS* markers is monitored separately through an independent fluorescent detection channel.

In a parallel workflow, the hemoglobin assay stool sample is prepared and analyzed in a quantitative Enzyme-Linked Immunosorbent Assay (ELISA) that determines the concentration of hemoglobin in the sample. Each sample is incubated in a single well of a 96-well plate coated with anti-hemoglobin antibody, which is then washed to remove any unbound material. A second anti-hemoglobin antibody conjugated to the enzyme horseradish peroxidase (HRP) is then added to the wells and incubated with a colorimetric substrate for HRP. After the reaction is stopped and the optical density read on a plate reader, the level of hemoglobin present in the stool sample is calculated using a calibration curve prepared from a set of calibrators with known hemoglobin concentrations.

Run control samples for both the molecular assays and hemoglobin assay are tested along with patient samples to show that the process has been performed appropriately. Run controls from the *Cologuard* DNA Control Kit (Exact Sciences, 100074) and *Cologuard* Hemoglobin Control Kit (Exact Sciences, 100073) are required in each run to obtain valid assay results. Results from the methylation, mutation, and hemoglobin assays are integrated by the Exact Sciences Analysis Software to determine a Positive or Negative reportable result or an Invalid result.

Reagents

Cologuard utilizes several reagent kits stored at different temperatures including DNA Capture Reagents (2 to 8°C), DNA Preparation Reagents (15 to 30°C), *QuARTS* Assay Reagents (-25 to -15°C), and Hemoglobin Assay Reagents (2 to 8°C).

Lots of reagents are matched for performing the assay. A Supplemental Lot Information sheet is supplied with the reagents. On the sheet is a 2D barcode or set of barcodes, the Supplemental Lot Information Barcode (SLIB), which includes calibration and lot matching information for that set of reagents. The SLIB is scanned into the Exact Sciences Analysis Software prior to performing any portion of the automated run using these reagents.

Use lot numbers of reagents and calibrators listed in Supplemental Lot Information together. DO NOT mix or substitute reagents from Supplemental Lot Information containing different lot groupings.

Ancillary and bulk assay reagents (stored at 15 to 30°C) are also required to run *Cologuard*. Bulk assay reagents are not lot matched to *Cologuard* reagents and may be used with any lot of reagent kits.

DNA Capture Reagents (Exact Sciences, 100028)

Part #	Component	Description	Amount	# provided
200150	CAP BDS, Capture Beads	Magnetic particles with covalently bound oligonucleotide probes	7 mL	10

DNA Preparation Reagents (Exact Sciences, 100029)

Part #	Component	Description	Amount	# provided
200123	DEN SLN, Denaturation Solution	0.1 M NaOH solution	14.5 mL	10
200124	BIS SLN, Bisulfite Conversion Solution	Ammonium bisulfite solution	8.5 mL	10
200125	NEU SLN, Neutralization Solution	Potassium Acetate solution	8.5 mL	10
200222	DES SLN, Desulphonation Solution (Concentrate)	0.35 M NaOH solution	7.5 mL	10
200127	BND BDS, Binding Beads	Magnetic silica particles	8.4 mL	10
200218	DNA and *QuARTS* Supplementary Lot Information	n/a	1 each	1

QuARTS Assay Reagents (Exact Sciences, 100030)

Part #	Component	Description	Amount	# provided
200235	CAR SLN, Carrier Solution	Bovine Serum Albumin, Tris, EDTA	1600 µL	10
200130	ELU BFR, Elution Buffer	Tris, EDTA Solution	12.5 mL	10
200131	MIX A, Oligo Mix A, Methylation	Oligonucleotides, FRET probes, dNTPs	1200 µL	10
200132	MIX B, Oligo Mix B, Mutation	Oligonucleotides, FRET probes, dNTPs	1200 µL	10
200133	ENZ, Enzyme Mix	Enzymes in a buffer with glycerol	250 µL	10
200134	D CAL 1, DNA Calibrator 1, High Methylation	NDRG4, BMP3, ACTB DNA in buffer with non-human DNA carrier	60 µL	10
200135	D CAL 2, DNA Calibrator 2, Low Methylation	NDRG4, BMP3, ACTB DNA in buffer with non-human DNA carrier	60 µL	10
200136	D CAL 3, DNA Calibrator 3, High Mutation	KRAS, ACTB DNA in buffer with non-human DNA carrier	60 µL	10
200137	D CAL 4, DNA Calibrator 4, Low Mutation	KRAS, ACTB DNA in buffer with non-human DNA carrier	60 µL	10

Cologuard DNA Control Kit (Exact Sciences, 100074)

Part #	Component	Description	Amount	# provided
200139	DNA Control 1, High	Oligonucleotides, Tris, EDTA with Carrier DNA	15 mL	10
200140	DNA Control 2, Low	Oligonucleotides, Tris, EDTA with Carrier DNA	15 mL	10
200141	DNA Control 3, Negative	Oligonucleotides, Tris, EDTA with Carrier DNA	15 mL	10
200315	DNA Control Kit Supplementary Lot Information Barcode	n/a	1 each	1

Hemoglobin Assay Reagents (Exact Sciences, 100031)

Part #	Component	Description	Amount	# provided
200142	Hb PLATE, Hemoglobin Assay Plate	Mouse anti-Human Hemoglobin Antibody coated plate	1 plate	5
200143	SMP BFR, Sample Buffer	Tris, NaCl, casein	12 mL	5
200144	CONJ, Antibody Conjugate	Mouse anti-Human Hemoglobin Antibody-HRP Conjugate	12 mL	5
200100	SUBS, Substrate	Tetramethylbenzidine in buffer	12 mL	5
200101	STP SLN, Stop Solution	Acidic Buffered Solution	12 mL	5
200146	Hb CAL, Hemoglobin Assay Calibrator	Human Hemoglobin, buffer (lyophilized)	1 each	5
200219	Hemoglobin Assay Supplementary Lot Information	n/a	1 each	1

Cologuard Hemoglobin Control Kit (Exact Sciences, 100073)

Part #	Component	Description	Amount	# provided
200147	Hemoglobin Control 1, High	Human Hemoglobin, buffer (lyophilized)	1 each	5
200148	Hemoglobin Control 2, Low	Human Hemoglobin, buffer (lyophilized)	1 each	5
200149	Hemoglobin Control 3, Negative	Human Hemoglobin, buffer (lyophilized)	1 each	5
20031e	Hemoglobin Control Kit Supplementary Lot Information Barcode	n/a	1 each	1

Ancillary Materials and Bulk Assay Reagents

Required ancillary materials, kits and bulk assay reagents are available using the part numbers listed below.

Part #	Component	Description	Amount per vessel
200204	STL BFR, Stool Buffer	Tris, EDTA Solution	20 L
200151	TABLT, Inhibitor Removal Tablet	Polyvinylpolypyrrolidone with excipient	95 ea
200138	FILT, Spin Filter	Spin filters for 50 mL tubes	46 ea
200152	TUBES, Barcoded Mixing Tubes	Empty barcoded tubes	50 ea
200120	PRE WSH, Capture Bead Pre-wash	Sodium Bicarbonate Buffer	350 mL
200121	CAP SLN, Capture Solution	Guanidine Thiocyanate	450 mL

200122	CAP WSH, Capture Wash	MOPS Buffer, NaCl	450 mL
200126	BND SLN, Binding Solution	Guanidine Hydrochloride	450 mL
200129	CNV WSH, Conversion Wash Concentrate	Tris Buffer	200 mL
200145	Hb WSH, Hemoglobin Assay Wash Concentrate	Phosphate Buffer with detergent	300 mL

Procedural Warnings and Precautions

Warnings and notes emphasize important reagent information and critical instructions for safely performing laboratory procedures.

- Use standard laboratory precautions in accordance with applicable federal, state, and local regulations.
- Laboratory areas should be cleaned and maintained according to good laboratory practices for clinical laboratories processing biological specimens. Appropriate procedures shall be defined by the laboratory director.
- Sodium hypochlorite may not be appropriate for decontamination of instruments and pipettes.
- Refer to user's manuals for complete decontamination procedures for instruments and equipment.
- Product components (residual product, packaging, waste) can be considered laboratory waste. Dispose in accordance with applicable federal, state, and local regulations.

Reagent Warnings and Precautions

IVD	For *In vitro* diagnostic use
	Users should familiarize themselves with the instructions contained in the *Cologuard* Patient Guide, this booklet and the equipment used to perform the *Cologuard* prior to use.
	Caution: Patients should avoid bringing preservative solution in contact with skin or eyes. Irritation could result.
	Some reagents or waste are potentially corrosive or flammable. Dispose of all reagents in accordance with local, state, and Federal regulations. (CLSI doc GP5-A2, EPA/530-SW-86-014)
	Warning, biological hazard. Specimens may be infectious. Only personnel adequately trained in handling infectious materials should be permitted to perform this diagnostic procedure. Human materials used in Hemoglobin Assay Calibrator were tested and found to be negative for HIV, Hepatitis B and Hepatitis C. As an additional precaution, Hemoglobin Assay Calibrator (Hb CAL) should be treated as potentially infectious. Dispose of all potentially biohazardous materials in accordance with local, state and Federal regulations.
	Danger, Corrosive. Skin and respiratory Irritant. Avoid contact with Denaturation Solution (DEN SLN), Bisulfite Conversion Solution (BIS SLN), and Desulphonation Solution (DES SLN) with skin, eyes, and mucous membranes. If these fluids come into contact with skin or eyes, wash with water. If spills of these fluids occur, dilute with water before wiping dry.

⚠	Warning, Irritant. Avoid contact with Binding Solution (BND SLN), Capture Solution (CAP SLN) and Stop Solution (STP SLN) with skin, eyes and mucous membranes. If these fluids come into contact with skin or eyes, wash with water. If swallowed, DO NOT induce vomiting unless directed by poison control center. If spills of these fluids occur, dilute with water before wiping dry.
⚠	Warning, Respiratory irritant. Avoid contact with Bisulfite Conversion Solution (BIS SLN) with skin, eyes and mucous membranes. If these fluids come into contact with skin or eyes, wash with water. If spills of these fluids occur, dilute with water before wiping dry. If swallowed, DO NOT induce vomiting unless directed by poison control center. If inhaled, move to fresh air. If breathing becomes difficult, give oxygen and consult physician.
⚠	The Antibody Conjugate (CONJ) may contain a mixture of 5-chloro-2-metyhl-4-isothiazolin-3-one and 2-methyl-4-isothiazolin-3-one which are components of ProClin®. The components are classified per applicable European Community (EC) Directives as: Irritants (Xi). Avoid contact with skin eyes and mucous membranes.
⚠	Warning. Do not use sodium hypochlorite (bleach) to decontaminate surfaces or dispose of waste from steps using Bisulfite Solution (BIS SLN) or Capture Solution (CAP SLN). Salts from these reagents are not compatible with cleaning solutions containing bleach.

Instrument Warnings and Precautions

i	Users should familiarize themselves with detailed user information contained in equipment user manuals prior to following the *Cologuard* laboratory procedure.
☞	Conduct instrument maintenance according to individual instrument user's manuals on all instruments to ensure safe and appropriate use.
☞	Conduct STARlet daily and weekly maintenance as required. Failure to empty liquid waste may result in release of hazardous materials into the environment. Failure to empty tip waste may result in an aborted run and/or contamination of the instrument deck.
☞	There are environmental specifications for Hamilton STARlet that are specific to Cologuard. Ensure that instruments are maintained in environments at 18-26 C with 20-85% RH.

Reagent Storage and Handling Requirements

Part #	Reagent Group	Storage Requirement	Additional Handling Requirements
100028	DNA Capture Reagents	2°C–8°C	DO NOT FREEZE CAP BDS.
100029	DNA Preparation Reagents	15°C–30°C	See Warnings and Precautions specified for DEN SLN, BIS SLN, and DES SLN. Add 17.5 mL 100% isopropanol to DES SLN concentrate before use. Store BIS SLN protected from light.
100030	*QuARTS* Assay Reagents	-25°C–-15°C	Reagents may be shipped at 2 to 8°C. Transfer to -25 to -15°C upon receipt.
100031	Hemoglobin Assay Reagents	2°C–8°C	See Warnings and Precautions specified for STP SLN.
200204	STL BFR, Stool Buffer	15°C–30°C	
200151	TABLT, Inhibitor Removal Tablet	15°C–30°C	

Part #	Reagent Group	Storage Requirement	Additional Handling Requirements
200138	FILT, Spin Filter	15°C–30°C	
200152	TUBES, Barcoded Mixing Tubes	15°C–30°C	
200120	PRE WSH, Capture Bead Pre-wash	15°C–30°C	
200121	CAP SLN, Capture Solution	15°C–30°C	See Warnings and Precautions specified for CAP SLN. If precipitate is observed, heat at 35°C to 50°C to solubilize.
200122	CAP WSH, Capture Wash	15°C–30°C	
200126	BND SLN, Binding Solution	15°C–30°C	If precipitate is observed, heat at 35°C to 50°C to solubilize.
200129	CNV WSH, Conversion Wash Concentrate	15°C–30°C	Prepare working solution before use according to instructions in the *Cologuard* Laboratory Procedure.
200145	Hb WSH, Hemoglobin Assay Wash Concentrate	2°C–8°C	Prepare working solution before use according to instructions in the *Cologuard* Laboratory Procedure. If precipitate is observed in concentrate, heat at 35°C to 50°C to solubilize.

Instrumentation

Part #	Instrument	Manufacturer (Supplier)
300810	Sample Mixer	Exact Sciences
11675200	MaxQ™ 2000 Open-Air Platform Shaker	Thermo Fisher
300551	Capture Shaker Rack	Exact Sciences
300546	Capture Incubator	Exact Sciences
100034	Capture Aspirator	Exact Sciences
100065	Hamilton Microlab® STARlet	Hamilton (Exact Sciences)
100066	STARlet Hemoglobin Package	Exact Sciences
4406984	7500 Fast Dx IVD w/laptop	Life Technologies
ELx808™	ELx808 Custom Plate Reader	BioTek
3100620	620 nm Filter assembly for ELx808	BioTek
200268	System Computer	Exact Sciences

The primary instruments required to perform the laboratory procedure for *Cologuard* are listed above. These instruments and supporting software are provided and installed separately through Exact Sciences service prior to training of laboratory personnel. The Cologuard *Laboratory Procedure* section outlines the specific use of these instruments for performing *Cologuard*.

The Sample Mixer is used to mix the stool sample for DNA testing (Container) received from the patient (refer to Cologuard *Laboratory Procedure, Preparation of Stool Homogenate*). The MaxQ™ Shaker is equipped with the Capture Shaker Rack for use in preparing the supernatant for the DNA Capture (refer to Cologuard *Laboratory Procedure, DNA Capture*). The Capture

Incubator and Capture Aspirator are equipment used during the DNA Capture steps of the assay (refer to Cologuard *Laboratory Procedure, DNA Capture*). The Capture Incubator performs sample heating, cooling and mixing of 50 mL tubes. The Capture Aspirator automates magnetic particle separation from supernatant in a 50 mL tube format.

Once DNA Capture steps are completed, the Hamilton Microlab® STARlet is used for automated DNA preparation and *QuARTS* plate setup as well as automated Hemoglobin plate setup (refer to Cologuard *Laboratory Procedure, DNA Preparation and* QuARTS *Assay* and *Hemoglobin Assay*). The 7500 Fast Dx instrument is used to perform the QuARTS reactions setup in the prepared *QuARTS* assay plate. The ELx808™ Custom Plate Reader is used to measure hemoglobin assay plate.

Specimen Collection and Preparation for Analysis

Specimens for use with *Cologuard* must be collected with the *Cologuard* Collection Kit (Exact Sciences, 100026), including a stool sample for DNA testing (Container) and a stool sample for Hemoglobin testing (Tube). Detailed instructions for sample receipt and processing are outlined in the *Cologuard* Laboratory procedure below. Known interfering substances that may impact the assay results are summarized in the *Performance Characteristics, Interfering Substances* section below.

	Patients should familiarize themselves with detailed information contained in *Cologuard* Patient Guide and collection instructions before completing sample collection.
	Stool samples must be collected with the *Cologuard* Collection Kit (Exact Sciences, 100026).
	The *Cologuard* Collection Kit should be stored protected from direct sunlight at ambient temperature.
	Patients should not provide a sample if they have diarrhea or blood in their urine or stool from bleeding hemorrhoids, bleeding cuts or wounds on their hands, rectal bleeding, or menstruation.
	To ensure the integrity of the sample, the laboratory must receive patient specimens within 72 hours of collection. Detailed instructions are outlined in the *Cologuard* Laboratory Procedure.
	Samples may be stored by the laboratory until processing. The Tube (hemoglobin sample) can be stored for up to 7 days after receipt at 2 to 8°C. The Container (DNA sample) can be stored ambiently or at 2 to 8°C and should be processed within 6 days of collection. Detailed processing instructions are outlined in the *Cologuard* Laboratory Procedure.
	Avoid cross-contamination during the specimen handling steps. If gloves come into contact with specimen, change gloves to avoid cross-contamination.

Cologuard Laboratory Procedure

Receipt of *Cologuard* Collection Kit

The patient collects a stool sample using the *Cologuard* Collection Kit (Exact Sciences, 100026). Stool samples are sent to the laboratory according to the *Cologuard* Patient Guide that accompanies the *Cologuard* Collection Kit. Laboratory processing begins with receipt of the collection kit and preparation of stool for DNA capture.

1. Check that both the Tube (hemoglobin sample) and Container (DNA sample) are present and review the information noted on the handwritten labels affixed by the patient.
   a. Confirm that collection date and time occurred less than 72 hours prior to receipt.

2. Remove the samples and discard packaging in accordance with local regulations.
3. Vortex hemoglobin sample tube at highest speed until grooves of the probe are void of stool.
4. Store the samples appropriately until processing:
   a. The hemoglobin sample can be stored for up to 7 days after receipt at 2 to 8°C.
   b. The DNA sample can be stored at ambient temperature or at 2 to 8°C and should be processed within 6 days of collection.

Preparation of Stool Homogenate for DNA Testing

1. Weigh the Container (containing sample) and record the Sample Weight.
2. Calculate Stool Weight:
   a. Stool Weight = Container (g) – 535 g (empty container + preservation weight).
3. Based on the calculated Stool Weight, adjust the stool: buffer ratio as follows:
   b. If stool weight is less than or equal to 0 g, sample is invalid. Discard sample and request a replacement.
   c. If stool weight is greater than 0 g and less than or equal to 72 g, proceed to Step 4.
   d. If stool weight is greater than 72 g and less than 300 g, calculate amount of Stool Buffer (Exact Sciences, 200204) to add. Stool Buffer can be added by volume **or** by weight (see table below). Open the Container and add the Stool Buffer. Proceed to Step 4.
   e. If stool weight is greater than or equal to 300 g, the sample is invalid. Discard sample and request a replacement.

Stool Weight (X)	Stool Buffer to Add*		Additional Information
	By Volume (mL)	By Weight (g)	
X ≤ 0 g	N/A	N/A	Invalid sample
0 g < X ≤ 72 g	N/A	N/A	Sample adequately buffered
72 g < X ≤ 280 g	4X – 290	1.04(4X – 290)	Dilution yields 1 g stool per 4 mL buffer
280 g < X < 300 g	(1143 – X)/1.04	1143 – X	Dilution maximized based on capacity of Container
X ≥ 300 g	N/A	N/A	Invalid Sample
*Stool Buffer may be added by volume or by weight. The density of the buffer (1.04 g/mL) is used in conversions between volume and weight.			

4. Tighten the Container lid following instructions in the Sample Mixer User's Manual.
5. Place the Container in the Sample Mixer (Exact Sciences, 300810), secure the container with the lid attachment, and close the mixer door. Initiate the mixing cycle.
6. When the mixing is complete, remove the Container from the mixer, and then remove the lid from the Container.
7. Prepare at least two 50 mL tubes (Corning, 430829) with barcoded labels to identify the sample.
8. Transfer homogenate sample up to the 45 mL graduation mark to each of the 50 mL tubes.

9. Place samples at -15°C or colder for at least 8 hours prior to use. Discard the processing pipette, any remaining homogenate, and the Container according to local regulations.

Assay Overview

Each *Cologuard* reagent kit contains sufficient materials for 480 tests. This includes reagents for 5 groups of 86 patient samples and the required controls and calibrators. The assay procedure includes steps for DNA Capture, DNA Preparation, *QuARTS* Assay, Hemoglobin Assay, and Data Analysis using the Exact Sciences System Software. DNA Capture steps are performed manually and are typically processed in sets of 23 patient and control samples. DNA Preparation and *QuARTS* Assay steps are performed using the Microlab® STARlet (STARlet), custom built for Exact Sciences, and are processed in batches of up to 46 samples, including 43 patient samples and the required controls. Hemoglobin Assay steps are performed in 96-well assay plates and typically include batches of up to 86 patient samples and required calibrators and controls. This assay uses the STARlet for plate setup, followed by a manual sandwich ELISA. Optimal usage of *Cologuard* is achieved with four full sets of DNA Capture, two full batches of DNA Preparation and *QuARTS* Assay, and one full batch for the Hemoglobin Assay.

DNA and Hemoglobin Control Samples are supplied in the *Cologuard* DNA Control Kit (Exact Sciences, 100074) and the *Cologuard* Hemoglobin Control Kit (Exact Sciences, 100073). Controls D CTRL 1, D CTRL 2, and D CTRL 3 are required for each batch of DNA Preparation and *QuARTS* assay samples processed on the STARlet. At least one positive DNA Control (D CTRL 1 or D CTRL 2) is required for every distinct set of DNA Capture tubes. Hemoglobin Controls (Hb CTRL 1, Hb CTRL 2, and Hb CTRL 3) are required for each plate of Hemoglobin Assay samples.

Reagents not denoted as ancillary material or bulk assay reagents are packaged for single use and the leftover reagents cannot be reused. Test samples should be stored and run in maximum batch sizes as defined in the “Optimal *Cologuard* Workflow” figure below to maximize use of the reagents. Each DNA Capture set **must** contain at least one positive control for every run performed, regardless of the number of test samples, and each DNA Preparation and *QuARTS* assay must contain all DNA controls. Hemoglobin Assay runs must contain all controls. An example of optimal setup is shown below.

Example *Cologuard* Setup



DNA Capture

Prepare Capture Beads

NOTE: For each DNA Capture set of 23 tubes, 3.25 mL of prepared Capture Beads are required. If desired, multiple tubes of Capture Beads from a single lot can be prepared simultaneously for processing additional sets.

1. Set the Capture Incubator (Exact Sciences, 300546) to preheat ("Bead Prep 1" program).
2. Allow Capture Beads (Exact Sciences, 200150) to sit at room temperature for a minimum of 30 minutes.
3. Vortex Capture Beads at highest setting for 30 seconds to suspend the beads.
4. Label the 50 mL conical tube(s) with Capture Bead preparation date and lot information.

Labels used in the Capture Incubator have specific requirements for size, material and thickness. Label the bead preparation tubes using permanent marker or refer to Capture Incubator User's Manual for detailed label specifications.

5. Transfer 3.25 mL of beads to the labeled 50 mL tube.
6. Add 10 mL of Capture Bead Pre-wash (Exact Sciences, 200120) and secure the 50 mL tube cap.
7. When the Capture Incubator has reached programmed temperature and display prompts user to insert test tubes, place tube(s) in the Capture Incubator. Close cover and press the 'Start/Select' button to proceed with the cycle.
8. When the cycle is complete, remove tube(s) from the incubator and place in the centrifuge with appropriate balance tube, if necessary. Centrifuge the tube(s) until the centrifuge reaches 500 × g for and hold for 1-10 seconds.
9. Remove cap(s) and transfer the tube(s) to the first row (left-most) of the Capture Aspirator (Exact Sciences, 300490). Execute the "Prep Beads" protocol to remove supernatant from the tube(s).

NOTE: If operator prefers to prepare greater than six tubes, see Procedural Notes and Precautions, Prepare Capture Beads (for >6 tubes).

10. When aspiration run is complete, remove tube(s) from Capture Aspirator and add 3.25 mL of fresh Capture Bead Pre-wash solution to each tube(s), replace cap(s) and vortex at highest setting until all beads are suspended.

NOTE: Once the Capture Beads have been prepared, they can be stored in closed tube for up to 7 days at 2-8°C before use.

Prepare Samples and Perform DNA Capture

Prepare and Label Sample Tubes

1. Remove stool aliquot samples and DNA controls (D CTRL 1-3) from storage.

a. Place frozen samples in racks to allow air to circulate around the tubes. Leave racks of frozen samples at 2 to 8°C for at least 13 hours, but no more than 80 hours, until use.
b. Equilibrate DNA controls at room temperature for at least 30 minutes before further processing.

2. Create three labels for subsequent processing steps for each tube. These steps will include Inhibitor Removal Tablet addition (denoted as "TAB" or equivalent), use of Spin Filter (denoted as "SPN" or equivalent), and Capture Incubation (denoted as "CAP" or equivalent).

Prepare Supernatant

NOTE: Ensure that an aliquot of Stool Buffer (Exact Sciences, 200204) is available for use in subsequent steps if needed for possible volume adjustment.

1. Centrifuge the stool sample aliquots for 45 minutes at a setting of 4500 × g. Ensure that the centrifuge is balanced.
2. When the centrifugation is complete, promptly and carefully remove the tubes and place in racks.

NOTE: If the interface between pellet and supernatant becomes obviously disrupted (e.g., tube is dropped or inverted), repeat Steps 1-2.

3. Confirm that the labels from the centrifuged stool sample aliquots and DNA control tubes match the labels on the prepared, clean tubes for the next step.
4. Add one Inhibitor Removal Tablet (Exact Sciences, 200151) to each "TAB" tube before transferring samples.
5. Transfer 14 mL of supernatant from the spun stool sample aliquots and the DNA control tubes into the respective, clean, labeled "TAB" tubes.

NOTE: Ensure that only clean tube caps are used and that appropriate steps are taken during sample transfer to minimize any risk of cross-contamination. Do not interchange caps between tubes once the caps have been exposed to a sample.

NOTE: Aspirate centrifuged stool samples slowly and avoid disturbing the solid/liquid interface. Avoid aspirating any material from the pellet or material floating on the surface of the supernatant.

a. If the volume of the supernatant is between 5 mL and 14 mL, bring the total volume of supernatant to 14 mL with Stool Buffer (Exact Sciences, 200204).
b. If the volume of the supernatant is less than 5 mL, store the supernatant at 2 to 8°C until it can be combined with more supernatant obtained by repeating Steps 1 – 5 with additional homogenate aliquots as necessary.

6. Transfer the capped tubes to the MaxQ Shaker (Shaker) (Thermo Fisher, 11675200) with the Capture Shaker Rack (Exact Sciences, 300551) and mix for 15 minutes at 400 RPM.
7. After this point, the used 50 mL tubes with stool pellet may be discarded according to local regulations.
8. After mixing sample supernatants with the Inhibitor Removal Tablet, confirm that the labels from these tubes match the labels on the prepared, clean "SPN" tubes for the next step.
9. Place one Spin Filter (Exact Sciences, 200138) into each "SPN" labeled tube before transferring samples from Step 6. Reserve tube caps for use in a future step.
10. Swirl each tube from Step 6 to suspend content, remove cap and pour the contents into the

spin filter of the respectively “SPN” labeled spin filter tube. Close the lid of the spin filter. Repeat for all samples.

11. After this point, the used “TAB” tube and cap may be discarded according to local regulations.
12. Once all samples are transferred to spin filters, place spin filter tubes into the centrifuge. Ensure that centrifuge is balanced and spin for 6 min at 3300 × g.
13. Remove the tubes from the centrifuge and confirm that the labels from these tubes match the labels on the prepared, clean “CAP” tubes for the next step.
14. For each tube, remove the spin filter from the tube, and then transfer 10 mL of supernatant to the capture tube (“CAP” label).
15. Confirm that the transferred supernatant contains 10 mL volume.
    a. If a tube contains between 5 and 10 mL supernatant, bring the volume up to 10 mL with Stool Buffer (Exact Sciences, 200204).
    b. If a tube contains less than 5 mL supernatant, store the filtered supernatant at 2 to 8°C until it can be combined with extra supernatant obtained by repeating Steps 1-14 with additional homogenate aliquots, as necessary.
16. At this point, the used 50 mL tubes with spin filters may be discarded according to local regulations.
17. Proceed to next step or store at 2-8°C for up to 6 days.

Capture Incubation

NOTE: If prepared supernatant was stored at 2-8°C, incubate at 15-30°C for 30 minutes.

NOTE: If applicable, remove prepared Capture Beads from 2-8°C and incubate at 15-30°C for 30 minutes before use.

1. Inspect Capture Solution (Exact Sciences, 200121) for precipitate. If precipitation is present, warm at 35°C-50°C for 20 minutes or until solubilized. Invert to mix as needed.
2. Add 7.25 mL of Capture Solution to each capture tube, ensuring that the Capture Solution runs down the inside of the tube to avoid foaming.

NOTE: Prepared Capture Beads must match the reagent lots listed on the DNA and* QuARTS *Supplementary Lot Information (Exact Sciences, 200218) that will be used for the assay run.

3. Vortex prepared Capture Beads for 30 seconds at the highest setting to suspend the beads.

☞ If Capture Beads are not suspended before transfer, DNA Capture may not work properly.

4. Add 125 µL of beads to each of the capture tubes and then tighten the tube caps.
5. Place all tubes into the Capture Incubator using the Capture Incubator Tube Lift (Exact Sciences, 300547) and then start the EXAS8 program.

NOTE: Place a 17.5 mL water-filled blank tube into each empty position of the Capture Incubator.

6. When the program reaches completion, remove tubes from the Capture Incubator, remove and discard caps, and place open tubes in the Capture Aspirator, 300490.
7. Perform capture aspiration using the “BIND 10 min” program.
8. Remove the tubes from the Capture Aspirator and inspect for complete aspiration.

NOTE: If incomplete aspiration is observed, bring tube volume to 10 mL using Capture Wash, mix by pulse vortexing and repeat the BIND 10 min program.

9. Promptly add 750 µL of Capture Wash (Exact Sciences, 200122) to each tube.
10. Cap the tubes using reserved caps from *Prepare Supernatant*, Step 9. Place tubes into the Shaker, and mix for 1 minute at 400 RPM. Confirm that the Capture Beads are suspended in each tube.

NOTE: If beads are not suspended, rotate tube and mix for 1 minute at 400 RPM.

11. Remove the tubes and store at 2 to 8°C if not proceeding to *Automated DNA Preparation and* QuARTS *Assay Setup* immediately. Closed tubes containing capture wash and beads can be stored for up to 4 days before use.

DNA Capture will need to be performed on two full sets (46 samples and controls total) to obtain a full batch of samples for the next steps.

DNA Preparation and *QuARTS* Assay

DNA Preparation and *QuARTS* Assay steps are processed in batches of up to 46 samples from the DNA Capture steps. Input samples include up to 43 patient samples in addition to D CTRL 1, D CTRL 2, and D CTRL 3 in each batch. DNA preparation and the *QuARTS* assay plate setup are only performed on the automated STARlet. The program for the Exact Sciences STARlet Interface Software guides the operator through loading the sample tubes, resources, and reagents onto the system.

The system uses barcodes to identify samples and reagents. The barcode on each sample tube is used to ensure tracking to the final result while the reagent barcode tracking ensures that the matched lots of reagents are used together and that the reagents have not expired. It is the responsibility of the assay operator to ensure that the Capture Beads used in the capture process match the reagent lots listed with DNA and *QuARTS* Supplementary Information (Exact Sciences, 200218) used for an assay run. Errors detected by the system are reported in the run results.

DNA and *QuARTS* Reagent Supplemental Lot Information (Exact Sciences, 200218) is used to transfer lot and calibrator information into the Exact Sciences System Software. The information needs to be entered only once for each unique Supplemental Lot Information lot number. Similarly, *Cologuard* DNA Control Kit Supplemental Lot Information (Exact Sciences, 200315) is used for transfer of the control values and acceptance limits for the particular kit lot of controls used in the procedure.

Detailed instructions are provided in the software screens on the correct positioning of each reagent and all consumables, samples, and racks. Each DNA Preparation and *QuARTS* batch consists of up to 43 patient samples and 3 controls. Users are instructed through the software to provide 2 calibrators for the methylation assay (D CAL 1 and D CAL 2 (Exact Sciences, 200134, 200135)), and 2 calibrators for the mutation assay (D CAL 3 and D CAL 4 (Exact Sciences, 200136, 200137)). Patient samples, controls, and calibrators for both methylation and mutation assays are set up in one 96-well *QuARTS* reaction plate. Additonal instructions can be found in the Exact Sciences System Software User's Manual.

The steps for *Automated DNA Preparation and* QuARTS *Plate Setup* are completed in about 7 hours. After the DNA preparation steps are completed, the instrument prompts the user to mix, uncap, and replace reagents for the *QuARTS* plate setup. When the *QuARTS* plate is ready to run, the user removes the plate, covers with a plate seal, centrifuges to ensure the reagents are at the bottom of all wells, and runs the plate on the 7500 Fast Dx Real-Time PCR Instrument (7500 Fast Dx; Life Technologies, 4406984). The *QuARTS* analytic run is completed in approximately 2.5 hours. Once complete, the data are exported to the Exact Sciences Analysis Software and the run results are calculated. Methylation and Mutation assay runs are considered valid if actual results from all DNA controls are within the expected ranges included in the *Cologuard* DNA Control Supplemental Lot Information, the calibration curve meets the acceptance criteria, and no fatal processing errors were detected by the system.

Automated DNA Preparation and *QuARTS* Assay Setup

Reagent Preparation

1. Assemble the following reagents. Equilibrate the Carrier Solution (Exact Sciences, 200235) to room temperature (may take up to 30 minutes). Steps 2-4 below may proceed while the Carrier Solution equilibrates.

Part #	Component Abreviation / Name
200122	CAP WSH, Capture Wash
200123	DEN SLN, Denaturation Solution
200124	BIS SLN, Bisulfite Conversion Solution
200125	NEU SLN, Neutralization Solution
200222	DES SLN, Desulphonation Solution
200127	BND BDS, Binding Beads
200126	BND SLN, Binding Solution
200129	CNV WSH, Conversion Wash
200235	CAR SLN, Carrier Solution

2. Add 17.5 mL of 100% isopropanol to the Desulphonation Solution (Exact Sciences, 200222) bottle, replace cap, and invert to mix 10 times.
3. To prepare Conversion Wash, add 800 mL of 100% ethanol to the Conversion Wash (Exact Sciences, 200129) bottle, replace cap, and invert to mix 10 times. Mark date on the prepared Conversion Wash bottle once ethanol is added. Prepared Conversion Wash can be used for up to 1 month.
4. Remove the captured DNA sample tubes resulting from *Capture Incubation* steps above from storage and allow samples to come to room temperature.
5. Place the tubes into the Shaker and mix for 1 minute at 400 RPM.

STARlet Setup

1. Check for outstanding maintenance at the beginning of each day prior to performing the run.
   a. Log into the Exact Sciences STARlet Interface Software.
   b. Under the Maintenance Monitor, see whether Maintenance Check is listed as ‘valid’ or ‘invalid’. The monitor also lists the last date checked.

c. If maintenance check is listed as ‘valid’, daily or weekly maintenance does not need to be performed.
d. If maintenance check is listed as ‘invalid’, select the prompt to run maintenance.
e. A new pop-up appears that lists both daily and weekly maintenance and when each was last performed.
f. Select the required maintenance type, select the green arrow to begin, and follow the software prompts to perform the maintenance.

NOTE: Weekly maintenance covers all of the daily maintenance tasks. If weekly maintenance is being executed, daily maintenance does not need to be run.

NOTE: Cologuard methods will not begin if required maintenance has not been completed successfully.

2. If SLIBs have not been previously scanned into the Analysis Software, follow instructions for *Entering Supplementary Lot Information* under the *Procedural Notes and Precautions* section.
3. Select the *Cologuard* test and the Methylation & Mutation setup run on the Exact Sciences STARlet Interface Software, then select ‘Load Setup’ to initiate the run.



4. The deck layout diagram appears as shown below. Select ‘Run’ to start the loading process.



5. Confirm that the loading tray positions in front of the carriers on the deck are clear and hit 'Next' in order to prompt the instrument to unload any carriers stored in the instrument.

NOTE: Carriers are unloaded from the deck left to right and loaded right to left.

6. Load the appropriate carriers on the loading tray according to the deck layout. Load two deep-well plates (Axygen, P-DW-20-C) for Capture Wash and for conversion/cleanup and two trays 1000 µL CORE tips (Hamilton, 235905) in the right hand tip carrier.

Load only full trays of tips, or an invalid run may result.

Do not store any new or unused tips for waste aspiration in the right carrier between runs. Any tips present in the righthand tip carrier from a previous run must be presumed to be used and discarded to prevent possible aborted runs or cross-contamination of samples.

7. Load one MicroAmp Fast 96-well plate (Life Technologies, 436906) with barcode to the front.
8. Load uncapped samples and controls into 50 mL Tube Carriers (Hamilton, 182045) working back to front, left to right.
   a. Place samples in the sample carriers back to front, left to right, with no empty positions between tubes.
   b. Place controls in the sample carriers with no empty positions between the first and last sample or control.
   c. Place all sample carriers on deck, even if empty.

NOTE: Unread sample barcodes may require repositioning or manual barcode entry to correct the error. Results for samples with manually entered barcodes are flagged in the reports.

For the Methylation & Mutation method, three controls (D CTRL1, D CTRL 2, and D CTRL 3) must be present within the sample carriers for the run to begin.

Push each tube to the bottom of the rack and ensure that the barcode is visible in the slot on the right.

Empty positions are not permitted in sample carriers, except after the last loaded sample. Always load samples from left to right with no empty spaces between samples. Load all carriers regardless of the number of samples, placing empty carriers at the end.

Carriers with unread sample barcodes will be unloaded. The barcodes must be adjusted and the carrier reloaded until the barcode is successfully rescanned, or barcode sample IDs may be entered manually by the operator.

d. When the sample carriers are successfully loaded, a prompt confirming the number of samples appears in the Instructions Box. Select 'Yes' if sample count is correct. Select 'No' if sample count is not correct.
e. Selecting 'No' will unload the carriers for the operator to correct the issue. Once corrected, select 'Next'.

9. Load Reagents as shown in the on-screen deck layout into the indicated carrier positions.
   a. If a SLIB needs to be entered to continue setup of the run, scan the appropriate supplemental lot information barcodes using the 2D barcode scanner into the Analysis Software prior to hitting 'Next' to reload the carrier.
   b. See the table below for components and special instructions for individual reagents.

NOTE: Reagents are loaded and checked to match lot numbers from a SLIB scanned into the Analysis software. Reagents are identified and confirmed to be the correct lot and location by scanning their barcodes as they are loaded on the instrument. A run will not proceed with incorrect reagent part numbers or reagents from mixed SLIB lots (master lot mismatch) or reagents with unknown lot numbers.

NOTE: If any reagents are not recognized or do not match to a SLIB in the system, the carrier containing the mismatched reagents is unloaded. The software prompts the user to correct the issue and then select 'Next' to reload the carrier.

Transfer peel-off barcodes to troughs so that barcode is placed on the curved edge of the trough, starting at the top of the trough with the curved edge of trough ion the left, as shown in the following figure. Human-readable barcode content should be perpendicular to the top of the trough as indicated below.

Part #	Component	Additional Instruction

Part #	Component	Additional Instruction
200122	CAP WSH, Capture Wash	On first use, transfer peel-off barcode to clean 200 mL trough. Mix by inverting bottle 10 times, then transfer 100 mL of CAP WSH to trough. Trough is rinsed with distilled water after each use and allowed to dry, then re-used for the remaining volume of the CAP WSH bottle.
200123	DEN SLN, Denaturation Solution	Transfer peel-off barcode to clean 50 mL trough. Mix by inverting bottle10 times, then transfer all contents to trough.
200235	CAR SLN, Carrier Solution	Ensure liquid is completely thawed. Vortex 3-5 seconds at highest speed, spin briefly to collect volume, then **UNCAPUNCAP** and place vial in the indicated carrier.
200124	BIS SLN, Bisulfite Conversion Solution	Mix by inverting 5 times, then **UNCAPUNCAP** and place vial in the indicated carrier.
200125	NEU SLN, Neutralization Solution	Mix by inverting 10 times, then **UNCAPUNCAP** and place vial in the indicated carrier.
200222	DES SLN, Desulphonation Solution (after isopropanol addition)	Transfer peel-off barcode to clean 50 mL trough. Transfer all contents to trough and cover with 50 mL Trough Lid (Exact Sciences, 100072).
200127	BND BDS, Binding Beads	Transfer peel-off barcode to clean 50 mL trough. Vortex bottle for 30 seconds, then transfer all contents to trough.
200126	BND SLN, Binding Solution	On first use, transfer peel-off barcode to clean 200 mL trough. Mix by inverting bottle 10 times, then transfer 100 mL to trough. Trough is rinsed with distilled water after each use and allowed to dry, then re-used the remaining volume of the BND SLN bottle.
200129	CNV WSH, Conversion Wash (after ethanol addition)	On first use, transfer peel-off barcode to clean 200 mL trough. Mix by inverting bottle 10 times, then transfer 200 mL to trough and cover with Trough Lid (Exact Sciences, 100071). Trough is rinsed with distilled water after each use and allowed to dry, then re-used for the remaining volume of the CNV WSH bottle.
200130	ELU BFR, Elution Buffer	Place **CAPPED** tube onto the instrument deck. Cap will be removed at a later step.
200131	MIX A, Oligo Mix A, Methylation	Place **CAPPED** tube onto the instrument deck. Cap will be removed at a later step.
200132	MIX B, Oligo Mix B, Mutation	Place **CAPPED** tube onto the instrument deck. Cap will be removed at a later step.
200133	ENZ, Enzyme Mix	Place **CAPPED** tube onto the instrument deck. Cap will be removed at a later step.
200134	D CAL 1, DNA Calibrator 1, High Methylation	Place **CAPPED** tube onto the instrument deck. Cap will be removed at a later step.
200135	D CAL 2, DNA Calibrator 2, Low Methylation	Place **CAPPED** tube onto the instrument deck. Cap will be removed at a later step.
200136	D CAL 3, DNA Calibrator 3, High Mutation	Place **CAPPED** tube onto the instrument deck. Cap will be removed at a later step.
200137	D CAL 4, DNA Calibrator 4, Low Mutation	Place **CAPPED** tube onto the instrument deck. Cap will be removed at a later step.

10. Load two capped, barcoded empty tubes (Exact Sciences, 200152).
11. Load three trays 1000 µL CORE (Hamilton, 235905) and two trays 50 µL CORE (Hamilton, 235948) pipette tips into the left hand tip carrier.

Load only full trays of tips, or an invalid run may result.

12. When all carriers on the loading tray have been properly loaded with reagents, samples, and

consumables, select 'Next'.

13. As carriers are loaded into the STARlet, a series of prompts will appear in the Instructions box, prompting the user to select tips, confirm sample counts, and correct reagent and sample loading issues such as mismatched lots or unread barcodes, if applicable. Select 'Next' after each action is performed.

NOTE: The system will check for the correct location or lot (when applicable) of tips, reagents, or samples as they are loaded by scanning and recording the item barcodes. Items with unread barcodes are indicated in red on the deck layout graphic.

14. When loading tips, Tip inventory should be updated when the tip tray becomes highlighted by a blinking red box on the screen.
    a. To update the tip count, select each tip tray so that the larger image appears, then select the blue +/- at the top of each column to auto-fill that column or select 'New Tray' to fill the entire tray with tips.
    b. A blue filled circle indicates that the user has verified that a tip is present in the corresponding tray location.
    c. Select 'Done' when the on-screen inventory matches the loaded tray.

NOTE: Ensure that all tip trays placed on the deck are completely filled and tip inventory is updated. Running without full trays of tips may lead to invalid assay results. If consolidating tips into a tray, make sure tip type matches tip barcode.

DNA Preparation

1. After all required reagents, samples, and controls have been loaded, the automated method begins. The STARlet records the plate barcode as the run identifier.
2. During the method, liquid transfer verification is used by the software to monitor the transfer of reagents and samples. Errors detected by the system are reported in the run results.
3. The Messages box displays status notifications during the run, such as approximate end time of incubation steps.

QuARTS Plate Setup

Plan time and resources accordingly. Once the *QuARTS* Plate Setup is complete (Steps 1-7 below), the run on the 7500 Fast Dx must be started within 30 minutes.

1. The user is prompted to uncap and prepare the *QuARTS* reagents near the end of the Methylation & Mutation run. The STARlet unloads the reagent carrier.

The *QuARTS* reagents must be reloaded and the prompt addressed within 60 minutes, or the run will abort.

2. Remove the capped reagents from the carrier, vortex all except TUBES, 200152 to mix, and then spin all, except ELU BFR, 200130, briefly in a centrifuge.

Part #	Component
200130	ELU BFR, Elution Buffer
200131	MIX A, Oligo Mix A, Methylation

200132	MIX B, Oligo Mix B, Mutation
200133	ENZ, Enzyme Mix
200134	D CAL 1, DNA Calibrator 1, High Methylation
200135	D CAL 2, DNA Calibrator 2, Low Methylation
200136	D CAL 3, DNA Calibrator 3, High Mutation
200137	D CAL 4, DNA Calibrator 4, Low Mutation
200152	2 × TUBES, Barcoded Mixing Tubes

3. Remove caps and replace each tube on the carrier in its original location. Select 'Next' to reload the carrier.

Failure to remove cap will abort the run and result in run failure.

4. As the carrier is reloaded, all reagents in the carrier are scanned and checked against the barcodes scanned at the start of the run. If the barcodes do not match, the carrier is unloaded and the user prompted to correct the reagent placement.
5. At the end of the run, the user is prompted to remove the 96-well *QuARTS* assay plate. Select 'Next' to unload carriers.
6. Seal the plate with adhesive seal (Life Technologies, 4311971).
7. Centrifuge the sealed plate at 1900 × g to 2000 × g for 1 minute.

Run the *QuARTS* Plate

NOTE: Preparation of the 7500 Fast Dx may be completed during the QuARTS* Plate Setup *steps.

1. Place plate into the 7500 Fast Dx.
2. Power on the instrument and computer and log into the 7500 Fast Dx Real-Time PCR instrument software.

3. Select 'Create a new document'.



4. Load the Template '*QuARTS* Assay Run Template'.



5. Scan the barcode on the plate into the filename as the “Plate Name”.


7500 Fast System SDS Software
File View Tools 21CFR11 Instrument Help
New Document Wizard
Define Document
Select the assay, container, and template for the document, and enter the operator name and comments.
Assay: Standard Curve (Absolute Quantitation)
Container: 96-Well Clear
Template: QuARTS Assay Run Template
Browse...
Run Mode: Standard 7500
Operator: LabAdmin
Comments: SDS v1.4
Plate Name: <PlateBarcode>
< Back
Next >
Finish
Cancel
Ready
Disconnected
User: LabAdmin

6. Close the 7500 Fast Dx drawer.

7. After selecting finish, save the run.



File name must match the plate barcode exactly in order to link assay runs for data analysis.

8. Place plate into the 7500 Fast Dx.
9. Start the run.



10. When the run is complete, using Windows Explorer, open the desktop shortcut labeled "Instrument SDS Files."
11. Copy the <barcode>.sds file, open the Desktop shortcut labeled "SDS Files for Analysis," and paste it into that folder. Alternatively, copy the file to a USB drive for manual transfer to the Exact Sciences System computer running the Analysis Software application.



Hemoglobin Assay

The Hemoglobin Assay plate is set up on the STARlet to ensure tracking of sample positions in a 96-well plate. The remaining assay steps are completed manually. Hemoglobin Assay Supplemental Lot Information (Exact Sciences, 200219) is used with Exact Sciences System Software to enter calibration and lot information for the Hemoglobin Assay. The assay also requires the hemoglobin controls from the *Cologuard* Hemoglobin Control Kit (Exact Sciences, 100073) and entry of the *Cologuard* Hemoglobin Control Kit Supplemental Lot Information (Exact Sciences, 200313).

Preparation of Hemoglobin Samples and Reagents

Prepare Samples

1. Remove the hemoglobin samples from storage.
2. Bring hemoglobin samples to room temperature prior to use. From 2 to 8°C storage, let

stand 60 minutes on bench top. Tubes may remain at room temperature before sampling for up to 3 hours.

Prepare Hemoglobin Assay Reagents and STARlet

1. Remove Hemoglobin Assay Reagents, including Hb Calibrator (Exact Sciences, 100031) and Hemoglobin controls from *Cologuard* Hemoglobin Control Kit (100073), from storage. Allow to equilibrate at room temperature for 60 minutes.
2. Log into the STARlet Interface Software.
3. Perform daily or weekly maintenance on the STARlet, if required. Refer to *Automated DNA Preparation and QuARTS Assay Setup STARlet Setup* for further instruction.
4. Select the *Cologuard* test and Hemoglobin setup run and select 'Load Setup'.
5. Select 'Run' to start the loading process. The Hemoglobin Assay deck layout appears on the screen.



6. Thirty minutes after equilibration begins, reconstitute the Hemoglobin Assay Calibrator (Exact Sciences, 200146) and Hemoglobin Assay Controls 1-3 (Exact Sciences 100073) each with 1.5 mL deionized or higher grade water.
7. Replace stoppers and invert to ensure any material on the rubber stopper is fully reconstituted.
8. Vortex to reconstitute Hemoglobin Assay Calibrator and Hemoglobin Assay Controls 1-3 (Exact Sciences, 100073) at highest setting for 10 seconds.
9. Continue to equilibrate to the end of the 60-minute period.
10. Inspect Hemoglobin Assay Wash Concentrate (Exact Sciences, 200145) for precipitate. If precipitation is present, warm at 35-50°C for 20 minutes or until solubilized. Invert to mix as needed.

11. Prepare Hemoglobin Assay Wash by performing a 10-fold dilution of the Hemoglobin Assay Wash Concentrate (Exact Sciences 200145). For each plate processed, combine 50 mL of Hemoglobin Assay Wash Concentrate with 450 mL of deionized or higher grade water.

Ensure Hemoglobin Assay Wash is prepared fresh the day of use.

STARlet Setup for Hemoglobin Assay

1. Confirm that Tube sample grooves are void of stool sample. If stool sample remains, vortex at highest speed until grooves of the probe are void of stool sample.
2. Place the Tubes with foil side up into the carriers working back to front, left to right. Orient the tubes so that the barcodes are facing the barcode reader.

Push the tubes completely into the rack positions. Verify that all tubes are in alignment and make adjustments, if needed.

Empty positions are not permitted, except after the last sample tube. Always load from left to right. Load all six carriers regardless of the number of samples, placing empty carriers to the right of loaded carriers.

Carriers with unread sample barcodes will be unloaded. The barcodes must be adjusted and the carrier reloaded until the barcode is successfully rescanned, or barcode sample IDs may be entered manually by the operator.

3. Load Sample Buffer (Exact Sciences, 200143).
   a. Transfer the peel-off barcode from the buffer bottle to a clean 50 mL trough, mix by inversion and transfer all contents to the trough.
   b. Place trough in the indicated carrier position.
4. Load three trays 1000 µL CORE (Hamilton, 235905) pipette tips into the left tip carrier.

Load only full trays of tips, or an invalid run may result.

5. Load 4 uncapped mixing tubes (Exact Sciences, 200152) for calibrator dilution into the appropriate positions on the deck.
6. Vortex Hb CAL vial and Hb CTRL 1-3 vials at highest setting for 10 seconds, remove caps, and place vials in the appropriate position on the Hemoglobin reagent carrier.
7. Using a clean swab for each vial, wipe the inside of the neck of the Hb CAL vial and Hb CTRL 1-3 vials to remove any residual liquid from the vial neck.

Residual liquid inside the neck of the Hb CAL can Hb CTRL vials can interfere with liquid detection and may result in an invalid run.

Wash Hemoglobin Assay Plate

When adding reagents to the hemoglobin plate, add to each column using an 8-channel pipette. Maintain the same order of addition for all subsequent reagent additions.

Automated Hemoglobin Plate Setup procedure must be started within 10 minutes after wash steps are completed.

1. After the end of the 60-minute equilibration period, immediately before automated hemoglobin plate setup, wash the Hemoglobin Assay Plate (Exact Sciences, 200142) five times using the prepared Hemoglobin Assay Wash.
   a. Add 250 µL of prepared Hemoglobin Assay Wash to each well using an 8-channel pipette.
   b. Quickly flip the plate to remove the contents of the plate.
   c. Repeat Steps 1a and 1b four more times for a total of five wash steps.
   d. After the fifth wash, remove residual wash by inverting and tapping on dry paper towels.
2. Inspect the plate to ensure that residual assay wash buffer has been removed.

Residual assay wash buffer could adversely affect assay performance. If residual buffer is present, tap plate upside down on paper towels until removed.

3. Load the washed Hemoglobin Assay Plate into the appropriate carrier and select 'Next'. Follow the instructions on the screen.
4. The instrument reads the plate barcode to ensure validity of the expiration date and that the lot matches a scanned SLIB.

Automated Hemoglobin Plate Setup

1. Once all required materials are loaded and checks are completed, plate setup begins with the first pickup of tips. For complete runs (i.e., 86 patient samples), a
2. When plate setup run is complete, immediately proceed with processing the plate following the *Hemoglobin Assay Procedure*.

Hemoglobin Assay Procedure

1. Remove the assay plate, cover with Sigma Titer Top (Sigma, T-TOPS-100), and incubate at room temperature for 60 minutes.

NOTE: Incubation time from end of Automated Hemoglobin Plate Setup to Step 2 below is 1 hourmust be 60 minutes ±5 minutes.

2. Remove cover and invert plate quickly to complete remove the contents.
3. Wash the plate five times.
   a. Add 250 µL of prepared Hemoglobin Assay Wash to each well using an 8-channel pipette.
   b. Quickly flip the plate to remove the contents of the plate.
   c. Repeat Steps 3a and 3b four more times for a total of five wash steps.
   d. After the fifth wash, remove residual wash by inverting and tapping on dry paper towels.

Residual assay wash buffer could adversely affect assay performance. If residual buffer is present, tap plate upside down on paper towels until removed.

4. Add 100 µL of Antibody Conjugate (Exact Sciences, 200144) to each well using the 8 channel pipette.
5. Cover with Sigma Titer Top and incubate at room temperature for 1 hour ± 5 minutes.

6. Remove cover and invert plate quickly to complete remove the contents.
7. Wash the plate five times.
   a. Add 250 µL per well of prepared Hemoglobin Assay Wash using an 8 channel pipette.
   b. Quickly flip the plate to remove the contents of the plate.
   c. Repeat Steps 7a and 7b four more times for a total of five wash steps.
   d. After the fifth wash, remove residual wash by inverting and tapping on dry paper towels.

Residual assay wash buffer could adversely affect assay performance. If residual buffer is present, tap plate upside down on paper towels until removed.

8. Add 100 µL Substrate (Exact Sciences, 200100) to each well of the plate using an 8-channel pipette.
9. Cover with Sigma Titer Top and incubate at room temperature for 15 minutes ± 1.5 minutes.
10. Remove cover and add 100 µL of Stop Solution (Exact Sciences, 200101) to each well.
11. Proceed immediately to Step 1 of *Read Hemoglobin Plate* below.

Read the Hemoglobin Assay Plate within 15 minutes of addition of the Stop Solution.

Read Hemoglobin Plate

NOTE: The Reader Control Software may be used during a STARlet run. If needed, leave the Exact Sciences STARlet Interface Software logged in and running and switch to the Reader Control Software.

1. Log into the Exact Sciences Reader Control Software. The Run screen appears:



2. Scan or enter the barcode of the Hemoglobin plate. For manual entry of barcode, repeat the barcode entry in the second field if required by site administrator. After typing the barcode(s), select the arrow button to continue.
3. Load plate into the BioTek reader when prompted by the software.

NOTE: Place the plate into the reader so that embossed "A1" is positioned in the back left corner of the reader tray. Incorrect placement may result in invalid or incorrect results.

4. Press 'Run' to read the plate.
5. Unload the plate when prompted to do so.
6. Press 'Done' to reset the software to be ready to read a new plate.
7. The assay data are automatically transmitted to the Analysis Software.
8. In the event that a connection to the Exact Sciences Analysis Software is not available, data may be transferred manually. Copy the data file from C:\ExactSciences\Reader\Runs with file name <plate barcode>.<checksum>.reader to a USB drive if data needs to be transferred between computers.

Hemoglobin Sample Storage

1. When the hemoglobin plate setup steps are completed, cover the used Tubes with water-

resistant wrap (e.g., Parafilm or plastic wrap) and store with foil side up at 2 to 8°C for up to 7 days or freeze at <-15 °C for longer storage.

2. If repeat testing is required, see *Retest Hemoglobin Sample Tubes* in *Procedural Notes and Precautions* for detailed instructions.

Data Handling and Analysis

Data from the Methylation, Mutation, and Hemoglobin Assays are integrated and analyzed by the Exact Sciences Analysis Software. The Exact Sciences Analysis Software maintains traceability of the sample to result through sample barcodes scanned during the hemoglobin and molecular (methylation and mutation assays collectively) assay plate setup runs. For the molecular assay, data from the thermocycler (7500 Fast Dx) are imported into the Exact Sciences Analysis Software and the fluorescent signal for each channel versus cycle time is analyzed to calculate a crossing point (Cp) where the detection threshold is exceeded. This value enables the calculation of detected concentration of each DNA marker using the respective calibrators. For the Hemoglobin Assay, the optical density data is imported from the reader and the hemoglobin concentration in each sample and control is calculated from respective calibrators.

The system uses the expected values and actual results of the calibrator and control samples to assign a run status (valid/invalid) for Methylation, Mutation and Hemoglobin Assay runs. Users review, comment upon, or invalidate sample or run data in the software as required to capture any errors or invalid samples that occur during the assay procedure.

The software calculates an overall *Cologuard* score for each sample by combining the released results of each marker for that sample (linked by sample ID). A Negative or Positive result is assigned for each sample based on the *Cologuard* score. Invalid *Cologuard* results occur if any constituent assay results are invalid. Details on the use of the software can be found in the Exact Sciences System Software User's Manual.

The barcoded identification numbers on the hemoglobin sample and the DNA sample must match for Hemoglobin and DNA assay results to be matched into an overall *Cologuard* result. If a different identification number is assigned to the DNA sample, the same identification number must also be assigned to the corresponding Tube.

The barcodes affixed to the hemoglobin sample and the DNA sample must follow the appropriate barcode format, resolution, placement and ANSI/ISO specifications as directed by the STARlet instrument manual.

Review and Release Methylation, Mutation, or Hemoglobin Assay Results

Once the assay runs are complete, the run data are imported into the Analysis Software and the assay run and individual sample assay results are calculated. Results of each assay are reviewed before the software incorporates the results into the calculations that generate the *Cologuard* result. Users must release assay results in the order that they are run.

1. Log into the Analysis Software.
2. The Runs screen displays a list of recent runs that have not been released. Apply filters to locate the run to be released in the Runs table.



NOTE: To navigate back to the Runs screen at any time, select the 'Runs' button on the top of the screen. Filters will need to be re-applied.

3. If run data files need to be manually imported (e.g., the computers running the other Exact Sciences System Software applications (STARlet Interface and Reader Control) are not connected to the computer running the Analysis Software), follow these steps:
   a. Select 'Import' on the bottom right of the screen to select a file to upload.
   b. Connect the USB drive containing the data and select the file to upload. Valid file types from STARlet Interface have a .plate extension, and files from the Reader Control have .reader extensions.

   NOTE: Altering file names may make files invalid for import.

4. If the 7500 Fast Dx Real-Time PCR Instrument computer is not networked to the computer running the Analysis Software, copy the SDS file from the ABI computer. (SDS files are typically saved in the "Instrument SDS Files" folder on the desktop or the D:\Applied Biosystems\SDS Documents\ folder). Save the SDS file on an USB drive, transfer memory stick to Analysis computer, and upload file to Analysis Software using the 'Import' button.
5. Once the run list includes the desired run, select the hyperlink under the Assay column to display Run details.

Plate ID	Date	Operator	Assay	Run Status	Released	Comments
QuartsPlate01	2012-06-21T06:01:00	7500DX	Methylation	Valid	No	
QuartsPlate01	2012-06-21T06:01:00	7500DX	Mutation	Valid	No	
QuartsPlate02	2012-06-21T06:02:00	7500DX	Methylation	Valid	No	
QuartsPlate02	2012-06-21T06:02:00	7500DX	Mutation	Valid	No	
HbPlate01	2012-07-03T07:35:25	hGlobin	Hemoglobin	Valid	No	

6. The run detail screen displays run information in the Summary section and on tabs for Calibration, Calibrator, Controls, Samples, and Reagents.



7. Review data as needed in Calibrator, Calibration, Controls, Samples, and Reagents tabs.
   a. Each tab has specific information about the results of the testing for the displayed run. Select each tab to review the data.
   b. Plate IDs match the barcode on the 96-well plate used in the setup run.
   c. For the Hemoglobin setup run, one Plate ID is linked to one “Hemoglobin” assay run in the Assay column.
   d. For the Methylation & Mutation setup run, one Plate ID is linked to one “Methylation” assay and one “Mutation” assay in different rows of the Assay column.
   e. The run status of the assays on a plate is marked as Pending until assay run data for the plate are uploaded from the 7500 Fast Dx Real-Time PCR Instrument SDS file or the Reader Control Software.
   f. When an assay run status is either Valid or Invalid, the assay run results are ready for user review.
8. If user comment is required, follow these steps.
   a. If user comment is needed for an individual control, calibrator, or sample, select the ‘Comment’ area for the sample. Enter the comment and select an area outside the comment area to save the comment.

b. To comment on an entire run, select the 'Run Comments' field in the Summary section and enter the comment.
c. Each calibrator, control, and sample has a status listed as Valid or Invalid.

9. In the event of errors in manual processing steps or other errors observed, a user may invalidate sample results.

NOTE: Ensure that invalidation of individual sample results is completed prior to releasing the run.

a. To mark individual sample assay results as invalid, go to the Samples tab, select the checkbox for the individual sample (s), and select the 'Invalidate Samples…' button. Enter comment, user name and password to complete the action.

Sample results invalidated by a user are permanently marked as invalid for that particular run once Run Disposition is performed and run data are released or closed. Invalidation/Undo Invalidation cannot be performed on released or closed sample results

b. To mark all the results in a run invalid, select the 'Invalidate Run…' button. Press yes to confirm the invalidate action and enter Username and Password to complete the action.

Runs invalidated by a user cannot be released through a Run Disposition. Invalid assay results may only be Closed.

10. To undo user invalidation, select the checkbox for the individual sample and select on the 'Undo Invalidate…' button. Enter Username and Password to complete the action.
11. To release a run of assay results, select the 'Run Disposition' button on the bottom of the screen and enter Username and password to complete the action.

Invalid or Pending runs cannot be released, they may only be dispositioned as Closed.

Results from Closed runs are not available for overall test result interpretation.

Runs that have been released or closed cannot be invalidated by any user, nor can invalidation be undone on a released or closed run.

12. Releasing an assay run makes the results available for calculation of the overall *Cologuard* test result for the samples in the run.
13. To generate an assay run report, select the 'Report' button on the bottom of the screen. The report may be printed or saved to a PDF file.

Review and Release Overall *Cologuard* Results

After users confirm and release valid Methylation, Mutation, and Hemoglobin run results, the software generates an overall *Cologuard* score for each sample using each of the marker results for that sample. A Negative or Positive result is assigned based on the *Cologuard* score. Invalid *Cologuard* results occur if any of the constituent assay results are invalid. Active users in the Supervisor or Administrator role may disposition (release or close) overall *Cologuard* test results. Once matched samples from Methylation, Mutation, and Hemoglobin Assay runs have been released, overall test results for each sample are available.

1. Log into the Exact Sciences Analysis Software.
2. Select the 'Overall' button to view the overall test results table. The filter defaults to show results that have not yet been released. Filters may be applied to narrow the table to show only specific samples or types of results.



   a. For each sample, ! (Flags), Released, Sample ID, Test Result, Score, and Methylation, Mutation, and Hemoglobin run status are displayed on the Overall screen.
   b. To select a group of samples for action, enter a checkmark in the samples selection box.
3. To review a summary of individual assay results for a sample, select the hyperlink for the Sample ID.
   a. The Sample Detail report with overall result and individual assay results for the selected sample are displayed.



b. The Report button displays a printable copy of the sample detail report for the selected rows that can also be saved to PDF file.



4. In the event of errors in manual processing steps or other errors observed, a Supervisor-

level user may invalidate overall test results before they are dispositioned.

a. To invalidate an overall test result for a sample, go to the Overall screen, select the checkbox next to the sample, and touch the Invalidate button. A confirmation screen appears. Select 'Yes' to confirm and enter your comment, username and password to complete the action.

Samples invalidated by a user are permanently marked as invalid once a Disposition is performed and sample results are released.

b. To undo user invalidation, select the checkbox for the individual sample and select the 'Undo Invalidate…' button and enter your username and password to complete the action.

5. To disposition the overall results for a sample or group of samples, select the checkbox(s) beside the sample(s) and select the 'Disposition' Button on the bottom of the screen. Enter Supervisor-level username and password, then:

NOTE: The column header can be used to select all the samples displayed in the Overall results table in order to release or close multiple results simultaneously. The user must double-click on the column heading.

NOTE: If all selected results cannot be released, the 'Release' button will not be available.

a. Select 'Release' to release overall test results for export. Released overall test results are automatically written to LIS Export file. Released overall test results are written to MFG Export files when the MFG Export function is selected.

b. Alternatively, select 'Close' to indicate that the results should not be exported. Closed results cannot be exported in LIS Export or MFG Export files.



6. Select the 'Disposition' button and enter user credentials to release or close the selected

sample results. If a button is not enabled, the corresponding action is not available for all of the selected samples.

Interpretation of Results

The Exact Sciences System Software imports run data into the Analysis Software. The software calculates assay results for controls and individual samples. Results of each assay, methylation, mutation, and hemoglobin, are reviewed prior to the release of the assay run data in the software. Prior to release assay run results, deviations to process in the assay set up are noted for affected samples. If a deviation (e.g., operator error, instrument error) occurs, the deviation may compromise the results of the test regardless of the control validity. In the event that a sample result is compromised, the individual result may be invalidated. After results are reviewed for deviations, results may be released for the entire assay run. If an assay control fails, or the operator invalidates an entire assay run, no sample results will be present for that assay and all samples in the run must be retested. Only valid sample results from valid assay runs are used to calculate an overall *Cologuard* score.

Users may also review overall *Cologuard* results and invalidate sample results as needed based on events and issues known to the user, such as individual sample or reagent contamination, process errors, or automation abort with unconfirmed completed steps. Users may enter comments in the software for any sample result that is invalidated by the user for a known technical error that cannot be detected by the software.

As users confirm and release valid methylation, mutation, and hemoglobin results for an assay run, the Analysis Software will link the constituent assay results by sample ID and calculate a *Cologuard* score. The score is used to assign the final *Cologuard* result: Positive, Negative, or Invalid. Valid *Cologuard* results may be released and reported. An invalid *Cologuard* result occurs if the sample result from any constituent assay is invalid.

Procedural Notes and Precautions

Additional Stool Homogenate Aliquots

If additional aliquots of stool homogenate are desired for testing or for sample archive, label and prepare additional tubes as in *Preparation of Stool Homogenate* and store as directed.

1. Do not store samples in a Stool Container. Any additional homogenate aliquots should be processed and stored in 50 mL tubes at the same time as testing aliquots.
2. Homogenate aliquot tubes can be stored frozen (< -15°C) for 1 month.
3. Homogenate aliquot tubes can be stored frozen (<-70°C) for 2 years.

Insufficient Supernatant

If less than 5 mL of supernatant are obtained from a sample during *Prepare Samples and Perform DNA Capture, Prepare Supernatant* and additional stool sample aliquot tubes are available, the prepared supernatant may be stored until additional material has been obtained, as described in Steps 5b and 15b of the *Prepare Supernatant* section. If 5 mL of supernatant are still not obtained or additional sample aliquot tubes are not available, discard supernatant and request a new sample.

Enter Supplementary Lot Information

Use the following procedures to support STARlet Setup when required.

1. Log in to the Analysis Software.
2. Scan all of the 2D barcodes on the SLIB sheet in any order. Each scan is acknowledged in a pop-up dialog. When all barcodes in a SLIB have been scanned, the entry status is indicated as “added” or “already been imported”.
3. Supplementary lot information loaded into the system can be viewed in the SLIB summary table by selecting the SLIB button in the Analysis Software.

Prepare Capture Beads (for >6 tubes)

The “Prep Beads” protocol referred to in Step 9 of *DNA Capture, Prepare Capture Beads* aspirates up to six tubes of beads. If operator prefers to prepare greater than six tubes, use the following steps.

1. Set the Capture Incubator (Exact Sciences, 300546) to preheat (“Bead Prep 1” program).
2. Allow Capture Beads (Exact Sciences, 200150) to sit at room temperature for a minimum of 30 minutes.
3. Vortex Capture Beads at highest setting for 30 seconds to suspend the beads.
4. Label the 50 mL conical tubes with Capture Bead preparation date and lot information.

Labels used in the Capture Incubator have specific requirements for size, material and thickness. Label the bead preparation tubes using permanent marker or refer to Capture Incubator User’s Manual for detailed label specifications.

5. Transfer 3.25 mL of beads to each of the labeled 50 mL tubes.
6. Add 10 mL of Capture Bead Pre-wash (Exact Sciences, 200120) to each tube and secure the 50 mL tube cap.
7. When the Capture Incubator has reached temperature and display prompts user to insert test tubes, place tubes in the Capture Incubator. Close cover and press the ‘Start/Select’ button to proceed with the cycle.
8. When the cycle is complete, remove tubes from the incubator and place in the centrifuge with appropriate balance tube, if necessary. Centrifuge the tubes until the centrifuge reaches 500 × g and hold for 1-10 seconds.
9. Remove caps and transfer the tubes to the Capture Aspirator (Exact Sciences, 300490). Execute the “**Bind 10 min**” protocol to remove supernatant from the tubes.
10. When aspiration run is complete, remove tubes from the Capture Aspirator and add 3.25 mL of fresh Capture Bead Pre-wash solution to each tube, replace caps, and vortex at highest setting until all Capture Beads are suspended.

NOTE: Once the Capture Beads have been prepared, they can be stored in closed tube for up to 7 days at 2-8°C before use.

Procedure for Performing Partial Runs

Each *Cologuard* kit contains sufficient materials to test 96 samples including 86 patient samples, required controls and calibrators. If fewer than 86 patient samples are processed, use the following guidelines to ensure valid results.

1. DNA Capture steps are performed manually and are processed in sets of up to 22 patient samples. At least one positive DNA Control (D CTRL 1 or D CTRL 2) is required for every distinct set of DNA Capture samples.
2. Input samples for automated processing include a maximum of 43 samples with corresponding controls for a total of 46 samples and controls per batch. In maximum batch sizes, two full sets of DNA capture samples (46 samples and controls) are used in each batch of automated processing on the STARlet (DNA Preparation and *QuARTS* Assay steps). To process the suggested 86 samples, two batches of DNA Preparation and *QuARTS* Assay steps are to be performed. Reagent fill volumes are designed to be sufficient for this full capacity of two automated batches. Leftover reagents must be discarded at end of run, even if less than a full run of samples was performed.
3. In maximum batch sizes, Hemoglobin Assay steps are performed in 96 well assay plates, using the STARlet for plate setup, followed by a 96 well ELISA based assay. Hemoglobin Controls (Hb CTRL 1, Hb CTRL 2, and Hb CTRL 3) are required for each batch of Hemoglobin Assay samples. Reagent fill volumes are designed for full capacity. Leftover reagents must be discarded at end of run, even if less than a full run of samples was performed.
4. All three DNA controls are required for each DNA plate setup run, regardless of the number of samples run.
5. Use full reagent containers and troughs for setup runs to avoid invalid results. The system does not adjust reagent usage for runs with less than the maximum number of samples.
6. Discard leftover reagents at the end of the run.

Quality Control

Process Controls

1. Required controls must be present in each assay plate setup run to achieve valid results. The system software will not proceed with the method if not all required controls are present.
   a. Input samples for *Automated DNA Preparation and* QuARTS *Assay Setup* must include D CTRL 1, D CTRL 2, and D CTRL 3 in each setup run.
   b. Hemoglobin Controls (Hb CTRL 1, Hb CTRL 2 and Hb CTRL 3) are required for each setup run of Hemoglobin Assay samples.
2. Process controls must yield expected results, or the assay run will be invalid. Allowed ranges for control results are defined by the *Cologuard* DNA Control Kit Supplemental Lot Information (Exact Sciences, 200315) and *Cologuard* Hemoglobin Control Kit Supplemental Lot Information (Exact Sciences, 200313) for each lot of controls.

Lot Matching and Sample Tracking during Processing

1. Users are responsible for ensuring that reagent lots used in manual processing steps are correctly lot matched to reagents used in automated processes. Refer to the DNA and *QuARTS* Reagents Supplemental Lot Information (Exact Sciences, 200218) and the Hemoglobin Assay Supplementary Lot Information (Exact Sciences, 200219).
2. Users are responsible for tracking sample IDs and documentation of processing errors from manual Processing and DNA Capture steps.

Review and Release *Cologuard* Results

1. As users confirm and release valid methylation, mutation, and hemoglobin results for a run,

the software will match the results by sample ID and generate an overall *Cologuard* result of Positive, Negative, or Invalid. Sample results invalidated during previous review steps will be called 'Invalid.'

2. Users should review and comment upon overall *Cologuard* results and invalidate sample results as needed based events and issues known to the user.

Troubleshooting Guide

DNA Capture

Centrifuged Stool Sample Abnormalities

1. If the solid/liquid interface of a centrifuged stool sample is unclear, remove 14 mL from the topmost portion of the sample during *Prepare Samples and Perform DNA Capture, Prepare Supernatant*, Step 5.
2. If a solid layer is present above the liquid layer of the centrifuged stool sample, hold the tube at an angle with the tip below the solid layer while pipetting to avoid aspiration of the solid material.

Incomplete Dispersion of Inhibitor Removal Tablet

If Inhibitor Removal Tablet does not immediately disperse during *Prepare Samples and Perform DNA Capture*, *Prepare Supernatant*, Step 5, use the following procedure.

1. Vortex the sample and inhibitor removal tablet at highest setting until tablet breaks apart.

If step above is unsuccessful in breaking up and dispersing the Inhibitor Removal Tablet, request a new sample, as this sample is considered invalid.

2. Proceed with Prepare Samples and Perform DNA Capture, Prepare Supernatant, Step 6.

Broken Spin Filter

Spin filter failure while centrifuging the "SPN" tubes during *Prepare Samples and Perform DNA Capture*, *Prepare Supernatant*, Step 12, is evident by the presence of white inhibitor removal tablet particles present in the filtrate. If this is noted, use the following procedure.

1. Remove the broken spin filter from the SPN tube and cap the tube using a clean cap.
2. Shake the tube to mix sample and the dispersed inhibitor removal tablet and then transfer to a new labeled tube fitted with a spin filter.
3. Place spin filter tube into the centrifuge. Ensure that centrifuge is balanced and spin for 6 min at 3300 × g.
4. Proceed with Prepare Samples and Perform DNA Capture, Prepare Supernatant, Step 13.

Capture Incubator Produces Error Message

If the Capture Incubator produces an error message while running the EXAS8 program during *Prepare Samples and Perform DNA Capture*, *Capture Incubation,* Step 7, refer to the Capture Incubator User's Manual. A description of each error code displayed on the Capture Incubator can be found in the Appendices.

Capture Aspirator Produces Error Message

If the Capture Aspirator produces an error message or a power outage while running the “BIND 10 min” program during *Prepare Samples and Perform DNA Capture*, *Capture Incubation,* Step 9, refer to the Capture Aspirator User’s Manual (Troubleshooting section).

Incomplete Aspiration of Supernatant

If liquid remains in the 50 mL conical tubes after aspiration during *Prepare Samples and Perform DNA Capture*, *Capture Incubation*, Step 8, use the following procedure.

1. Bring tube volume to 10 mL using Capture Wash.
2. Mix to ensure that beads are suspended in liquid.
3. Place tube in Capture Aspirator and repeat the “BIND 10 min” program.

NOTE: Empty positions in rows that contain sample tubes should be occupied to ensure optimal aspiration. Place a 50 mL tube filled with 17.5 mL of water into each empty space.

4. Proceed with Prepare Samples and Perform DNA Capture, Prepare Supernatant, Step 10.

Sample Appears Gelatinous After Aspiration

If a sample appears gelatinous after aspiration during *Prepare Samples and Perform DNA Capture*, *Capture Incubation*, Steps 10-11, observe one of the following recommendations.

1. Re-aspirate sample following Troubleshooting Procedure for Incomplete Aspiration of Supernatant.
2. Continue to process sample if it appears that it will not cause an issue on the automated platform.
3. If sample still appears too gelatinous to run on the automated platform, discard and request a new sample.

Beads Not Fully Suspended After Shaking

If beads remain on the sides of the 50 mL conical tubes after shaking during *Prepare Samples and Perform DNA Capture*, *Capture Incubation*, Step 12, use the following procedure.

1. Rotate tube in Shaker rack, and mix for 1 minute at 400 RPM.
2. Confirm that Capture Beads are suspended in tube.
3. If beads are not suspended, rotate tube and repeat Steps 1 and 2.

DNA Preparation and *QuARTS* Assay

Methylation/Mutation Setup Run Aborts

1. If samples have not been transferred from the 50 mL conical tubes to the deep-well plate, place a new cap on the sample tubes and store at 2 to 8°C until ready to rerun in the methylation/mutation setup.
2. If samples have been transferred to the deep-well plate, discard all samples and reagents on deck. Samples must be repeated from Step 1 of *Prepare and Label Sample Tubes, Prepare Samples and Perform DNA Capture.*

3. .

Hemoglobin Assay

Hemoglobin Setup Run Aborts

1. If samples have not been punctured, place at 2 to 8°C until ready to rerun in the Hemoglobin setup. Samples may remain at room temperature for up to three hours. To rerun the setup, begin with *Prepare Hemoglobin Assay Reagents and STARlet*, Step 1.
2. If samples have been punctured, rerun within 3 hours or cover with water-resistant cover and store foil-side up at 2 to 8°C until ready to rerun the Hemoglobin setup. Before rerunning, confirm that all sample tubes are correctly placed into sample racks.

Retest Hemoglobin Sample Tubes

If repeat testing is required due to a Hemoglobin Assay run failure or individual sample failure in a Hemoglobin Assay, the following procedure should be used for repeat testing.

1. When initial hemoglobin plate setup steps are completed, cover hemoglobin sample tubes with a water-resistant cover and store in racks with foil side up at 2 to 8°C for up to 7 days.
2. Remove punctured hemoglobin sample tubes from 2 to 8°C and equilibrate to room temperature.
3. Remove plastic wrap and proceed with Hemoglobin Assay, Preparation of Samples and Reagents, Prepare Samples, Step 4.

Procedural Limitations

- DO NOT mix or substitute reagents from Supplemental Lot Information containing different lot groupings.
- DO NOT use any reagent after its expiration date.
- DO NOT store reagents in “frost-free” freezers.
- Only use with specimens collected with the *Cologuard* Collection Kit (Exact Sciences, 100026).
- *Cologuard* reagents are intended to be used only with the Exact Sciences System Software and instrumentation.
- To ensure the integrity of the sample, the laboratory must receive patient specimens within 72 hours of collection. Refer to *Cologuard Laboratory Procedure, Receipt of Cologuard Collection Kit* in this document.
- The barcoded identification numbers on the hemoglobin sample and the DNA sample from a Collection Kit must match for Hemoglobin and DNA assay results to be matched into an overall *Cologuard* result.
- Instrument and assay procedures reduce the risk of contamination during the laboratory procedure. However, good laboratory practice and careful adherence to the procedures specified in this document are important to reduce further risk of nucleic acid contamination from calibrators, positive controls, or specimens.
- Invalid results could occur from improper handling or storage, technical error, or sample mix up. Ensure that only trained personnel perform the laboratory procedure.
- *Cologuard* results are qualitative. The numeric value of the *Cologuard* Score is not indicative of extent of disease.

- A negative test result does not exclude the possibility that the patient may have a pre-cancerous or cancerous polyp. A false negative result with *Cologuard* could potentially delay colonoscopy and a potentially delayed diagnosis of disease.
- A positive *Cologuard* test suggests the presence of pre-cancerous polyps and/or cancer. A false positive result could result in an additional invasive screening procedure for the patient, such as colonoscopy, and thus expose patients to the attendant risks associated with such a procedure.
- Results from the *Cologuard* cross-contamination analysis indicated no observed cross-contamination from automated equipment or repeated testing of manual steps.However, operator-induced cross-contamination can occur if procedures are not carefully followed.
- Cancers in organs connected to the digestive tract (i.e., pancreas and liver) may shed markers that could be detected by *Cologuard*. As such, it is expected that a certain level of reactivity will be observed in cases of these cancers. Refer to *Performance Characteristics, Sensitivity and Cross-Reactivity* in this document.

Performance Characteristics

Clinical Cutoff

The cut-offs and the algorithm for the *Cologuard* sDNA-based colorectal cancer screening test were established based on an evaluation of a panel of donor samples that were categorized by colonoscopy. Variable selection for the *Cologuard* model was performed as a stepwise selection with the main variables assessed one at a time based on their respective statistical significance. The total sample size of the dataset for algorithm development included 953 samples, including 794 normal samples, 73 advanced adenomas and 86 cancers. The derived *Cologuard* algorithm sensitivity and specificity compared to colonoscopy outcome demonstrated a sensitivity of approximately 98% for cancer and approximately 57% for advanced adenoma.

Analytical Sensitivity

Sensitivity: Limit of Blank (LoB), Limit of Detection (LoD), Limit of Quantification (LoQ) and Linearity.

LoB, LoD, and LoQ studies were performed for both the methylation and mutation component (i.e., molecular assay) and the hemoglobin assay component of *Cologuard* based on guidance from the CLSI Standard: EP17-A (*Protocols for Determination of Limits of Detection and Limits of Quantitation; Approved Guideline*). For molecular assays, such as the *QuARTS* component of *Cologuard*, the signal from the blank wells is absent. Therefore, the LoD and LoQ were established through means independent of a Limit of Blank (LoB) measurement. For *NDRG4*, *BMP3*, *KRAS* 38A, *KRAS* 35T, and *ACTB*, a minimum of 60 replicates per marker near the LoD concentration were tested across 6 samples at the expected LoD concentration within a dilution series in order to use probit analysis to predict LoD. For LoQ, a minimum of 60 replicates per marker near the anticipated LoQ concentration was tested across 6 samples and the lowest concentration with total error less than that of the total error goal of 20% CV on log strands was the determined LoQ.

Linearity and Linear Range studies using concentrations above and below the anticipated linear range were tested in the molecular assay and hemoglobin assay components of *Cologuard*.

Linearity studies were performed based on guidance from CLSI Standard: EP6-A (*Evaluation of the Linearity of Quantitative Measurement Procedures: A Statistical Approach; Approved Guideline).* All markers were individually assessed at 9 levels spanning 5 logs including concentrations 30% above and below the anticipated upper LoQ and LoD, respectively. The markers were tested at 8 replicates total per level per marker, 2 replicates per plate, 2 instruments with 1 operator per instrument. The 9 concentrations were (in log strands per reaction): 5.59, 5.48, 5.30, 4.30, 3.30, 2.30, 1.30, 1.00, and 0.85

In summary, analytical sensitivity characteristics for *Cologuard* were observed as follows (**Table 1**): The methylation markers NDRG4, BMP3, and BT-ACT have a LoD at 0.702-0.738 log strands. KRAS was assigned a LoD value to that of KRAS 35T, the KRAS mutant with the highest LoD. The KRAS LoD is 1.058 log strands with a CI range that encompasses the lower cutoff used in the *Cologuard* software. The established LoD meets criteria of the ability to detect one percent mutation or methylation when 3.000 log strands of ACTB are present and the LoD is less than or equal to 1.300 log strands. The molecular assay LoQ is 1.176 log strands per reaction. This exceeds the input requirement that the LoQ must be less than 2.000 log strands. The molecular assays demonstrate good linearity over at least 5 logs and that R^2 is = $\geq$ 0.996 for all targets.

Analytical sensitivity characteristics for *Cologuard* were observed as follows:

Table 1: Analytical Sensitivity Characteristics Summary

Performance Characteristic	Molecular Assay	Hemoglobin Assay
Limit of Blank	Not Applicable	0.4 ng/mL
Limit of Detection	Methylation Markers: *NDRG4, BMP3* and *ACTB* 0.702 to 0.738 log strands Mutation Markers: *KRAS* 1.058 log strands	1.3 ng/mL
Limit of Quantification	LoQ ≤ 1.176 log strands	4.8 ng/mL
Assay linearity	R^2 = $\geq$ 0.996 Linear range = 1.1760 to 5.591 log strands	Linear range = 4.8 ng/mL to 500 ng/mL No hook effect observed for concentrations up to 100 µg/mL

Specificity and Cross-Reactivity

***Cologuard* Molecular Assay Cross-Reactivity with Wild-Type KRAS**

The potential for cross-reactivity with wild-type *KRAS* was evaluated by testing two levels of *KRAS* wild-type DNA in the *Cologuard QuARTS* methylation and mutation assays. *KRAS* wild-type DNA was assessed at levels of 20,000 copies of wild-type *KRAS*, which is greater than the average expected to be seen in normal human stool samples, and 200,000 copies of wild-type *KRAS*, 10 times higher.

Results from this study indicated that cross-reactivity for wild-type *KRAS* at 200,000 copies was nearly 0% for the methylation assay and 0.01% for the mutation assay.

***Cologuard QuARTS* Partial Methylation Testing**

Many genes have elevated methylation in their promoter region in colorectal cancer, whereas the same genes have low levels of methylation in normal colon epithelial cells. The DNA oligonucleotides used in the *Cologuard* methylation assay are designed to be a perfect match to fully methylated DNA in *NDGR4* and *BMP3*.

The analytical specificity of the DNA methylation assay component of *Cologuard* was tested against partially methylated *BMP3* and *NDRG4* DNA targets using the *QuARTS* assay. The testing utilized synthetic DNA targets that contained all possible permutations of partial methylations in the *QuARTS* assay footprint region of *BMP3* and *NDRG4*.

The study results demonstrated that *Cologuard* is specific for highly methylated DNA, specifically highly methylated *NDRG4* and *BMP3*. At least five of eight potential methylation sites for *BMP3* and five of nine potential methylation sites for *NDRG4* have to be methylated for any reactivity in *Cologuard*. With respect to *NDRG4*, the percent cross-reactivity was 2.5%, indicating that the analytical specificity for total methylations in *NDRG4* is 97.5%. With respect to *BMP3*, the percent cross-reactivity was 1.8%, indicating that the analytical specificity for total methylations in *BMP3* is 98.2%, above the 95% specificity outlined in the pre-defined acceptance criteria.

***Cologuard* Hemoglobin Assay Cross-Reactivity and Specificity**

The ability of the Hemoglobin Assay to detect hemoglobin in specimens heterozygous for Hemoglobin S (HbS) and Hemoglobin C (HbC) was evaluated. Samples used for testing Hb variants consisted of a stool sample background spiked with normal, HbS heterozygous, or HbC heterozygous whole blood. The Hemoglobin Assay detected both HbS and HbC variants, when comparing equivalent volumes of blood from normal and heterozygous variant specimens.

Additionally, cross-reactivity of *Cologuard* Hemoglobin Assay with animal hemoglobin and myoglobin was evaluated. Samples used for testing animal blood cross-reactivity consisted of a stool sample background spiked with animal whole blood. Samples used for testing myoglobin cross-reactivity consisted of a stool sample background spiked with prepared meat extracts or purified myoglobin. Thirteen replicates of each sample type were tested with the *Cologuard* Hemoglobin Assay. Included in the study were 21 samples: 6 human blood samples, 8 animal blood samples, 4 animal myoglobin samples prepared from meat extracts, 2 animal myoglobin samples from purified myoglobin, and 1 negative sample.

Mean Hb concentrations for all animal hemoglobin and myoglobin samples were less than the limit of detection (LoD) of the assay (1.3 ng/mL) after the mean concentration of the Hb Negative Stool Sample was subtracted, indicating that no cross-reactivity was detected.

***Cologuard* Cross-Reactivity with Non-Colorectal Cancers and Diseases**

The potential for cross-reactivity with non-colorectal cancers was evaluated by testing 151 specimens from subjects with cancers or diseases other than CRC that have a potential association with the GI tract, or inflammatory conditions that could affect the screening population for *Cologuard*. Samples were tested with both the Molecular and Hemoglobin Assay components of *Cologuard*. Overall *Cologuard* Scores were then generated to assess whether reactivity was found with any of these non-CRC samples.

Cancers in organs connected to the digestive tract (i.e., pancreas and liver) may shed markers

that could be detected by *Cologuard*. As such, it is expected that a certain level of reactivity will be observed in cases of these cancers. The results of this testing are included in Table 2 below.

Table 2: Incident Rates and Contribution to *Cologuard* Positivity for Non-CRC Diseases and Cancers

Disease or Cancer*	Number of specimens tested	Incident rate per 10,000**	% Positivity of Cologuard	Number positive Cologuard calls in 10,000 subjects
Bladder Cancer	17	2.3	17.6%***	0.4
Breast Cancer	14	12.4	0.0%***	0.0
Esophagus Cancer	11	0.5	18.2%***	0.1
Gynecologic Cancer	11	2.0	36.4%	0.7
Hepatic Cancer	6	0.8	50%	0.4
IBD	18	1.0	38.9%	0.4
Lung Cancer	10	6.5	20.0%***	1.3
Lupus	17	0.2-0.8	11.8%***	0.1
Pancreas Cancer	12	1.2	41.6%	0.5
Prostate Cancer	12	15.5	8.3%***	1.3
Rheumatoid Arthritis	15	4.1	26.7%***	1.1
Stomach Cancer	8	0.8	25.0%***	0.2
Total per 10,000 subject		NA	NA	6.5

*Listed value for gynecologic cancer is the sum of ovarian and cervix uteri cancers.
**For cancers, figures were obtained from the National Cancer Institute (http://seer.cancer.gov/statfacts/index.html). For other diseases, figures were obtained from the Centers for Disease Control and Prevention (http://www.cdc.gov).
*** Not significantly greater than what would be expected in a "normal" population.

Based on the results of this study, considering the non-CRC diseases and cancers where the percent positivity was slightly higher than would be expected in a normal population, the expected positivity for the tested diseases would result in only a minimal (0.02%) decrease in specificity for Cologuard (or two positive calls per 10,000 screening patients tested). .

Precision and Reproducibility

A laboratory-to-laboratory precision and reproducibility study was performed to assess variation of the *Cologuard* assay measurement system based on guidance from the CLSI Standard: EP15-A2 (*User Verification of Performance for Precision and Trueness; Approved Guideline*). As part of the study, a variance component analysis was performed by sample type for the *Cologuard* system to estimate the components of precision for each source of variation (operator, run, site, and replicate) as well as total variation for each individual marker and the overall *Cologuard* Score.

The study was performed at three sites (100, 200, 300), with a minimum of two operators at each site. A total of 22 *Cologuard* runs were performed at each site, 11 per operator. Each run involved 42 samples, including six replicates of each of the following: four stool pool samples (negative, high negative, low positive and high positive) and three control samples (negative, low positive and high positive).

For the molecular assay component of *Cologuard*, the stool sample types were prepared by

combining characterized residual stool samples. The samples were characterized as positive or negative for CRC based on colonoscopy results. Subsequently, these residual clinical stool specimens were tested with the *Cologuard* assay and combined to establish the planned DNA content of samples for use in this study. Spiked synthetic DNA was used to create the contrived control samples.

For the hemoglobin assay component of *Cologuard*, the clinical stool pools were prepared by adding fresh whole blood to normal patient stool pools. Specifically, whole blood was spiked into stool samples and diluted to the appropriate concentration. Control samples (including negative, low, and high controls) were provided to each testing site in lyophilized form for reconstitution prior to testing.

Percent agreement between sites was evaluated by generating two-by-two (2 x 2) contingency tables for negative and positive results for all site pairs, calculating the average positive agreement (APA) and average negative agreement (ANA), and calculating the exact two-sided lower 95% confidence interval by the Clopper-Pearson method. The resulting lower confidence limit was then compared to the target agreement rate of 0.95. The lower confidence interval for percent agreement of all site pairs was ≥0.95. Inter-site agreement is shown in Table 3 and shows minimal variation.

Table 3: Inter-site Agreement

Site Comparison	Number Agreed	Total Compared	Agreement Rate	95% CI Lower Bound***
ANA* – Site 100 and Site 200	768	777	0.988	0.978
APA** – Site 100 and Site 200	1026	1035	0.991	0.983
Site Agreement – Site 100 and Site 200	897	906	0.990	0.982
ANA – Site 100 and Site 300	744	746	0.997	0.990
APA – Site 100 and Site 300	1012	1014	0.998	0.993
Site Agreement – Site 100 and Site 300	878	880	0.998	0.992
ANA – Site 200 and Site 300	756	764	0.990	0.979
APA – Site 200 and Site 300	1004	1012	0.992	0.984
Site Agreement – Site 200 and Site 300	880	888	0.991	0.982

*ANA = Average negative agreement
**APA = Average positive agreement
***Clopper-Pearson Confidence Interval

Descriptive statistics were separately calculated for all marker/sample combinations. %CV was calculated only for samples with an expected positive result. Inter-site descriptive statistics are provided below (Table 4).

Table 4: Inter-Site Descriptive Statistics for the *Cologuard* Score

Sample	N	Mean	Lower 95% CL for Mean	Upper 95% CL for Mean	Std Dev	Total %CV
Negative Stool Pool	387	9.98	9.65	10.31	3.31	NA
High Negative Stool Pool	394	62.92	60.24	65.61	27.14	NA
Low Positive Stool Pool	393	391.11	383.66	398.36	74.13	18.96

High Positive Stool Pool	394	978.34	977.44	979.24	9.13	0.93
Negative Control	392	6.35	6.26	6.44	0.90	NA
Low Positive Control	393	626.24	621.39	631.09	48.91	7.81
High Positive Control	393	963.38	962.30	964.46	10.89	1.13

Overall, the assay was highly reproducible with inter-site agreement values of the lower confidence interval of >95% (Table 4) and all of the positive *Cologuard* Scores had inter-site CVs of less than 20% (Table 5).

An additional multi-operator prospective study was conducted to further evaluate the intermediate precision and repeatability of the processes developed for use with the *Cologuard* assay with high negative and low positive stool samples containing levels of DNA or hemoglobin that together, give a *Cologuard* Score near the cut-off of the *Cologuard* assay. The results from the supplemental precision and reproducibility study demonstrated acceptable %CV for samples near the assay cut-off.

The study was performed at one site with two operator teams. A total of 22 *Cologuard* runs were performed during the study, in which each operator team performed 11 complete runs, with each run requiring 2 shifts to complete. Each run involved 12 samples, including six replicates of each of the high negative and low positive stool samples. A single lot of *Cologuard* reagents and controls was used throughout the study.

Percent agreement between operators was evaluated by generating two-by-two (2 x 2) contingency tables for negative and positive results, calculating the weighted average negative agreement (ANA) and average positive agreement (APA), and calculating the exact two-sided lower 95% confidence interval. The lower confidence interval for total agreement of all pairs was >0.95% agreement of all site pairs. Descriptive statistics were calculated and are shown in Table 5 below.

Table 5: Inter-Operator Descriptive Statistics for the *Cologuard* Score

Sample	N	Median	Mean	Lower 95% CL for Mean	Upper 95% CL for Mean	Std Dev	%CV
High Negative Stool Pool	132	141.9	142.0	139.1	145.0	17.4	12.2
Low Positive Stool Pool	132	238.5	236.5	232.7	240.2	21.8	9.2

Overall the additional testing continued to demonstrate that the assay was highly reproducible with inter-operator agreement values of the lower confidence interval of >95% and all of the positive *Cologuard* Scores had inter-operator CVs of less than 20%.

Lot-to-Lot Reproducibility

Lot-to-Lot reproducibility was evaluated for *Cologuard* based on guidance from the CLSI Standards: EP5-A2 (Evaluation of Precision Performance of Quantitative Measurement Methods; Approved Guideline); EP15-A2 (User Verification of Performance for Precision and Trueness; Approved Guideline); EP12-A2 (User Protocol for Evaluation of Qualitative Test Performance; Approved Guideline); and I/LA28-A2 (Quality Assurance for Design Control and Implementation of Immunohistochemistry Assays; Approved Guideline).

Lot-to-Lot reproducibility was assessed by testing a sample panel comprised of seven samples containing various levels of DNA and hemoglobin, using three lots of *Cologuard* reagents and controls.

For the molecular assay component of *Cologuard*, the stool sample types were prepared by combining characterized residual stool samples available to Exact Sciences. The samples were characterized as positive or negative for CRC based on colonoscopy results. Subsequently, these residual clinical stool specimens were tested with the *Cologuard* assay and combined to establish the planned DNA content of samples for use in this study. Spiked synthetic DNA was used to create the contrived control samples.

For each sample in the panel, there were 24 sample results per lot and 72 sample results for the entire study. Across the seven samples in the panel, there were 168 results per lot, and 504 results for the entire study.

The mean, SD, %CV, N, minimum value and maximum value were calculated for each marker or each lot and test sample. Additionally, *Cologuard* Scores were determined. Percent positive results for the *Cologuard* Score were analyzed across lots and for lot to lot. Variance component analyses were also conducted.

Descriptive statistics were calculated for all marker/sample combinations, including median, mean, mean upper and lower 95% confidence intervals, standard deviation, and coefficient of variation values. %CV was calculated only for controls with expected result of positive. Descriptive statistics were calculated both within and across lots. Descriptive statistics for this study are shown below (Table 5). The *Cologuard* Score %CV values for positive samples were within the pre-specified acceptance criteria, ranging between 0% and 16.8%.

Table 6: Descriptive Statistics for Lot-to-Lot Cologuard Score

Sample Name	N	Median	Mean	Lower 95% CL for Mean	Upper 95% CL for Mean	Std Dev	CV
Negative Stool Pool	72	9.47	11.39	10.19	12.58	5.07	NA
High Negative Stool Pool	72	64.46	57.74	51.12	64.36	28.18	NA
Low Positive Stool Pool	71	380.75	373.93	359.03	388.84	62.98	16.84
High Positive Stool Pool	71	973.92	972.88	970.36	975.40	10.64	1.09
Negative Control	70	6.33	6.40	6.21	6.59	0.79	NA
Low Positive Control	71	584.09	579.52	570.09	588.95	39.85	6.88
High Positive Control	71	1000	1000	1000	1000	0	0

Percent agreement between lots was evaluated by generating 2 x 2 tables for negative and positive results for all lot pairs, calculating the average positive agreement (APA) and average negative agreement (ANA). Testing of samples with various levels of hemoglobin and DNA markers demonstrated a percent agreement for positive and negative samples across multiple lots between 98.6% and 100%, with a lower confidence limit above 95% (Table 6).

Table 7: Lot-to-Lot Percent Agreement

Lot Comparison	Number Agreed	Total Compared	Agreement Rate *	95% CI Lower Bound***
ANA* - Lot1 and Lot2	142	142	1.0000	0.9744
APA** - Lot1 and Lot2	188	188	1.0000	0.9806
Lot Agree - Lot1 and Lot2	165	165	1.0000	0.9779
ANA - Lot1 and Lot3	140	142	0.9859	0.9501
APA - Lot1 and Lot3	180	182	0.9890	0.9609
Lot Agree - Lot1 and Lot3	160	162	0.9877	0.9561
ANA - Lot2 and Lot3	142	144	0.9861	0.9507
APA - Lot2 and Lot3	184	186	0.9893	0.9617
Lot Agree - Lot2 and Lot3	163	165	0.9879	0.9569

NOTE: Proportion values are point estimates used to determine the Clopper-Pearson 2-sided Confidence Interval. Only Clopper-Pearson Lower Limit values are shown in the above table.
*ANA = Average negative agreement
**APA = Average positive agreement
***Clopper-Pearson Confidence Interval

The study demonstrated that *Cologuard* results are reproducible across multiple reagent lots.

Robustness

The *Cologuard* performance was assessed in response to defined variable factors (see below) at specific steps in the test procedure, using both the molecular assay and hemoglobin assay components of *Cologuard*. The processing steps analyzed in this study are the steps at which operator variability or error are most likely to occur. Three total instrument and operator sets were used for the study.

Cologuard Molecular Assay Robustness

Results when these various factors were introduced into the processing steps were compared to the expected results for a positive stool sample, a control sample with high levels of mutation and methylation markers, and a control sample with moderate levels of mutation and methylation markers. Fourteen replicates of each sample type were used. Analysis of these samples assumed a hemoglobin value of zero, when calculating overall *Cologuard* score. Factors tested included the following:

- Factors related to DNA capture, including wait times between processing steps, amount of reagents added, and duration of storage at the appropriate temperatures;
- Factors related to the amount of time various instruments are paused during the automated DNA preparation and *QuARTS* assay steps of the *Cologuard* process; and

- Factors related to the amount of time between plate assembly and processing during the *QuARTS* assay step.

The results for the molecular assay component of *Cologuard* showed that time between plate assembly and processing during the *QuARTS*™ assay step and the number of days the captured DNA was stored at the appropriate temperatures could have a detectable effect on assay response. Testing demonstrated that the prepared *QuARTS*™ plate should be processed within 30 minutes and captured DNA could be tested for up to four days.

Cologuard Hemoglobin Assay Robustness

Results when these factors were introduced into the processing steps were compared to the expected results for a stool sample with a known level of endogenous hemoglobin and a high and low control sample with high and low levels of hemoglobin. The study tested 16 replicates of each sample type. Analysis of these results involved comparing the resulting hemoglobin concentration with the expected hemoglobin concentration. Factors tested include the following:

- Time between steps during plate preparation;
- Incubation times for antibodies and substrates; and
- Time between steps during plate reading phase.

Results for the hemoglobin assay component of *Cologuard* showed that substrate incubation time had a detectable effect on assay performance. Testing demonstrated that a substrate incubation time of 15 ± 1.5 minutes would result in acceptable assay performance.

Interfering Substances

***Cologuard* Molecular Assay Interference Testing**

Interference with the molecular assay component of *Cologuard* was evaluated using 55 common substances that potentially could be present in stool materials including potential interfering substances in the following categories:

- Common lotions, creams, and feminine over-the-counter products;
- Stool softeners, anti-diarrhea, and laxative products;
- Anti-acids and upset stomach relief products;
- Animal genomic DNA of commonly edible animals (both high and low levels);
- Urine and alcohol;
- A mixture of common vegetables and fruits; and
- Fecal Fats (fatty acids and cholesterol).

No interference with the molecular assay component of *Cologuard* was observed for any of the tested substances.

***Cologuard* Hemoglobin Assay Interference Testing**

Interference with the Hemoglobin Assay component of *Cologuard* was evaluated using 49 common substances that potentially could be present in stool materials including potential interfering substances in the following categories:

- Common lotions, creams, and feminine over-the-counter products;

- Urine;
- Stool softeners, anti-diarrhea, and laxative products;
- Anti-acids and upset stomach relief products;
- Antibiotics, anti-inflammatories, anti-fungal drugs, pain relievers, and decongestants;
- A mixture of common vegetables and fruits;
- Fats and lipids;
- Alcohol;
- Iron sulfate (as found in oral supplements);
- Vitamin C; and
- DNA Stabilization Buffer (preservative solution provided in the *Cologuard* Collection Kit for the whole stool sample used in the molecular assay).

None of the substances tested interfered with the *Cologuard* hemoglobin assay.

Carry-over and Cross-contamination Cologuard Testing

Carry-over Evaluation

Sequential runs of high positive and negative samples were used to evaluate carry-over contamination for each assay component of *Cologuard*. Testing of the molecular assay and hemoglobin assay components was conducted in two separate studies.

For the molecular assay (methylation/mutation assay), the testing involved two consecutive runs of high positive DNA samples, composed of 10x high level run controls diluted in Tris, EDTA and non-human DNA, followed by a run of negative samples composed of Tris, EDTA and non-human DNA. A total of 43 high positive samples and 3 run controls were used in each high positive run. A total of 43 negative samples and 3 run controls were used for the negative run.

For the hemoglobin assay, the testing involved two consecutive runs of high positive hemoglobin samples, composed of 100,000 ng/mL hemoglobin, followed by a run of negative samples composed solely of the protein preservative solution from the hemoglobin sample collection tube. The high positive samples consisted of a hemoglobin level that is much higher than the quantitative range of the assay, which identifies all samples >500 ng/mL as greater than the maximum range of the assay. For the high positive runs, a total of 86 high positive hemoglobin samples were used. For the negative run, 86 negative samples were used. In each run, the signal obtained on the controls was utilized to ensure the validity of the run.

Results from the molecular assay and hemoglobin assay carry-over analyses demonstrated that there was no carry-over in the *Cologuard* assay.

Cross-contamination Evaluation

Cross-contamination testing of *Cologuard* was based on a checkerboard study design, alternating high positive and negative samples, to evaluate the potential for contamination from the positive to the negative samples within a run. Testing of the molecular assay and hemoglobin assay components was conducted in two separate studies.

For the molecular assay, 22 high positive samples, 21 negative samples, and three run control samples were used. The high positive samples for this study were composed of 10x high level run controls diluted in Tris, EDTA and non-human DNA, and the negative samples were composed of Tris, EDTA and non-human DNA. One run was performed and samples were processed using the *Cologuard* molecular process from the semi-automated front end sample processing through the automated processing.

For the hemoglobin assay, a total of 43 high hemoglobin and 43 negative hemoglobin samples were used. As in the carry-over study, the high positive samples contained 100,000 ng/mL hemoglobin, while the negative samples consisted solely of the protein preservative solution from the hemoglobin sample collection tube. Three runs were performed and samples were processed using the *Cologuard* hemoglobin process.

Results from the cross-contamination analysis for the molecular assay demonstrated that the molecular assay component of *Cologuard* and the associated instruments needed to run the assay performed as intended and met the study acceptance criteria. Specifically, one well experienced some cross-contamination (52 strands of ACTB), however, this was within the pre-specified acceptance criteria, which dictated that no more than three wells could exhibit 10-100 strands of ACTB and no single well could exhibit more than 100 strands.

The high hemoglobin samples utilized in this study contained hemoglobin levels that are approximately 50 times higher than the median positive hemoglobin values observed in colorectal cancer subjects (Levi et. al, 2007). The high hemoglobin concentrations tested in this study are much higher than would be expected in use of *Cologuard*.

Testing of the Hemoglobin Assay cross-contamination showed a low level of contamination (~0.01%). Signal was observed in 4 out of 43 negative samples with an average detectable hemoglobin level of 11 ng/mL (0.011%). This calculates to a 0.011% contamination level in those four samples. As the hemoglobin assay involves several manual steps (e.g., manual washing and reagent addition), repeat testing was conducted, in which no cross contamination was observed.

Under normal use conditions, low level contamination observed in this study would be neglibilble. However, this study provides evidence that cross contamination is possible due to the manual steps in the assay processing.

Stability Studies

In-Use Stability: Molecular Assay Stability Under Standard Operating Conditions

The stability of reagents used in the molecular assay component of *Cologuard* was evaluated following guidance from CLSI standard: EP25-A (*Evaluation of Stability of In Vitro Diagnostic Reagents; Approved Guideline*). The purpose of this testing was to determine reagent stability after opening the containers and using them under potential user operating conditions. All reagents required for the molecular assay were tested.

Samples were processed with the molecular assay component of *Cologuard*, using these reagents, to determine the in-use stability of the reagents and the effect of the various factors on *Cologuard* results. The samples used in the in-use stability study for the various *Cologuard* reagent groups included DNA calibrators; High Positive and Low Positive control samples

consisting of synthetic targets in stool collection buffer; a Negative DNA control sample; DNA positive and negative run controls; and a positive stool sample.

The study demonstrated that *Cologuard* reagents are stable when opened or stored for variable times before use under standard operating conditions. Specifically:

- Multiple-use reagents stored at room temperature are stable for up to six weeks from the open date.
- Capture Beads that have been pre-washed and stored at 2-8°C are stable for up to 13 days.
- Pre-washed Capture Beads are stable for up to six hours at room temperature prior to use.

Single-use reagents that are used on the automated system are stable on the Hamilton Microlab® STARlet deck for up to 4 hours prior to the start of the run.

Freeze-Thaw Stability

A freeze-thaw stability study was conducted to evaluate the stability of the *QuARTS*™ assay reagents when subjected to repeated freeze/thaw events. The *QuARTS*™ assay reagents tested included only those assay components normally stored frozen (-25 to -15°C):

1) Oligo Mix A, Methylation;

2) Oligo Mix B, Mutation;

3) Enzyme Mix;

4) DNA Calibrator 1 High Methylation;

6) DNA Calibrator 2 Low Methylation;

7) DNA Calibrator 3 High Mutation; and

8) DNA Calibrator 4, Low Mutation.

Materials from one lot of each assay component were subjected to 0, 2, 4, and 6 freeze-thaw cycles. Each component was then tested in the *Cologuard* molecular assay component using the *Cologuard* DNA Controls (i.e., DNA Control 1, High Positive and DNA Control 2, Low Positive), which did not undergo freeze-thaw cycling. The study tested 16 replicates for each component and each freeze-thaw cycle. Calibrators used during testing to assess assay validity and to generate curves for sample concentration assessment were not subjected to freeze-thaw cycling. Log strands for each marker were compared to those for samples where the reagents did not undergo freeze thaw cycling.

All log strand results for all samples were statistically equivalent to those that did not undergo freeze thaw cycling, thereby demonstrating that the *Cologuard QuARTS*™ assay reagents are stable for six freeze thaw events.

Real-Time Stability

Real-time stability testing of *Cologuard* was conducted by evaluating the functional performance of three reagent lots over a period of 41 weeks. Each lot was comprised of unique batches of reagents, which were tested at various time points over 41 weeks.

Samples that were used to evaluate hemoglobin assay reagent stability consisted of negative stool matrix spiked with whole blood to create samples with a low and high hemoglobin concentration. Samples for evaluation of molecular assay reagent stability consisted of negative stool matrix spiked with oligonucleotides that contain the marker sequences. Oligonucleotides for *NDRG4, BMP3, BTACT, KRAS1, KRAS2,* and *ACT* were spiked into the negative stool samples to create samples with a low and high level of sDNA samples. At each time point, seven replicates of samples and controls were tested.

The results of the real time stability studies demonstrated that overall the components of the *Cologuard* assay gave similar results through the 41 week study. These data supported the 6 month shelf life currently assigned to the *Cologuard* assay reagents.

Clinical Sensitivity and Specificity

The pivotal study ("Multi-Target Colorectal Cancer Screening Test for the Detection of Colorectal Advanced Adenomatous Polyps and Cancer: DeeP-C Study") was conducted to generate data to support the safety and effectiveness of *Cologuard* as a screening test for the detection of markers associated with the presence of colorectal cancer (CRC) and advanced adenoma (AA). To evaluate the performance of *Cologuard*, the *Cologuard* test result (negative or positive) was compared with the histopathological result from optical colonoscopic examination and histopathological diagnosis of all significant lesions discovered during the colonoscopy and either biopsied or removed. Based on this comparison, *Cologuard* sensitivity (true positive fraction) was 92.3% (60/65) for subjects with a histopathological diagnosis of CRC and 42.4% (322/760) for subjects with a diagnosis of AA. For subjects without a diagnosis of CRC or AA, *Cologuard* specificity (true negative fraction) was 86.6% (7967/9198). Furthermore, among subjects having a valid *Cologuard* test result and also a valid test result from a commercially available FIT (OC FIT-CHEK, Polymedco, Inc.) ("FIT"), both sensitivity for CRC and sensitivity for AA were higher for *Cologuard* (92.3%, 42.4%) than for FIT (73.8%, 23.8%), and both differences (18.5%, 18.6%) were significantly different from zero (p=0.002, 0.001). However, for subjects without CRC or AA, specificity was lower for *Cologuard* (86.6%) than for FIT (94.9%), and the difference (-8.3%) was significantly different from zero ($p < 0.0001$).

An overview of the study design and results is provided below.

Study Design

The *Cologuard* pivotal study was a prospective, cross-sectional, multi-center study (DeeP-C study) that began enrollment of study participants on June 30, 2011. A total of 12,776 patients were enrolled from 90 sites in the U.S. and Canada, including both colonoscopy centers and primary care sites, with study participation concluding on February 4, 2013. Subjects were provided with a collection kit, which they used to collect stool samples for *Cologuard* and FIT testing. Subjects subsequently underwent colonoscopy within 90 days of study enrollment.

The stool samples for analysis with *Cologuard* were sent to a central biorepository for batch testing at one of three laboratories while the stool samples for the FIT were sent to a single

laboratory for testing. Samples tested with *Cologuard* were assayed by laboratory technicians blinded to the results of colonoscopy and the FIT results. Results from *Cologuard* and the FIT test were compared to the results of an optical colonoscopic examination, and histopathological diagnosis of all significant lesions discovered during the colonoscopy and either biopsied or removed.

Colonoscopy findings were recorded per site specific standard of practice. Subjects with no findings were categorized as negative by colonoscopy. Histopathological results from biopsied tissue or excised lesions were categorized based on the most clinically significant lesion present (i.e., the index lesion) by a central pathologist according to the pre-specified standards outlined in **Table 8**.

Table 8: Histopathological category definitions

Category	Findings
1	CRC, all stages (I-IV)
2	Advance adenoma, including the following subcategories: 2.1 – Adenoma with carcinoma *in situ*/high grade dysplasia, any size 2.2 – Adenoma, villous growth pattern (≥25%), any size 2.3 – Adenoma ≥ 1.0 cm in size, or 2.4 – Serrated lesion, ≥ 1.0 cm in size
3	1 or 2 adenoma (s), >5 mm in size, or < 10 mm size, non-advanced
4	≥ 3 adenomas, <10mm, non-advanced
5	1 or 2 adenoma(s), ≤5 mm in size, non-advanced
6	Negative – No neoplastic findings 6.1 – negative upon histopathological review 6.2 – no findings on colonoscopy, no histopathological review

Inclusion and Exclusion Criteria

Subjects eligible for enrollment in the study were of both genders between the ages of 50 and 84 years (inclusive), who were at average risk for development of colorectal cancer and asymptomatic for gastrointestinal symptoms warranting diagnostic colonoscopy. In addition, subject enrollment was age-weighted toward a slightly older population to increase the point prevalence of colorectal cancer in this study. An effort was made to enroll the majority of subjects of age 65-84; 64% of subjects in the actual study population were of age 65-84.

Clinical Performance Measures

The performance of Cologuard was evaluated based on comparison of the test result with the histopathological category (Table 9 above). Two co-primary performance measures were pre-specified: Cologuard sensitivity for subjects diagnosed with CRC (histopathological category 1), and Cologuard specificity for subjects without a diagnosis of CRC or AA (categories 3-6). For subjects with CRC, Cologuard sensitivity is the fraction of CRC subjects called positive by the Cologuard test (true positive fraction). Defining advanced neoplasia (AN) as CRC or AA, Cologuard specificity for AN is the fraction of non-AN subjects called negative by the Cologuard test (true negative fraction). For the study to be successful, the co-primary analysis required that

the lower bound of the 95% one-sided confidence interval was greater than 65% for Cologuard sensitivity for CRC and greater than 85% for Cologuard specificity for AN. It should be noted that sensitivity for CRC and specificity for AN are not complimentary in that advanced adenoma (AA, histopathological category 2) is left out of the definition of both measures.

Two secondary performance evaluations were pre-specified: Cologuard was evaluated for non-inferiority to FIT in sensitivity for CRC and for superiority to FIT in sensitivity for AA (fraction of AA subjects testing positive). Per the pre-specified protocol, Cologuard would be declared non-inferior to FIT in sensitivity for CRC if the one-sided 95% confidence interval lower bound on the Cologuard – FIT difference exceeded -5%. If Cologuard were to be declared non-inferior to FIT in CRC sensitivity, then evaluation for superiority to FIT in CRC sensitivity was permitted and declared if the difference was positive and its one-sided p-value (based on exact McNemar test) was less than 0.025. Likewise, per protocol, Cologuard would be declared superior to FIT in AA sensitivity if the Cologuard – FIT difference was positive and the one-sided McNemar p-value was less than 0.025.

Accountability of PMA Cohort

The study enrolled a total of 12,766 subjects at 90 sites, including both primary care point-of-referral (POR) sites and colonoscopy centers. A total of 2,753 subjects were excluded from the primary analysis population due to unusable data (e.g., no colonoscopy). A total of 10,023 subjects were included in the primary analysis population. This population included 65 subjects with CRC. Analysis was conducted to rule out bias associated with the subjects excluded from the analysis population.

Study Population Demographics and Baseline Parameters

The baseline demographic characteristics for the Primary Effectiveness Population are presented in Table 9 below. As shown in the table, the average age of subjects was 64.2 years old, and there was a slightly higher percentage of female subjects (5,378/10,023, 53.7%) as compared with male subjects (4,645/10,023, 46.3%). The majority of subjects were White (8,422/10,017, 84.1%), although 10.7% of the population were Black or African American subjects (1,071/10,017). Nearly 10% of subjects were Hispanic or Latino (991/10,019, 9.9%). Average BMI was 28.83 and the majority of subjects never smoked (5,531 /10,019, 55.2%). It should be noted that two 49-year-old subjects and one 44-year-old subject were included in the study, which is inconsistent with the intended use population. Each of these subjects was a true negative and their inclusion did not notably impact data analyses.

Subjects that were enrolled at POR sites were similar to those enrolled at non-POR sites and to the population as a whole.

Table 9: Baseline Demographics – Primary Effectiveness Subjects

Parameter Statistic	All Enrolled (N=10023)	Specificity Subset (Cat. 2-6) (N=9958)	Specificity Subset (Cat. 3-6) (N=9198)	CRC Subset (Cat. 1) (N=65)	AA Subset (Cat. 2) (N=760)	FIT Secondary Effectiveness (N=65)
Age (years) at Screening						
n	10023	9958	9198	65	760	65
Mean (SD)	64.2 (8.42)	64.1 (8.41)	64.0 (8.44)	70.2 (7.92)	65.4 (7.93)	70.2 (7.92)
Median	66	66	66	70	66	70
Min, Max	44, 84	44, 84	44, 84	50, 84	50, 84	50, 84
Gender, n (%)						
Male	4645 (46.3)	4611 (46.3)	4161 (45.2)	34 (52.3)	450 (59.2)	34 (52.3)
Female	5378 (53.7)	5347 (53.7)	5037 (54.8)	31 (47.7)	310 (40.8)	31 (47.7)
Race, n (%)						
White	8422 (84.1)	8367 (84.1)	7726 (84.0)	55 (84.6)	641 (84.5)	55 (84.6)
Black or African American	1071 (10.7)	1063 (10.7)	978 (10.6)	8 (12.3)	85 (11.2)	8 (12.3)
Asian	259 (2.6)	258 (2.6)	245 (2.7)	1 (1.5)	13 (1.7)	1 (1.5)
American Indian or Alaska Native	36 (0.4)	36 (0.4)	32 (0.3)	0 (0.0)	4 (0.5)	0 (0.0)
Native Hawaiian or Other Pacific Islander	23 (0.2)	23 (0.2)	23 (0.3)	0 (0.0)	0 (0.0)	0 (0.0)
Other	206 (2.1)	205 (2.1)	189 (2.1)	1 (1.5)	16 (2.1)	1 (1.5)
Missing	6	6	5	0	1	0
Ethnicity, n (%)						
Hispanic or Latino	991 (9.9)	982 (9.9)	923 (10.0)	9 (13.8)	59 (7.8)	9 (13.8)
Not Hispanic or Latino	9028 (90.1)	8972 (90.1)	8272 (90.0)	56 (86.2)	700 (92.2)	56 (86.2)
Missing	4	4	3	0	1	0
BMI (kg/m2) at Baseline						
n	10015	9950	9190	65	760	65
Mean (SD)	28.83 (5.836)	28.84 (5.841)	28.77 (5.817)	27.55 (4.861)	29.67 (6.068)	27.55 (4.861)
Median	28.0	28.0	27.9	26.8	29.0	26.8
Min, Max	13.3, 68.2	13.3, 68.2	13.3, 68.2	19.3, 42.4	16.3, 59.9	19.3, 42.4
Smoking History, n (%)						
Never Smoked	5531 (55.2)	5498 (55.2)	5157 (56.1)	33 (50.8)	341 (44.9)	33 (50.8)
Former Smoker	3589 (35.8)	3564 (35.8)	3279 (35.6)	25 (38.5)	285 (37.5)	25 (38.5)
Current Smoker	903 (9.0)	896 (9.0)	762 (8.3)	7 (10.8)	134 (17.6)	7 (10.8)
If Former or Current Smoker, Daily Use, n (%)						

Parameter Statistic	All Enrolled (N=10023)	Specificity Subset (Cat. 2-6) (N=9958)	Specificity Subset (Cat. 3-6) (N=9198)	CRC Subset (Cat. 1) (N=65)	AA Subset (Cat. 2) (N=760)	FIT Secondary Effectiveness (N=65)
<1/2 Pack Per Day	2162 (48.3)	2154 (48.4)	1970 (48.9)	8 (25.0)	184 (44.0)	8 (25.0)
1 Pack Per Day	1585 (35.4)	1569 (35.3)	1418 (35.2)	16 (50.0)	151 (36.1)	16 (50.0)
>1 Pack Per Day	732 (16.3)	724 (16.3)	641 (15.9)	8 (25.0)	83 (19.9)	8 (25.0)
Missing	13	13	12	0	1	0
If Former or Current Smoker, # Years Smoking						
n	4480	4448	4029	32	419	32
Mean (SD)	21.82 (14.733)	21.77 (14.732)	21.13 (14.450)	28.47 (13.488)	27.93 (15.959)	28.47 (13.488)
Median	20.0	20.0	20.0	29.0	30.0	29.0
Min, Max	0.0, 70.0	0.0, 70.0	0.0, 70.0	1.0, 60.0	1.0, 65.0	1.0, 60.0

Safety and Effectiveness Results

Primary Effectiveness Evaluations (Sensitivity for CRC/Specificity for AN)

The primary effectiveness population consisted of 10,023 subjects with a valid histopathological diagnosis and a valid Cologuard result. The basic data table for primary effectiveness evaluation is provided (Table 11).

Table 10: Distribution of *Cologuard* Result by Histological Category (%), n = 10,023

Cologuard Result	CRC (Category 1)	AA (Category 2)	Categories 3-6
Negative	5 (7.7)	438 (57.6)	7967 (86.6)
Positive	60 (92.3)	322 (42.4)	1231 (13.4)

The primary objectives of the DeeP-C study – demonstration of greater than 65% Cologuard sensitivity for CRC and greater than 85% Cologuard specificity for AN – were successfully met. Specifically, Cologuard sensitivity for CRC was 92.3% (60/65) with a one-sided 95% confidence interval lower bound of 84.5% (Table 11). Cologuard specificity for AN was 86.6%, with a one-sided 95% confidence interval lower bound of 86.0% (Table 12). Further, the two-sided 95% confidence interval was 83.0-97.5% for Cologuard CRC sensitivity and 85.9-87.3% for Cologuard AN specificity.

Table 11: Sensitivity for CRC – Primary Effectiveness Subjects with Valid *Cologuard* Positive Result (N=65)

Case Category	n/N (%)
1: CRC Stages 1-4	60/65 (92.3%)
Sensitivity Based on Category 1: Primary (one-sided 95% CI lower bound)	92.3% (>84.5%)
Sensitivity Based on Category 1: Supportive (one-sided 97.5% CI lower bound)	92.3% (>83.0%)

[1] Percentages based on valid test results within a category.
[2] Lower bounds calculated using an exact one-sided binomial test.

Table 12: Specificity for AN – Primary Effectiveness Subjects with Valid *Cologuard* Negative Result (N=9198)

Case Category	n/N (%)
3: 1-2 Adenomas 5 – <10 mm	607/749 (81.0%)
4: ≥3 Adenomas <10 mm, Non-advanced	302/419 (72.1%)
5: 1-2 Adenomas <5 mm, Non-advanced	1496/1735 (86.2%)
6.1: Negative upon histopathological review	1543/1821 (84.7%)
6.2: No findings on colonoscopy, no histopathological review	4019/4474 (89.8%)
Specificity Based on Categories 3-6: Primary (one-sided 95% lower bound)	86.6% (>86.0%)
Specificity Based on Categories 3-6: Supportive (one-sided 97. 5% lower bound)	86.6% (>85.9%)

[1] Percentages based on valid test results within a category.
[2] Lower bounds calculates using an exact one-sided binomial test.
[3] As noted above, one 44-year-old and two 49-year-old true negative subjects were included in the analysis population, although they would not be included in the intended user population.

Secondary Effectiveness Evaluations

The secondary effectiveness population consisted of 9,989 subjects with a valid histopathological diagnosis, a valid Cologuard result, and a valid FIT result. The basic data table for secondary effectiveness evaluation is provided (Table 13).

Table 13: Distribution of *Cologuard* and FIT Results by Histological Category, n = 9,989

CRC (Category 1)

	FIT	
Cologuard	**Negative**	**Positive**
Negative	4	1
Positive	13	47

AA (Category 2)

	FIT	
Cologuard	Negative	Positive
Negative	407	29
Positive	170	151

Categories 3-6

	FIT	
Cologuard	Negative	Positive
Negative	7787	149
Positive	908	323

Cologuard was compared with FIT for non-inferiority in sensitivity for CRC with respect to margin 5% and for superiority in sensitivity for advanced adenoma (AA). Secondary performance goals comparing *Cologuard* with FIT were evaluated in subjects having valid results from both tests.

Sensitivity for CRC was greater for *Cologuard* (92.3%, 60/65) than for FIT (73.8%, 48/65) (**Table 14 and Figure 1**), for a difference of 18.5%. *Cologuard* identified 13 CRCs that were missed by FIT. FIT identified one CRC that was missed by *Cologuard*. The one-sided 95% confidence interval lower bound on the *Cologuard* – FIT difference was 8.0%, which exceeds -5%, indicating that *Cologuard* is non-inferior to FIT in sensitivity for CRC with respect to the pre-defined non-inferiority margin 5%. Because *Cologuard* was declared non-inferior to FIT in sensitivity for CRC, it is statistically justifiable and was permissible per protocol to evaluate it for superiority to FIT as well. Sensitivity for CRC was significantly greater for *Cologuard* than for FIT (two-sided McNemar exact p value 0.0018), indicating that *Cologuard* is superior to FIT in sensitivity for CRC. Finally, for the *Cologuard* – FIT difference of 18.5% in CRC sensitivity, the two-sided 95% exact confidence interval was 7.2-30.4% (**Table 15**).

Table 14: Overall Sensitivity: CRC Subset (Category 1) - Secondary Effectiveness Subjects with Valid Results from Both *Cologuard* and FIT Tests (N=65)

	Cologuard	**FIT**
1: CRC Stages 1-4 **(n/N (%))**	60/65 (92.3%)	48/65 (73.8%)
Sensitivity Based on Category 1: Primary (one-sided 95% lower bound)	92.3% (>84.5%)	73.8% (>63.4%)
Sensitivity Based on Categories 1: Supportive (one-sided 97. 5% lower bound)	92.3% (>83.0%)	73.8% (>61.5%)

[1] Percentages based on valid test results within a category.
[2] Lower bounds calculated using an exact one-sided binomial test.

Table 15: Sensitivity Non-Inferiority and Superiority Test – CRC Subset (Category 1)

	FIT Outcome			**McNemar test p-value**
***Cologuard* Outcome**	**Negative**	**Positive**	**Totals**	
Negative, n (%)	4 (80.0)	1 (20.0)	5	0.0018
Positive, n (%)	13 (21.7)	47 (78.3)	60	
Totals	17	48	65	

[1] p-value is from a McNemar paired comparison test of the discordant pairs.

Figure 1: CRC Sensitivity



For histopathologically-confirmed AAs, sensitivity was greater for *Cologuard* (42.4%, 321/757) than for FIT (23.8%, 180/757) (**Table 16**). The difference of 18.6% was significantly different from zero (two-sided exact McNemar p value < 0.001), indicating that *Cologuard* is superior to FIT in sensitivity for AA. *Cologuard* identified 170 AA cases that were not called positive by the FIT test, while FIT identified 29 AA cases that were not called positive by *Cologuard* test. Finally, the two-sided 95% CI for the *Cologuard* – FIT difference of 18.6% was 15.3-22.1% (**Figure 2**).

Table 16: Sensitivity Superiority Test – AA Subset (Category 2)

	FIT Outcome			**McNemar test p-value**
***Cologuard* Outcome**	**Negative**	**Positive**	**Totals**	
Negative, n (%)	407 (93.3)	29 (6.7)	436	<0.0001
Positive, n (%)	170 (53.0)	151 (47.0)	321	
Totals	577	180	757	

[1] p-value is from a McNemar paired comparison test of the discordant pairs.

Figure 2: AA Sensitivity



The combined sensitivity for CRC and AA subjects was also analyzed *post hoc*. The sensitivity for CRC/AA was 46.3% (381/822) for *Cologuard* and 27.7% (228/822) for FIT, for a difference of 18.6% (**Table 17**).

Table 17: Sensitivity for Advanced Neoplasia (CRC + AA)

Category	*Cologuard* (N=822)	FIT (N=822)
Category 1 Only	92.3% (60/65)	73.8% (48/65)
Categories 1-2	46.4% (381/822)	27.7% (228/822)

Numerically greater sensitivity for *Cologuard* compared to FIT was observed across all sub-categories of AA. For example, sensitivity for adenoma with carcinoma in situ or high grade dysplasia (Category 2.1) was 69.2% for *Cologuard*, compared to 46.2% for FIT. *Cologuard* identified 43.0% of serrated lesions, whereas FIT sensitivity for these lesions was 5.1%.

For subjects without CRC or AA, the specificity (fraction of subjects called negative) was smaller for *Cologuard* (86.6%, 7936/9167) than for FIT (94.9%, 8695/9167) (**Table 18**). The difference in specificity (–8.3%) was significantly different from zero ($p < 0.0001$). The two-sided 95% confidence interval on the difference was (–9.0%, –7.6%).

For subjects without CRC or AA (categories 3-6), a positive test result is considered a false positive. The false positive fraction is 1 – specificity and was significantly higher for Cologuard (13.4%) than for FIT (5.1%) ($p < 0.0001$). On the other hand, for subjects with CRC or AA, the true positive fraction was higher for Cologuard (46.3%) than for FIT (27.7%) (**Table 18**Table).

For subjects without CRC (categories 2-6), the specificity (fraction of subjects called negative) was smaller for *Cologuard* (84.4%, 8372/9924) than for FIT (93.4%, 9272/9924). The difference was –9.1% with two-sided 95% confidence interval (–9.8%, –8.4%). The *Cologuard* specificity for CRC (84.4%) together with its sensitivity for CRC (92.3%) form a complimentary pair spanning the entire study population. By comparison, the FIT specificity for CRC was higher (93.4%) while its sensitivity for CRC was lower (73.8%) than for *Cologuard*.

Table 18: Specificity – Specificity Subset (Categories 3-6)

		FIT Outcome		
	***Cologuard* Outcome**	**Negative**	**Positive**	**Totals**
Categories 3-6	**Negative, n (%)**	7787 (98.1%)	149 (1.9%)	7936
	Positive, n (%)	908 (73.8%)	323 (26.2%)	1231
	Totals	8695	472	9167

Cologuard was also compared with FIT on the Receiver Operating Characteristic curve ("ROC curve"). For tests yielding (but not necessarily reporting) a continuous or ordinal value (measurement or score), a threshold or cut-off may be applied to define test positive and test negative results. The ROC curve is a plot of all possible pairs of sensitivity and 1 – specificity (true and false positive fractions) generated by varying the cut-off over the entire range of observed values.

For CRC, the ROC curves are displayed for *Cologuard* and FIT (**Figure 3**). In the figure, the false positive and true positive fractions associated with cut-offs used by the tests are superimposed. Also displayed in the figure is the area under the ROC curve (AUC) for each test. For *Cologuard* the AUC was 93.0%, indicating that a probability is 93.0% that a randomly chosen CRC subject has a higher underlying *Cologuard* composite score than a randomly chosen non-CRC subject. For FIT this probability was 88.0%.



Figure 3: ROC curves for CRC: Cologuard and FIT. The ROC curve plots sensitivity for CRC (Category 1) vs. 1 – specificity for non-CRC (Categories 2-6).

For AN, the ROC curves are displayed for *Cologuard* and FIT (**Figure 4**). For *Cologuard*, the AUC for AN was 73.3%, indicating that a probability is 73.3% that a randomly chosen subject with CRC or AA has a higher underlying *Cologuard* composite score than a randomly chosen subject without CRC or AA. For FIT this probability was 66.7%.



Figure 4*: ROC curves for AN: Cologuard and FIT. The ROC curve plots sensitivity for advanced neoplasia (AN, categories 1-2) vs. 1 – specificity for non-AN (Categories 3-6).*

Additional Effectiveness Analyses

In addition to the sensitivity and specificity for CRC and AA, the positive and negative likelihood ratios for *Cologuard* were calculated from the study data. Results demonstrated a positive likelihood ratio of 5.9 for CRC, indicating that a person with CRC would be 5.9 times more likely to have a positive *Cologuard* result than someone without CRC. The negative likelihood ratio for CRC was 0.09, indicating that someone without CRC is approximately 11 times (1/0.09) more likely to test negative on *Cologuard* compared to someone with CRC.

Results also demonstrated a positive likelihood ratio of 3.2 for AA (**Table 19**), indicating that a person with AA would be 3.2 times more likely to have a positive *Cologuard* results than someone without AA or CRC. The negative likelihood ratio for AA was 0.67, indicating that someone without AA or CRC is approximately 1.5 times (1/0.67) more likely to test negative on *Cologuard* compared to someone with AA.

Table 19: Positive and Negative Likelihood Ratios

	Category 1 (CRC) vs Categories 2-6	Category 2 (AA) vs Categories 3-6
Positive Likelihood Ratio (PLR)		
Sensitivity, %	92.3 (60/65)	42.4 (322/760)
1-Specificity, %	15.6 (1553/9958)	13.4 (1231/9198)
PLR	5.9	3.2
95% Confidence Interval	(5.4, 6.4)	(2.9, 3.5)
Negative Likelihood Ratio (NLR)		
1-Sensitivity, %	7.7 (5/65)	57.6 (438/760)
Specificity, %	84.4 (8405/9958)	86.6 (7967/9198)
NLR	0.09	0.67
95% Confidence Interval	(0.04, 0.21)	(0.63, 0.71)

Analysis was also performed to calculate the positive and negative predictive values ("PPV" and "NPV") for *Cologuard* (**Table 20**). As with any CRC screening test, the PPV is impacted by the very low prevalence of CRC in the general population. The PPV was calculated to be 3.72% (60/1613) for CRC and 19.86% (322/1613) for AA. Meanwhile, the NPV was 94.73%.

Table 20: Positive Predictive Value – Primary Effectiveness Subjects

Cologuard	**Category 1 (CRC)**	**Category 2 (AA)**	**Categories 3-6**
Negative	0.06, 0.02-0.14 (5/8410)	5.2, 4.7- 5.7 (438/8410)	94.7, 94.2-95.2 (7967/8410)
Positive	3.72, 2.85-4.76 (60/1613)	20.0, 18.0-22.0 (322/1613)	76.3, 74.2-78.4 (1231/1613)

*2-Sided 95% CIs

<u>Sub-Group Analyses</u>

The DeeP-C study results were also analyzed according to various demographic characteristics, as well as lesion size and location. Performance goals were not pre-specified for subgroup analysis. No attempt was made to adjust p values for multiple subgroup analyses.

<u>Results by Gender</u>

Sensitivity of *Cologuard* was higher for males than for females, both for CRC and AA. *Cologuard* sensitivity for CRC was 100.0% for males, compared with 83.9% for females (**Table 21)**. *Cologuard* sensitivity for AA was 44.7% for males, compared with 39.0% for females.

Table 21: *Cologuard* Sensitivity by Gender (Categories 1 and 2)

Gender, n/N (%)	**Category 1 (CRC)**	**Category 2 (AA)**
Male	34/34 (100.0)	201/450 (44.7)
Female	26/31 (83.9)	121/310 (39.0)

[1] Sensitivity calculated as number of positives (CRC or AA) divided by subjects with CRC or AA, respectively.

Meanwhile, specificity of *Cologuard* for subjects with neither CRC nor AA (AN) was very similar for females as compared with males. As shown in **Table 23** below, Specificity for non-AN was 87.3% (4,398/5,037) for females compared with 85.8% (3,569/4,161) for males.

Table 22: *Cologuard* Specificity by Gender

Gender, n/N (%)	**Categories 3-6[1]**
Male	3569/4161 (85.8)
Female	4398/5037 (87.3)

[1] Specificity calculated as number of negatives among subjects without CRC or AA.

Results by Race and Ethnicity

With respect to race, *Cologuard* sensitivity for CRC was higher among White subjects (53/55, 96.4%) than among Black or African-American subjects (5/8, 62.5%). There was only one Asian CRC case in the study (1/1, 100.0%) (**Table 23)**. However, the results observed in Black or African-American and Asian subjects may well have been driven by the low overall number of cancer cases in that subpopulation. Sensitivity among Hispanic or Latino subjects (8/9, 88.9%) was also high, although again the sample size was small. Sensitivity for AA was similar for White (271/641 42.3%) and Black/African-American (36/85, 42.4%) subjects. Sensitivity was also similar among Hispanic/Latino subjects (23/59, 39.0%). *Cologuard* sensitivity for AA was lower among Asian subjects (4/13, 30.8%) and higher for American Indian or Alaskan Natives (3/4, 75.0%), compared with other groups.

Table 23: *Cologuard* Sensitivity by Race and Ethnicity, CRC and AA Subsets (Categories 1 and 2)

Subgroup	Category 1 (CRC)	Category 2 (AA)
Race, n/N (%)		
White	53/55 (96.4)	271/641 (42.3)
Black or African American	5/8 (62.5)	36/85 (42.4)
Asian	1/1 (100.0)	4/13 (30.8)
American Indian or Alaska Native	0/0	3/4 (75.0)
Native Hawaiian or Other Pacific Islander	0/0	0/0
Other	1/1 (100.0)	7/16 (43.8)
Ethnicity, n/N (%)		
Hispanic or Latino	8/9 (88.9)	23/59 (39.0)
Not Hispanic or Latino	52/56 (92.9)	298/700 (42.6)

[1] Sensitivity calculated as number of positives (CRC or AA) divided by subjects with CRC or AA.

Cologuard specificity for subjects without CRC or AA (categories 3-6) varied among race groups (p-value = 0.001) (Table 27). Specificity was highest for Asian (93.5%, 229/245) and Native Hawaiian/Pacific Islander subjects (91.3%, 21/23) and lowest for American Indian/Alaska Native subjects (75.0%, 24/32).

Table 24: *Cologuard* Specificity by Race and Ethnicity – Primary Effectiveness Subjects

Subgroup	Categories 3-6
Race, n/N (%)	
White	6639/7726 (85.9)
Black or African American	879/978 (89.9)
Asian	229/245 (93.5)
American Indian or Alaska Native	24/32 (75.0)
Native Hawaiian or Other Pacific Islander	21/23 (91.3)
Other	171/189 (90.5)
Ethnicity, n/N (%)	
Hispanic or Latino	837/923 (90.7)
Not Hispanic or Latino	7127/8272 (86.2)

[1] Specificity calculated as number of negatives among subjects without CRC or AA.

Results by Age

Cologuard sensitivity for CRC was consistently high across all age groups (**Table 25**), ranging from 88.9-100.0% for age groups with at least six subjects. Although sensitivity was 75% for

subjects of age 60-64, the number of CRC cases was particularly small in this age group (n = 4), and only one CRC case was not detected by *Cologuard*.

Cologuard sensitivity for AA increased from 38.0% for subjects of age < 60 to 46.8% for subjects between age 70-79 (**Table 25**).

Table 25: *Cologuard* Sensitivity by Age

Age, n/N (%)	**Category 1 (CRC)**	**Category 2 (AA)**
<60 years	7/7 (100.0)	65/171 (38.0)
60-64 years	3/4 (75.0)	24/57 (42.1)
65-69 years	19/20 (95.0)	125/301 (41.5)
70-74 years	16/18 (88.9)	72/154 (46.8)
75-79 years	6/6 (100.0)	29/62 (46.8)
80-84 years	9/10 (90.0)	7/15 (46.7)

[1] Sensitivity calculated as number of positives (CRC or AA) divided by subjects with CRC or AA.
[2] Two 49-year-old subjects and one 44-year-old subject were included in the analysis population, although they would not be included in the intended use population.

Cologuard specificity for subjects without CRC or AA (categories 3-6) was highest for younger subjects and lowest for older subjects, ranging from 77.8-92.2% (**Table 26**).

Table 26: *Cologuard* Specificity by Age

Age, n/N (%)	**Categories 3-6**
<60 years	2491/2703 (92.2)
60-64 years	681/765 (89.0)
65-69 years	2871/3352 (85.7)
70-74 years	1292/1566 (82.5)
75-79 years	480/617 (77.8)
80-84 years	152/195 (77.9)

[1] Specificity calculated as number of negatives among subjects without CRC or AA.
[2] Two 49-year-old subjects and one 44-year-old subject were included in the analysis population, although they would not be included in the intended use population.

Results by Lesion Size and Cancer Stage

Cologuard results were evaluated by lesion size, as well as cancer stage (Table 27). Sensitivity of *Cologuard* increased with lesion size, as would be expected for a stool-based DNA test of this type. The amount of DNA shed from cancerous or pre-cancerous tissue in the colon is generally expected to increase with increased mass or lesion size.

As shown in the table below, sensitivity was > 90% for most lesion sizes. Sensitivity for CRC was highest for subjects with CRCs ≥ 30 mm (32/34, 94.1%) and lowest for subjects with CRCs 5-9 mm in size (4/5, 80.0%). Sensitivity by cancer stage was generally high and was the highest for subjects with Stage II cancers (21/21, 100.0%) and Stage III cancers (9/10, 90%). Sensitivity of *Cologuard* for AA was higher among subjects with AAs of larger lesion sizes.

Table 27: *Cologuard* Sensitivity within Lesion Subgroups

Subgroup	Category 1 (CRC)	Category 2 (AA)
Largest Lesion Size, n/N (%)		
<5 mm	0/0	2/10 (20.0)
5-9 mm	4/5 (80.0)	18/56 (32.1)
10-19 mm	13/14 (92.9)	225/577 (39.0)
20-29 mm	11/12 (91.7)	51/79 (64.6)
>=30 mm	32/34 (94.1)	26/38 (68.4)
Stage, n/N (%)		
I	26/29 (89.7)	N/A
II	21/21 (100.0)	N/A
III	9/10 (90.0)	N/A
IV	3/4 (75.0)	N/A
Unknown*	1/1 (100.0)	N/A

[1] Sensitivity calculated as number of positives (CRC or AA) divided by subjects with CRC or AA.

Specificity of *Cologuard* for subjects without AA or CRC was stratified by lesion size (**Table 28)**. Specificity of *Cologuard* for CRC was 86.2% (1,847/2,142), for subjects with CRCs < 5 mm in size, and 79.7% (1,523/1,912) for subjects with CRCs 5-9 mm in size.

Table 28: *Cologuard* Specificity by Lesion Size – Primary Effectiveness Subjects

Largest Lesion Size, n/N (%)	Categories 3-6
<5 mm	1847/2142 (86.2)
5-9 mm	1523/1912 (79.7)
10-19 mm	0/0
20-29 mm	0/0
>=30 mm	0/0

[1] Specificity calculated as number of negatives among subjects without CRC or AA.

Results by Lesion Location

Cologuard sensitivity was also assessed by lesion location (**Table 29**). Sensitivity of *Cologuard* for CRC was 90% or greater, regardless of lesion location. Sensitivity of *Cologuard* for AA was

higher among subjects with distal AAs (133/238, 55.9%) and lower among subjects with proximal AAs (143/433, 33.0%).

Table 29: *Cologuard* Sensitivity by Lesion Location

Lesion Location, n/N (%)	Category 1 (CRC)	Category 2 (AA)
Proximal	27/30 (90.0)	143/433 (33.0)
Distal	22/24 (91.7)	133/238 (55.9)
Rectal	11/11 (100.0)	45/88 (51.1)

[1] Sensitivity calculated as number of positives (CRC or AA) divided by subjects with CRC or AA.

Specificity of *Cologuard* for subjects without CRC or AA (categories 3-6) was high, regardless of lesion location. Specificity of *Cologuard* was 83.4% for subjects with proximal CRCs, 82.1% for subjects with distal CRCs, and 84.5% for subjects with rectal CRCs (**Table 30**).

Table 30: *Cologuard* Specificity by Lesion Location – Primary Effectiveness Subjects

Lesion Location, n/N (%)	Categories 3-6
Proximal	1723/2066 (83.4)
Distal	1131/1377 (82.1)
Rectal	517/612 (84.5)

[1] Specificity calculated as number of negatives among subjects without CRC or AA.

Safety Analyses

With respect to safety, due to the design of the study and the nature of the stool collection process, Adverse Effects (AEs) caused by or related to the stool collection procedure were not expected. As a result, events associated with potential errors in use of the collection kit and any product complaints were captured in the safety analyses. There were no cases in which the study investigator believed the product contributed to a serious adverse event, and only 4 adverse events were reported. None of the AEs experienced in the study were deemed serious, all were categorized as “mild” events. None of the events led to the subject discontinuing the study. Additionally, one subject died of unrelated causes prior to undergoing colonoscopy. The subject met all eligibility criteria and successfully collected a stool sample, but did not present for the subsequent colonoscopy.

Abbreviations Used

CRC: Colorectal Cancer
AA: Advanced Adenoma
QuARTS: Quantitative Allele-specific Real-time Target and Signal amplification
KRAS: v-Ki-ras2 Kirsten rat sarcoma viral oncogene homolog
ACTB: Beta actin
NDRG4: N-Myc Downstream Regulated Gene 4
BMP3: Bone Morphogenetic Protein 3
Mutation *QuARTS*: Triplex *QuARTS* assay containing wild-type *ACTB* (as a reference gene) and 7 *KRAS* point mutation markers
Methylation *QuARTS*: Triplex *QuARTS* assay containing *ACTB* (as a reference gene), in addition to *NDRG4* and *BMP3* methylation markers

Key Symbols Used

Symbol	Description
IVD	*In vitro* diagnostic medical device
	Consult instructions for use
Σ 480	Contains sufficient reagents for 480 Tests
	Manufacturer
REF	Part number
	Important information for proper operation
NOTE:	Followed by additional information required for the procedure
	Indicates upper and lower temperature limits for storage
	Warning symbol used with specific hazards noted
	Danger, warning, or caution symbol followed by specific precautions

References

1. CLSI EP6-A; Evaluation of the Linearity of Quantitative Measurement Procedures: A Statistical Approach; Approved Guidelines
2. CLSI EP15-A2, Vol. 21 No 25. User Verification of Performance for Precision and Trueness; Approved Guideline - Second Edition (2005)
3. NCCLS EP17-A, Vol. 24 No. 34. Protocols for Determination of Limits of Detection and Limits of Quantitation; Approved Guideline (2012)
4. Ahlquist DA, H Zou, M Domanico, et al. 2010. Next generation stool DNA testing for detection of colorectal neoplasia: early marker evaluation. Presented at American Association of Cancer Research, Oct. 29, 2010, Philadelphia, PA.
5. Jemal A, R Siegel, E Ward, et al. 2010. Cancer statistics, 2010. CA Cancer J Clin, 60: 277-300.
6. Melotte, V, Lentjes, MHFM, van den Bosch SM ,et al. 2009.N-Myc Downstream-Regulated Gene 4 (*NDRG4*): A Candidate Tumor Suppressor Gene and Potential Biomarker for Colorectal Cancer V, JNCI J Natl Cancer Inst, Volume101, Issue13.916-927.
7. Zou H, Harrington JH, Abdirashid M, et al. 2007.Highly Methylated Genes in Colorectal Neoplasia: Implications for Screening, Cancer Epidemiology, Biomarkers and Prevention;16:2686.
8. Bos, J.L., Fearon, E.R., Hamilton, S.R., Verlaan-de Vries, M. and Vogelstein B. (1987) Nature 327, 293-297. Prevalence of ras gene mutations in human colorectal cancers.
9. Malumbres M, Barbacid M. RAS oncogenes: the first 30 years. Nat Rev Cancer. 2003;3:459-65.
10. Downward J. Targeting RAS signaling pathways in cancer therapy. Nat Rev Cancer. 2003;3:11-22.
11. Ahlquist D, et al. Gastroenterology. 2012;142:2, p 248-256.
12. miRBase Sequence database, release 13.0. ftp://ftp.sanger.ac.uk/pub/mirbase/sequences/.

Contact Information

Items supplied through Exact Sciences, including consumable or customer orderable replacement parts, may be ordered through Exact Sciences Customer Service at 800-XXX-XXXX.

Exact Sciences Corporation
441 Charmany Drive
Madison, WI USA

Trademarks

QuARTS and *Cologuard* are trademarks of Exact Sciences Corporation. MicroAmp® and Applied Biosystems® are registered trademarks of Life Technologies Corporation. Microlab® is a registered trademark of Hamilton Robotics. ELx808™ is a trademark of BioTek® Instruments Inc. Titer Tops® is a registered trademark of Diversified Biotech. Parafilm® is a registered trademark of Pecheney Plastic Packaging. ProClin® is a registered trademark of Rohm and Haas Company.

Patents

This product and its use are covered by one or more of the following patents, associated pending patent applications, or non-U.S. international equivalents: U.S. 5,741,650, 5,952,178, 6,268,136, 6,303,304, 6,406,857, 6,551,777, and 8,361,720. This product and its use are also covered by patents pending in the U.S. and around the world.

Dye compounds in this product are sold under license from Biosearch Technologies, Inc. and may be protected by U.S. and worldwide patents or patent applications. The license covers human in vitro diagnostic (IVD) applications and monitoring of treatment effectiveness.

The purchase of this product grants the purchaser rights under certain Hologic patents to use it

solely for providing human *in vitro* diagnostic services. No general patent or other license of any kind other than this specific right of use from purchase is granted hereby.

Limited License

The receipt of this product from Exact Sciences Corporation or its authorized distributor includes a limited, non-exclusive license under patent rights held by Exact Sciences. Acquisition of this product constitutes acceptance by the recipient of this limited license. Recipients unwilling to accept the limited license must return the product for a full refund. Such license is solely for the purposes of using this product to [FDA ACCEPTABLE STATEMENT OF PRODUCT USE] . For avoidance of doubt, the foregoing license does not include the rights to use this product for [EXCLUSIONS]. The foregoing license does not include a license to use the product for new product research or development, product manufacture, or any reverse-engineering purposes. The purchase of this product is not authorized to transfer this product to any third party for any purpose without the express written consent of Exact Sciences Corporation. Except as expressly provided in this paragraph, no other license is granted expressly, impliedly, or by estoppel.

For information concerning availability of additional licenses to practice the patent methodologies, contact:

Exact Sciences
441 Charmany Dr.
Madison, WI 53719
(608) 284-5700

Warranty

Exact Sciences Corporation warrants that this product will meet the specifications stated on the product information sheet. If any component of this product does not conform to these specifications, Exact Sciences Corporation will at its sole discretion, as its sole and exclusive liability and as the users sole and exclusive remedy, replace the product at no charge or refund the cost of the product; provided that notice of nonconformance is given to Exact Sciences Corporation within sixty (60) days of receipt of the product.

This warranty limits Exact Sciences Corporation liability to the replacement of this product or refund of the cost of the product. NO OTHER WARRANTIES OF ANY KIND, EXPRESS OR IMPLIED, INCLUDING WITHOUT LIMITATION IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE OR NON-INFRINGMENT, ARE PROVIDED BY EXACT SCIENCES CORPORATION. Exact Sciences Corporation shall have no liability for any direct, indirect, consequential or incidental damages arising out of the use, the results of the use or the inability to use this product and its components.

Copyright

©2013 Exact Sciences Corporation