

Colorectal Cancer Screening by Detection of Altered Human DNA in Stool: Feasibility of a Multitarget Assay Panel

DAVID A. AHLQUIST,* JOEL E. SKOLETSKY,† KEVIN A. BOYNTON,† JONATHAN J. HARRINGTON,* DOUGLAS W. MAHONEY,* WILLIAM E. PIERCEALL,† STEPHEN N. THIBODEAU,* and ANTHONY P. SHUBER†

*Division of Gastroenterology and Hepatology, Department of Biostatistics, and Division of Molecular Genetics, Mayo Clinic, Rochester, Minnesota; and †Applied Research Group, EXACT Laboratories, Maynard, Massachusetts

Background & Aims: Assay of altered DNA exfoliated into stool represents an intriguing approach to screen for colorectal neoplasia, but multiple markers must be targeted because of genetic heterogeneity. We explored the feasibility of a stool assay panel of selected DNA alterations in discriminating subjects with colorectal neoplasia from those without. **Methods:** Freezer-archived stools were analyzed in blinded fashion from 22 patients with colorectal cancer, 11 with adenomas ≥ 1 cm, and 28 with endoscopically normal colons. After isolation of human DNA from stool by sequence-specific hybrid capture, assay targets included point mutations at any of 15 sites on *K-ras*, *p53*, and *APC* genes; *Bat-26*, a microsatellite instability marker; and highly amplifiable DNA. **Results:** Analyzable human DNA was recovered from all stools. Sensitivity was 91% (95% confidence interval, 71%–99%) for cancer and 82% (48%–98%) for adenomas ≥ 1 cm with a specificity of 93% (76%–99%). Excluding *K-ras* from the panel, sensitivities for cancer were unchanged but decreased slightly for adenomas to 73% (39%–94%), while specificity increased to 100% (88%–100%). **Conclusions:** Assay of altered DNA holds promise as a stool screening approach for colorectal neoplasia. Larger clinical investigations are indicated.

Colorectal cancer remains the second leading cause of malignant mortality in industrialized nations, accounting for more than 10% of all cancer deaths.¹ Because of its orderly natural history and nonsurgical accessibility, colorectal neoplasia appears ideally suited for preventive interventions. However, screening efforts have had relatively little impact partly because of performance limitations and low compliance rates with current screening tools. More optimally tailored screening tools are needed that would exhibit the combined features of high sensitivity and specificity for early-stage cancers and large premalignant adenomas, broad acceptability by the general population, affordability, and safety.

Stool testing merits further consideration because its theoretical potential has not yet been achieved. Stool tests are noninvasive, require no cathartic preparation, can be performed on mailed-in specimens without a mandated health center visit, and may reflect the full length of the colorectum. Fecal occult blood testing regularly performed over a decade or more may lower colorectal cancer mortality by 15%–33%, primarily by detecting cancers at an earlier stage.^{2–4} However, many cancers and most premalignant adenomas do not bleed and are missed.^{5,6} Furthermore, high false-positive rates as a result of frequent and trivial sources of occult bleeding lead to unnecessary colonoscopies, which drive up programmatic costs.^{7–9} More sensitive and specific markers would improve the effectiveness and efficiency of stool screening.

Neoplasm-specific DNA alterations have been well-characterized^{10,11} and represent intriguing candidate markers for stool screening. In contrast to blood, altered DNA arises from the neoplasm rather than from the circulation and is released into the lumen continuously via exfoliation rather than intermittently via bleeding.¹² Furthermore, DNA appears to be stable in stool¹³ and amplification techniques permit detection of minute amounts of analyte. Several investigators have recovered mutant DNA in stools from patients with colorectal cancer or adenomas.^{14–20} Assays used have typically analyzed mutations on a single gene, especially *K-ras* because of its small number of mutational sites. However, colorectal neoplasms are genetically heterogeneous^{10,11}; no single mutation has been identified that is expressed across all colorectal neoplasms. Mutant *K-ras*, for example, is expressed by fewer than half.^{10,11,21} Thus, multiple

Abbreviations used in this paper: L-DNA, “long” DNA; PCR, polymerase chain reaction.

© 2000 by the American Gastroenterological Association

0016-5085/00/\$10.00

doi:10.1053/gast.2000.19580

DNA alterations must be targeted to achieve high neoplasm detection rates.

An assay system was developed which targets a spectrum of DNA alterations that occur with colorectal neoplasia. This multicomponent assay panel targets point-mutations at any of 15 mutational hot spots on *K-ras*, *APC*, and *p53* genes; Bat-26, a marker of microsatellite instability^{22,23}; and highly-amplifiable or “long” DNA (L-DNA). Observations by other investigators¹⁷ of higher fecal DNA yields from patients with colorectal cancer than from controls provided the basis to include this latter marker in the assay panel.

The aim of this blinded clinical pilot investigation was to assess the potential of a multitarget fecal DNA assay panel to discriminate selected patients with colorectal neoplasia from those without neoplasia using colonoscopy as the criterion standard.

Materials and Methods

Design and Subjects

The investigation was approved by the Mayo Clinic Institutional Review Board and comprised 2 clinical pilot studies. Stools for each were selected from a freezer archive to yield subject groups with verified colorectal adenocarcinoma, colorectal adenomas ≥ 1.0 cm, and colonoscopically normal colons. Subjects were chosen to provide a balanced age and gender representation across groups and a mixed distribution of neoplasms from both proximal and distal colorectal sites (Table 1). Most patients with cancer had been referred with a known diagnosis or with a radiographically suspicious mass. All subjects with adenomas and normal colons were asymptomatic and undergoing surveillance because of either a family history of colorectal neoplasia or a previous personal history of neoplasia.

Pilot study 1 was conducted to explore the diagnostic discrimination of the fecal DNA assay panel by confirming the appropriateness of preset positivity levels for point mutations and BAT-26, by establishing a cutoff level of positivity for L-DNA, and by correlating specific mutations found in stool with those in matched tumors. In pilot study 2, the assay panel was applied using assay parameters established from pilot study 1, and the major focus of study 2 was to examine specificity in a separate group. As indicated in Table 1, 7 of the 10 patients with neoplasms in pilot study 2 had also been tested in pilot study 1. Because separate unfrozen fecal aliquots were assayed on these 7 patients, within-stool reproducibility of assay results could be determined. All assays were performed by technicians blinded to the clinical data. On pathology review, 1 of the adenomas originally in pilot study 1 was found to have high-grade dysplasia with a focus of invasive cancer and was reassigned from the adenoma to the cancer group. In pilot study 2, 2 of the original 20 normal controls were excluded, one because of an inadequate colonoscopic examination due to a poor preparation and the other

Table 1. Demographic and Colorectal Neoplasm Characteristics of Subject Groups

	Pilot 1	Pilot 2	Combined
Cancers			
n	21	8	22 ^b
Sex (M/F)	11/10	4/4	11/11
Age (yr) ^a	69 (38–88)	73 (54–83)	70 (38–88)
Tumor site, prox/ dist	10/11	4/4	14/15
Tumor size (cm) ^a	4.9 (2.6–11)	3.9 (2.5–11)	4.1 (2.5–11)
Tumor stage, Dukes AB/CD	13/8	5/3	13/9
Adenomas			
n	9	2	11
Sex (M/F)	4/5	1/1	5/6
Age (yr) ^a	69 (61–76)	74 (72–76)	73 (61–76)
Polyp site, prox/ dist	5/4	1/1	6/5
Polyp size (cm) ^a	1.5 (1–5)	4 (1–7)	2 (1–7)
Normal colons			
n	10	18	28
Sex (M/F)	5/5	9/9	14/14
Age (yr) ^a	69 (53–77)	67 (50–74)	68 (50–77)

prox, proximal to splenic flexure; dist, splenic flexure or distal.

^aMedian (range).

^bBecause stools from 7 cancer patients from pilot 1 were repeated in pilot 2, total number of unique cancer patients was 22.

because of the subsequent finding of a malignant ileal carcinoid tumor.

Stool Collection, Processing, and Storage

All stools had been collected within days before cathartic preparation for a scheduled colonoscopy, which served as the criterion standard. Any previous instrumentation had occurred ≥ 2 weeks before stool collections from colorectal cancer patients and ≥ 1 year for patients with adenomas and normal controls. To prevent toilet water artifact,²⁴ a plastic bucket device was used to collect whole stools. Stools in sealed buckets were received within 12 hours of defecation at the on-site processing laboratory where they were tested by Hemocult (see below) and promptly frozen at -80°C in multiple aliquots. Frozen single fecal aliquots of at least 6 g per subject were sent in batches on dry ice for blinded DNA analyses at EXACT Laboratories (Maynard, MA).

Multitarget DNA Assay Panel

Total nucleic acid preparation. All stool samples were thawed at room temperature and homogenized in an excess volume ($>1:10$, wt, vol) of EXACT buffer A (EXACT Laboratories) using an EXACTOR stool shaker (EXACT Laboratories). After homogenization, a 4-g stool equivalent of each sample was centrifuged to remove all particulate matter, and the supernatants were incubated at 37°C after addition of proteinase K (0.5 $\mu\text{g}/\mu\text{L}$) and sodium dodecyl sulfate (0.5%). The supernatants were subsequently extracted with Tris-saturated phenol (GIBCO BRL, Grand Island, NY), phenol/chloroform/isoamyl alcohol (25:24:1), and chloroform. Total nu-

cleic acid was then precipitated (1/10 volume 3 mol/L NaAc and an equal-volume isopropanol), removed from solution by centrifugation, and resuspended in TE (0.01 mol/L Tris [pH 7.4] and 0.001 mol/L EDTA) buffer containing RNase A (2.5 µg/mL). For each group of samples prepared, process positive control samples as well as component negative controls were included.

Sequence-specific purification and amplification.

Sequence-specific DNA fragments were purified from the total nucleic acid preparations by performing oligonucleotide-based hybrid captures. For each sample, 7 unique hybrid capture reactions were performed in duplicate. Each capture reaction was carried out by adding 300 µL of sample preparation to an equal volume of 6 mol/L guanidine isothiocyanate solution (GIBCO BRL) containing biotinylated sequence-specific oligonucleotides (20 pmol; Midland Certified Reagent Co., Midland, TX). After a 2-hour incubation at 25°C, streptavidin-coated magnetic beads were added to the solution, and the tubes were incubated for an additional hour at room temperature. The bead/hybrid capture complexes were then washed 4 times with 1× B+W buffer (1 mol/L NaCl, 0.01 mol/L Tris-HCl [pH 7.2], 0.001 mol/L EDTA, and 0.1% Tween 20), and the sequence-specific captured DNA was eluted into 35 µL L-TE (1 mmol/L Tris [pH 7.4] and 0.1 mol/L EDTA) by heat denaturation.

Polymerase chain reaction (PCR) amplifications (50 µL) were performed on MJ Research Tetrad Cyclers (Watertown, MA) using 10 µL of captured DNA, 1× GeneAmp PCR buffer (PE Biosystems, Foster City, CA), 0.2 mmol/L dNTPs (Promega, Madison, WI), 0.5 µmol/L sequence-specific primers (Midland Certified Reagent Co., Midland, TX), and 5 U AmpliTaq DNA polymerase (PE Applied Biosystems, Norwalk, CT). All sequence-specific amplification reactions were performed in identical thermocycler conditions. After an initial denaturation of 94°C for 5 minutes, PCR amplification was performed for 40 cycles consisting of 1 minute at 94°C, 1 minute at 60°C, and 1 minute at 72°C, with a final extension of 5 minutes at 72°C. For PCR product analysis, 8 µL of each amplification reaction was loaded and electrophoresed on a 4% ethidium bromide-stained NuSieve 3:1 agarose gel (FMC, Rockland, ME) and visualized with a Stratagene EagleEye II (Stratagene, La Jolla, CA) still image system (Figure 1).




Figure 1. Agarose gel electrophoretic analysis of K-ras PCR products. Amplification results representing 6 unique stool DNA samples (1–6) amplified in duplicate (*lanes a + b*) with appropriate negative (*lane 7*) and positive (*lane 8*) control amplifications. Similar results were obtained for other amplicons including *p53*, *APC*, and *Bat-26*.




Figure 2. Point mutational results for *APC* codon 1378 position 1. Point mutation results representing 6 unique stool DNA samples (5–10 and 15–20) analyzed in duplicate (*lanes a + b*). Wild-type reactions (*lanes 1–10*) and corresponding mutant reactions (*lanes 11–20*) were analyzed by polyacrylamide gel electrophoresis. Positive wild-type results (*lanes 1–10*) served as internal sample specific controls. Each set of reactions (*lanes 11–20*) included mutation-specific positive controls representing 1% (*lane 13*) and 5% (*lane 14*) mutant DNA populations and negative control samples (*lanes 1, 2, 11, and 12*) containing wild-type DNA only. Within this analysis, a single-stool DNA sample was positive for both wild-type (*lane 8, a + b*) and mutation-specific reactions (*lane 18, a + b*). All other stool DNA samples were positive for the wild-type reaction (*lanes 5–7, 9, and 10, a + b*) and negative for the *APC* 1378 mutation (*lanes 15–17, 19, and 20, a + b*).

Point mutation and Bat-26 analysis. The presence or absence of point mutations or Bat-26-associated mutations was determined by using a modified solid-phase minisequencing method.²⁵ Point mutation targets included codons K12p1, K12p2, and K13p2 on the *K-ras* gene; codons 1309 delta 5, 1367p1, 1378p1, and 1450p1 on the *APC* gene; and codons 175p2, 245p1, 245p2, 248p1, 248p2, 273p1, 273p2, and 282p1 on the *p53* gene. These targets were selected for assay because they correspond to the highest frequency mutational sites observed in available tissue databases. From these databases, theoretical diagnostic yields for cancer were estimated to be 44% for *p53*,²⁶ 41% for *K-ras*,²⁷ and 19% for *APC*.²⁸ For all gene targets, both wild-type and mutant-specific reactions were performed. Within the wild-type reactions, radionucleotide bases complementary to the wild-type base were added (Figure 2). For each point mutation-specific reaction, radionucleotide bases complementary to the expected mutant bases were added in addition to unlabeled dideoxy nucleotides complementary to the wild-type base (Figure 2). Bat-26 mutations associated with a deletion of 4–15 base pairs (bp) were identified by size discrimination of reaction products (Figure 3). We estimated that the theoretical yield by this microsatellite instability marker for cancer detection would be at least 15% based on reported observations in tissue.²⁹

L-DNA analysis. L-DNA was performed by analyzing the relative intensity of each sample-specific PCR product. For each stool sample analyzed, 7 unique PCR amplification




Figure 3. Polyacrylamide gel electrophoresis analysis for deletions within Bat-26. Results representing 10 unique stool DNA samples (lanes 1–10) analyzed for deletions within Bat-26. The upper (U) region of the gel contains reaction products representing the wild-type full-length product. Presence of reaction products within the lower (L) region of the gel is indicative of deletions within the Bat-26 polyA tract sequence. In addition to the stool samples analyzed (lanes 1–10), a no-DNA negative control (lane 11) and a wild-type-only DNA positive control (lane 12) were analyzed. Deletion-positive controls containing 1% (lane 13) and 5% (lane 14) mutant DNA (15 base deletions) were also analyzed within each assay to assure that resolution between wild-type and deleted sequences was achieved. Note that 2 of the samples (lanes 2 and 9) contain deletions within Bat-26, while all remaining samples contain the wild-type number of bases.

products were generated in duplicate (or 14 amplifications per subject) and independently scored by 2 technicians. PCR product intensities were scored as high (A), medium (B), or low (C) by visual examination of the gel image. Figure 1 illustrates examples of samples scored as A amplifications (lanes 1ab + 4ab), B amplifications (lanes 2ab, 5ab, and 6ab), and C amplifications (lane 3ab). The cutoff score to indicate a positive result was determined in pilot study 1.

Sequence information for all capture probes and primers will be available on request to the corresponding author.

Fecal Occult Blood Testing

All stools collected from subjects with adenomas and normal colonoscopy had been tested in blinded fashion by Hemoccult II (SmithKline Diagnostics, Sunnyvale, CA) immediately on receipt and before freezing. From a single stool, 2 aliquots sampled from opposite ends of the specimen were each smeared onto 2 windows of a Hemoccult II test card for a total of 4 windows. A single drop of peroxide catalyst was promptly added to each window, and a blue color reaction within 60 seconds on at least 1 of the 4 test windows was called a positive result for that stool. Positive and negative controls were tested with each run. Stools were selected from the archive for study without knowledge of Hemoccult status.

Tissue Processing and Assay

DNA from colorectal cancers in pilot study 1 was obtained at the Mayo Clinic after microdissection of sections from the original paraffin-embedded, formalin-fixed tumor

blocks using a previously described extraction technique.³⁰ This DNA was sent to EXACT Laboratories where point mutation assays by the single-base extension method (see above) were performed in a blinded fashion.

Statistical Analysis

Sensitivity and specificity were estimated relative to the results of colonoscopy in the usual manner; 95% confidence intervals (CIs) for these estimated parameters were based on the exact binomial distribution. Comparisons of proportions between various subgroups were based on the Fisher exact test, and McNemar's matched-pairs test for proportions was used to compare sensitivities between the fecal DNA assay panel (with and without *K-ras*) and Hemoccult. Interobserver variability for L-DNA was assessed in pilot study 1 using a weighted κ statistic because the individual scores were ordinal in nature. The weighted κ statistic was estimated for each of the 7 PCR amplification products separately and for the pooled observations across all PCR amplifications. Among the 560 PCR amplifications scored, there were 7 instances in which 1 of the duplicate amplifications was noninformative; such instances were considered discordant so as to obtain a conservative estimate of interobserver variability.

Results

Analyzable human DNA was recovered in all subjects. When detected, mutant DNA accounted for 1%–24% of total human DNA recovered in stools from cancer patients and for 1%–7% from those with large adenomas.

Pilot Study 1

Cancers. At least 1 point mutation among the 15 targeted sites on *K-ras*, *APC*, and *p53* genes was present in stools from 11 (52%) of the 21 cancers. Bat-26 was positive in 4 cases, 1 of which also had a point mutation, yielding a detection rate of 14/21 (67%) when combined with point mutation components. Samples that were assigned an "A" score on >8 amplifications were considered L-DNA positive, because all of the colonoscopically normal controls fell below this cutoff. Using this definition, L-DNA alone was positive in stools from 14 (67%) of the 21 cancers. With all component markers together, the fecal DNA assay panel detected 19 (90%) of the 21 cancers (Table 2).

Tissue was available for DNA extraction and point mutation analyses on 19 of the 21 cancers. Point mutation results on tissue and stool were concordant in 12 cases (63%); at least one pair of identical mutations was found in stools and matched tumors in 7 cases, and all targeted mutations were absent in both stool and tumors in 5 cases. In the 7 discordant cases, a mutation was

Table 2. Positivity Rates of the Fecal DNA Assay Panel by Subject Group From Each Pilot Study and From Combined Studies: Component Markers and All Markers Together

	K-ras	APC	p53	Bat-26	L-DNA ^a	All markers	
						n (%)	95% CI
Pilot 1							
Cancers (21)	4	5	3	4	14	19 (90)	70%–99%
Adenomas (9)	0	3	0	0	5	7 (78)	40%–97%
Normals (10)	0	0	0	0	0	0 (0)	0%–31%
Pilot 2							
Cancers (8)	2	2	1	2	4	8 (100)	63%–100%
Adenomas (2)	1	0	0	0	1	2 (100)	16%–100%
Normals (18)	2	0	0	0	0	2 (11)	1%–35%
Combined studies							
Cancers (22) ^b	4	5	3	5	14	20 (91)	71%–99%
Adenomas (11)	1	3	0	0	6	9 (82)	48%–98%
Normals (28)	2	0	0	0	0	2 (7)	1%–24%
Combined studies (excluding K-ras)							
Cancers (22) ^b	—	5	3	5	14	20 (91)	71%–99%
Adenomas (11)	—	3	0	0	6	8 (73)	39%–94%
Normals (28)	—	0	0	0	0	0 (0)	0%–12%

^aL-DNA refers to “long” or nonapoptotic DNA.

^bBecause stools from 7 cancer patients from pilot 1 were repeated in pilot 2, total number of unique cancer patients was 22.

detected in the stool only in 4 cases and the tissue only in 3 cases.

Adenomas and normal controls. The fecal DNA panel detected 7 (78%) of the 9 adenomas ≥ 1 cm. All positives were the result of APC mutations or elevated L-DNA (Table 2). The assay panel was negative in all 10 colonoscopically normal patients.

Interobserver variability for L-DNA. Perfect agreement between the 2 technicians scoring L-DNA ranged from 88% to 96% for duplicate testing on the 7 PCR amplification products; perfect agreement was $>93\%$ for 6 of the 7 products. Weighted κ values for individual PCR products ranged from 0.36 ± 0.09 to 0.74 ± 0.07 ; the pooled weighted κ value across all 7 PCR products was 0.58 ± 0.12 .

Pilot Study 2

All 8 of the cancers and both of the adenomas were detected by the fecal DNA assay panel using the same assay parameters as in pilot study 1 (Table 2). Of the 7 cancer patients in this series who had also been evaluated by separate fecal aliquots in pilot study 1, at least 1 identical DNA alteration was reproduced in all 7 instances; all component markers of the assay panel were concordant in 5 instances and at least 1 marker was discordant in the other 2 instances. The assay panel was positive in 2 (11%) of the 18 colonoscopically normal controls. Both of these false-positive cases were a result of K-ras mutations.

Both Studies Combined

Sensitivity and specificity. Using the full panel of component markers, the sensitivity of the fecal DNA assay panel for the 22 cancers was 91% (95% CI, 71%–99%) and for the 11 adenomas ≥ 1 cm was 82% (95% CI, 48%–98%) with a specificity of 93% (95% CI, 76%–99%). If K-ras markers were excluded from the panel, then sensitivity for cancer was unaffected at 91% (95% CI, 71%–99%) but decreased slightly for adenomas to 73% (95% CI, 39%–94%) while specificity increased to 100% (95% CI, 88%–100%).

For all neoplasms (cancers and adenomas), L-DNA proved to be the most informative marker and alone detected 20 (61%) of the 33 unique lesions. Bat-26 and p53 markers were positive with cancer but not with adenomas in this initial series. In stools from the 20 cancer patients with a positive DNA assay panel, 3 component markers were positive in 3 cases, 2 markers positive in 5 cases, and a single marker positive in 12 cases. In stools from the 9 test-positive adenoma patients, 2 component markers were positive in 2 cases and a single marker was positive in the other 7.

In this highly selected subject group, the positive predictive value for colorectal neoplasia by the fecal DNA panel (excluding K-ras markers) was 100% (28/28) and the negative predictive value was 85% (28/33).

Clinical correlates. Positive results by the component marker Bat-26 were significantly associated with proximal colorectal cancer site, and an association of positive L-DNA results with distal tumor site was sug-

gested (Table 3). All 5 Bat-26–positive tumors were located proximal to the splenic flexure (4 in the cecum or ascending colon). Conversely, 10 of the 14 L-DNA–positive cancers were located distal to the splenic flexure.

Outcomes by the fecal DNA assay panel were not associated with subject age or gender or with lesion size or stage. Because all adenomas were classified as low-grade dysplasia and all but one as tubular, neither histologic grades nor architectural type could be evaluated in this small series.

Comparison of the fecal DNA assay panel and Hemoccult tests. Hemoccult testing had been performed only on the asymptomatic adenoma and control subjects. Whether or not *K-ras* was included, the DNA assay panel detected substantially and significantly more of the adenomas ≥ 1 cm than did the Hemoccult test based on a single stool per subject (Table 4). Hemoccult was negative in stools from all 28 normal controls. Thus, at matched specificities of 100% in this small series, the DNA assay panel (minus the *K-ras* marker) detected 8 (73%) of the 11 adenomas compared with none (0%) by the Hemoccult test ($P < 0.008$).

Discussion

The rationale and appeal of stool screening for colorectal cancer would be strengthened if markers substantially more accurate than occult blood were used. On the basis of this pilot investigation, the aggregate assay of multiple genetic markers exfoliated into stool represents a promising alternative. The panel of neoplasm-associated DNA alterations targeted in this study highly discriminated patients with colorectal cancer or large adenomas from those with endoscopically normal colons.

The fecal DNA assay panel detected 20 (91%) of the 22 colorectal cancers in this blinded pilot. A likely factor contributing to high neoplasm detection rates by the

Table 4. Comparison in Adenoma Detection Rates by the Fecal DNA Assay Panel and Hemoccult Test

	Adenoma detection rate
DNA assay panel (all markers)	9/11 (82%) ^a
DNA assay panel minus <i>K-ras</i>	8/11 (73%) ^b
Hemoccult test	0/11 (0%)

NOTE. Both tests were performed on same single stool per patient.

^a $P = 0.004$.

^b $P = 0.008$.

DNA assay panel, in addition to the multiple markers targeted, was the efficient isolation of human DNA from stool. Amplifiable human DNA was recovered from all stools in this pilot compared with recovery rates of 45%–90% by others.^{14–20} This difference may, in part, be accounted for by the more effective removal of PCR inhibitors, known to be present in stool,³¹ by the sequence-specific hybrid capture method used. However, it must be emphasized that the cancers included in this pilot study were largely symptomatic and may behave differently from asymptomatic cancers with respect to marker shedding. Another variable that could influence sensitivity is the representativeness of the fecal aliquot tested, and the degree of tissue-stool discordance observed would suggest that specimen sampling was suboptimal in this pilot study. The effect of different sampling techniques on test performance will require further evaluation.

A major limitation of stool screening has related to the very low sensitivity of occult blood as a marker for premalignant adenomas.^{5,6} The functional inadequacy of fecal occult blood tests to detect adenomas is evidenced practically by their failure in controlled trials to meaningfully reduce the cumulative incidence of colorectal cancer.^{2–4} In the present small initial series, the fecal DNA assay panel detected a remarkably high proportion (73%–82%) of adenomas ≥ 1 cm, whereas single-stool Hemoccult testing in the same asymptomatic patients detected none. The results by this multitarget DNA assay approach must be corroborated in larger clinical studies, but the clear implications are for improved cancer prevention and, perhaps, less frequent screening.

The presence of high-integrity DNA, or L-DNA, proved to be the most informative component marker of the fecal DNA assay panel and alone detected 61% of neoplasms. We speculate that longer template DNA is an epigenetic phenomenon consistent with the known abrogation of apoptosis that occurs with colorectal cancer.^{32,33} There appears to be abundant exfoliation of nonapoptotic cells from neoplasms³⁴; in contrast, colonocyte shedding from normal mucosa is

Table 3. Components of the Fecal DNA Assay Panel Associated With Colorectal Cancer Site

Marker	Positive markers by colorectal site	
	Proximal to splenic flexure	Distal to splenic flexure
Bat-26		
Pilot 1	4/9	0/12
Pilot 2	2/4	0/4
Combined studies ^a	5/10 (50%)	0/12 (0%)
L-DNA		
Pilot 1	4/9	10/12
Pilot 2	0/4	4/4
Combined studies ^b	4/10 (40%)	10/12 (83%)

^a $P = 0.01$.

^b $P = 0.07$.

relatively sparse, and sloughed cells appear to be largely apoptotic.^{34,35} Furthermore, normal cells rapidly undergo apoptosis after detachment from their basement membrane.³⁶ Because a hallmark of apoptosis is the cleavage of DNA by endonucleases into fragments of 180–200 bp,^{37,38} it follows that human DNA in normal stools would exist primarily in fragmented or “short” forms. However, stools from patients with colorectal neoplasia should contain subsets of both nonapoptotic or “long” DNA arising from dysplastic cells and “short” DNA from normal mucosa. Because the primers used in the L-DNA assay were set to amplify human DNA sequences longer than 200 bp, an increase in the number of amplifiable template molecules serves as a logical marker for exfoliated dysplastic cells. In preliminary experiments, we have found that amplification of even longer human DNA sequences (e.g., 1800–2400 bp) from stool may allow for a more dramatic and objective separation of cancer patients from controls.³⁹ Further optimization and validation of this marker are needed.

High specificity is a requisite for cost-effective screening. Two false-positives were identified in this small series, both a result of *K-ras* mutations. Mutant *K-ras* can potentially arise from nonneoplastic sources, such as pancreatic hyperplasia⁴⁰ or hyperplastic aberrant crypt foci throughout the colorectum,⁴¹ or even normal-appearing colonic mucosa.⁴² Exclusion of mutant *K-ras* improved assay specificity to 100% with no loss in the cancer detection rate and only minimal loss in adenoma detection. The finding of a fecal DNA alteration in a patient with normal colonoscopic findings could indicate a supracolonial neoplasm. Indeed, the control patient disqualified from the present study because a malignant ileal carcinoid tumor was subsequently found to have a positive fecal DNA assay panel due to the L-DNA component. Further observations are needed to determine the positive predictive value of each component marker of the fecal DNA assay panel for supracolonial neoplasms in colonoscopy-negative subjects and the implications on screening algorithms. Finally, 2 recent studies have shown that Bat-26 polymorphisms occur in some African Americans,^{43,44} which could potentially yield false-positive fecal Bat-26 results. As such, alternative markers of microsatellite instability that are less affected by these or other polymorphisms might be considered.

Stool findings in the present study were consistent with certain biological features of colorectal neoplasia previously observed in tissue. Multiple DNA alterations were more frequently recovered in stools from patients with cancer than from those with adenomas, in keeping with the known

accumulation of gene mutations that occurs during the adenoma-to-carcinoma progression.^{10,11} The microsatellite instability marker, Bat-26, was positive only in stools of patients with proximally located colon cancer, reflecting a site-association well established from tissue studies.⁴⁵ Finally, neither Bat-26 nor p53 markers were positive with adenomas in this stool study, which would be in accord with previous tissue studies reporting that these 2 markers are rarely expressed before adenomatous transformation to high-grade dysplasia or frank cancer has occurred.^{46–48} Results from these 2 component markers of the fecal DNA assay panel may have been different had adenomas with high-grade dysplasia been included.

Detection of altered human DNA in stool using a multitarget assay panel appears to be a feasible method to screen both cancers and premalignant adenomas of the colorectum. Larger clinical studies are clearly indicated using optimal sample preparation and marker selection. The current labor-intensive assay techniques will need to be streamlined so that multitarget assay systems become practical for high-volume application, and exciting technologies are emerging that may allow for such.⁴⁹ If these pilot observations hold up in further clinical studies and the testing can be simplified, then the long-held goal of a noninvasive, sensitive, and specific fecal screening tool may be within reach.

References

1. Landis SH, Murray T, Bolden S, Wingo PA. Cancer statistics, 1998. *CA Cancer J Clin* 1998;48:6–29.
2. Hardcastle JD, Chamberlain JO, Robinson MH, Moss SM, Amar SS, Balfour TW, James PD, Mangham CM. Randomised controlled trial of faecal-occult blood screening for colorectal cancer. *Lancet* 1996;348:1472–1477.
3. Mandel JS, Bond JH, Church TR, Snover DC, Bradley GM, Schuman LM, Ederer F. Reducing mortality from colorectal cancer by screening for fecal occult blood. Minnesota Colon Cancer Control Study. *N Engl J Med* 1993;328:1365–1371.
4. Kronborg O, Fenger C, Olsen J, Jorgensen OD, Sondergaard O. Randomised study of screening for colorectal cancer with faecal-occult blood test. *Lancet* 1996;348:1467–1471.
5. Ahlquist DA, Wieand HS, Moertel CG, McGill DB, Loprinzi CL, O'Connell MJ, Mailliard JA, Gerstner JB, Pandya K, Ellefson RD. Accuracy of fecal occult blood screening for colorectal neoplasia. A prospective study using Hemoccult and HemoQuant tests. *JAMA* 1993;269:1262–1267.
6. Ransohoff DF, Lang CA. Small adenomas detected during fecal occult blood test screening for colorectal cancer: the impact of serendipity. *JAMA* 1990;263:76–78.
7. Simon JB. Fecal occult blood testing: clinical value and limitations. *Gastroenterologist* 1998;6:66–78.
8. Frank JW. Occult blood screening for colorectal cancer: the yield and costs. *Am J Prev Med* 1985;1:18–24.
9. Ahlquist DA. Fecal occult blood testing for colorectal cancer: can we afford to do this? *Gastroenterol Clin North Am* 1997;26:41–55.
10. Fearon ER, Vogelstein B. A genetic model for colorectal tumorigenesis. *Cell* 1990;61:759–767.
11. Boland CR, Sato J, Saito K, Carethers JM, Marra G, Laghi L,

- Chauhan DP. Genetic instability and chromosomal aberrations in colorectal cancer: a review of current models. *Cancer Detect Prev* 1998;22:377–382.
12. Ahlquist DA, Gilbert JA. Stool markers for colorectal cancer screening: future considerations. *Dig Dis* 1996;14:132–144.
 13. Poinar HN, Hofreiter M, Spaulding WG, Martin PS, Stankiewicz A, Bland H, Evershed RP, Possnert G, Paabo S. Molecular coproscopy: dung and diet of the extinct ground sloth *Nototheriops shastensis*. *Science* 1998;281:402–406.
 14. Sidransky D, Tokino T, Hamilton SR, Kinzler KW, Levin B, Frost P, Vogelstein B. Identification of ras oncogene mutations in the stool of patients with curable colorectal tumors. *Science* 1992;256:102–105.
 15. Smith-Ravin J, England J, Talbot IC, Bodmer W. Detection of c-Ki-ras mutations in faecal samples from sporadic colorectal cancer patients. *Gut* 1995;36:81–86.
 16. Nollau P, Moser C, Weinland G, Wagener C. Detection of K-ras mutations in stools of patients with colorectal cancer by mutant-enriched PCR. *Int J Cancer* 1996;66:332–336.
 17. Villa E, Dugani A, Rebecchi AM, Vignoli A, Grottola A, Buttafoco P, Losi L, Perini M, Trande P, Merighi A, Lerose R, Manenti F. Identification of subjects at risk for colorectal carcinoma through a test based on K-ras determination in the stool. *Gastroenterology* 1996;110:1346–1353.
 18. Hasegawa Y, Takeda S, Ichii S, Kaizumi K, Maruyama M, Fujii A, Ohta H, Nakajima T, Okuda M, Baba S, Nakamura Y. Detection of K-ras mutations in DNAs isolated from feces of patients with colorectal tumors by mutant-allele-specific amplification (MASA). *Oncogene* 1995;10:1441–1445.
 19. Ratto C, Flamini G, Sofo L, Nucera P, Ippoliti M, Curigliano G, Ferretti G, Sgambato A, Merico M, Doglietto GB, Cittandi A, Crucitti F. Detection of oncogene mutation from neoplastic colonic cells exfoliated in feces. *Dis Colon Rectum* 1996;39:1238–1244.
 20. Eguchi S, Kohara N, Komuta K, Kanematsu T. Mutations of the p53 gene in stool of patients with resectable colorectal cancer. *Cancer* 1996;77:1707–1710.
 21. Scott N, Quirke P. Molecular biology of colorectal neoplasia. *Gut* 1993;34:289–292.
 22. Samowitz WS, Slattery ML, Potter JD, Leppert MF. BAT-26 and BAT-40 instability in colorectal adenomas and carcinomas and germline polymorphisms. *Am J Pathol* 1999;154:1637–1641.
 23. Hoang JM, Cottu PH, Thuille B, Salmon RJ, Thomas G, Hamelin R. BAT-26, an indicator of the replication error phenotype in colorectal cancers and cell lines. *Cancer Res* 1997;57:300–303.
 24. Ahlquist DA, Schwartz S, Isaacson J, Ellefson M. A stool collection device: the first step in occult blood testing. *Ann Intern Med* 1988;108:609–612.
 25. Syvanen AC, Aalto-Setälä K, Harju L, Kontula K, Soderlund H. A primer-guided nucleotide incorporation assay in the genotyping of apolipoprotein E. *Genomics* 1990;8:684–692.
 26. Soussi T. p53 mutation in colorectal cancer. Retrieved in revised version from World Wide Web 1999. http://perso.curie.fr/Thierry.Soussi/p53_mutation_in_colon.htm.
 27. Vogelstein B, Fearon ER, Hamilton SR, Kern SE, Preisenger AC, Leppert M, Natasmura Y, White R, Smits AM, Bos JL. Genetic alterations during colorectal tumor development. *N Engl J Med* 1988;319:525–532.
 28. Soussi T. Frequency of APC mutations. Retrieved in revised version from World Wide Web 1999. <http://perso.curie.fr/Thierry.Soussi/APC.htm>.
 29. Aaltonen LA, Peltomäki P, Leach FS, Siltanen P, Pylkkonen L, Mecklin LP, Jarvinen H, Powell SM, Jen J, Hamilton SR, Peterson G, Kinzler KW, Vogelstein B, de la Chapelle A. Clues to the pathogenesis of familial colorectal cancer. *Science* 1993;260:812–816.
 30. Thibodeau SN, French AJ, Cunningham JM, Tester D, Burgart LJ, Roche PC, McDonnell SK, Schaid DJ, Vockley CW, Michels VV, Farr GH, O'Connell MJ. Microsatellite instability in colorectal cancer: different mutator phenotypes and the principle involvement of hMLH1. *Cancer Res* 1998;58:1713–1718.
 31. Wilde J, Eiden J, Yolken R. Removal of inhibitory substances from human fecal specimens for detection of group A rotaviruses by reverse transcriptase and polymerase chain reactions. *J Clin Microbiol* 1990;28:1300–1307.
 32. Tsujitani S, Shirai H, Tatebe S, Sugamura K, Ohfugi S, Gomyo Y, Malta M, Ito H, Kaibara N. Apoptotic cell death and its relationship to carcinogenesis in colorectal carcinoma. *Cancer* 1996;77(suppl):1711–1716.
 33. Bedi A, Pasricha PJ, Akhtar AJ, Barber JP, Bedi GC, Giardiello FM, Zehnbauser BA, Hamilton SR, Jones RJ. Inhibition of apoptosis during development of colorectal cancer. *Cancer Res* 1995;55:1811–1816.
 34. Ahlquist DA, Harrington JJ, Burgart LJ, Roche PC. Morphometric analysis of the “mucocellular layer” overlying colorectal cancer and normal mucosa: relevance to exfoliation and stool screening. *Hum Pathol* 2000;31:51–57.
 35. Potten CS, Allen TD. Ultrastructure of cell loss in intestinal mucosa. *J Ultrastructure Res* 1977;60:272–277.
 36. Stratton J, Wedding U, Barth TFE, Koretz K, Elsing C, Moller P. Rapid onset of apoptosis in vitro follows disruption of β 1-integrin/matrix interactions in human colonic crypt cells. *Gastroenterology* 1996;110:1776–1784.
 37. Wyllie AH. Glucocorticoid-induced thymocyte apoptosis is associated with endogenous endonuclease activation. *Nature* 1980;284:555–557.
 38. Sen S. Programmed cell death: concept, mechanism and control. *Biol Rev* 1992;67:287–319.
 39. Ahlquist DA, Cameron AJ, Jett J, Pearson RK, de Groen PC, Edell ES, Huyber ML, Devens ME, Shuber A. Universal detection of aerodigestive cancers by assay of nonapoptotic human DNA in stool (abstr). *Gastroenterology* 2000;118:A855.
 40. Caldas C, Hahn SA, Hruban RH, et al. Detection of K-ras mutations in the stool of patients with pancreatic adenocarcinoma and pancreatic ductal hyperplasia. *Cancer Res* 1994;54:3568–3573.
 41. Yamashita N, Minamoto T, Ochiai A, Onda M, Esumi H. Frequent and characteristic K-ras activation and absence of p53 protein accumulation in aberrant crypt foci of the colon. *Gastroenterology* 1995;108:434–440.
 42. Zhu P, Keohavong P, Finkelstein SD, Swalsky P, Bakker A, Weissfeld J, Srivastava S, Whiteside TL. K-ras gene mutation in normal colorectal tissues from K-ras mutation-positive colorectal cancer patients. *Cancer Res* 1997;57:2485–2492.
 43. Samowitz WS, Slattery ML, Potter JD, Leppert MF. Bat-26 and Bat-40 instability in colorectal adenomas and carcinomas and germline polymorphisms. *Am J Pathol* 1999;154:1637–1641.
 44. Pyatt R, Chadwick RB, Johnson CK, Adebamawo C, de la Chapelle A, Prior TW. Polymorphic variation at the Bat-25 and Bat-26 loci in individuals of African origin. *Am J Pathol* 1999;155:349–353.
 45. Thibodeau SN, Bren G, Schaid D. Microsatellite instability in cancer of the proximal colon. *Science* 1993;260:816–819.
 46. Samowitz WS, Slattery ML. Microsatellite instability in colorectal adenomas. *Gastroenterology* 1997;112:1515–1519.
 47. Young J, Leggett B, Gustafson C, Ward M, Searle J, Thomas L, Buttenshaw R, Chenevix-Trench G. Genomic instability occurs in colorectal carcinomas but not in adenomas. *Hum Mutat* 1993;2:351–354.
 48. Kikuchi-Yanoshita R, Konishi M, Ito S, Seki M, Tanaka K, Maeda Y, Iiono H, Fukayama M, Koike M, Mori T, Sakuraba H, Fukunari H, Iwama T, Miyaki M. Genetic changes of both p53 alleles associated with the conversion from colorectal adenoma to early carcinoma in familial adenomatous polyposis and nonfamilial

- adenomatous polyposis patients. *Cancer Res* 1992;52:3965–3971.
49. Service RF. Microchip arrays put DNA on the spot. *Science* 1998; 282:396–401.

Received October 7, 2000. Accepted June 21, 2000.

Address requests for reprints to: David Ahlquist, M.D., Division of Gastroenterology and Hepatology, Mayo Clinic, Rochester, Minnesota 55905. e-mail: ahlquist.david@mayo.edu; fax: (507) 266-0350.

Supported by grants from the National Cancer Institute of National Institutes of Health (R01 CA 7160), EXACT Laboratories (Maynard, MA), and Mayo Foundation (Rochester, MN) and by the philanthropic support of Charles W. Oswald.

The authors thank Drs. Bert Vogelstein and Kenneth W. Kinzler of Johns Hopkins University Oncology Center and Howard Hughes Medical Institute for helpful comments through the course of this study; Joy Yucaitis for extensive logistical support; Jaci McCormick, Mary Devens, Joel Swenson, and Nancy Diehl for valuable database support; Dr.

Lawrence Burgart for pathology review; and Drs. Steven J. Laken and Barry Berger for contributions to technical aspects of assay interpretation and critical scientific input. The authors also thank the following individuals for expertise and dedicated efforts: Kristina Stuhr Brennan (sample preparation); Cynthia LeSage Proulx, Jean A. Smith, J. Kent Moore, David Morrissey, and Sharon Magnolo (DNA purifications and amplifications); Pamela Shaw and Beverly M. Vucson (mutation detection and analyses); and Lauren A. Robison (quality assurance/quality control support).

Mayo Foundation is a minor equity investor in EXACT Laboratories. D. Ahlquist is a member of EXACT Laboratories' Scientific Advisory Board but holds no stock and has received no consulting fees. J. Harrington and D. Mahoney have no affiliation with EXACT Laboratories. A. Shuber, J. Skoletsky, W. Pierceall, and K. Boynton are employees of EXACT Laboratories.

Presented in part at the annual meeting of the American Gastroenterological Association in Orlando, Florida, in May 1999 (*Gastroenterology* 1999;116:A369).