

Implant Report



TMJ Report

IS I-CAT Specifications:

X-ray Source:	High Frequency, Constant Pot anode 120 kVp, 1-3 mA
X-ray Beam:	Cone
Focal Spot:	0.5mm
Image Detector:	Amorphous Silicon Flat Panel 25cm
Gray Scale:	12 bit
Voxel Size:	0.4mm (typical), 0.2mm (minir
Image Acquisition:	Single 360 degree rotation
Scan Time:	40 seconds or less
Patient Position:	Seated
Scan Dimensions:	17cm (diameter) x 11cm (heig 17cm (diameter) x 17cm (heig available in Phase 2
Primary Reconstruction:	About 2 minutes
Secondary Reconstruction:	Real Time

Copyright © 2004 Imaging Sciences International
1910 North Penn Rd., Hatfield, PA 19440
email: info@imagingsciences.com

Site by Ravine Design

Osseo 2023
U.S. Patent No. 6,944,262