





ResMed Pty Ltd
1 Elizabeth Macarthur Drive
Bella Vista NSW 2153 Australia
See www.resmed.com for other
ResMed locations worldwide

EC REP ResMed SAS
Parc Technologique de Lyon
292 Allée Jacques Monod
69791 Saint Priest Cedex France

STORAGE & TRANSPORT RANGES

-20C / 60C
-4F / 140F

95%
5

2591m
8500ft
ALTITUDE
LIMITATION

OPERATING RANGES

5C / 35C
41F / 95F

95%
10

RTCA/DO-160 Section 21
Category M Compliant
& FAA Compliant

DC24V
3.75A

IP22

1.3kg
2.9lb

For patent information, see www.resmed.com/ip
See symbols glossary at www.resmed.com/symbols

MD

R370-4146/2

Rx Only



This device complies with Part 15 of the FCC Rules and with ICSED Canada Rules.
Operation is subject to the following two conditions: (1) This device may not cause
harmful interference, and (2) This device must accept any interference received,
including interference that may cause undesired operation.

R370-438
VER. 1



ResMed

